



### Notice of Proposed Reciprocity Activity JAN - 7 2000


<b>Licensee Name</b> Varian Medical Systems 911 Hansen Way, MS C-080 Palo Alto, CA 94304	<b>Licensee Contact</b> Tony Sorensen, Radiation Safety Officer 650-424-6662 650-855-7375 (fax)
<b>License Information</b> State of California, Department of Health Services # 1025-43 Expiration Date: 11/25/97 (timely renewal)	<b>Person Authorized to Perform Activity</b> Richard Nusspickel Contact #: 800-864-1672


<b>Description of Device</b> Varian Model VariSource High Dose Rate Remote Afterloader (CA661D103S)	<b>Activity to be Performed</b> Non Routine Service (see comments)
<b>Description of Source</b> Omnitron International Model SL-777/777V (LA0760S102S) Iridium-192, Special Form, Sealed Source Maximum Activities: 13 curies (shipped) 10 curies (installed)	

<b>Site Name and Address</b> Divine Providence Hospital 1100 Grampian Blvd Williamsport, PA 17740	<b>Site Contact Person</b> Name John V. Calce, MD Telephone 570-321-2400
<b>Licensee/Site Identification Number</b> NA 000151	<b>Dates on which work will performed</b> From 1/10/00 to 1/11/00

**Comments** RTS # 388 On site to repair unit. Unresettable error code 78 - - WIRE NOT PARKED, TREAT NOT ENABLED

I hereby certify that the above information is true and complete.

Signed  Date 1/7/00

  
 Signature M. C. Hernandez  
 Title: Radiation Specialist  
 Date: 1/7/00

copy to RN