

Notice of Proposed Reciprocity Activity

JAN 26 2000

| | |
|---|--|
| Licensee Name Varian Medical Systems 911 Hansen Way, MS C-080 Palo Alto, CA 94304 | Licensee Contact Tony Sorensen, Radiation Safety Officer 650-424-6662 650-855-7375 (fax) |
| License Information State of California, Department of Health Services # 1025-43 Expiration Date: 11/25/97 (timely renewal) | Person Authorized to Perform Activity Richard Nusspickel Contact #: 800-864-1672 |

| | |
|---|---|
| Description of Device Varian Model VariSource High Dose Rate Remote Afterloader (CA661D103S) | Activity to be Performed Non Routine Service (see comments) |
| Description of Source Omnitron International Model SL-777/777V (LA0760S102S) Iridium-192, Special Form, Sealed Source Maximum Activities: 13 curies (shipped) 10 curies (installed) | |

| | |
|--|--|
| Site Name and Address Divine Providence Hospital 1100 Grampian Blvd Williamsport, PA 17740 | Site Contact Person Name John V. Calce, MD Telephone 570-321-2400 |
| Licensee/Site Identification Number N/A | Dates on which work will performed From 1/31/00 to 2/4/00 |

Comments RTS # 388 On site to repair unit. Unresettable error code 78 -- WIRE NOT PARKED, TREAT NOT ENABLED

I hereby certify that the above information is true and complete.

Signed _____ Date _____

M. C. Hernandez
Signature M. C. Hernandez
Title: Radiation Specialist
1/26/00

