

package received at the hospital; failure to perform tests of the radiation dose calibrator and to measure the constituents of a radioactive pharmaceutical; and failure to provide annual refresher training.

The hospital has until June 13 to pay the fine or to protest it. If the fine is protested and subsequently imposed by the NRC staff, the hospital may request a hearing.

#