

30-323



PECO PowerLabsSM
Testing and Solutions, Guaranteed.

PECO Energy Company
175 North Caln Road
Coatesville, PA 19320

June 20, 2000

Mr. George Tasick
James A. Fitzpatrick Nuclear Plant
268 Lake Road
Lycoming, NY 13093

Subject: Valve Operation Test & Evaluation System (VOTES)
Re-calibrations

Reference: Letter - G.C. Bell (PECO PowerLabs) to George Tasick (IP3),
dated 5/26/2000

Dear Mr. Tasick:

The Votes clamps listed below, as originally identified in the referenced letter, were re-calibrated and determined to be in tolerance. Two clamps (JAF E-299 and JAF E-147) were found to have exceeded their lifetime change limits. This does not affect accuracy, but is an indicator for possible rework or replacement by the manufacturer.

- JAF E-181 Liberty Technology Model No. DCC-100
- JAF E-244 Liberty Technology Model No. MUC-100
- JAF E-221 Liberty Technology Model No. DCC-100
- JAF E-300 Liberty Technology Model No. DCC-350
- JAF E-132 Liberty Technology Model No. CCL-100
- JAF E-148 Liberty Technology Model No. MCC-100
- JAF E-219 Liberty Technology Model No. CCL-100
- JAF E-299 Liberty Technology Model No. DCC-350
- JAF E-360 Liberty Technology Model No. EMC-125
- JAF E-180 Liberty Technology Model No. DCC-100
- JAF E-182 Liberty Technology Model No. MUC-100
- JAF E- 1058 Liberty Technology Model No. VVC-100
- JAF E- 129 Liberty Technology Model No. UCL-100
- JAF E-130 Liberty Technology Model No. UCL-100
- JAF E-232 Liberty Technology Model No. UCL-100
- JAF E-134 Liberty Technology Model No. MCC-100
- JAF E-1018 Liberty Technology Model No. CCL-100
- JAF E-133 Liberty Technology Model No. SCL-100
- JAF E-297 Liberty Technology Model No. ECC-225
- JAF E-2023 Liberty Technology Model No. CCL-100
- JAF E-2025 Liberty Technology Model No. EMC-125
- JAF E-147 Liberty Technology Model No. MUC-100
- JAF E-296 Liberty Technology Model No. ECC-225

Acc

Mr. George Tasick
June 20, 2000
Page 2

We regret any inconvenience the event, as described in the referenced letter, may have had on your operation. We believe that the actions we have taken in response to the event will serve to prevent its recurrence.

Sincerely,



George C. Bell
Director

Copy to:

H. J. Miller
USNRC

USNRC
Document Control Desk