

## Application and Review Checklist for Acceptance Review for SSD 00-013

<b>SUMMARY DATA</b>									
<b>Name and Complete Mailing Address of the Applicant:</b> DraxImage, Inc. 16751 Trans Canada Highway Kirkland, QC 9H9 4J4	<b>Name, Title, and Telephone Number of the Individual to Be Contacted If Additional Information or Clarification Is Needed by the NRC:</b>  Dr. R. J. Flanagan, Exe. Vice President (514) 694-9295, rflanagan@draximage.com								
<b>The Applicant is (check one):</b> <table border="1" style="width: 100%; border-collapse: collapse; margin-top: 5px;"> <tr> <td style="width: 20px; text-align: center;"><input type="checkbox"/></td> <td>Custom User</td> </tr> <tr> <td style="text-align: center;"><input type="checkbox"/></td> <td>Manufacturer</td> </tr> <tr> <td style="text-align: center;"><input type="checkbox"/></td> <td>Distributor</td> </tr> <tr> <td style="text-align: center;"><input checked="" type="checkbox"/></td> <td>Manufacturer and Distributor</td> </tr> </table>	<input type="checkbox"/>	Custom User	<input type="checkbox"/>	Manufacturer	<input type="checkbox"/>	Distributor	<input checked="" type="checkbox"/>	Manufacturer and Distributor	<b>If the Applicant Is Not the Manufacturer, Provide the Name and Complete Mailing Address of the Manufacturer:</b>
<input type="checkbox"/>	Custom User								
<input type="checkbox"/>	Manufacturer								
<input type="checkbox"/>	Distributor								
<input checked="" type="checkbox"/>	Manufacturer and Distributor								
<b>If the Applicant Is a Custom User, Provide the Name and Complete Mailing Address of the Distributor:</b>	<b>Provide the Name, Complete Mailing Address, and Function of Other Companies Involved:</b> Isogenic Science Ltd 16751 Trans Canada Highway Kirkland, QC H9H 4J4 Product Development								
<b>Model Number: LS-1</b>	<b>Principal Use Code (see Appendix E): (V) General Medical use</b>								
<b>Name Used by the Industry to Identify the Product (e.g., Radiography Exposure Device, Teletherapy Source, Calibration Source, etc.):</b>  <b>Brachytherapy Source</b>	<b>For Use by:</b> <table border="1" style="width: 100%; border-collapse: collapse; margin-top: 5px;"> <tr> <td style="width: 20px; text-align: center;"><input checked="" type="checkbox"/></td> <td>Specific Licensees Only</td> </tr> <tr> <td style="text-align: center;"><input type="checkbox"/></td> <td>General Licensees Only</td> </tr> <tr> <td style="text-align: center;"><input type="checkbox"/></td> <td>Both Specific and General Licensees</td> </tr> <tr> <td style="text-align: center;"><input type="checkbox"/></td> <td>Persons Exempt from Licensing</td> </tr> </table>	<input checked="" type="checkbox"/>	Specific Licensees Only	<input type="checkbox"/>	General Licensees Only	<input type="checkbox"/>	Both Specific and General Licensees	<input type="checkbox"/>	Persons Exempt from Licensing
<input checked="" type="checkbox"/>	Specific Licensees Only								
<input type="checkbox"/>	General Licensees Only								
<input type="checkbox"/>	Both Specific and General Licensees								
<input type="checkbox"/>	Persons Exempt from Licensing								
<b>Leak-Test Frequency:</b> <table border="1" style="width: 100%; border-collapse: collapse; margin-top: 5px;"> <tr> <td style="width: 20px; text-align: center;"><input type="checkbox"/></td> <td>Periodic Leak-Testing is Not Required</td> </tr> <tr> <td style="text-align: center;"><input checked="" type="checkbox"/></td> <td>6 Months</td> </tr> <tr> <td style="text-align: center;"><input type="checkbox"/></td> <td>Attached is justification for a leak test frequency of greater than 6 months</td> </tr> </table>	<input type="checkbox"/>	Periodic Leak-Testing is Not Required	<input checked="" type="checkbox"/>	6 Months	<input type="checkbox"/>	Attached is justification for a leak test frequency of greater than 6 months	<b>Principal Section of the 10 CFR that Applies to the User (e.g., General Licensees under 10 CFR 31.5): 10 CFR 32.74, 10 CFR 35</b>  <b>Radionuclides and Maximum Activities (including loading tolerance):</b> I-125, 75 mCi		
<input type="checkbox"/>	Periodic Leak-Testing is Not Required								
<input checked="" type="checkbox"/>	6 Months								
<input type="checkbox"/>	Attached is justification for a leak test frequency of greater than 6 months								
<b>CERTIFICATION:</b>  THE APPLICANT UNDERSTANDS THAT ALL STATEMENTS AND REPRESENTATIONS MADE IN THIS APPLICATION ARE BINDING UPON THE APPLICANT.  THE APPLICANT AND ANY OFFICIAL EXECUTING THIS CERTIFICATION ON BEHALF OF THE APPLICANT, NAMED IN ITEM 2, CERTIFY THAT THIS APPLICATION IS PREPARED IN CONFORMITY WITH TITLE 10, CODE OF FEDERAL REGULATIONS, PARTS 30 AND 32 AND THAT ALL INFORMATION CONTAINED HEREIN IS TRUE AND CORRECT TO THE BEST OF THEIR KNOWLEDGE AND BELIEF.  WARNING: 18 U.S.C. SECTION 1001 ACT OF JUNE 25, 1948 62 STAT. 749 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.									
<b>Certifying Officer — Typed Name and Title</b>									
<b>Signature:</b>	<b>Date:</b>								

**CHECKLIST**

**Registration Certificate Holder:**

**Model:**

DESCRIPTION	OK/DEF	COMMENTS
DESCRIPTION/CONSTRUCTION		
If registration certificate holder is requesting to register more than one source/device on a certificate, are designs similar enough to do so?	N/A	
Device/source design with complete engineering drawings (dimensions, tolerances, list of materials)	✓	
Assembly methods (screw, welds, etc.); verify integrity	✓	
Source mounting (size and integrity) and security	N/A	
Is source classification sufficient (ANSI N43.6 or ISO 2919)?		Brachytherapy source ISO 2919
Radiography - Unprotected ..... 43515	53211	
Radiography - In Device ..... 43313		
Medical - Radiography ..... 32312		
Medical - $\gamma$ Teletherapy ..... 53524		
Medical - Brachytherapy ..... 53211		
Medical - Source Applicators ..... 43312		
$\gamma$ Gauges - Unprotected ..... 43333		
$\gamma$ Gauges - In Device ..... 43232		
$\beta$ Gauges, Low Energy $\gamma$ Gauges, or X-ray fluorescence ..... 33222		
Oil Well Logging ..... 56522		
Portable Moist/Density ..... 43333		
Neutron Applications ..... 43323		
Calibration source activity > 30 $\mu$ Ci (1 MBq) ..... 22212		
$\gamma$ Irradiators (I) ..... 43323		
$\gamma$ Irradiators (II, III) ..... 43424		
$\gamma$ Irradiators (II, III, IV) ..... 53424		
Chromatography ..... 32211		
Static Eliminators ..... 22222		
Smoke Detectors ..... 32222		
Definition of shutter operation (locked in Off position, not locked in On position), Fail safe, spacing and tolerances	N/A	
On-Off indicators (description, qty., location)	N/A	
Safety interlocks, guards, etc. to prevent access to beam or high radiation levels	N/A	
Corrosion between unlike materials (e.g., aluminum & steel, depleted uranium & steel, etc.)	✓	Titanium
Shielding efficiency and integrity	N/A	
For medical devices: Was a 510(k) provided? (provide written notification to FDA)	✓	in application
Well logging sources must be nondispersible and nonsoluble. (see App. B for a list of approved well logging sources as of Nov. 1991)	N/A	
See "ANSI and Other Standards" list for references for particular source/device designs (e.g. radiography, Brachytherapy, etc.)	✓	ISO 2919 ANSI 43.6-1997

## CHECKLIST

**Registration Certificate Holder:**

**Model:**

DESCRIPTION	OK/DEF	COMMENTS
<b>LABELING</b>		
Copy of label	✓	
Materials, dimensions, colors (note on registration certificate if labeling is exempt from the color requirements of 10 CFR Part 20)	✓	
Permanent attachment and location(s) - visible to users?	✓	shipping pot, container
Contents: Model#, Serial#, Isotope, Activity, Manufacturer, Date of Assay, Trefoil, "CAUTION - RADIOACTIVE MATERIAL" (Depleted Uranium information must be included)	✓	
<b>CONDITIONS OF USE</b>		
Expected working life of the source/device (years, operations)	1 yr	
Actions to be taken when product reaches end of its working life.	✓	
Maximum allowable temperature, vibration, shock, corrosion, etc. (during use, handling, storage, and transport)	✓	
How the device will be used	N/A	
Meets dose limits of Part 32 for distribution general licensees or persons exempt from licensing	N/A	
<b>PROTOTYPE TESTING/HISTORICAL USE</b>		
Tests methods and conditions (for source and device)	✓	
Tests results	✓	
Years of use (incidents, failures, etc.)	N/A	
Similarities to other sources/devices if they are used as basis.	N/A	
<b>RADIATION PROFILES</b>		
Survey instrument used (type, window thickness, sensitivity, etc.)	✓	see p13
Conditions: including environments, scatter (product in beam), and use of guards and shields	✓	
Distance from source/surface (per ANSI 538-1979)	✓	
Shutter Open and Closed/Source Shielded	N/A	
Verify radiation surveys for $\gamma$ radiation meet $inv^2$ law.	N/A	
Verify radiation surveys for non- $\gamma$ radiation have not been calculated using $inv^2$ law.	N/A	

---

---

**CHECKLIST**

---

---

**Registration Certificate Holder:****Model:**

<b>DESCRIPTION</b>	<b>OK/DEF</b>	<b>COMMENTS</b>
<b>QUALITY ASSURANCE</b>		
Materials, subassemblies, services	✓	
Assembly methods (screws, welding, etc.)	✓	
Dimensions and tolerances	✓	
Activity, radiation levels, leak tests	✓	
QA Manual and comparison of manual to Regulatory Guide 6.9		
<b>INSTALLATION</b>		
Fixed, portable, movable, fixed installation but portable source housing	N/A	
Inherent shielding, inaccessibility	N/A	
Beam access: size of air gap/opening to beam and use of interlocks, locks, additional shielding or barriers	N/A	
Mounting integrity	N/A	
<b>SAFETY INSTRUCTIONS</b>		
Operation, maintenance, calibration, damage/failure, specific warnings, leak test, and radiation surveys	✓	
<b>ACCOMPANYING DOCUMENTATION</b>		
Leak tests results and radiation surveys	✓	
Transportation documents	✓	
Operation, maintenance, calibration, damage/failure, specific warnings, leak test, and radiation survey instructions if applicable	✓	
For Distribution to General Licensees: Verify NRC Regions and Agreement State listing is up-to-date and copies of all pertinent regulations	N/A	

**CHECKLIST**

**Registration Certificate Holder:**

**Model:**

DESCRIPTION				OK/DEF	COMMENTS
SERVICING					
The following activities may be performed by the persons indicated:					
Activity	by a General Licensee	Only by a Specific Licensee	Will be Offered by the Applicant		
Installation		✓			
Relocation		✓			
Maintenance					
Repair					
Source Exchange					
Calibration					
Leak Testing		✓			
Radiation Survey		✓			
Training		✓			
FOREIGN VENDORS					
Drop ship					
Who and where is source installed					Hospital
Leak test and radiation surveys					Hospital
QA in the U.S.					Hospital

**Signature:**     /RA/

**Date:** 03/01/00