



September 21, 2000  
NMP1L 1537

U. S. Nuclear Regulatory Commission  
Attn: Document Control Desk  
Washington, DC 20555

RE: Nine Mile Point Unit 1  
Docket No. 50-220  
        DPR-63        

***Subject: Special Report***

Gentlemen:

In accordance with Nine Mile Point Unit 1 Technical Specification Table 3.6.11-1, "Accident Monitoring Instrumentation," Action Statement 4, Niagara Mohawk Power Corporation is submitting the following Special Report concerning the inoperability of the #11 Containment Hydrogen Monitoring System.

**Description of Event**

On September 11, 2000, at 1050, operators declared the #11 Containment Hydrogen Monitoring System inoperable for scheduled maintenance. The scheduled maintenance was to perform Preventive Maintenance Procedure N1-IPM-201-008, "H2O2 Analyzers Support Instrumentation Calibration." The procedure was completed and the Control Room operators declared the #11 Containment Hydrogen Monitoring System operable on September 16, 2000.

The #12 Containment Hydrogen Monitoring System was in service during the time that the #11 Containment Hydrogen Monitoring System was inoperable.

**Cause of Event**

The #11 Containment Hydrogen Monitoring System was declared inoperable in order to perform instrumentation calibration.

JE22

**Corrective Actions**

The preventive maintenance was completed and the Control Room operators declared the #11 Containment Hydrogen Monitoring System operable on September 16, 2000.

Very truly yours,



Lawrence A. Hopkins  
Plant Manager - NMP1

LAH/KLE/jlb

xc: Mr. H. J. Miller, Regional Administrator, Region I  
Mr. G. K. Hunegs, NRC Senior Resident Inspector  
Records Management