

ATLANTIC RICHFIELD

HARVEY, JIM

7-7-70

Met with Dr. Snow

[REDACTED] 1-1-60 to 3-31-70 for 100 men. Ex. 6

[REDACTED] manual

[REDACTED] for quarter 20 men.

Ex. 6

20 bridge now reserved, not all assigned

R.S. behavior falls bridge, bin-walking & 13 water

AFC-5 height

Committee

now only three Committee members

Snow, Billings & GNL

All questions are thru Dr. Snow.

All requests to returning are brought to office of Snow.

George Macdonald, Mgr of Process Plant in Snow's signature

All engine test work has been discontinued -

Most of the work without participation of [REDACTED] Harvey and

BJA

(2)

transport to field locations where experimental work is done
by the three committee personnel.

The 1963 year the large preparation of H^3 compound
& unincorporation was made.

H^3 compounds & their solutions all are stored in hood which
remains at all times.

N - building

Co^{60} work is at a standstill at present but facility
is still available for use.

3 - Se^{75} sources were shipped on 9-23-69 30mc
to Atomic Disposal Co. - No other waste shipped since last
inspection.

1st leak ~~test~~ 6-24-70 all $< 2 \times 10^{-5}$ μ ci.

2 - Co^{137} sealed sources were (300mc ea.)

1 - Cs^{137} " " 10mc

Leak test 3 times each year

Also 3 times a year smear survey of lab & lab shows
 $\approx 5 \times 10^{-5}$ μ ci activity