

**DEPARTMENT OF HEALTH SERVICES**  
**RADIOLOGIC HEALTH BRANCH**  
P.O. BOX 942732, MS-178  
SACRAMENTO, CA 94234-7320  
(916) 327-6214



February 22, 2001

John Harrison  
140 Sutherland Dr  
Auburn, CA 95603

Dear Mr. Harrison:

Enclosed you will find a corrected page for your sealed source and device registry certificate, number CA0414D102S. Please look it over carefully for errors and/or omissions as it will be distributed nationwide to all NRC Regional Offices and Agreement States, and used for determination of licensing requirements.

If you have any questions, please call me at (916) 327-6214.

Sincerely,

A handwritten signature in black ink, appearing to read "Xiaosong Yin".

Xiaosong Yin  
Associate Health Physicist  
Radiologic Health Branch

Nm 5512

REGISTRY OF RADIOACTIVE SEALED SOURCES AND DEVICES  
SAFETY EVALUATION OF DEVICE.  
(CORRECTED PAGE 1-FEBRUARY 22, 2001)

NO.: CA0414D102S

DATE: June 28, 1999

PAGE: 1 of 5

DEVICE TYPE: Sealed Source Holder

MODEL: 0112 and 0122

MANUFACTURER/DISTRIBUTOR:

John Harrison  
(Formerly Kevex & Fison Instruments)  
140 Sutherland Dr.  
Auburn, CA 95603  
(916) 782-8899 (voice)

SEALED SOURCE MODEL DESIGNATION: Various

ISOTOPE:

MAXIMUM ACTIVITY:

Iron 55  
Americium 241

100 millicuries (3.7 GBq)  
200 millicuries (7.4 GBq)

LEAK TEST FREQUENCY: Six (6) months

PRINCIPAL USE: (U) X-Ray Fluorescence

CUSTOM SOURCE: \_\_\_\_\_ YES  NO