

**FOUR SEASONS NUCLEAR
MEDICINE**

Nadeem G. Qazi, M.D.
Nuclear Medicine

Naeem A. Qazi, M.D.
Cardiology

Four seasons nuclear medicine Inc.
1279-Stadium Drive

TEL: (304)-323-2588
FAX: (304)-323-2589

FAX FACE SHEET

Person Sending: Nadeem Qazi

Person Receiving: Bryan Parker

Of Pages (Not Including Cover Sheet) 1

Date/Time: 4/3/01 12:00 PM

259405

Four Seasons Nuclear Medicine, Inc.
1279 Stadium Drive
Bluefield, WV 24701
(304) 323-2588

Nuclear Regulatory Commission
Region II
61 Forsyth Street, S.W. Suite 23T85
Atlanta, GA 30303

Attn: Bryan Parker


As a follow up to Dr. Anthony telephone conversation with you on this date, we wish to amend our license.

To provide for the ownership, possession and use of Gd-153 sources as attenuation correction devices for our Nuclear Medicine camera. We plan to use sources obtained from North American Scientific or isotopes laboratories as source model numbers. NAS: MED3601 or IPL:A3410 or equivalent. We anticipate that the activity per source will not exceed 300 millicuries. Since the attenuation device requires 2 sources, we request possession limit of 1 curie to allow for source exchange.

These sealed sources will either be exchanged every 6 months, or will have leak test performed every 6 months.

We would appreciate your expedited review of this amendment request, in order that we may begin using this attenuation device as soon as possible.

Sincerely,

Nadeem G. , M.D.
Radiation Safety Officer

Received in Region II NE 5
April 3, 2001 cc: _____

259405