

## CONVERSATION RECORD

**TYPE:**

Outgoing Telephone x  
Incoming Telephone  
Meeting

**NAME OF PERSON CONTACTED:**

Ann Warbick Cerone

**ORGANIZATION:**

MDS Nordion

**TIME:**

2:30pm

**DATE:**

8/11/99

**SUBJECT:**

Request for additional information, Application No. 99-44

**SUMMARY:**

I called Ms. Cerone, after completing the acceptance review of their application, dated July 2, 1999. Douglas Beatty/Theratronics was in her office and was also on the phone. I asked for the following information not contained in the application:

1. Demonstration that the specimen which had been subjected to the leach test (dated 1993) are identical to those which are manufactured presently;
2. Information on Nordion's QA program, e.g. program summary, ISO certification, list of procedures;
3. Differences between ISO 2919-1998 and ANSI 43.10-1997.

She said that she will mail the information shortly.

**ACTION REQUIRED:**


NRC to conduct safety evaluation when information arrives.

**PLACE THIS RECORD IN:**

Registration File NR-  
QA File  
Incident File  
General File 99-44

**PERSON DOCUMENTING THE CONVERSATION:**

John Jankovich

**SIGNATURE:****DATE:**

8/11/99