

Region I  
NMSS Licensee Event Report

License No. 07-28762-02G

Docket No. 03002988

MLER-RI 2001-038

LICENSEE: AGILENT TECHNOLOGIES

EVENT DESCRIPTION NOTICE OF GENERAL LICENSE DEVICES W/HIGH REMOVABLE ACTIVITY

EVENT DATE \_\_\_\_\_ REPORT DATE 7/5/2001

1. REPORTING REQUIREMENT

- |                                                              |                                                         |
|--------------------------------------------------------------|---------------------------------------------------------|
| <input type="checkbox"/> 10 CFR 20.2201 Theft or Loss        | <input type="checkbox"/> 10 CFR 35.33 Misadministration |
| <input type="checkbox"/> 10 CFR 20.2203 30 Day Report        | <input type="checkbox"/> License Condition              |
| <input type="checkbox"/> 10 CFR 30.50 Report                 |                                                         |
| <input checked="" type="checkbox"/> Other <u>10 CFR 31.5</u> |                                                         |

2. REGION I RESPONSE

- |                                                                       |                                       |
|-----------------------------------------------------------------------|---------------------------------------|
| <input type="checkbox"/> Immediate Site Inspection                    | Inspector/Date _____                  |
| <input type="checkbox"/> Special Inspection                           | Inspector/Date _____                  |
| <input type="checkbox"/> Telephone Inquiry                            | Inspector/Date _____                  |
| <input type="checkbox"/> Preliminary Notification                     | <input type="checkbox"/> Daily Report |
| <input checked="" type="checkbox"/> Information Entered on the RI Log |                                       |
| <input checked="" type="checkbox"/> Review at Next Routine Inspection |                                       |
| <input type="checkbox"/> Report Referred to _____                     |                                       |

3. REPORT EVALUATION

- |                                                            |                                                                                  |
|------------------------------------------------------------|----------------------------------------------------------------------------------|
| <input checked="" type="checkbox"/> Description of Event   | <input type="checkbox"/> Corrective Actions                                      |
| <input checked="" type="checkbox"/> Levels of RAM Involved | <input type="checkbox"/> Calculation Adequate                                    |
| <input type="checkbox"/> Cause of Event                    | <input type="checkbox"/> Letter to Licensee Requesting<br>Additional Information |

4. SPECIAL INSTRUCTIONS OR COMMENTS

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

Completed By: Butz Ullrich Date 7/24/01  
Reviewed By: F. Henderson for JDK Date 7/24/01

RECEIVED  
REGION I

2001 JUL -5 AM 10: 24

 **Agilent Technologies**  
Innovating the HP Way  
2850 Centerville Road  
Wilmington, DE 19808

Division of Nuclear Materials Safety  
U.S. Nuclear Regulatory Commission  
Region I  
475 Allendale Road,  
King of Prussia, PA 19406

Re: Notice of General License devices with high removable activity (NRC License 07-28762-02G)

June 28, 2001

Dear Sir or Madam,

Per the requirements of 10 CFR 31.5, I am notifying your office of general license devices from which we obtained removable activity wipes in excess of 0.005  $\mu\text{Ci}$ . The cells were new cells from one of our plating sources, AEA Technology. Engineering evaluation revealed no visible reason for the high wipes. The cells are being returned to AEA with notification of the incident.

We have noticed higher incoming wipes on AEA parts recently and our manufacturing support team is working in concert with our parts buying group and AEA in an effort to resolve the situation.

<u>Serial #</u>	<u>Model #</u>	<u>Activity (<math>\mu\text{Ci}</math>)</u>
HI956	19233	0.0097
HI957	19233	0.0056

Please contact me at 302-633-8262 if there are any questions.

Thank you,



David S. Bennett  
Radiation Safety Officer