

NRC FORM 314 (6 B) 10 CFR 30.26 (a)(1)(iv) 10 CFR 40.42 (a)(1)(iv) 10 CFR 70.38 (a)(1)(iv)	U.S. NUCLEAR REGULATORY COMMISSION <h2 style="margin: 0;">CERTIFICATE OF DISPOSITION OF MATERIALS</h2>	APPROVED BY OMB: 3150-0028 EXPIRES: 4/30/92 ESTIMATED BURDEN PER RESPONSE TO COMPLY WITH THIS INFORMATION COLLECTION REQUEST: 5 HRS. FORWARD COMMENTS REGARDING BURDEN ESTIMATE TO THE RECORDS AND REPORTS MANAGEMENT BRANCH (P-630), U.S. NUCLEAR REGULATORY COMMISSION, WASHINGTON, DC 20555, AND TO THE PAPERWORK REDUCTION PROJECT (3150-0028), OFFICE OF MANAGEMENT AND BUDGET, WASHINGTON, DC 20503
INSTRUCTIONS: SEND THE COMPLETED CERTIFICATE TO THE NRC OFFICE SPECIFIED ON THE REVERSE. (All items MUST be completed - print or type)		

LICENSEE NAME AND ADDRESS <p style="text-align: center;">Supelco Supelco Park Bellefonte, PA 16823</p>	LICENSE NUMBER LICENSE EXPIRATION DATE
---	---

THE LICENSEE OR ANY INDIVIDUAL EXECUTING THIS CERTIFICATE ON BEHALF OF THE LICENSEE CERTIFIES THAT:
(Check and/or complete the appropriate item(s) below.)
Wayne Fishburn

A. MATERIALS DATA *(Check one and complete as necessary)*

1. NO MATERIALS HAVE EVER BEEN PROCURED OR POSSESSED BY THE LICENSEE UNDER THIS LICENSE.

General License

2. ALL MATERIALS PROCURED AND/OR POSSESSED BY THE LICENSEE UNDER THE LICENSE NUMBER CITED ABOVE HAVE BEEN DISPOSED OF IN THE FOLLOWING MANNER. *(If additional space is needed, use the reverse side or provide attachments.)*

Describe specific material transfer actions and, if there were radioactive wastes generated in terminating this license, the disposal actions, including the disposition of low-level radioactive waste, mixed waste, Greater-than-Class-C waste, and sealed sources, if applicable.

ECD Model #: 02-001972-00 ECD S/N: A2618
 Manufacturer: Varian N163 15mCi Max.

For transfers, specify the date of the transfer, the name of the licensed recipient, and the recipient's NRC license number or Agreement State name and license number.

9/8/94 C.J. Bruyn & Co.
 License # 5670-07

If materials were disposed of directly by the licensee rather than transferred to another licensee, licensed disposal site or waste contractor, describe the specific disposal procedures (e.g., decay in storage).

Allied Ecology Inc.
 7066 Commerce Circle
 Pleasanton, CA. 94588 (510) 463-9280

B. OTHER DATA

1. OUR LICENSE HAS NOT YET EXPIRED; PLEASE TERMINATE IT.

2. WAS A RADIATION SURVEY CONDUCTED TO CONFIRM THE ABSENCE OF LICENSED RADIOACTIVE MATERIALS AND TO DETERMINE WHETHER ANY CONTAMINATION REMAINS ON THE PREMISES COVERED BY THE LICENSE? *(Check one)*

NO *(Attach explanation)*

YES, THE RESULTS *(Check one)*

ARE ATTACHED, or

WERE FORWARDED TO NRC ON *(Date)*

3. THE PERSON TO BE CONTACTED REGARDING THE INFORMATION PROVIDED ON THIS FORM

NAME C.J. Bruyn & Co., Safety Administrator	TELEPHONE NUMBER (800) 252-7896
--	---

4. MAIL ALL FUTURE CORRESPONDENCE REGARDING THIS LICENSE TO

C.J. Bruyn & Co.
 156 Belle Ave.
 Pleasant Hill, CA. 94523

CERTIFYING OFFICIAL

I CERTIFY UNDER PENALTY OF PERJURY THAT THE FOREGOING IS TRUE AND CORRECT.

SIGNATURE 	DATE 9/20/94
PRINTED NAME AND TITLE C.J. Bruyn, Safety Administrator	

WARNING: FALSE STATEMENTS IN THIS CERTIFICATE MAY BE SUBJECT TO CIVIL AND/OR CRIMINAL PENALTIES. NRC REGULATIONS REQUIRE THAT SUBMISSIONS TO THE NRC BE COMPLETE AND ACCURATE IN ALL MATERIAL RESPECTS. 18 U.S.C. SECTION 1001 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.

PLEASE FILE THIS DOCUMENT CAREFULLY

This and all other survey records must be displayed on request of a Radiologic Health Inspector.

Failure to do so may result in a citation, suspension of your license, and/or fine.

CERTIFICATE OF LEAK TEST

With this document C.J. Bruyn & Co. certifies that the Electron Capture Detector (ECD) indicated below has been surveyed.

*The amount of removable activity was at least 50 times less than
the maximum allowable
.005 micro. Curie.*


Date: 9/14/1994 Manufacturer: Varian Serial No: A2618

Isotope: <15 milliCurie - Nickel 63. ECD Model Nr: 02-001972-00

NOTE: This is to remind you that the Radiation Authorities require that the Ni 63 ECD must be leak tested every *36 months, or in accordance with your license, whatever is more frequent.

*The Nebraska Authorities require a survey every 6 months

C.J. Bruyn & Co. can, on request mail you wipe-kits.



C.J. Bruyn

Environmental Safety administrator.

156 Belle Ave.

Pleasant Hill, CA. 94523

Phone: 800-252-7896 Fax: 510-256-9503