

**From:** ~~Doddcv@~~  
**To:** Ian Barnes  
**Date:** Tue, Jul 25, 2000 10:51 AM  
**Subject:** Phase shifts

Exp  
[Redacted]  
[Redacted]  
Exp

Ian:  
The values for the different notches with the 100% set to 40 are:

	300 kHz	400 kHz
100	40	40
60	27	23
40	23	19
20	15	11

These can vary a degree or so. A number of calibrations on different standards should be done to average out the results. Talk to you later.  
Caius

**CC:** <WLS@nrc.gov>

Information in this record was deleted  
in accordance with the Freedom of Information  
Act, exemptions  
FOIA- 2001-0226

6/47