

Region I
NMSS Licensee Event Report

License No. 07-28762-02
G

Docket No. 03002988

MLER-RI 2002-008

LICENSEE: AGILENT

EVENT DESCRIPTION NOTICE OF GENERAL LICENSE DEVICES WITH HIGH REMOVABLE ACTIVITY

EVENT DATE 2/2002

REPORT DATE 2/15/2002

1. REPORTING REQUIREMENT
- 10 CFR 20.2201 Theft or Loss
 - 10 CFR 20.2203 30 Day Report
 - 10 CFR 30.50 Report
 - Other 10 CFR 31.5
 - 10 CFR 35.33 Misadministration
 - License Condition

2. REGION I RESPONSE
- Immediate Site Inspection
 - Special Inspection
 - Telephone Inquiry
 - Preliminary Notification
 - Information Entered on the RI Log
 - Review at Next Routine Inspection
 - Report Referred to _____
 - Daily Report
 - Inspector/Date _____
 - Inspector/Date _____
 - Inspector/Date _____

3. REPORT EVALUATION
- Description of Event
 - Levels of RAM Involved
 - Cause of Event
 - Corrective Actions
 - Calculation Adequate
 - Letter to Licensee Requesting Additional Information

4. SPECIAL INSTRUCTIONS OR COMMENTS

Completed By: Judith A. Jouston Date 2/20/02

Reviewed By: J. Jouston Date 2/20/02

 **Agilent Technologies**
2850 Centerville Road
Wilmington, DE 19808

RECEIVED
REGION 1

2002 FEB 15 PM 3: 24

Division of Nuclear Materials Safety
U.S. Nuclear Regulatory Commission
Region I
475 Allendale Road,
King of Prussia, PA 19406

Re: Notice of General License devices with high removable activity (NRC License 07-28762-02G)

February 14, 2002

Dear Sir or Madam,

Per the requirements of 10 CFR 31.5, I am notifying your office of a general license device from which we have obtained removable activity wipes in excess of 0.005 μCi . The cell was returned from an overseas customer. The cells were evaluated by support engineering and its disposition is identified below. I will notify the cell supplier of the high wipe.

<u>Serial #</u>	<u>Model #</u>	<u>Activity (μCi)</u>	<u>Disposition</u>
F3927	19233	0.008	Scrapped

Please contact me at 302-633-8262 if there are any questions.

Thank you,


David S. Bennett
Radiation Safety Officer