

Region I
NMSS Licensee Event Report

License No. 07-28762-02G

Docket No. 03002988

MLER-RI 2002-010

LICENSEE: AGILENT TECHNOLOGIES

EVENT DESCRIPTION NOTICE OF GENERAL LICENSE DEVICES WITH HIGH REMOVABLE ACTIVITY

EVENT DATE 03/02 REPORT DATE 3-4-02

1. REPORTING REQUIREMENT

- | | |
|--|---|
| <input type="checkbox"/> 10 CFR 20.2201 Theft or Loss | <input type="checkbox"/> 10 CFR 35.33 Misadministration |
| <input type="checkbox"/> 10 CFR 20.2203 30 Day Report | <input type="checkbox"/> License Condition |
| <input type="checkbox"/> 10 CFR 30.50 Report | |
| <input checked="" type="checkbox"/> Other <u>10 CFR 31.5</u> | |

2. REGION I RESPONSE

- | | |
|---|---------------------------------------|
| <input type="checkbox"/> Immediate Site Inspection | Inspector/Date _____ |
| <input type="checkbox"/> Special Inspection | Inspector/Date _____ |
| <input type="checkbox"/> Telephone Inquiry | Inspector/Date _____ |
| <input type="checkbox"/> Preliminary Notification | <input type="checkbox"/> Daily Report |
| <input checked="" type="checkbox"/> Information Entered on the RI Log | |
| <input checked="" type="checkbox"/> Review at Next Routine Inspection | |
| <input type="checkbox"/> Report Referred to _____ | |


3. REPORT EVALUATION

- | | |
|--|---|
| <input checked="" type="checkbox"/> Description of Event | <input checked="" type="checkbox"/> Corrective Actions |
| <input checked="" type="checkbox"/> Levels of RAM Involved | <input type="checkbox"/> Calculation Adequate |
| <input type="checkbox"/> Cause of Event | <input type="checkbox"/> Letter to Licensee Requesting Additional Information |

4. SPECIAL INSTRUCTIONS OR COMMENTS

Completed By: Betsy Williams Date 3/5/02

Reviewed By: John D. [Signature] Date 3.5.2002

 **Agilent Technologies**
2850 Centerville Road
Wilmington, DE 19808

RECEIVED
REGION 1

2002 MAR -4 PM 2: 49

Division of Nuclear Materials Safety
U.S. Nuclear Regulatory Commission
Region I
475 Allendale Road,
King of Prussia, PA 19406

Re: Notice of General License devices with high removable activity (NRC License 07-28762-02G)

March 1, 2002

Dear Sir or Madam,

Per the requirements of 10 CFR 31.5, I am notifying your office of a general license device from which we have obtained removable activity wipes in excess of 0.005 μCi . The cell was returned from an overseas customer. The cell was evaluated by support engineering and its disposition is identified below. I will notify the cell supplier of the high wipe.

<u>Serial #</u>	<u>Model #</u>	<u>Activity (μCi)</u>	<u>Disposition</u>
L7705	19233	0.016	Scrapped

Please contact me at 302-633-8262 if there are any questions.

Thank you,



David S. Bennett
Radiation Safety Officer