

Region I
NMSS Licensee Event Report

Licensee
Event Description

AGILENT TECHNOLOGIES

NOTICE OF GEN LICENSE DEVICE W/HIGH REMOVABLE ACTIVITY

License # 07-28762-02G

Docket #

03002988

MLER-RI

2002-031

Event Date

Report Date

5-29-02

1. REPORTING REQUIREMENT

☐

10 CFR 20.2201 Theft or Loss

☐

10 CFR 35.33 Misadministration

☐

10 CFR 20.2203 30 Day Report

☐

License Condition

☐

10 CFR 30.50 Report

☒

Other 10 CFR 31.5

2. REGION I RESPONSE

☐

Immediate Site Inspection

Inspector/Date

☐

Special Inspection

Inspector/Date

☐

Telephone Inquiry

Inspector/Date

☐

Preliminary Report

☐

Daily Report

☒

Information Entered in RI Log

☒

Review at next inspection

☐

Report referred to:

3. REPORT EVALUATION

☒

Description of Event

☐

Corrective Actions

☒

Levels of RAM Involved

☐

Calculations Adequate

☐

Cause of Event

☐

Additional Information Requested from Licensee

4. SPECIAL INSTRUCTIONS OR COMMENTS

Public

X BY 6/6/02

Non-Public

Initials/Date

Completed by:

Beth Ullrich

Date

Initials/Date


6/6/02

Reviewed by:

John D. Krum

Date

6/10/2002

 **Agilent Technologies**
2850 Centerville Road
Wilmington, DE 19808

RECEIVED
REGION 1

2002 MAY 31 PM 12:45

Division of Nuclear Materials Safety
U.S. Nuclear Regulatory Commission
Region I
475 Allendale Road,
King of Prussia, PA 19406

Re: Notice of General License device with high removable activity (NRC License 07-28762-02G)

May 29, 2002

Dear Sir or Madam,

Per the requirements of 10 CFR 31.5, I am notifying your office of a general license device returned to our site by customer from which we have obtained a removable activity wipe in excess of 0.005 μCi . The cell was evaluated by support engineering and its disposition is identified below. I will notify the cell supplier of the high wipe.


Serial #	Model #	Activity (μCi)	Disposition
F4118	19233	0.032	Scrapped

Comments:

Cell 10 years old, inside dark and dirty, customer has not performed wipes since receiving the source in March of 2000.

Please contact me at 302-633-8262 if there are any questions.

Thank you,


David S. Bennett
Radiation Safety Officer

NMSS-11