

UNITED STATES
NUCLEAR REGULATORY COMMISSION
OFFICE OF NUCLEAR MATERIAL SAFETY AND SAFEGUARDS
WASHINGTON, D.C. 20555

December 13, 1996

NRC INFORMATION NOTICE 96-66: RECENT MISADMINISTRATIONS CAUSED BY INCORRECT CALIBRATIONS OF STRONTIUM-90 EYE APPLICATORS

Addressees

All U.S. Nuclear Regulatory Commission Medical Use Licensees authorized to use strontium-90 (Sr-90) eye applicators.

Purpose

The U.S. Nuclear Regulatory Commission (NRC) is issuing this information notice to alert addressees to recent misadministrations caused by incorrect source strength determinations of Sr-90 eye applicators. It is expected that recipients will review the information for applicability to their facilities and consider actions, as appropriate, to avoid similar problems. However, suggestions contained in this information notice are not NRC requirements; therefore, no specific action nor written response is required.

Description of Circumstances

The primary causes of two recent events have been problems with calibration and calculation of the dose rates from Sr-90 applicators. As a result, the administered doses were not within 20 percent of the prescribed dose.

Case 1. In October 1995, during an inspection of a licensee authorized to use a Sr-90 eye applicator in Mayagüez, Puerto Rico, it was determined that the calibration record for the licensee's Sr-90 eye applicator was missing. A previous owner of the source lost the original calibration certificate, and his medical physics consultant performed a check of the source strength. This check was performed with an inappropriate measurement instrument and resulted in an incorrect determination of the source strength as 0.24 Gray (Gy)/sec (24 rad/sec). The licensee obtained the source in 1994 and relied on the erroneous source strength determination during treatments. Also, the source had not been decay-corrected by either owner since 1990. NRC urged the licensee to have the eye applicator recalibrated and to be instructed in the proper method to calculate the decay of the source.

The licensee subsequently had the eye applicator recalibrated by the National Institute of Standards and Technology. The resultant recalibration revealed a dose rate of 0.53 Gray (53 rad)/sec rather than the 0.24 Gy (24 rad)/sec used by the licensee. The licensee had prescribed a total dose of between 10 to 25 Gy (1000 to 2500 rad) for approximately

9612090360

IN 96-66
December 13, 1996
Page 2 of 4

70 patients, but actually administered about twice this amount. An NRC medical consultant has reviewed the cases and has recommended that the patients be evaluated for any adverse effects.

Case 2. In November 1995, while reviewing treatment records during a routine NRC inspection of a licensee authorized to use a Sr-90 eye applicator in Honolulu, Hawaii, an NRC inspector determined that the licensee had incorrectly calculated the decay of the strength of a Sr-90 eye applicator and subsequently delivered doses 20 percent over the prescribed doses to 16 patients.

The root cause of this event was the licensee improperly calculating the decay of the Sr-90 source. Specifically, the licensee linearly decayed the strength of the source instead of logarithmically decaying the source strength of the eye applicator. Before May 1995, a technologist had correctly calculated the decay of the source; however, when this technologist left, the physician incorrectly calculated the source decay based on a linear decay from the previous calculations. In addition, the incorrect half-life of Sr-90 was used in the calculation.

Following a correct determination of the source strength, the licensee determined that 17 misadministrations involving 16 patients occurred, from May 6, 1995 to November 16, 1995, including one patient who was treated on both eyes. The delivered doses range from 21.1 to 22.7 percent greater than the prescribed dose of 40 Gy (4000 rad).

Discussion

10 CFR 35.32, "Quality Management Program," requires licensees to develop written procedures to meet five objectives associated with any brachytherapy dose, including assurance that the prescribed dose is the administered dose. A necessary part of this is to ensure that the dose rate emitted from an applicator is correct. If the manufacturer's certificate of calibration or original activity/dose rate name plate is missing, the licensee should arrange with a qualified expert to determine the dose rate from the Sr-90 source. (For additional details, see NRC IN 94-17, "Sr-90 Eye Applicators: Submission of Quality Management Plan, Calibration and Use," dated March 11, 1994.) Failure to implement the Quality Management Rule and make required reports may subject a licensee to orders, civil penalties, and notices of violation.

In view of the nature of the cited events, it appears worthwhile to review some of the properties of Sr-90 eye applicators and the related processes for decay correction over time. New Sr-90 eye applicators typically contain a 2 gigabecquerel (GBq) [54 millicurie(mCi)] source, exhibiting a surface dose rate of about 0.50 Gy (50 rad)/sec. The half-life of the parent Sr-90 is 28.5 yrs [maximum beta energy equal to 0.54 mega-electron volts (MeV)], and the yttrium-90 (Y-90) daughter half-life is 64.2 hrs (beta-max, 2.27 MeV); therefore, both isotopes are in equilibrium on the eye applicator. Since Sr-90 and Y-90 are in equilibrium, emissions from both isotopes must be accounted for in dosimetry calculations.

The dose rate (D_t) at a time (t) can be calculated from the initial dose rate (D_o) at the time of the most recent calibration with the following formula:

$$D_t = D_o \cdot (e^{-\lambda t})$$

(F1)

where e , the base of the natural logarithm (2.718), is raised to the power $-\lambda t$ and is referred to as the "decay factor" (df), where $-\lambda = -0.693/T_{1/2}$, $T_{1/2}$ is the half-life of the isotope, and t is the elapsed time. For Sr-90, $T_{1/2}=28.5$ years; therefore, $\lambda = (0.693)/(28.5 \text{ yrs}) = 0.0243 \text{ yr}^{-1}$. The values for df can be used to determine the fraction of original activity remaining after t years. These values can then be used to calculate dose rates by using the following formula:

$$D_t = D_o \cdot df$$

(F2)

The fraction of activity remaining after a given number of years from the original measurement date is given in Table 1 (Attachment 1).

For example

An eye applicator calibrated by the manufacturer on January 1, 1978, delivered an initial dose rate (D_o) of 0.75 Gy/sec (75 rad/sec) on contact. If, on January 1, 1996, the eye applicator needs to be used, the elapsed time (in years) since calibration is 18, and from Table 1, $df = 0.646$ for Sr-90. Using formula F2:

$$D_t = D_o \cdot df$$

$$D_{18 \text{ years}} = (0.75 \text{ Gy/sec})(0.646)$$

$$D_{18 \text{ years}} = 0.485 \text{ Gy/sec} \quad (48.5 \text{ rad/sec})$$

IN 96-66
December 13, 1996
Page 4 of 4

Table 1 may be extended using formula F1 and the value $\lambda = 0.0243 \text{ yr}^{-1}$ for Sr-90. Other values of λ that need to be evaluated by licensees will need to be re-calculated by using the appropriate half-life for that isotope. Alternatively, a semi-logarithmic plot of the above data will yield a straight line, which may be extended beyond 25 years.

This information notice requires no specific action nor written response. If you have any questions about the information in this notice, please contact one of the technical contacts listed below or the appropriate regional office.

signed by

Donald A. Cool, Director
Division of Industrial and
Medical Nuclear Safety
Office of Nuclear Material Safety
and Safeguards

Technical contacts: Jose M. Diaz-Velez, RII
(404) 331-7438
Email: jxd2@nrc.gov

Emilio M. Garcia, RIV
(510) 975-0239
Email: emg@nrc.gov

James A. Smith, NMSS
(301) 415-7904
Email: jas4@nrc.gov

Attachments:
1. Table 1

TABLE 1

FRACTION (EXPRESSED AS DECIMAL) OF ORIGINAL
 SR-90 ACTIVITY REMAINING AFTER (t) YEARS

| Years (t) | df |
|-----------|-------|-----------|-------|-----------|-------|-----------|-------|
| .25 | 0.994 | 6.5 | 0.854 | 12.75 | 0.734 | 19 | 0.63 |
| .5 | 0.988 | 6.75 | 0.849 | 13 | 0.729 | 19.25 | 0.626 |
| .75 | 0.982 | 7 | 0.844 | 13.25 | 0.725 | 19.5 | 0.623 |
| 1 | 0.976 | 7.25 | 0.838 | 13.5 | 0.72 | 19.75 | 0.619 |
| 1.25 | 0.97 | 7.5 | 0.833 | 13.75 | 0.716 | 20 | 0.615 |
| 1.5 | 0.964 | 7.75 | 0.828 | 14 | 0.712 | 20.25 | 0.611 |
| 1.75 | 0.958 | 8 | 0.823 | 14.25 | 0.707 | 20.5 | 0.608 |
| 2 | 0.953 | 8.25 | 0.818 | 14.5 | 0.703 | 20.75 | 0.604 |
| 2.25 | 0.947 | 8.5 | 0.813 | 14.75 | 0.699 | 21 | 0.6 |
| 2.5 | 0.941 | 8.75 | 0.808 | 15 | 0.695 | 21.25 | 0.597 |
| 2.75 | 0.935 | 9 | 0.804 | 15.25 | 0.69 | 21.5 | 0.593 |
| 3 | 0.93 | 9.25 | 0.799 | 15.5 | 0.686 | 21.75 | 0.589 |
| 3.25 | 0.924 | 9.5 | 0.794 | 15.75 | 0.682 | 22 | 0.586 |
| 3.5 | 0.918 | 9.75 | 0.789 | 16 | 0.678 | 22.25 | 0.582 |
| 3.75 | 0.913 | 10 | 0.784 | 16.25 | 0.674 | 22.5 | 0.579 |
| 4 | 0.907 | 10.25 | 0.78 | 16.5 | 0.67 | 22.75 | 0.575 |
| 4.25 | 0.902 | 10.5 | 0.775 | 16.75 | 0.666 | 23 | 0.572 |
| 4.5 | 0.896 | 10.75 | 0.77 | 17 | 0.662 | 23.25 | 0.568 |
| 4.75 | 0.891 | 11 | 0.765 | 17.25 | 0.658 | 23.5 | 0.565 |
| 5 | 0.886 | 11.25 | 0.761 | 17.5 | 0.654 | 23.75 | 0.562 |
| 5.25 | 0.88 | 11.5 | 0.756 | 17.75 | 0.65 | 24 | 0.558 |
| 5.5 | 0.875 | 11.75 | 0.752 | 18 | 0.646 | 24.25 | 0.555 |
| 5.75 | 0.87 | 12 | 0.747 | 18.25 | 0.642 | 24.5 | 0.551 |
| 6 | 0.864 | 12.25 | 0.743 | 18.5 | 0.638 | 24.75 | 0.548 |
| 6.25 | 0.859 | 12.5 | 0.738 | 18.75 | 0.634 | 25 | 0.545 |