

Region I NMSS Licensee Event Report

Licensee	Agilent		
Event Description	Notice of General License Device w/High Removable		
	026	Activity	
License #	07-28762-	Docket #	03002988
Event Date	7-02	Report Date	7-18-02
		MLER-RI	2002-039

1. REPORTING REQUIREMENT

<input type="checkbox"/> 10 CFR 20.2201 Theft or Loss	<input type="checkbox"/> 10 CFR 35.33 Misadministration
<input type="checkbox"/> 10 CFR 20.2203 30 Day Report	<input type="checkbox"/> License Condition
<input type="checkbox"/> 10 CFR 30.50 Report	
<input checked="" type="checkbox"/> Other <u>10 CFR 31.5</u>	

2. REGION I RESPONSE

<input type="checkbox"/> Immediate Site Inspection	Inspector/Date	
<input type="checkbox"/> Special Inspection	Inspector/Date	
<input type="checkbox"/> Telephone Inquiry	Inspector/Date	
<input type="checkbox"/> Preliminary Report	<input type="checkbox"/> Daily Report	
<input checked="" type="checkbox"/> Information Entered in RI Log	<input checked="" type="checkbox"/> Review at next inspection	
Report referred to: _____		


3. REPORT EVALUATION

<input checked="" type="checkbox"/> Description of Event	<input type="checkbox"/> Corrective Actions
<input checked="" type="checkbox"/> Levels of RAM Involved	<input type="checkbox"/> Calculations Adequate
<input type="checkbox"/> Cause of Event	<input type="checkbox"/> Additional Information Requested from Licensee

4. SPECIAL INSTRUCTIONS OR COMMENTS

Public BY 8/2/02 Non-Public _____

	Initials/Date	Initials/Date
Completed by:	<u>Rety Ulling</u>	Date <u>8/2/02</u>
Reviewed by:	<u>[Signature]</u>	Date <u>8/9/2002</u>

 **Agilent Technologies**
2850 Centerville Road
Wilmington, DE 19808

RECEIVED
REGION 1

2002 JUL 19 PM 2: 53

Division of Nuclear Materials Safety
U.S. Nuclear Regulatory Commission
Region I
475 Allendale Road,
King of Prussia, PA 19406

Re: Notice of General License device with high removable activity (NRC License 07-28762-02G)

July 18, 2002

Dear Sir or Madam,

Per the requirements of 10 CFR 31.5, I am notifying your office of a general license device returned to our site by customer from which we have obtained a removable activity wipe in excess of 0.005 μCi . The cell was evaluated by support engineering and its disposition is identified below. I have contacted the customer to ascertain the conditions under which the cell was used and will notify the cell supplier of the high wipe.

<u>Serial #</u>	<u>Model #</u>	<u>Activity (μCi)</u>	<u>Disposition</u>
K3883	G1533A	0.025	Refurbished and reused

Comments:

The cell is 5 years old and the source area was fairly clean.

Please contact me at 302-633-8262 if there are any questions.

Thank you,



David S. Bennett

NMSS/RGNI MATERIALS-002

Radiation Safety Officer