

Region I
NMSS Licensee Event Report

Licensee	AGILENT TECHNOLOGIES		
Event Description	Notice of General License device w/high removable activity		
License #	07-28762-02G	Docket #	03002988
Event Date	2002	Report Date	8/8/02
		MLER-RI	2002- 052

1. REPORTING REQUIREMENT

<input type="checkbox"/>	10 CFR 20.2201 Theft or Loss	<input type="checkbox"/>	10 CFR 35.33 Misadministration
<input type="checkbox"/>	10 CFR 20.2203 30 Day Report	<input type="checkbox"/>	License Condition
<input type="checkbox"/>	10 CFR 30.50 Report		
XX	Other 10 CFR 31.5		

2. REGION I RESPONSE

<input type="checkbox"/>	Immediate Site Inspection	Inspector/Date	
<input type="checkbox"/>	Special Inspection	Inspector/Date	
<input type="checkbox"/>	Telephone Inquiry	Inspector/Date	
<input type="checkbox"/>	Preliminary Report		
XX	Information Entered in RI Log	<input type="checkbox"/>	Daily Report
<input type="checkbox"/>	Report referred to:	<input checked="" type="checkbox"/>	Review at next inspection



3. REPORT EVALUATION


<input type="checkbox"/>	Description of Event	<input type="checkbox"/>	Corrective Actions
<input type="checkbox"/>	Levels of RAM Involved	<input type="checkbox"/>	Calculations Adequate
<input type="checkbox"/>	Cause of Event	<input type="checkbox"/>	Additional Information Requested from Licensee

4. SPECIAL INSTRUCTIONS OR COMMENTS

Public ☒

Non-Public ☐

	Initials/Date		Initials/Date
Completed by:		Date	8/8/02
Reviewed by:		Date	8/27/2002

 **Agilent Technologies**
2850 Centerville Road
Wilmington, DE 19808

RECEIVED
REGION 1

2002 AUG -8 PM 1:05

Division of Nuclear Materials Safety
U.S. Nuclear Regulatory Commission
Region I
475 Allendale Road,
King of Prussia, PA 19406

Re: Notice of General License device with high removable activity (NRC License 07-28762-02G)

August 7, 2002

Dear Sir or Madam,

Per the requirements of 10 CFR 31.5, I am notifying your office of a general license device recently received from a source supplier from which we have obtained a removable activity wipe in excess of 0.005 μCi . The cell was evaluated by support engineering and its disposition is identified below. I will notify the cell supplier of the high wipe.

<u>Serial #</u>	<u>Model #</u>	<u>Activity (μCi)</u>	<u>Disposition</u>
U4646	G2397A	0.018	Cleaned and processed for customer use

Comments:

No problems observed by support engineer.

Please contact me at 302-633-8262 if there are any questions.

Thank you,


David S. Bennett
Radiation Safety Officer