

Region I NMSS Licensee Event Report

Licensee	Agilent Technologies
Event Description	Notice of General License Device With High Removable
	Activity

License #	07-28762- ⁹²⁶	Docket #	03002988	MLER-RI	2002-056
Event Date	2002	Report Date	8-28-02		

1. REPORTING REQUIREMENT

<input type="checkbox"/> 10 CFR 20.2201 Theft or Loss	<input type="checkbox"/> 10 CFR 35.33 Misadministration
<input type="checkbox"/> 10 CFR 20.2203 30 Day Report	<input type="checkbox"/> License Condition
<input type="checkbox"/> 10 CFR 30.50 Report	
<input checked="" type="checkbox"/> Other <u>31.5</u>	

2. REGION I RESPONSE

<input type="checkbox"/> Immediate Site Inspection	Inspector/Date	
<input type="checkbox"/> Special Inspection	Inspector/Date	
<input type="checkbox"/> Telephone Inquiry	Inspector/Date	
<input type="checkbox"/> Preliminary Report	<input type="checkbox"/> Daily Report	
<input checked="" type="checkbox"/> Information Entered in RI Log	<input checked="" type="checkbox"/> Review at next inspection	
<input type="checkbox"/> Report referred to:		

3. REPORT EVALUATION

<input checked="" type="checkbox"/> Description of Event	<input checked="" type="checkbox"/> Corrective Actions
<input checked="" type="checkbox"/> Levels of RAM Involved	<input type="checkbox"/> Calculations Adequate
<input type="checkbox"/> Cause of Event	<input type="checkbox"/> Additional Information Requested from Licensee

4. SPECIAL INSTRUCTIONS OR COMMENTS

Public

Non-Public

	Initials/Date		Initials/Date
Completed by:	Betsy Ullrich	Date	8/29/02
Reviewed by:	Joe D. K...	Date	8/29/2002

RECEIVED
REGION 1

 **Agilent Technologies**
2850 Centerville Road
Wilmington, DE 19808

2002 AUG 28 PM 1: 37

Division of Nuclear Materials Safety
U.S. Nuclear Regulatory Commission
Region I
475 Allendale Road,
King of Prussia, PA 19406

Re: Notice of General License device with high removable activity (NRC License 07-28762-02G)

August 27, 2002

Dear Sir or Madam,

Per the requirements of 10 CFR 31.5, I am notifying your office of the following devices returned to us by customers from which we have obtained removable activity wipes in excess of 0.005 μCi . Cells. The cells were evaluated by support engineering and their disposition is identified below. I have contacted the customers for further information and will notify the cell supplier of the high wipes.

<u>Serial #</u>	<u>Model #</u>	<u>Activity (μCi)</u>	<u>Disposition</u>	<u>Comment</u>
F2621	19233	0.006	Refurbish	Cell 11 years old and source dirty
L6603	19233	0.046	Scrap	Top broken. Cell 13 years old.
K2219	G1533A	0.332	Refurbish	Cell 6 years old. Transferred to other location at customer's request

Please contact me at 302-633-8262 if there are any questions.

Thank you,



David S. Bennett
Radiation Safety Officer