



YOUR REF:

OUR REF: August 23, 2002

DATE:

ISLAND ROADS CORPORATION
6501 RED HOOK PLAZA
SUITE 201
ST. THOMAS
VI 00802-1306

TELEPHONE: (340) 775-5704
FACSIMILE: (340) 775-5663

FAX COVER SHEET

TO: DIVISION OF NUCLEAR MATERIALS SAFETY

ATTN: ANDY MILLER

FAX NO: 1-404-562-4955

SUBJECT: MISCELLANEOUS

FROM: ANDY SMITH

No. of Sheets Including this Cover Page 1+3+1

PLEASE FIND ATTACHED, ANDREW SMITH (MYSELF), GEOFF FLEMING JR. AND MARTIN IORIO'S CERTIFICATIONS FROM A RECENTLY HELD COURSE.

ALSO, PLEASE FIND ENCLOSED A COPY OF THE LATEST LEAK TEST ANALYSIS WHICH WAS DONE ON JANUARY 22, 2002.

WE HAVE ORDERED A REPLACEMENT TEST KIT AND WILL CARRY OUT THE NEXT TEST AS SOON AS IT ARRIVES.

FINALLY, WE DID LOCATE OUR RADIATION SURVEY METER.

PLEASE PHONE ME ON 340-690-6094 TO CONFIRM RECEIPT OF THIS FAX.

THANKING YOU, ANDY SMITH

Aug. 21 2002 9:17PM CRMI-David Rhee

(787)791-9454

p. 1

CRMI

Central Institute for Radiation Monitoring & Information
NRC License 52-1470-043

This Certifies That

Andrew Smith

Has successfully completed the courses entitled:

RADIATION SAFETY OFFICER Certification,
NUCLEAR GAUGE Certification, and
HAZMAT for RADIATION Certification

August 18, 2002

This course provides the continuing education requirements in compliance with the 49 Code of Federal Regulations Part 172.704 and the Nuclear Regulatory Commission NUREG-1536

Nikola E. Uffink, M.P.
NRC-Certified Safety Instructor

David Rhee
DAVID RHEE, HP/MP
NRC License 52-1470-043
NRC-Certified Safety Instructor

CRMI

Continuing Education Programs
1000 West Street, Suite 100
Cruz Bay, St. John, VI 00801

This Certifies That

Gregory Fleming Jr

Has successfully completed the courses entitled:

NUCLEAR GAUGE Certification and
HAZMAT for RADIATION Certification

August 18, 2002

This course provides the continuing education requirements in
compliance with the 49 Code of Federal Regulations Part 172.704 and
the Nuclear Regulatory Commission NUREG-1556.

William C. O'Brien, Director, CRMI
NSI-Certified Safety Instructor



William C. O'Brien, Director
NSI-Certified Safety Instructor

CRMI

Continuing Education Activities Department
NRC License #22-131601

This Certifies That

Martin Lewis

Has successfully completed the courses entitled:

NUCLEAR GAUGE Certification, and HAZMAT for RADIATION Certification

August 18, 2002

This course provides the continuing education requirements in compliance with the 49 Code of Federal Regulations, Part 172.704 and the Nuclear Regulatory Commission NUREG 1556.

Virginia C. Crockett, Director
NRC Certified Activity Provider


David M. Knoe, HPWR
NSO/CHME
NRC Certified Activity Provider



Troxler Electronic Laboratories, Inc.

3008 Cornwallis Rd., P.O.Box 12067
Research Triangle Park, NC 27709
Tel: (919) 549-8661 Fax: (919) 485-2250
License: NC 032-0182-1

LEAK TEST CERTIFICATE

DEVICE:

Model: 3430 Serial No: 22594 Sample Date: 1-22-2

SEALED SOURCES:

NUCLIDE	SERIAL NO.	INITIAL ACTIVITY (uCi)	CURRENT ACTIVITY (uCi)
CS-137	76 - 4350	8.0	6.6
AM-241SE	47 - 18437	40.0	39.4

LEAK TEST ANALYSIS:

The sample has been analyzed using a counting system capable of measuring beta, gamma, and alpha contamination which was last calibrated on: 6/14/02

	ALPHA	BETA-GAMMA
Conversion factor (cpm/uCi)	4.75E+05	7.79E+05
Background measurement (cpm)	0	31
Sample measurement (cpm)	2	93
Activity (uCi)	< MDA	< MDA
Min. Detectable Activity (uCi)	8.7E-06	1.7E-05

This certifies that the above leak test results are:

Less than 0.005 uCi (185 Bq) Greater than 0.005 uCi (185 Bq)

Measurement Date: 2/7/02 Certified by: Greg Wells

If greater than 0.005 uCi (185 Bq):

Person Notified _____ Date _____

Phone _____ and/or Fax _____