

**Region I
NMSS Licensee Event Report**

Licensee
Event Description

Agilent Technologies
Notice of General License device w/High Removable Activity

License #

07-28762-02^S

Docket #

03002988

MLER-RI

2002-062

Event Date

2002

Report Date

9/17/02

1. REPORTING REQUIREMENT

<input type="checkbox"/>
<input type="checkbox"/>
<input type="checkbox"/>
<input checked="" type="checkbox"/>

10 CFR 20.2201 Theft or Loss

10 CFR 20.2203 30 Day Report

10 CFR 30.50 Report

Other 10 CFR 31.5

<input type="checkbox"/>
<input type="checkbox"/>

10 CFR 35.33 Misadministration

License Condition

2. REGION I RESPONSE

<input type="checkbox"/>
<input type="checkbox"/>
<input type="checkbox"/>
<input type="checkbox"/>
<input checked="" type="checkbox"/>
<input type="checkbox"/>

Immediate Site Inspection

Special Inspection

Telephone Inquiry

Preliminary Report

Information Entered in RI Log

Report referred to: _____

Inspector/Date

Inspector/Date

Inspector/Date

Daily Report

☒ Review at next inspection

3. REPORT EVALUATION

<input checked="" type="checkbox"/>
<input checked="" type="checkbox"/>
<input checked="" type="checkbox"/>

Description of Event

Levels of RAM Involved

Cause of Event

<input checked="" type="checkbox"/>
<input type="checkbox"/>
<input type="checkbox"/>

Corrective Actions

Calculations Adequate

Additional Information Requested from Licensee

4. SPECIAL INSTRUCTIONS OR COMMENTS

Public

X

Non-Public

Initials/Date

<u>Batz Ullrich</u>
<u>[Signature]</u>

Completed by:


Date

Initials/Date

<u>9/19/02</u>
<u>9/20/2002</u>

Reviewed by:

Date

 **Agilent Technologies**
2850 Centerville Road
Wilmington, DE 19808

RECEIVED
REGION I
2002 SEP 17 PM 1:25

Division of Nuclear Materials Safety
U.S. Nuclear Regulatory Commission
Region I
475 Allendale Road,
King of Prussia, PA 19406

Re: Notice of General License device with high removable activity (NRC License 07-28762-02G)

September 16, 2002


Dear Sir or Madam,

Per the requirements of 10 CFR 31.5, I am notifying your office of the following devices returned to us by customers from which we have obtained removable activity wipes in excess of 0.005 μCi . Cell serial # F3281 was returned from overseas and U1135 was domestic. The cells were evaluated by support engineering and their disposition is identified below. I have contacted the U.S. customer for further information and will notify the cell supplier of the high wipes.

Serial #	Model #	Activity (μCi)	Disposition	Comments
U1135	G2397A	0.035	Scrapped	Plated ring was dislodged inside cell. Cell was 3 years old
F3281	19233	0.245	Scrapped	Source dirty. Cell was 11 years old

Please contact me at 302-633-8262 if there are any questions.

Thank you,


David S. Bennett
Radiation Safety Officer