

Region I NMSS Licensee Event Report

Licensee	AGILENT TECHNOLOGIES
Event Description	Notice of Gen License Device w/High Removable Activity

License #	07-28762-02G	Docket #	03002988	MLER-RI	2002-067
Event Date	2002	Report Date	9/30/02		

1. REPORTING REQUIREMENT

<input type="checkbox"/> 10 CFR 20.2201 Theft or Loss	<input type="checkbox"/> 10 CFR 35.33 Misadministration
<input type="checkbox"/> 10 CFR 20.2203 30 Day Report	<input type="checkbox"/> License Condition
<input type="checkbox"/> 10 CFR 30.50 Report	
<input checked="" type="checkbox"/> Other <u>10 CFR 31.5</u>	

2. REGION I RESPONSE

<input type="checkbox"/> Immediate Site Inspection	Inspector/Date	
<input type="checkbox"/> Special Inspection	Inspector/Date	
<input type="checkbox"/> Telephone Inquiry	Inspector/Date	
<input type="checkbox"/> Preliminary Report	<input type="checkbox"/> Daily Report	
<input checked="" type="checkbox"/> Information Entered in RI Log	<input checked="" type="checkbox"/> Review at next inspection	
<input type="checkbox"/> Report referred to:		


3. REPORT EVALUATION

<input checked="" type="checkbox"/> Description of Event	<input type="checkbox"/> Corrective Actions
<input checked="" type="checkbox"/> Levels of RAM Involved	<input type="checkbox"/> Calculations Adequate
<input checked="" type="checkbox"/> Cause of Event	<input type="checkbox"/> Additional Information Requested from Licensee

4. SPECIAL INSTRUCTIONS OR COMMENTS

Public Non-Public

Completed by:	<div style="text-align: center;">Initials/Date</div> <p style="font-size: 1.5em; text-align: center;">BETSY ULLRICH</p>	Date	<div style="text-align: center;">Initials/Date</div> <p style="font-size: 1.5em; text-align: center;">BU 10/3/02</p>
Reviewed by:	<div style="text-align: center;">Initials/Date</div> <p style="font-size: 1.5em; text-align: center;">[Signature]</p>	Date	<div style="text-align: center;">Initials/Date</div> <p style="font-size: 1.5em; text-align: center;">10/7/02</p>

 **Agilent Technologies**
2850 Centerville Road
Wilmington, DE 19808

RECEIVED
REGION I
2002 SEP 30 PM 1:57

Division of Nuclear Materials Safety
U.S. Nuclear Regulatory Commission
Region I
475 Allendale Road,
King of Prussia, PA 19406

Re: Notice of General License device with high removable activity (NRC License 07-28762-02G)

September 26, 2002

Dear Sir or Madam,

Per the requirements of 10 CFR 31.5, I am notifying your office of the following devices returned to us by customers from which we have obtained removable activity wipes in excess of 0.005 μCi . Cell serial # U2215 was returned from overseas and U2075 was domestic. The cells were evaluated by support engineering and their disposition is identified below. I have contacted the U.S. customer for further information and will notify the cell supplier of the high wipes.

Serial #	Model #	Activity (μCi)	Disposition	Comments
U2215	G2397A	0.045	Scrapped	Cell was 2 years old and source was dirty
U2057	G2397a	0.023	Scrapped	Cell was 1 year old and source was clean

Please contact me at 302-633-8262 if there are any questions.

Thank you,



David S. Bennett
Radiation Safety Officer