

<b>TELEPHONE CONVERSATION RECORD</b>	<b>Date:</b> October 9, 2002	<b>Time:</b> 11:00
<b>Mail Control No.:</b> N/A <b>Insp. No.:</b> 2002-004	<b>License No.:</b> General Licensee	<b>Docket No.:</b> 99990001
<b>Person Called:</b> Floyd Graham, Operations Manager	<b>Licensee:</b> Chemtech Consulting Group	<b>Telephone No.:</b> 908-789-8900
<b>Person Calling:</b> Steven Courtemanche/(610) 337-5075		
<b>Subject:</b> Clarification of wording in Response to NOV.		
<p><u>Summary:</u> I was informed by the receptionist that Mr. Divya Mehta was not in the building but that I could speak with Mr. Graham. I questioned Mr. Graham as to what was meant by dismantling the ECD. He stated that he had spoken with the manufacturer, i.e., Agilent Technologies, who described the process he would have to take in order to determine what the ECD serial number. The process would only involve removing the ECD from the gas chromatograph and removing ancillary equipment so that the Serial Number could be visible. I informed Mr. Graham that the Commission would not require him to send the leak test reports to the Region I Office unless the results indicated that an ECD had more removable contamination than 0.005 microcuries. Review of the leak test records would be part of a future inspection.</p>		
<b>Action Required/Taken:</b> Place into ADAMS.		
<b>Prepared by</b> Steven Courtemanche	<b>Date:</b> October 10, 2002	