

Region I NMSS Licensee Event Report

Licensee
Event Description

AGILENT TECH

License #	07-28762-026	Docket #	03002988	MLER-RI	2002-070
Event Date	2002	Report Date	10/7/02		

1. REPORTING REQUIREMENT

<input type="checkbox"/> 10 CFR 20.2201 Theft or Loss <input type="checkbox"/> 10 CFR 20.2203 30 Day Report <input type="checkbox"/> 10 CFR 30.50 Report <input checked="" type="checkbox"/> Other <u>10 CFR 31.5</u>	<input type="checkbox"/> 10 CFR 35.33 Misadministration <input type="checkbox"/> License Condition
--	---

2. REGION I RESPONSE

<input type="checkbox"/> Immediate Site Inspection <input type="checkbox"/> Special Inspection <input type="checkbox"/> Telephone Inquiry <input type="checkbox"/> Preliminary Report <input checked="" type="checkbox"/> Information Entered in RI Log <input type="checkbox"/> Report referred to: _____	<table border="1" style="width: 100%; border-collapse: collapse;"> <tr><td style="width: 50%;">Inspector/Date</td><td> </td></tr> <tr><td>Inspector/Date</td><td> </td></tr> <tr><td>Inspector/Date</td><td> </td></tr> </table> <input type="checkbox"/> Daily Report <input checked="" type="checkbox"/> Review at next inspection	Inspector/Date		Inspector/Date		Inspector/Date	
Inspector/Date							
Inspector/Date							
Inspector/Date							

3. REPORT EVALUATION


<input checked="" type="checkbox"/> Description of Event <input checked="" type="checkbox"/> Levels of RAM Involved <input type="checkbox"/> Cause of Event	<input type="checkbox"/> Corrective Actions <input type="checkbox"/> Calculations Adequate <input type="checkbox"/> Additional Information Requested from Licensee
---	--

4. SPECIAL INSTRUCTIONS OR COMMENTS

Public

Non-Public

Completed by:	<table border="1" style="width: 100%; border-collapse: collapse;"> <tr><td style="width: 50%;">Initials/Date</td><td> </td></tr> <tr><td colspan="2" style="text-align: center; padding: 5px;">BETSY ULRICH</td></tr> </table>	Initials/Date		BETSY ULRICH		Date	<table border="1" style="width: 100%; border-collapse: collapse;"> <tr><td style="width: 50%;">Initials/Date</td><td> </td></tr> <tr><td colspan="2" style="text-align: center; padding: 5px;">10/8/02</td></tr> </table>	Initials/Date		10/8/02	
Initials/Date											
BETSY ULRICH											
Initials/Date											
10/8/02											
Reviewed by:	<table border="1" style="width: 100%; border-collapse: collapse;"> <tr><td style="width: 50%;">Initials/Date</td><td> </td></tr> <tr><td colspan="2" style="text-align: center; padding: 5px;">[Signature]</td></tr> </table>	Initials/Date		[Signature]		Date	<table border="1" style="width: 100%; border-collapse: collapse;"> <tr><td style="width: 50%;">Initials/Date</td><td> </td></tr> <tr><td colspan="2" style="text-align: center; padding: 5px;">10/9/2002</td></tr> </table>	Initials/Date		10/9/2002	
Initials/Date											
[Signature]											
Initials/Date											
10/9/2002											

 **Agilent Technologies**
2850 Centerville Road
Wilmington, DE 19808

RECEIVED
REGION I
2002 OCT -7 PM 2:50

Division of Nuclear Materials Safety
U.S. Nuclear Regulatory Commission
Region I
475 Allendale Road,
King of Prussia, PA 19406

Re: Notice of General License device with high removable activity (NRC License 07-28762-02G)

October 4, 2002


Dear Sir or Madam,

Per the requirements of 10 CFR 31.5, I am notifying your office of the following devices returned to us by customers from which we have obtained removable activity wipes in excess of 0.005 μCi . Cell serial # K4505 was returned from overseas and the others were domestic. The cells were evaluated by support engineering and their disposition is identified below. I have contacted the U.S. customers for further information and will notify the cell supplier of the high wipes.

Serial #	Model #	Activity (μCi)	Disposition	Comments
C0732	18803	0.006	Scrapped	Cell 24 years old. Source dirty and dark.
S10736	19312	0.009	Scrapped	Cell 14 years old. Indication of excessive temperature. Source dark.
L5298	19235	2.489	Scrapped	Cell 13 years old. Cell is rusty and dark
K4505	G1533A	0.027	Refurbished	Cell is 4 years old. Source lightly darkened.

Please contact me at 302-633-8262 if there are any questions.

Thank you,


David S. Bennett
Radiation Safety Officer