

# Region I NMSS Licensee Event Report

Licensee	AGILENT		
Event Description	Notice of General License Device with High		
	Removable Activity		

License #	07-28762- <sup>026</sup>	Docket #	03002988	MLER-RI	2002-071
Event Date		Report Date			

**1. REPORTING REQUIREMENT**

<input type="checkbox"/> 10 CFR 20.2201 Theft or Loss <input type="checkbox"/> 10 CFR 20.2203 30 Day Report <input type="checkbox"/> 10 CFR 30.50 Report <input checked="" type="checkbox"/> Other <u>10 CFR 31.5</u>	<input type="checkbox"/> 10 CFR 35.33 Misadministration <input type="checkbox"/> License Condition
--	---

**2. REGION I RESPONSE**

<input type="checkbox"/> Immediate Site Inspection <input type="checkbox"/> Special Inspection <input type="checkbox"/> Telephone Inquiry <input type="checkbox"/> Preliminary Report <input checked="" type="checkbox"/> Information Entered in RI Log <input type="checkbox"/> Report referred to: _____	<table border="1" style="width: 100%; border-collapse: collapse;"> <tr><td style="width: 50%;">Inspector/Date</td><td></td></tr> <tr><td>Inspector/Date</td><td></td></tr> <tr><td>Inspector/Date</td><td></td></tr> </table> <input type="checkbox"/> Daily Report <input checked="" type="checkbox"/> Review at next inspection	Inspector/Date		Inspector/Date		Inspector/Date	
Inspector/Date							
Inspector/Date							
Inspector/Date							

**3. REPORT EVALUATION**

<input checked="" type="checkbox"/> Description of Event <input checked="" type="checkbox"/> Levels of RAM Involved <input type="checkbox"/> Cause of Event	<input checked="" type="checkbox"/> Corrective Actions <input type="checkbox"/> Calculations Adequate <input type="checkbox"/> Additional Information Requested from Licensee
---	---

**4. SPECIAL INSTRUCTIONS OR COMMENTS**

Public

Non-Public

Completed by:	Betz Ullrich	Date	10/21/02
Reviewed by:	[Signature]	Date	10/24/2002

 **Agilent Technologies**  
2850 Centerville Road  
Wilmington, DE 19808

RECEIVED  
REGION 1  
2002 OCT 11 PM 1:34

Division of Nuclear Materials Safety  
U.S. Nuclear Regulatory Commission  
Region I  
475 Allendale Road,  
King of Prussia, PA 19406

Re: Notice of General License device with high removable activity (NRC License 07-28762-02G)

October 10, 2002

Dear Sir or Madam,

Per the requirements of 10 CFR 31.5, I am notifying your office of a general license device recently received from a source supplier from which we have obtained a removable activity wipe in excess of 0.005  $\mu\text{Ci}$ . The cell was evaluated by support engineering and its disposition is identified below. I will notify the cell supplier of the high wipe.

<u>Serial #</u>	<u>Model #</u>	<u>Activity (<math>\mu\text{Ci}</math>)</u>	<u>Disposition</u>
K1998	G1533A	0.025	Refurbish

Comments:

Source was slightly discolored but not corroded.

Please contact me at 302-633-8262 if there are any questions.

Thank you,



David S. Bennett  
Radiation Safety Officer