

**NUREG-1556**  
**Vol. 9**

---

---

# **Consolidated Guidance About Materials Licenses**

## **Program-Specific Guidance About Medical Use Licenses**

**Final Report**

---

---

**U.S. Nuclear Regulatory Commission**

**Office of Nuclear Material Safety and Safeguards**

**R. W. Broseus, P. A. Lanzisera, A. R. Jones, R. G. Gattone, R. D. Reid**



## AVAILABILITY OF REFERENCE MATERIALS IN NRC PUBLICATIONS

### NRC Reference Material

As of November 1999, you may electronically access NUREG-series publications and other NRC records at NRC's Public Electronic Reading Room at <http://www.nrc.gov/reading-rm.html>.

Publicly released records include, to name a few, NUREG-series publications; *Federal Register* notices; applicant, licensee, and vendor documents and correspondence; NRC correspondence and internal memoranda; bulletins and information notices; inspection and investigative reports; licensee event reports; and Commission papers and their attachments.

NRC publications in the NUREG series, NRC regulations, and *Title 10, Energy*, in the Code of *Federal Regulations* may also be purchased from one of these two sources.

1. The Superintendent of Documents  
U.S. Government Printing Office  
Mail Stop SSOP  
Washington, DC 20402-0001  
Internet: [bookstore.gpo.gov](http://bookstore.gpo.gov)  
Telephone: 202-512-1800  
Fax: 202-512-2250
2. The National Technical Information Service  
Springfield, VA 22161-0002  
[www.ntis.gov](http://www.ntis.gov)  
1-800-553-6847 or, locally, 703-605-6000

A single copy of each NRC draft report for comment is available free, to the extent of supply, upon written request as follows:

Address: Office of the Chief Information Officer,  
Reproduction and Distribution  
Services Section  
U.S. Nuclear Regulatory Commission  
Washington, DC 20555-0001  
E-mail: [DISTRIBUTION@nrc.gov](mailto:DISTRIBUTION@nrc.gov)  
Facsimile: 301-415-2289

Some publications in the NUREG series that are posted at NRC's Web site address <http://www.nrc.gov/reading-rm/doc-collections/nuregs> are updated periodically and may differ from the last printed version. Although references to material found on a Web site bear the date the material was accessed, the material available on the date cited may subsequently be removed from the site.

### Non-NRC Reference Material

Documents available from public and special technical libraries include all open literature items, such as books, journal articles, and transactions, *Federal Register* notices, Federal and State legislation, and congressional reports. Such documents as theses, dissertations, foreign reports and translations, and non-NRC conference proceedings may be purchased from their sponsoring organization.

Copies of industry codes and standards used in a substantive manner in the NRC regulatory process are maintained at—

The NRC Technical Library  
Two White Flint North  
11545 Rockville Pike  
Rockville, MD 20852-2738

These standards are available in the library for reference use by the public. Codes and standards are usually copyrighted and may be purchased from the originating organization or, if they are American National Standards, from—

American National Standards Institute  
11 West 42<sup>nd</sup> Street  
New York, NY 10036-8002  
[www.ansi.org](http://www.ansi.org)  
212-642-4900

Legally binding regulatory requirements are stated only in laws; NRC regulations; licenses, including technical specifications; or orders, not in NUREG-series publications. The views expressed in contractor-prepared publications in this series are not necessarily those of the NRC.

The NUREG series comprises (1) technical and administrative reports and books prepared by the staff (NUREG-XXXX) or agency contractors (NUREG/CR-XXXX), (2) proceedings of conferences (NUREG/CP-XXXX), (3) reports resulting from international agreements (NUREG/IA-XXXX), (4) brochures (NUREG/BR-XXXX), and (5) compilations of legal decisions and orders of the Commission and Atomic and Safety Licensing Boards and of Directors' decisions under Section 2.206 of NRC's regulations (NUREG-0750).

NUREG-1556  
Vol. 9

---

---

# Consolidated Guidance About Materials Licenses

Program-Specific Guidance About  
Medical Use Licenses

Final Report

---

---

Manuscript Completed: October 2002  
Date Published: October 2002

Prepared by  
R. W. Broseus, P. A. Lanzisera, A. R. Jones, R. G. Gattone, R. D. Reid

**Division of Industrial and Medical Nuclear Safety  
Office of Nuclear Material Safety and Safeguards  
U.S. Nuclear Regulatory Commission  
Washington, DC 20555-0001**



## ABSTRACT

As part of its redesign of the materials licensing process, the United States Nuclear Regulatory Commission (NRC) is consolidating and updating numerous guidance documents into a single comprehensive repository as described in NUREG-1539, "Methodology and Findings of the NRC's Materials Licensing Process Redesign," dated April 1996, and draft NUREG-1541, "Process and Design for Consolidating and Updating Materials Licensing Guidance," dated April 1996. NUREG-1556, Vol. 9, "Consolidated Guidance about Materials Licenses: Program-Specific Guidance about Medical Use Licenses," dated October 2002, is the ninth program-specific guidance document developed for the new process and is intended for use by applicants, licensees, and NRC staff and will also be available to Agreement States.

This document contains information that is intended to assist applicants for licenses for the medical use of byproduct material in preparing their license applications. In particular, it describes the types of information needed to complete NRC Form 313, "Application for Material License" and NRC Form 313A, "Training and Experience and Preceptor Statement." The document provides an overview of the types of licenses issued by the NRC; the commitments and responsibilities that must be undertaken by a licensee; applicable regulations; the process for filing a license application; and the contents of applications for different types of medical uses of byproduct material. In particular, this document provides a description, on an item-by-item basis, of the information to be provided by an applicant on NRC Form 313. Because of the wide variety in the types of medical uses of byproduct material, indicators have been placed in the document to alert applicants for particular types of medical uses to material that pertains to those types of uses.

The document also contains appendices that include (1) copies of necessary forms; (2) a sample license application and completed licenses for some different types of medical uses of byproduct materials; and (3) examples of the types of supporting documents, such as implementing procedures, that may need to be prepared by applicants. NRC is placing added emphasis on conducting its regulatory activities in a risk-informed and performance-based manner. This approach is intended to be less prescriptive and to allow for the implementation by licensees that may be specific to their needs while meeting the regulatory requirements. By supplying examples, NRC seeks to provide information to meet the needs of applicants for licensure, without being prescriptive. Guidance in this document represents one means acceptable to NRC staff of complying with NRC regulations and is not intended to be the only means of satisfying requirements for a license.

Volume 9 of NUREG-1556 provides guidance for licensure under revised Title 10, Part 35, "Medical Use of Byproduct Material." It is also available for use by Agreement States and will combine and supercede guidance found in the documents listed below:

- Regulatory Guide (RG) 10.8, Revision 2, "Guide for the Preparation of Applications for Medical Use Programs;"

## ABSTRACT

- Appendix X to RG 10.8, Revision 2, “Guidance on Complying With New Part 20 Requirements;”
- Draft RG DG-0009, “Supplement to Regulatory Guide 10.8, Revision 2, Guide for the Preparation of Applications for Medical Use Programs;”
- Draft RG FC 414-4, “Guide for the Preparation of Applications for Licenses for Medical Teletherapy Programs;”
- RG 8.23, “Radiation Safety Surveys at Medical Institutions, Revision 1;”
- RG 8.33, “Quality Management Program;”
- RG 8.39, “Release of Patients Administered Radioactive Materials;”
- Policy and Guidance Directive (P&GD) 03-02, “Licensing Lixiscope and BMA;”
- Policy and Guidance Directive (P&GD) 03-08, “Standard Review Plan for Teletherapy;”
- Policy and Guidance Directive (P&GD) 3-17, “Review of Training and Experience Documentation Submitted by Proposed Physician User Applicants;”
- Policy and Guidance Directive (P&GD) FC 87-2, “Standard Review Plan for License Applications for the Medical Use of Byproduct Material;”
- Policy and Guidance Directive (P&GD) FC 86-4, Revision 1, “Information Required for Licensing Remote Afterloading Devices;”
- Addendum to Revision 1 to P&GD FC 86-4, “Information Required for Licensing Remote Afterloading Devices-Increased Source Possession Limits;”
- Policy and Guidance Directive (P&GD) FC 92-01 “Information Required for Licensing Mobile Nuclear Medicine Services,” and
- Policy and Guidance Directive (P&GD) 3-15, “Standard Review Plan for Review of Quality Management Programs.”

# CONTENTS

ABSTRACT .....	iii
FOREWORD .....	xi
ACKNOWLEDGMENTS .....	xv
ABBREVIATIONS .....	xvii
1 OVERVIEW .....	1-1
1.1 PURPOSE OF REPORT .....	1-1
1.2 TYPES OF LICENSES .....	1-6
1.2.1 SPECIFIC LICENSE OF LIMITED SCOPE .....	1-7
1.2.2 SPECIFIC LICENSE OF BROAD SCOPE .....	1-8
1.2.3 RESEARCH INVOLVING HUMAN SUBJECTS .....	1-8
1.2.4 GENERAL <i>IN VITRO</i> LICENSE .....	1-8
1.3 OTHER REQUIREMENTS .....	1-9
1.3.1 THE "AS LOW AS IS REASONABLY ACHIEVABLE (ALARA)" CONCEPT .....	1-9
1.3.2 WRITTEN DIRECTIVE (WD) PROCEDURES .....	1-10
1.3.3 TIMELY NOTIFICATION OF TRANSFER OF CONTROL .....	1-10
1.3.4 TIMELY NOTIFICATION OF BANKRUPTCY PROCEEDINGS .....	1-11
1.4 OMB CLEARANCES .....	1-11
2 AGREEMENT STATES .....	2-1
3 MANAGEMENT RESPONSIBILITY .....	3-1
4 APPLICABLE REGULATIONS .....	4-1
5 HOW TO FILE .....	5-1
5.1 PREPARING AN APPLICATION .....	5-1
5.2 ELECTRONIC APPLICATION .....	5-2
6 WHERE TO FILE .....	6-1
7 LICENSE FEES .....	7-1
8 CONTENTS OF AN APPLICATION .....	8-1
ITEMS FOR WHICH A RESPONSE FROM APPLICANT IS REQUIRED ON NRC FORM 313 .....	8-3
8.1 ITEM 1: LICENSE ACTION TYPE .....	8-5
8.2 ITEM 2: APPLICANT'S NAME AND MAILING ADDRESS .....	8-5
8.3 ITEM 3: ADDRESS(ES) WHERE LICENSED MATERIAL WILL BE USED OR POSSESSED .....	8-6
8.4 ITEM 4: PERSON TO BE CONTACTED ABOUT THIS APPLICATION .....	8-6
8.5 ITEM 5: RADIOACTIVE MATERIAL .....	8-7
8.6 ITEM 5: SEALED SOURCES AND DEVICES .....	8-10
8.7 ITEM 5: RECORDKEEPING FOR DECOMMISSIONING AND FINANCIAL ASSURANCE .....	8-11
8.8 ITEM 6: PURPOSE(S) FOR WHICH LICENSED MATERIAL WILL BE USED .....	8-13
8.9 ITEM 7: INDIVIDUAL(S) RESPONSIBLE FOR RADIATION SAFETY PROGRAM AND THEIR TRAINING AND EXPERIENCE .....	8-15
8.10 ITEM 7: RADIATION SAFETY OFFICER (RSO) .....	8-17
8.11 ITEM 7: AUTHORIZED USERS (AUs) .....	8-20
8.12 ITEM 7: AUTHORIZED NUCLEAR PHARMACIST (ANP) .....	8-22
8.13 ITEM 7: AUTHORIZED MEDICAL PHYSICIST (AMP) .....	8-24
8.14 ITEM 9: FACILITIES AND EQUIPMENT .....	8-26
8.15 ITEM 9: FACILITY DIAGRAM .....	8-26
8.16 ITEM 9: RADIATION MONITORING INSTRUMENTS .....	8-29
8.17 ITEM 9: DOSE CALIBRATOR AND OTHER EQUIPMENT USED TO MEASURE DOSAGES OF UNSEALED BYPRODUCT MATERIAL .....	8-31
8.18 ITEM 9: THERAPY UNIT — CALIBRATION AND USE .....	8-32
8.19 ITEM 9: OTHER EQUIPMENT AND FACILITIES .....	8-34
8.20 ITEM 10: RADIATION PROTECTION PROGRAM .....	8-37
8.21 ITEM 10: SAFETY PROCEDURES AND INSTRUCTIONS .....	8-37
8.22 ITEM 10: OCCUPATIONAL DOSE .....	8-39

CONTENTS

8.23	ITEM 10: AREA SURVEYS .....	8-42
8.24	ITEM 10: SAFE USE OF UNSEALED LICENSED MATERIAL .....	8-45
8.25	ITEM 10: SPILL PROCEDURES .....	8-46
8.26	ITEM 10: INSTALLATION, MAINTENANCE, ADJUSTMENT, REPAIR, AND INSPECTION OF THERAPY DEVICES CONTAINING SEALED SOURCES .....	8-46
8.27	ITEM 10: MINIMIZATION OF CONTAMINATION .....	8-48
8.28	ITEM 11: WASTE MANAGEMENT .....	8-48
8.29	ITEM 12: FEES .....	8-51
8.30	ITEM 13: CERTIFICATION .....	8-51
	<b>PROGRAM-RELATED GUIDANCE -- NO RESPONSE REQUIRED FROM APPLICANTS</b>	
	<b>ON NRC FORM 313 .....</b>	<b>8-53</b>
8.31	SAFETY INSTRUCTION FOR INDIVIDUALS WORKING IN OR FREQUENTING RESTRICTED AREAS .....	8-55
8.32	PUBLIC DOSE .....	8-56
8.33	OPENING PACKAGES .....	8-57
8.34	PROCEDURES FOR ADMINISTRATIONS WHEN A WRITTEN DIRECTIVE IS REQUIRED .....	8-58
8.35	RELEASE OF PATIENTS OR HUMAN RESEARCH SUBJECTS .....	8-58
8.36	MOBILE MEDICAL SERVICE .....	8-59
8.37	AUDIT PROGRAM .....	8-60
8.38	OPERATING AND EMERGENCY PROCEDURES .....	8-62
8.39	MATERIAL RECEIPT AND ACCOUNTABILITY .....	8-64
8.40	ORDERING AND RECEIVING .....	8-64
8.41	SEALED SOURCE INVENTORY .....	8-65
8.42	RECORDS OF DOSAGES AND USE OF BRACHYTHERAPY SOURCE .....	8-65
8.43	RECORDKEEPING .....	8-67
8.44	REPORTING .....	8-67
8.45	LEAK TESTS .....	8-68
8.46	SAFETY PROCEDURES FOR TREATMENTS WHEN PATIENTS ARE HOSPITALIZED ...	8-69
8.47	TRANSPORTATION .....	8-70
9	AMENDMENTS AND RENEWALS TO A LICENSE .....	9-1
10	APPLICATIONS FOR EXEMPTIONS .....	10-1
11	TERMINATION OF ACTIVITIES .....	11-1

**APPENDICES**

**APPENDICES A-H: FORMS AND SAMPLES**

A NRC Form 313“Application for Material License” ..... A-1  
 B NRC Form 313A“Training and Experience and Preceptor Statement” ..... B-1  
 C License Application Checklists ..... C-1  
 D Documentation of Training and Experience for Authorized User, Radiation Safety Officer, Authorized Nuclear Pharmacist, and Authorized Medical Physicist ..... D-1  
 E Sample License Application ..... E-1  
 F Sample Licenses ..... F-1  
 G Information Needed for Transfer of Control ..... G-1  
 H NRC Form 314 “Certificate of Disposition of Materials” ..... H-1

**APPENDICES I-W: MODEL PROCEDURES**

I Typical Duties and Responsibilities of the Radiation Safety Officer and Sample Delegation of Authority . . I-1  
 J Model Training Program ..... J-1  
 K Radiation Monitoring Instrument Specifications and Model Survey Instrument Calibration Program .... K-1  
 L Model Medical Licensee Audit ..... L-1  
 M Model Procedures for an Occupational Dose Program ..... M-1  
 N Model Emergency Procedures ..... N-1  
 O Model Procedures for Ordering and Receiving Packages ..... O-1  
 P Model Procedure for Safely Opening Packages Containing Radioactive Material ..... P-1  
 Q Model Leak Test Program ..... Q-1  
 R Model Procedure for Area Surveys ..... R-1  
 S Model Procedures for Developing, Maintaining, and Implementing Written Directives ..... S-1  
 T Model Procedures for Safe Use of Unsealed Licensed Material ..... T-1  
 U Model Procedure for Release of Patients or Human Research Subjects Administered Radioactive Materials ..... U-1  
 V Guidance for Mobile Medical Services ..... V-1  
 W Model Procedure for Waste Disposal by Decay-In-Storage, Generator Return, and Licensed Material Return ..... W-1

**APPENDICES X-BB: Recordkeeping and Reporting Requirements; DOT Rules for Shipping; References; Public Comments on Drafts and NRC Responses**

X Recordkeeping ..... X-1  
 Y Reporting ..... Y-1  
 Z Summary of DOT Requirements for Transportation of Type A or Type B Quantities of Licensed Material Z-1  
 AA List of Documents Considered in Development of this NUREG ..... AA-1

The following Appendix will be published as a separate document:

**APPENDIX BB: Summary of Public Comments on Drafts and NRC Responses ..... BB-1**

CONTENTS

**FIGURES**

Figure 2.1 U.S. Map. *Location of NRC Offices and Agreement States* ..... 2-2  
Figure 8.1 Facility Diagram for Nuclear Medicine Suite ..... 8-27  
Figure 8.2 Annual Occupational Dose Limits for Adults ..... 8-39

**TABLES**

Table 1.1 Sections of NUREG-1556, Volume 9 that Applicants for a Particular Type of Use Should Review ..... 1-2

Table 2.1 Who Regulates the Activity? ..... 2-1

Table 8.1 Worksheet for Determining Need for Financial Assurance for Sealed Sources ..... 8-12

Table C.1 Applicability Table ..... C-1

Table C.2 Items 5 and 6 on NRC Form 313: Radioactive Material And Use ..... C-4

Table C.3 Items 7 through 11 on NRC Form 313: Training & Experience, Facilities & Equipment, Radiation Protection Program, and Waste Disposal ..... C-7

Table K.1 Typical Survey Instruments ..... K-2

Table M.1 Investigational Levels ..... M-4

Table N.1 Relative Hazards of Common Radionuclides ..... N-2

Table R.1 Ambient Dose Rate Trigger Levels ..... R-2

Table R.2 Surface Contamination Levels in Restricted Areas (dpm/100 cm<sup>2</sup>) ..... R-3

Table R.3 Surface Contamination Levels in Unrestricted Areas (dpm/100 cm<sup>2</sup>) ..... R-4

Table R.4 Grouping of Radioisotopes for Alternate Survey Frequency ..... R-5

Table R.5 Classification of Laboratories for Alternate Survey Frequency ..... R-5

Table R.6 Modifying Factors for Alternate Survey Frequency ..... R-6

Table U.1 Activities and Dose Rates for Authorizing Patient Release ..... U-4

Table U.2 Activities and Dose Rates Above Which Instructions Should Be Given When Authorizing Patient Release ..... U-7

Table U.3 Activities of Radiopharmaceuticals that Require Instructions and Records When Administered to Patients Who are Breast-Feeding an Infant or Child ..... U-8

Table U.4 Summary of Release Criteria, Required Instructions to Patients, and Records to Be Maintained . U-14

Table U.5 Half-Lives and Exposure Rate Constants of Radionuclides Used in Medicine ..... U-16

Table U.6 Uptake Fractions and Effective Half-Lives for Iodine-131 Treatments ..... U-23

## FOREWORD

This report, NUREG-1556, Vol. 9, "Consolidated Guidance About Materials Licenses: Program-Specific Guidance About Medical Use Licenses," dated October 2002, is one of twenty volumes in NRC's NUREG-1556 series addressing its materials licensing process. This report is intended for use by applicants, licensees, NRC license reviewers, and other NRC license personnel addressing the medical use of byproduct material. Below is a list of volumes currently included in the NUREG-1556 series:

Vol. No.	Volume Title	Status
1	Program-Specific Guidance About Portable Gauge Licenses	Final Report
2	Program-Specific Guidance About Industrial Radiography Licenses	Final Report
3	Applications for Sealed Source and Device Evaluation and Registration	Final Report
4	Program-Specific Guidance About Fixed Gauge Licenses	Final Report
5	Program-Specific Guidance About Self-Shielded Irradiator Licenses	Final Report
6	Program-Specific Guidance About 10 CFR Part 36 Irradiator Licenses	Final Report
7	Program-Specific Guidance About Academic, Research and Development, and Other Licenses of Limited Scope	Final Report
8	Program-Specific Guidance About Exempt Distribution Licenses	Final Report
9	Program-Specific Guidance About Medical Use Licenses	Final Report
10	Program-Specific Guidance About Master Materials Licenses	Final Report
11	Program-Specific Guidance About Licenses of Broad Scope	Final Report
12	Program-Specific Guidance About Possession Licenses for Manufacturing and Distribution	Final Report
13	Program-Specific Guidance About Commercial Radiopharmacy Licenses	Final Report
14	Program-Specific Guidance About Well Logging, Tracer, and Field Flood Study Licenses	Final Report
15	Guidance About Changes of Control and About Bankruptcy Involving Byproduct, Source, or Special Nuclear Materials Licenses	Final Report
16	Program-Specific Guidance About Licenses Authorizing Distribution To General Licensees	Final Report
17	Program-Specific Guidance About Licenses for Special Nuclear Material of Less Than Critical Mass	Final Report
18	Program-Specific Guidance About Service Provider Licenses	Final Report
19	Guidance For Agreement State Licensees About NRC Form 241 "Report of Proposed Activities in Non-Agreement States, Areas of Exclusive Federal Jurisdiction, or Offshore Waters" and Guidance For NRC Licensees Proposing to Work in Agreement State Jurisdiction (Reciprocity)	Final Report
20	Guidance About Administrative Licensing Procedures	Final Report

A team composed of NRC staff and staff from state departments of health prepared the initial draft of this document, which was published for public comment in August 1998. The NRC staff members were P.A. Lanzisera, A.R. Jones, R.G. Gattone, R.D. Reid. A revised draft was

## FOREWORD

published in March 2002. Appendix Z of the March 2002 draft included a summary of comments on the 1998 draft and NRC responses. The NRC held two public workshops, on April 25 and April 30, 2002, to receive stakeholder comments on the March 2002 draft. The NRC also received written public comments during a 60-day comment period (April 5 to June 4, 2002). A summary and analysis of both sets of comments will be published as a separate document: Appendix BB to NUREG-1556 Vol. 9. This document will be available as noted inside the front cover of this document. Appendix BB to NUREG-1556, Vol. 9 will also be available on the NRC's web site <<http://www.nrc.gov>> in the Electronic Reading Room.

NRC staff is also producing a set of Questions and Answers (Q&As) on implementation of Part 35. Applicants may review the Q&As posted on the NRC's web page on the Medical Use of Byproduct Material <<http://www.nrc.gov/materials/miau/miau-reg-initiatives/by-product.html>> as another source of guidance about implementation of revised Part 35.

In addition to combining and updating the guidance for applicants and licensees previously found in numerous Regulatory Guides, Policy and Guidance Directives, draft Regulatory Guides, Standard Review Plans, and Information Notices, this guidance incorporates input from stakeholders received in the public workshops and comments.

This report follows the risk-informed, performance-based approach adopted for revisions to 10 CFR Part 35. It reduces the amount of information submitted by an applicant seeking to possess and use certain quantities of byproduct material for medical use. In a number of instances, the regulations found in 10 CFR Part 35 and reflected in this report do not require the submission of detailed procedures. Instead, applicants are requested to confirm that they have developed and will implement and maintain procedures required by Part 35, but they are not required to submit those procedures as part of their license application. This report contains appendices containing suggested procedures that applicants may consider. The risk-informed, performance-based approach to the regulation of NRC licensed materials is also being emphasized in the inspection and enforcement arena.

This document addresses those topics that an applicant must provide in preparing a license application on NRC Form 313. The report also includes descriptions of certain key elements of a medical use program that do not require a response on Form 313. This material is presented for clarification only.

NUREG-1556, Vol. 9, is not a substitute for NRC regulations. The approaches and methods described in this report are provided for information only. Guidance in this document represents one means acceptable to the staff of complying with NRC regulations and is not intended to be the only means of satisfying the requirements for licensing.

Complementary guidance on inspection procedures for inspections of medical use licensees is contained in the following documents available at the NRC's web page on the Medical Use of Byproduct Material <<http://www.nrc.gov/materials/miau/miau-reg-initiatives/by-product.html>>.

- Inspection Procedures in the 87100 series:

- "Nuclear Medicine Programs — Written Directive Not Required,"
- "Nuclear Medicine Programs — Written Directive Required,"
- "Brachytherapy Programs,"
- "Medical Gamma Stereotactic Radiosurgery and Teletherapy Programs," and
- "Medical Broad Scope Programs."

*Susan M. Hand for*

---

Donald A. Cool, Director  
Division of Industrial and Medical Nuclear Safety  
Office of Nuclear Material Safety and Safeguards

## ACKNOWLEDGMENTS

The guidance development team thanks the individuals listed below for assisting in the development and review of the report. All participants provided valuable insights, observations, and recommendations.

### The Participants

Brown, Carrie	Howe, Allen G.
Bolling, Lloyd A.	Howe, Donna-Beth
Brown, Frederick D.	Jones, Samuel Z.
Chidakel, Susan S.	Lohaus, Paul H.
Cool, Donald A.	Merchant, Sally L.
Decker, Thomas R.	Null, Kevin G.
DelMedico, Joseph R.	Psyk, Linda M.
Euchner, Jennifer	Rothschild, Marjorie U.
Flack, Diane S.	Roe, Mary Louise
Frant, Susan M.	Siegel, M.D., Barry
Haney, Catherine	Torres, Betty Ann
Henderson, Pamela J.	Treby, Stuart A.
Henson, Jay L.	Turner, Anita L.
Hill, Thomas E.	Walter, David
Holahan, Patricia K.	Whitten, Jack E.
Holden, Cornelius F., Jr.	Young, Thomas F.
Howard, Marcia	

We acknowledge the assistance of Francis "Chip" X. Cameron for leading two, facilitated round-table discussions and the participation of stakeholders in public meetings held at NRC headquarters on April 25 and 30, 2002.

The team also thanks Helen Chang, Dianne Geshen, Rolonda Jackson, Tamra King, and Agi Seaton of Computer Sciences Corporation and Craig Dean, Brantley Fry, Margo Brown, Vanessa Brown, and Kyle Jones of ICF Consulting Incorporated for their assistance in the preparation of this document.

The following individuals are recognized for their contribution to supporting documents that formed a basis for this report:

Ayres, Robert	Minnick, Sheri A.
Bhalla, Neelam	Paperiello, Carl A.
Brown, Carrie	Schlueter, Janet R.
Brown, Keith D.	Smith, James A.
Frazier, Cassandra F.	Taylor, Torre M.
Fuller, Mike L.	Trottier, Cheryl A.
Merchant, Sally L.	

## ABBREVIATIONS

AAPM	American Association of Physicists in Medicine
ACMUI	Advisory Committee on the Medical Use of Isotopes
ACR	American College of Radiology
ALARA	as low as is reasonably achievable
ALI	annual limit on intake
AMP	Authorized Medical Physicist
ANP	Authorized Nuclear Pharmacist
ANSI	American National Standards Institute
AU	Authorized User
bkg	background
BPR	Business Process Redesign
Bq	Becquerel
CFR	Code of Federal Regulations
Ci	Curie
cc	centimeter cubed
cm <sup>2</sup>	square centimeter
Co-57	cobalt-57
Co-60	cobalt-60
cpm	counts per minute
Cs-137	cesium-137
DAC	derived air concentration
DOT	United States Department of Transportation
dpm	disintegrations per minute
FDA	United States Food and Drug Administration
GM	Geiger-Mueller
GPO	Government Printing Office
GSR	gamma stereotactic radiosurgery
HDR	high dose-rate
I-125	iodine-125
I-131	iodine-131

## ABBREVIATIONS

IN	Information Notice
IP	Inspection Procedure
Ir-192	iridium-192
LDR	low dose-rate
mCi	millicurie
ml	milliliter
mo-99	molybdenum-99
mR	milliroentgen
mrem	millirem
mSv	millisievert
NaI(Tl)	sodium iodide (thallium doped)
NCRP	National Council on Radiation Protection and Measurements
NIST	National Institute of Standards and Technology
NRC	United States Nuclear Regulatory Commission
NVLAP	National Voluntary Laboratory Accreditation Program
OCFO	Office of the Chief Financial Officer
OCR	optical character reader
OMB	Office of Management and Budget
OSL	optically stimulated luminescence dosimeters
P-32	phosphorus-32
Pd-103	palladium-103
PDR	pulsed dose-rate
P&GD	Policy and Guidance Directive
QA	quality assurance
Ra-226	radium-226
RG	Regulatory Guide
RIS	Regulatory Issue Summary
RSC	Radiation Safety Committee
RSO	Radiation Safety Officer
SDE	shallow-dose equivalent
SI	International System of Units (abbreviated SI from the French Le Système Internationale d'Unites)

## ABBREVIATIONS

Sr-90	strontium-90
SSDR	Sealed Source and Device Registration
std	standard
Sv	Sievert
TAR	Technical Assistance Request
Tc-99m	technetium-99m
TEDE	total effective dose equivalent
TLD	thermoluminescent dosimeters
U-235	uranium-235
WD	written directive
Xe-133	xenon-133
$\mu$ Ci	microcurie
%	percent

# 1 OVERVIEW

## 1.1 PURPOSE OF REPORT

This report is intended to provide guidance on two topics to individuals who are preparing an application for a license for the medical use of byproduct material as well as NRC staff who review applications:

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓

- (1) Preparation of a license application using NRC Form 313 "Application for Material License," including supplemental NRC Form 313A, "Training and Experience and Preceptor Statement;" and
- (2) NRC's criteria for evaluating a medical use license application.

This report provides guidance for the following types of medical uses of byproduct material:

- Use of unsealed byproduct material for uptake, dilution, and excretion studies for which a written directive is not required under 10 CFR 35.40 (see Subpart D, 10 CFR 35.100-190);
- Use of unsealed byproduct material for imaging and localization studies for which a written directive is not required under 10 CFR 35.40 (see Subpart D, 10 CFR 35.200-290);
- Use of unsealed byproduct material for which a written directive is required under 10 CFR 35.40 (see Subpart E, 10 CFR 35.300-390);
- Use of sources for manual brachytherapy (see Subpart F, 10 CFR 35.400-490);
- Use of sealed sources for diagnosis (see Subpart G, 10 CFR 35.500);
- Use of a sealed source in a photon emitting remote afterloader unit, teletherapy unit, or gamma stereotactic radiosurgery unit (see Subpart H, 10 CFR 35.600-690); and
- Other medical uses of byproduct material or radiation from byproduct material not specifically covered by 10 CFR Part 35, Subparts 35.100 through 35.600 (see Subpart K, 10 CFR 35.1000).

To assist license applicants, this guide includes text boxes at the beginning of each section to indicate the type of use to which the guidance pertains (identified by the pertinent section of 10 CFR Part 35). These boxes are intended to guide the applicant through the sections of the guidance that are relevant to the applicant's particular type of use of byproduct material. A check indicates that applicants for that type of use should review the guidance section. Some of the checks have asterisks next to them. These asterisks indicate that there are conditions or limitations in that particular section of the guidance relating to the applicants who are subject to the checked section of the rule. Table 1.1 summarizes the material in the text boxes.

OVERVIEW

**Table 1.1 Sections of NUREG-1556, Volume 9 that Applicants for a Particular Type of Use Should Review**

NUREG-1556 - Volume 9, Section:	Type of Use						
	100	200	300	400	500	600	1000
8.1 License Action Type	•	•	•	•	•	•	•
8.2 Applicant's Name and Mailing Address	•	•	•	•	•	•	•
8.3 Address(es) Where Licensed Material Will Be Used or Possessed	•	•	•	•	•	•	•
8.4 Person to Be Contacted about This Application	•	•	•	•	•	•	•
8.5 Radioactive Material	•	•	•	•	•	•	•
8.6 Sealed Sources and Devices				•	•	•	•
8.7 Recordkeeping for Decommissioning and Financial Assurance	•	•	•	•	•	•	•
8.8 Purpose(s) for which Licensed Material Will Be Used	•	•	•	•	•	•	•
8.9 Individual(s) Responsible for Radiation Safety Program and their Training and Experience	•	•	•	•	•	•	•
8.10 Radiation Safety Officer (RSO)	•	•	•	•	•	•	•
8.11 Authorized User (AU)	•	•	•	•	•	•	•
8.12 Authorized Nuclear Pharmacist	•	•	•				•
8.13 Authorized Medical Physicist (AMP)				•	•	•	•
8.14 Facilities and Equipment	•	•	•	•	•	•	•
8.15 Facility Diagram	•	•	•	•	•	•	•
8.16 Radiation Monitoring Instruments	•	•	•	•	•	•	•
8.17 Dose Calibrator and Other Equipment Used to Measure Dosages of Unsealed Material	•	•	•				•
8.18 Therapy Unit - Calibration and Use				•	•	•	•
8.19 Other Equipment and Facilities	•	•	•	•	•	•	•
8.20 Radiation Protection Program	•	•	•	•	•	•	•
8.21 Safety Procedures and Instructions						•	•
8.22 Occupational Dose	•	•	•	•	•	•	•
8.23 Area Surveys	•	•	•	•	•	•	•
8.24 Safe Use of Unsealed Licensed Material	•	•	•				•
8.25 Spill Procedures	•	•	•	•	•		•

Table 1.1 (continued)

NUREG-1556 - Volume 9, Section	Type of Use						
	100	200	300	400	500	600	1000
8.26 Installation, Maintenance, Adjustment, Repair, and Inspection of Therapy Devices Containing Sealed Sources	•		•		•	•	•
8.27 Minimization of Contamination	•	•	•	•	•	•	•
8.28 Waste Management	•	•	•	•	•	•	•
8.29 Fees	•	•	•	•	•	•	•
8.30 Certification	•	•	•	•	•	•	•
<b>PROGRAM-RELATED GUIDANCE - NO RESPONSE FROM APPLICANTS ON NRC FORM 313</b>							
8.31 Safety Instruction for Individuals Working In or Frequenting Restricted Areas	•	•	•	•	•	•	•
8.32 Public Dose	•	•	•	•	•	•	•
8.33 Opening Packages	•	•	•	•	•	•	•
8.34 Procedures for Administrations Requiring Written Directive	•		•	•	•	•	•
8.35 Patient or Human Research Subject Release	•		•	•	•		•
8.36 Mobile Medical Service	•	•	•	•	•	•	•
8.37 Audit Program	•	•	•	•	•	•	•
8.38 Operating and Emergency Procedures	•	•	•	•	•	•	•
8.39 Material Receipt and Accountability	•	•	•	•	•	•	•
8.40 Ordering and Receiving	•	•	•	•	•	•	•
8.41 Sealed Source Inventory	•	•	•	•	•	•	•
8.42 Records of Dosages and Use of Brachytherapy Source	•	•	•	•	•		•
8.43 Recordkeeping	•	•	•	•	•	•	•
8.44 Reporting	•	•	•	•	•	•	•
8.45 Leak Tests	•	•	•	•	•	•	•
8.46 Safety Procedures for Treatments When Patients are Hospitalized	•		•	•	•	•	•
8.47 Transportation	•	•	•	•	•	•	•

Applicants also should be aware that 10 CFR Part 35 contains general information, administrative requirements, and technical requirements that are pertinent to some or all of the types of use listed above (see 10 CFR 35.1 through 35.92).

## OVERVIEW

This report is intended to consolidate into one document guidance that relates to satisfying regulations other than 10 CFR Part 35 that apply to medical use licensees, including the following:

- Provisions of 10 CFR Part 20 that relate to radiation safety.
- Provisions of 10 CFR Part 30 that relate to licensing (e.g., §30.33).

This report does not address certain aspects of licensing and radiation safety for the medical use of byproduct materials. In particular, applicants and licensees should consider the following:

- NUREG-1556, Volume 11, "Consolidated Guidance about Materials Licenses: Program-Specific Guidance About Licenses of Broad Scope," dated April 1999, provides additional licensing guidance on medical use programs of broad scope. Section 1.2.1 below provides a general discussion on specific licenses of broad scope.
- 10 CFR Part 19, "Notices, Instructions and Reports to Workers: Inspection and Investigations."
- 10 CFR Part 20, "Standards for Protection Against Radiation," and other regulatory requirements potentially applicable to medical use licensees listed in Section 4 below.
- 10 CFR Part 21, "Reporting of Defects of Noncompliance."
- This report does not address the commercial aspects of manufacturing, distribution, and service of sources containing byproduct material in devices. NUREG-1556, Volumes 12, 13, and 18 provide additional licensing guidance.
- This report does not describe the licensing, possession, or use of pacemakers, which are licensed under 10 CFR Part 70, "Domestic Licensing of Special Nuclear Material." However, a sample pacemaker license is included in Appendix F.

As a guidance document intended to assist a wide variety of applicants, this report contains a considerable amount of information about how licensees may choose to implement their programs to meet NRC regulatory requirements. The information in this document is not intended to impose any conditions beyond those required by the regulations in 10 CFR. This report provides specific guidance on what information should be submitted in an application to satisfy NRC requirements. Except for procedures required by Subpart H of 10 CFR Part 35, written procedures do not need to be submitted as part of the license application.

Guidance and model procedures provided in this NUREG that are not required to be submitted are for illustrative purposes to guide licensees in developing their programs. Use of the word "should" implies "may" and is not intended to mean "must" or "shall;" the procedures provided in this guidance are intended to serve only as examples.

Sections 1 through 7 of this document provide background information. Section 8 describes, item-by-item, the information that should be provided in Items 1 through 11 of NRC Form 313,

in completing a license application. The format within this document for each item of technical information is as follows:

- **Regulations** – references the regulations applicable to the item;
- **Criteria** – outlines the criteria used to judge the adequacy of the applicant's response;
- **Discussion** – provides additional information on the topic sufficient to meet the needs of most readers; and
- **Response from Applicant** – provides suggested response(s) or indicates that no response is needed on that topic during the initial licensing process.

Some sections of the guidance include references to other documents that may be useful to the applicant. Appendix AA provides a complete list of documents used to prepare or referenced in the guidance. While specific availability information is included for some reference documents, the documents also may be accessed at the NRC Public Document Room, which is located at NRC Headquarters in Rockville, Maryland, or the NRC Electronic Reading Room at <http://www.nrc.gov>. See the Notice of Availability on the inside front cover of this report for more information.

When NRC Form 313 does not have sufficient space to provide full responses to Items 5 through 11, provide the information on separate attachments, label the attachments to indicate which item is being addressed, and submit the attachments with the completed NRC Form 313.

Appendices to this report provide the following supplementary information:

- Appendices A and B provide sample application forms;
- Appendix C provides license application checklists for responding to Items 5-11;
- Appendix D describes how to fill out NRC Form 313A;
- Appendix E includes a sample application;
- Appendix F provides sample licenses;
- Appendices G and H provide information regarding required submissions;
- Appendices I through W provide model procedures;
- Appendices X through AA provide reference materials; and
- Appendix BB, to be published as a separate document, will provide a summary of public comments on drafts and NRC responses.

In this document, "dose" or "radiation dose" means absorbed dose, dose equivalent, effective dose equivalent, committed dose equivalent, committed effective dose equivalent, or total effective dose equivalent (TEDE). These quantities are defined in 10 CFR Part 20 and are expressed in units of rem and its SI equivalent, the Sievert (Sv) (1 rem = 0.01 Sv). (The quantities absorbed dose and exposure, and their associated units, the rad and the roentgen, are

## OVERVIEW

not used in 10 CFR Part 20 to specify dose limits.) Furthermore, the byproduct materials commonly used in medicine emit beta and photon radiation, for which the quality factor is 1; a useful rule of thumb is an exposure of 1 roentgen is equivalent to an absorbed dose of 1 rad and dose equivalent of 1 rem.

This NUREG not only updates the information and guidance provided in Revision 2 of RG 10.8, "Guide for the Preparation of Applications for Medical Use Programs," but also revises the format in which it is presented to assist with the preparation of a medical use license. Revision 2 was issued in August 1987 to provide guidance for the revised 10 CFR Part 35, which became effective April 1, 1987. Since then, 10 CFR Part 35 has been amended a number of times. Technology-specific information has been revised and expanded to include technologies that are now more commonly used, for example, computerized remote afterloading brachytherapy and gamma stereotactic radiosurgery (GSR).

## 1.2 TYPES OF LICENSES

NRC defines "Medical use" as "the intentional internal or external administration of byproduct material, or the radiation from byproduct material, to patients or human research subjects under the supervision of an authorized user" (10 CFR 35.2). An Authorized User is defined as "a physician, dentist, or podiatrist" who meets the training and experience requirements specified in 10 CFR Part 35 and who is identified as an authorized user on an NRC or Agreement State license, on a permit issued by a Commission master material licensee or a Commission master material permittee that is authorized to permit the medical use of byproduct material; a permit issued by a Commission or Agreement State broad scope licensee authorized to permit the medical use of byproduct material that is authorized to permit the medical use of byproduct material (10 CFR 35.2).

NRC issues two types of specific licenses for the medical use of byproduct material in medical practices and facilities:

- the specific license of limited scope (see Section 1.2.1), and
- the specific license of broad scope (see Section 1.2.2).

Medical use includes research involving human subjects, which may occur under either limited scope or broad scope specific licenses (see Section 1.2.3).

NRC also issues a general license pursuant to 10 CFR 31.11, under which a physician, veterinarian in the practice of veterinary medicine, clinical laboratory, or hospital may use byproduct material for certain *in vitro* clinical or laboratory testing. Such testing may not involve internal or external administration of byproduct material, or the radiation therefrom, to human beings or animals (see Section 1.2.4).

NRC usually issues a single byproduct material license to cover an entire radionuclide program. (Note, however, that nuclear-powered pacemakers are licensed separately under 10 CFR Part 70.) A license including teletherapy may also contain the authorization for source material (i.e., depleted uranium) used as shielding in many teletherapy units and a license may include authorization for possession of sealed sources to be used to calibrate dose calibration devices.

NRC may issue separate licenses to individual licensees for different medical uses. However, NRC does not usually issue separate licenses to different departments in a medical facility or to individuals employed by a medical facility or with whom the medical facility has contracted. Only the facility's management may sign the license application.

Applicants should study this report, related guidance, and all applicable regulations carefully before completing NRC Forms 313 and 313A. NRC expects licensees to provide information on specific aspects of the proposed radiation protection program in attachments to NRC Form 313. When necessary, NRC may ask the applicant for additional information in order to gain reasonable assurance that an adequate radiation protection program has been established.

After a license is issued, the licensee must conduct its program in accordance with the following:

- Statements, representations, and procedures contained in the application and in correspondence with NRC, when incorporated into a license by reference;
- Terms and conditions of the license; and
- NRC regulations.

In 10 CFR 30.9, NRC requires that the information in the application be complete and accurate in all material aspects. Information is considered material if it has the ability to change or affect an agency decision on issuing the license.

### 1.2.1 SPECIFIC LICENSE OF LIMITED SCOPE

NRC issues specific medical licenses of limited scope to private or group medical practices and to medical institutions. A medical institution is an organization in which more than one medical discipline is practiced. In general, individual physicians or physician groups located within a licensed medical facility (e.g., hospital) may not apply for a separate license because 10 CFR 30.33(a)(2) refers to the applicant's facilities. Since a physician group does not normally have control over the facilities, the hospital remains responsible for activities conducted on its premises and must apply for the license. On specific licenses of limited scope, the authorized users are specifically listed in the license.

Byproduct material may be administered to patients on an inpatient (i.e., hospitalized) or outpatient basis. For patients to whom byproduct material is administered and who are not releasable under 10 CFR 35.75, inpatient facilities are required. In general, facilities for private and group practices do not include inpatient rooms and, therefore, procedures requiring hospitalization of the patient under 10 CFR 35.75 cannot be performed.

A specific license of limited scope may also be issued to an entity requesting to perform mobile medical services (10 CFR 35.80, 10 CFR 35.647). A medical institution or a private or group practice may apply for authorization to use byproduct material in a mobile medical service.

## OVERVIEW

### 1.2.2 SPECIFIC LICENSE OF BROAD SCOPE

Medical institutions that provide patient care and conduct research programs that use radionuclides for *in vitro*, animal, and medical procedures may request a specific license of broad scope in accordance with 10 CFR Part 33. No medical use of byproduct material, including research involving human subjects, may be conducted without an authorization in a license from the NRC or an Agreement State as provided in 10 CFR Part 35. The criteria for the various types of broad scope licenses are found in 10 CFR 33.13 through 10 CFR 33.17. Generally, NRC issues specific licenses of broad scope for medical use (i.e., licenses authorizing multiple quantities and types of byproduct material for medical use under Part 35 as well as other uses) to institutions that (1) have experience successfully operating under a specific license of limited scope; and (2) are engaged in medical research and routine diagnostic and therapeutic uses of byproduct material. NUREG-1556, Vol. 11, offers additional guidance to applicants for a specific license of broad scope.

### 1.2.3 RESEARCH INVOLVING HUMAN SUBJECTS

10 CFR 35.2 defines “medical use” to include the administration of byproduct material or radiation therefrom to human research subjects. Furthermore, 10 CFR 35.6, “Provisions for the protection of human research subjects,” addresses the protection of the rights of human subjects involved in research by medical use licensees. For these licensees, prior NRC approval is not necessary if the research is conducted, funded, supported, or regulated by another Federal Agency that has implemented the Federal Policy for the Protection of Human Subjects. Otherwise, the licensee must apply for a specific amendment and receive approval for the amendment before conducting such research. Whether or not a license amendment is required, licensees must obtain informed consent from human subjects and prior review and approval of the research activities by an Institutional Review Board in accordance with the meaning of those terms under the Federal Policy. In accordance with 10 CFR 35.6(a), research involving human subjects shall be conducted only with byproduct materials listed in the license for the uses authorized in the license.

### 1.2.4 GENERAL *IN VITRO* LICENSE

In 10 CFR 31.11, “General License for Use of Byproduct Material for Certain *In Vitro* Clinical or Laboratory Testing,” NRC establishes a general license authorizing physicians, veterinarians, clinical laboratories, and hospitals to receive, acquire, possess, or use small quantities of certain byproduct material for *in vitro* clinical or laboratory tests not involving “medical use” (i.e., not involving administration to humans). Section 31.11 explains the requirements for using the materials listed. If the general license alone meets the applicant’s needs, only NRC Form 483, “Registration Certificate – *In Vitro* Testing With Byproduct Material Under General License,” need be filed. Medical-use licensees authorized pursuant to 10 CFR Part 35 do not need to file the form.

NRC limits possession to a total of 200 microcuries of photon-emitting materials listed in 10 CFR 31.11 at any one time, at any one location of storage or use. The use of materials listed in 10 CFR 31.11 within the inventory limits of that section is subject only to the requirements of

that section and not to the requirements of 10 CFR Parts 19, 20, and 21, except as set forth in 10 CFR 31.11.

An applicant needing more than 200 microcuries of these materials must apply for a specific license and may request the increased inventory limit as a separate line item on NRC Form 313. This type of applicant generally requests an increased limit of 3 millicuries. If requesting an increased inventory limit, the applicant will be subject to the requirements of 10 CFR Parts 19, 20, and 21, including the requirements for waste disposal.

## 1.3 OTHER REQUIREMENTS

### 1.3.1 THE "AS LOW AS IS REASONABLY ACHIEVABLE (ALARA)" CONCEPT

10 CFR 20.1101, "Radiation Protection Programs," states that "each licensee shall develop, document, and implement a radiation protection program commensurate with the scope and extent of licensed activities ..." and "the licensee shall use, to the extent practical, procedures and engineering controls based upon sound radiation protection principles to achieve occupational doses and doses to members of the public that are...ALARA." This section also requires that licensees review the content of the radiation protection program and its implementation at least annually. The RSO is responsible for the day-to-day operation of the radiation protection program.

**References:** The following documents contain information, methods, and references useful to those who are establishing radiation protection programs to maintain radiation exposures at ALARA levels in medical facilities:

- RG 8.10, "Operating Philosophy for Maintaining Occupational Radiation Exposures ALARA."
- RG 8.18, "Information Relevant to Ensuring That Occupational Radiation Exposures at Medical Institutions Will Be ALARA."
- NUREG-0267, "Principles and Practices for Keeping Occupational Radiation Exposures at Medical Institutions ALARA."
- NUREG-1134, "Radiation Protection Training for Personnel Employed in Medical Facilities."
- Information directly related to radiation protection standards in 10 CFR Part 20 is contained in NUREG 1736, "Consolidated Guidance: 10 CFR Part 20 - Standards for Protection Against Radiation."

Applicants should consider the ALARA philosophy detailed in these reports when developing plans to work with licensed radioactive materials.

## OVERVIEW

### 1.3.2 WRITTEN DIRECTIVE (WD) PROCEDURES

10 CFR 35.41 requires certain medical use licensees to develop, implement, and maintain written procedures to provide high confidence that before each administration requiring a WD, the patient's identity is verified and the administration is in accordance with the WD. This regulation also specifies what an applicant must, at a minimum, address in these procedures. Appendix S provides further information on developing these procedures.

### 1.3.3 TIMELY NOTIFICATION OF TRANSFER OF CONTROL

Under 10 CFR 30.34(b) and 10 CFR 35.14(b) licensees must provide full information and obtain NRC's *written consent* before transferring control of the license, or, as some licensees refer to the process, "transferring the license."

Control may be transferred as a result of mergers, buyouts, or majority stock transfers. Although it is not NRC's intent to interfere with the business decisions of licensees, it is necessary for licensees to obtain NRC's written consent before transferring control of the license. This is to ensure the following:

- Radioactive materials are possessed, used, or controlled only by persons who have valid NRC licenses;
- Materials are properly handled and secured;
- Persons using these materials are competent and committed to implementing appropriate radiological controls;
- A clear chain of custody is established to identify who is responsible for final disposal of the material;
- Public health and safety are not compromised by the use of such materials.

As provided in 10 CFR 35.14(b), if only the licensee's name or mailing address changes, and the name change does not constitute a transfer of control of the license as described in 10 CFR 30.34(b), a licensee must file a written notification with NRC no later than 30 days after the dates of the change(s). Otherwise, prior NRC written consent must be given prior to the transfer.

Guidance on information to be supplied to the NRC when seeking approval for transfer of control of licensed material is available in Appendix G.

**Reference:** See the Notice of Availability on the inside front cover of this report to obtain copies of IN 97-30, "Control of Licensed Material during Reorganizations, Employee-Management Disagreements, and Financial Crises," dated June 3, 1997, and NUREG-1556, Vol. 15, "Program-Specific Guidance About Changes of Control and About Bankruptcy Involving Byproduct, Source, or Special Nuclear Material Licenses," dated November 2000.

These documents can also be accessed at NRC's web site, in the Electronic Reading Room at <<http://www.nrc.gov/reading-rm/doc-collections/gen-comm/info-notices/1997/in97030.html>> and <<http://www.nrc.gov/reading-rm/doc-collections/nuregs/staff/>>. Appendix G, excerpted from Appendix F of NUREG-1556, Vol. 15, identifies the information to be provided about transferring control.

#### **1.3.4 TIMELY NOTIFICATION OF BANKRUPTCY PROCEEDINGS**

Immediately following filing of a voluntary or involuntary petition for bankruptcy for or against a licensee, the licensee is required by 10 CFR 30.34(h) to notify the appropriate NRC Regional Administrator, in writing, identifying the bankruptcy court in which the petition was filed and the date of the filing.

Even though the licensee may have filed for bankruptcy, the licensee remains responsible for compliance with all regulatory requirements. NRC needs to know when licensees are in bankruptcy proceedings in order to determine whether all licensed material is accounted for and adequately controlled and whether there are any public health and safety concerns (e.g., contaminated facility). NRC shares the results of its determinations with other entities involved (e.g., trustees) so that health and safety issues can be resolved before bankruptcy actions are completed.

**Reference:** See the Notice of Availability on the inside front cover of this report to obtain copies of NUREG-1556, Vol. 15, "Program-Specific Guidance About Changes of Control and About Bankruptcy Involving Byproduct, Source, or Special Nuclear Material Licenses," dated November 2000.

#### **1.4 OMB CLEARANCES**

The information collection requirements in 10 CFR Parts 30 and 35 and NRC Form 313 have been approved under the Office of Management and Budget (OMB) Clearance Numbers 3150-0017, 3150-0010, and 3150-0120, respectively.

## 2 AGREEMENT STATES

Certain states, called Agreement States (see Figure 2.1), have entered into agreements with NRC that give them the authority to license and inspect byproduct, source, or special nuclear materials used or possessed within their borders. Any applicant other than a Federal Agency who wishes to possess or use licensed material in one of these Agreement States needs to contact the responsible officials in that state for guidance on preparing an application. These applications are filed with state officials, not with NRC.

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓


In the special situation of work at Federally-controlled sites in Agreement States, it is necessary to know the jurisdictional status of the land to determine whether NRC or the Agreement State has regulatory authority. NRC has regulatory authority over land determined to be under "exclusive Federal jurisdiction," while the Agreement State has jurisdiction over non-exclusive Federal jurisdiction land. Applicants and licensees are responsible for finding out, in advance, the jurisdictional status of the specific areas where they plan to conduct licensed operations. NRC recommends that applicants and licensees ask their local contacts for the Federal Agency controlling the site (e.g., contract officer, base environmental health officer, district office staff) to help determine the jurisdictional status of the land and to provide the information in writing, so that licensees can comply with NRC or Agreement State regulatory requirements, as appropriate. Additional guidance on determining jurisdictional status is found in All Agreement States Letter, SP-96-022, dated February 16, 1996, which is available from NRC upon request.

Table 2.1 provides a quick way to check on which agency has regulatory authority.

**Table 2.1 Who Regulates the Activity?**

Applicant and Proposed Location of Work	Regulatory Agency
Federal Agency regardless of location (except that Department of Energy [DOE] and, under most circumstances, its prime contractors are exempt from licensing [10 CFR 30.12])	NRC
Non-Federal entity in non-Agreement State, U.S. territory, or possession	NRC
Non-Federal entity in Agreement State at non-Federally controlled site	Agreement State
Non-Federal entity in Agreement State at Federally-controlled site <i>not</i> subject to exclusive Federal jurisdiction	Agreement State
Non-Federal entity in Agreement State at Federally-controlled site subject to exclusive Federal jurisdiction	NRC

## Locations of NRC Offices and Agreement States



\* The 19 Non-Agreement States include the District of Columbia and three states that have filed letters of intent - Minnesota, Pennsylvania, and Wisconsin

Figure 2.1 U.S. Map. Location of NRC Offices and Agreement States

**Reference:** The identity of Agreement States shown in the map in Figure 2.1 may change over time. A current list of Agreement States (including names, addresses, and telephone numbers of responsible officials) may be obtained upon request from NRC's Regional Offices. NRC Office of State and Tribal Programs (OSTP) also provides the current list of Agreement States at web site <<http://www.hsr.gov/nrc>>, under "Directories" and then under "State Program Directors."

The All Agreement States Letter, SP-96-022, dated February 16, 1996, is available by calling NRC's toll-free number at (800) 368-5642 and asking for OSTP. OSTP also provides this information at web site <<http://www.hsr.gov/nrc>>, under "NRC-State Letters."

### 3 MANAGEMENT RESPONSIBILITY

Regulations: 10 CFR 30.9; 10 CFR 35.12; 10 CFR 35.24.

NRC endorses the philosophy that effective radiation protection program management is vital to safe operations that comply with NRC regulatory requirements (see 10 CFR 35.24).

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓

“Management” refers to the chief executive officer or other individual having the authority to *manage, direct, or administer the licensee’s activities* or that person’s delegate or delegates (see 10 CFR 35.2).

To ensure adequate management involvement in accordance with 10 CFR 35.12(a) and 35.24(a), a management representative (i.e., chief executive officer or delegate) must sign the submitted application acknowledging management’s commitments to and responsibility for the following:

- Radiation protection, security, and control of radioactive materials, and compliance with regulations;
- Completeness and accuracy of the radiation protection records and all information provided to NRC (10 CFR 30.9);
- Knowledge about the contents of the license application;
- Compliance with current NRC and United States Department of Transportation (DOT) regulations and the licensee’s operating and emergency procedures;
- Provision of adequate financial and other resources (including space, equipment, personnel, time, and, if needed, contractors) to the radiation protection program to ensure that patients, the public, and workers are protected from radiation hazards;
- Appointment of a qualified individual who has agreed in writing to work as the RSO;
- Approval of qualified individual(s) to serve as Authorized Medical Physicists (AMPs), Authorized Nuclear Pharmacists (ANPs), and Authorized Users (AUs) for licensed activities.

For information on NRC inspection, investigation, enforcement, and other compliance programs, see the following:

- “General Statement of Policy and Procedures for NRC Enforcement Actions,” NUREG-1600;
- NRC Inspection Manual, Chapter 2800 “Materials Inspection Program;” and

## MANAGEMENT RESPONSIBILITY

- Inspection Procedures in the 87100 series:
  - “Nuclear Medicine Programs — Written Directive Not Required,”
  - “Nuclear Medicine Programs — Written Directive Required,”
  - “Brachytherapy Programs,”
  - “Medical Gamma Stereotactic Radiosurgery and Teletherapy Programs,” and
  - “Medical Broad Scope Programs.”

For availability of these documents see the Notice of Availability on the inside front cover of this report. NUREG-1600 is also available at NRC’s web site, <<http://www.nrc.gov/reading-rm/doc-collections/nuregs/staff/sr1600/>>.

## 4 APPLICABLE REGULATIONS

Regulations applicable to medical use licensees are listed below. Applicants should be sure to refer to up-to-date versions of regulations, which are available at NRC's web site at <http://www.nrc.gov/reading-rm/doc-collections/cfr/> in the "Electronic Reading Room"; printed copies available from the U.S. Government Printing Office are updated annually.

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓

- 10 CFR Part 2, "Rules of Practice for Domestic Licensing Proceedings and Issuance of Orders"
- 10 CFR Part 19, "Notices, Instructions and Reports to Workers: Inspection and Investigations"
- 10 CFR Part 20, "Standards for Protection Against Radiation"
- 10 CFR Part 21, "Reporting of Defects and Noncompliance"
- 10 CFR Part 30, "Rules of General Applicability to Domestic Licensing of Byproduct Material"
- 10 CFR Part 31, "General Domestic Licenses for Byproduct Material"
- 10 CFR Part 32, "Specific Domestic Licenses to Manufacture or Transfer Certain Items Containing Byproduct Material"
- 10 CFR Part 33, "Specific Domestic Licenses of Broad Scope for Byproduct Material"
- 10 CFR Part 35, "Medical Use of Byproduct Material"
- 10 CFR Part 40, "Domestic Licensing of Source Material"
- 10 CFR Part 70, "Domestic Licensing of Special Nuclear Material" (for pacemaker devices)
- 10 CFR Part 71, "Packaging and Transportation of Radioactive Material"

Part 71 requires that licensees or applicants who transport licensed material or who may offer such material to a carrier for transport must comply with the applicable requirements of the DOT that are found in 49 CFR Parts 170 through 189.

- 10 CFR Part 150, "Exemptions and Continued Regulatory Authority in Agreement States and in Offshore Waters Under Section 274"
- 10 CFR Part 170, "Fees for Facilities, Materials, Import and Export Licenses, and Other Regulatory Services Under the Atomic Energy Act of 1954, as Amended"

## APPLICABLE REGULATIONS

- 10 CFR Part 171, “Annual Fees for Reactor Licenses and Fuel Cycle Licenses and Materials Licenses, Including Holders of Certificates of Compliance, Registrations, and Quality Assurance Program Approvals and Government Agencies Licensed by the NRC.”

**Availability:** The Notice of Availability on the inside front cover of this report provides information on how to request copies of the above documents. Applicants also may call the Government Printing Office (GPO) order desk in Washington, DC at (202) 512-1800. Order the two-volume bound version of Title 10, Code of Federal Regulations, Parts 0-50 and 51-199, from the GPO, Superintendent of Documents, Post Office Box 371954, Pittsburgh, Pennsylvania 15250-7954. You may also contact the GPO electronically through its web site at <<http://www.gpo.gov>>. Request single copies of the above documents from NRC’s Regional or Field Offices (see Figure 2.1 for addresses and telephone numbers).

NRC publishes amendments to its regulations in the *Federal Register*. These updates may be requested from the appropriate Regional Office before they are included in the bound version of Title 10. Title 10 is also available on NRC’s web site at <<http://www.nrc.gov/reading-rm/doc-collections/cfr/>>.

## 5 HOW TO FILE

### 5.1 PREPARING AN APPLICATION

Applicants for an NRC materials license should do the following:

- Be sure to use the most recent guidance in preparing an application;
- Complete NRC Form 313 (Appendix A) Items 1 through 4, 12, and 13 on the form itself;
- Complete NRC Form 313 Items 5 through 11 on supplementary pages, or use Appendix C;
- Complete NRC Form 313A (Appendix B) to document training and experience;
- Provide sufficient detail for NRC to determine that equipment, facilities, training, experience, and the radiation safety program are adequate to protect health and safety and minimize danger to life and property;
- For each separate sheet, other than NRC Form 313A and Appendix C, that is submitted with the application, identify and cross-reference it to the item number on the application or the topic to which it refers;
- Submit all documents, typed, on 8-1/2 x 11-inch paper;
- If submitted, proprietary information must be clearly identified;
- Avoid submitting proprietary information unless it is absolutely necessary;
- Submit an original, signed application and one copy; and
- Retain one copy of the license application for future reference.

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓

Applications must be signed by the applicant's or licensee's management as required by 10 CFR 35.12(a), see Section 8.30, "Certification."

All license applications will be made available for review by the general public in NRC's Public Document Rooms and electronically at the Public Electronic Reading Room. For more information on the Public Electronic Reading Room, visit <<http://www.nrc.gov>>. If it is necessary to submit proprietary information, follow the procedure in 10 CFR 2.790. Failure to follow this procedure could result in disclosure of the proprietary information to the public or substantial delays in processing the application. Employee personal information, i.e., home address, home telephone number, social security number, date of birth, and radiation dose information, should not be submitted unless specifically requested by NRC.

## HOW TO FILE

NRC's new licensing process will be faster and more efficient, in part, through acceptance and processing of electronic applications at some future date. NRC will continue to accept paper applications. However, these will be scanned through an optical character reader (OCR) to convert them to electronic format. To ensure a smooth transition to electronic applications, applicants should:

- Submit printed or typewritten – not handwritten – text on smooth, crisp paper that will feed easily into the scanner;
- Choose typeface designs that are sans serif, such as Arial, Helvetica, Futura, Univers; the text of this document is in a serif font called Times New Roman;
- Use 12-point or larger font;
- Avoid stylized characters such as script, italic, etc.;
- Be sure the print is clear and sharp;
- Be sure there is high contrast between the ink and paper (black ink on white paper is best).

## 5.2 ELECTRONIC APPLICATION

As the electronic licensing process develops, it is anticipated that NRC may provide mechanisms for filing applications via diskettes or CD-ROM and through the Internet. Additional filing instructions will be provided as NRC implements these new mechanisms. When the electronic process becomes available, applicants may file electronically instead of on paper.

## 6 WHERE TO FILE

Applicants that wish to possess or use licensed material in any State or U.S. territory or possession subject to NRC jurisdiction must file an application with NRC Regional Office for the locale in which the material will be possessed and/or used. Section 8.36 and Appendix V provide further information on filing procedures for applicants that wish to perform mobile medical services. Figure 2.1 shows NRC's four Regional Offices and their respective areas for licensing purposes, and identifies Agreement States.

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓

In general, applicants for possession or use of byproduct material in an Agreement State must file an application with the Agreement State, not NRC. However, if work will be conducted at Federally-controlled sites in Agreement States, applicants must first determine the jurisdictional status of the land in order to determine whether NRC or the Agreement State has regulatory authority. Section 2, "Agreement States," has additional information.

## 7 LICENSE FEES

Application fees are required for new license applications and some other licensing actions. Each application for which a fee is specified must be accompanied by the appropriate fee. Refer to 10 CFR 170.31 to determine the amount of the fee. NRC will not issue the licensing action before it receives the appropriate payment. Consult 10 CFR 170.11 for information on exemptions from fees. Once technical review has begun, no fees will be refunded. Application fees will be charged regardless of NRC's disposition of an application or the withdrawal of an application.

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓

Most NRC licensees are also subject to annual fees; refer to 10 CFR 171.16. Consult 10 CFR 171.11 for information on exemptions from annual fees and 10 CFR 171.16(c) on reduced annual fees for licensees that qualify as "small entities."

Direct all questions about NRC fees or completion of Item 12 of NRC Form 313 (Appendix A) to the Office of the Chief Financial Officer (OCFO) at NRC Headquarters in Rockville, Maryland, (301) 415-7554 (or toll free at (800) 368-5642, extension 415-7554). Information about fees may also be obtained by calling this NRC toll-free number or by sending e-mail to [fees@nrc.gov](mailto:fees@nrc.gov).

Enter the fee category and the amount of the fee enclosed with the application on NRC Form 313.

## 8 CONTENTS OF AN APPLICATION

This section explains, item by item, the information the applicants must provide on NRC Form 313 (see Appendix A) and NRC Form 313A (see Appendices B and D). The information needed to complete Items 5 through 11 on Form 313 describes the applicant's proposed radiation safety program. To assist the applicant in submitting complete information on these items, the applicable regulations are referenced in the discussion of each item.

Table 1 in Appendix C is provided to help applicants determine which procedures must be developed, implemented, and maintained for the type of medical use requested. Several appendices in this report present sample procedures that applicants may use in developing their procedures. Suggested responses for each block on the NRC Form 313 appear under "Response from Applicant" in this guide.

If a particular part of a section does not apply, simply note "N/A" for "not applicable." If a particular section applies, but a procedure does not have to be developed, simply note "N" for "no response required." N/A, N, or short sentence responses to Items 5 through 10 should run consecutively on one or more sheets separate from responses provided on NRC Form 313. Lengthy responses should be appended as attachments.

As indicated on NRC Form 313 (Appendix A), responses to Items 5 through 11 should be submitted on separate sheets of paper. Applicants should use NRC Form 313A (Appendix B) to document training and experience for new authorized users, medical physicists, nuclear pharmacists, and radiation safety officers. NRC Form 313A also may be used by experienced individuals seeking additional authorizations. Applicants may use Appendix C to assist with completion of the application.

**ITEMS FOR WHICH A RESPONSE FROM APPLICANT IS REQUIRED  
ON NRC FORM 313**

**8.1 ITEM 1: LICENSE ACTION TYPE**

**THIS IS AN APPLICATION FOR** (Check appropriate item)

Type of Action	License No.
<input type="checkbox"/> A. New License	Not Applicable
<input type="checkbox"/> B. Amendment to License No.	XX-XXXXXX-XX
<input type="checkbox"/> C. Renewal of License No.	XX-XXXXXX-XX

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓

Check A if the application is for a new license.

Check B for an amendment<sup>1</sup> to an existing license, and provide license number.

Check C for a renewal of an existing license, and provide license number.

**8.2 ITEM 2: APPLICANT'S NAME AND MAILING ADDRESS**

**Regulations:** 10 CFR 30.34(b); 10 CFR 35.14(b); 10 CFR 30.34(h).

List the legal name of the applicant's corporation or other legal entity with direct control over use of the radioactive material; a division or department within a legal entity may not be a licensee. An individual may be designated as the applicant only if the individual is acting in a private capacity and the use of the radioactive material is not connected with employment in a corporation or other legal entity. Provide the mailing address where correspondence should be sent. A post office box number is an acceptable mailing address. See Section 8.30, "Certification."

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓

*Note:* NRC must be notified before control of the license is transferred or whenever bankruptcy proceedings are initiated. See Sections 1.3.3 and 1.3.4 for more details. NRC IN 97-30, "Control of Licensed Material During Reorganizations, Employee-Management Disagreements, and Financial Crises," dated June 3, 1997, discusses the potential for the security and control of licensed material to be compromised during periods of organizational instability.

---

<sup>1</sup>See Section 9, "Amendments and Renewals to a License," in this document. Licensees may request an amendment to an existing license to add authorization for other uses of byproduct material.

**8.3 ITEM 3: ADDRESS(ES) WHERE LICENSED MATERIAL WILL BE USED OR POSSESSED**

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓

**Regulations:** 10 CFR 30.33(a)(2); 10 CFR 35.14(b)(2).

In order to ensure compliance with 10 CFR 30.33(a)(2) and as referenced in NRC Form 313 Item 3, specify the street address, city, and state or other descriptive address (e.g., on Highway 10, 5 miles east of the intersection of Highway 10 and State Route 234, Anytown, State) for each facility. The descriptive address should be sufficient to allow an NRC inspector to find the facility location. A post office box address is not acceptable. If byproduct material is to be used at more than one location under the license, the specific address (e.g., street and building) must be provided for each facility. If applying for a license for mobile medical services as authorized pursuant to 10 CFR 35.18(b), the applicant should refer to Section 8.36 and Appendix V of this report for specific licensing guidance. NRC must be notified if the mailing address changes.

**Being granted an NRC license does not relieve a licensee from complying with other applicable Federal, State, or local regulations (e.g., local zoning requirements; a local ordinance requiring registration of a radiation-producing device).**

*Note:* As discussed in Section 8.7 “Recordkeeping for Decommissioning and Financial Assurance,” licensees must maintain permanent records on where the licensed material was used or stored while the license was in effect. These records are important for making future determinations about the release of these locations for unrestricted use (e.g., before the license is terminated). For medical use licensees, acceptable records include sketches and written descriptions of the specific locations where material is (or was) used or stored and any information relevant to spills (e.g., where contamination remains after cleanup procedures or when there is reasonable likelihood that contaminants may have spread), damaged devices, or leaking radioactive sources.

**8.4 ITEM 4: PERSON TO BE CONTACTED ABOUT THIS APPLICATION**

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓

Identify the individual who can answer questions about the application and include his or her telephone number. This is typically the proposed RSO, unless the applicant has named a different person as the contact. NRC will contact this individual if there are questions about the application.

Notify NRC of changes of contact name or telephone number so that NRC can contact the applicant or licensee in the future with questions, concerns, or information. This notice is for “information only” and does not require a license amendment or a fee.

The individual named in Item 4 may or may not be the same individual who signs the application as the “certifying officer” on behalf of the licensee with the authority to make commitments to NRC (see Item 13 on NRC Form 313).

NRC recognizes that licensees may use a consultant or consultant group to help prepare the license application and provide support to the radiation protection program. However, NRC reminds licensees that regardless of the role of the consultant in radiation protection program management, the licensee remains responsible for all aspects of the licensed program, including the services performed by the consultant.

**8.5 ITEM 5: RADIOACTIVE MATERIAL**

**Regulations:** 10 CFR 30.32; 10 CFR 32.210; 10 CFR 35.65; 10 CFR 35.100; 10 CFR 35.200; 10 CFR 35.300; 10 CFR 35.400; 10 CFR 35.500; 10 CFR 35.600; 10 CFR 35.1000.

**Criteria:** 10 CFR Part 35 divides byproduct material for medical use into seven types of use (10 CFR 35.100, 35.200, 35.300, 35.400, 35.500, 35.600, and 35.1000).

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓

**Discussion:** The applicant should indicate the byproduct material requested. The amount and type of information necessary will vary according to the type of use requested.

**35.100 and 35.200 Use:** For 35.100 and 35.200 use, the chemical/physical form may be “Any” unsealed byproduct material permitted by 10 CFR 35.100 or 35.200, as appropriate. For 35.100 and 35.200 use, the total amount requested may be “As Needed.” The following format may be used:

Byproduct Material	Chemical/Physical Form	Maximum Amount
Any byproduct material permitted by 10 CFR 35.100	Any	As needed
Any byproduct material permitted by 10 CFR 35.200	Any	As needed

**35.300 Use:** For 35.300 use, the chemical/physical form may be “Any” unsealed byproduct material permitted by 10 CFR 35.300. The total amount requested must be specified. The following format may be used:

Byproduct Material	Chemical/Physical Form	Maximum Amount
Any byproduct material permitted by 10 CFR 35.300	Any	300 millicuries

CONTENTS OF AN APPLICATION

**35.400, 35.500, 35.600, and 35.1000 Use:** For 35.400, 35.500, 35.600, and 35.1000 use, the radionuclide, the chemical/physical form (i.e., sealed source or device identified by manufacturer and model number), the total amount in Becquerels (Bq), microcuries ( $\mu\text{Ci}$ ), millicuries (mCi), or curies (Ci), and maximum number of sources or activity possessed at any one time must be specified. Applicants should include all possible new sources they might use, in order to minimize the need for license amendments if they change model or vendor. The following format may be used:

Byproduct Material	Chemical/Physical Form	Maximum Amount
I-125 (specific radiation therapy system liquid brachytherapy source)	Liquid source (Manufacturer Name, Model #XYZ)	2 curies total
Cesium 137 (i.e., specific brachytherapy radionuclide)	Sealed source or device (Manufacturer Name, Model #XYZ)	2 curies total
Gadolinium 153 (i.e., specific diagnostic sealed source radionuclide)	Sealed source or device (Manufacturer Name, Model #XYZ)	Not to exceed 500 millicuries per source and 1 curie total
Cobalt 60 (i.e., specific teletherapy sealed source radionuclide)	Sealed source or device (Manufacturer Name, Model #XYZ)	Not to exceed 9,000 curies per source and 18,000 curies total
Iridium 192 (i.e., specific afterloader sealed source radionuclide)	Sealed source or device (Manufacturer Name, Model #XYZ)	Not to exceed 10 curies per source and 20 curies total
Cobalt 60 (i.e., specific gamma stereotactic radiosurgery sealed source radionuclide)	Sealed source or device (Manufacturer Name, Model #XYZ)	Not to exceed 36 curies per source and 6,600 curies total

For sealed sources used in devices, an applicant may wish to request a possession limit adequate to allow for the possession of a spare source, to accommodate the total quantity of material in the licensee's possession during replacement of the source in the device. The maximum activity for a single source or source loading may not exceed the activity specified by the manufacturer for the specific device and source combination as stated in the Sealed Source and Device Registration (SSDR) Certificate. However, an applicant may request a maximum activity for the source in the shipping container that exceeds the maximum activity allowed in the device. To request this authorization, applicants should provide certification that the source transport container is approved for the requested activity. A source that is received with a higher activity than permitted in the device must be allowed to decay to or below the licensed activity limit prior to installation in the device.

**Calibration, Transmission, and Reference Sources:** For calibration, transmission, and reference sources covered under 10 CFR 35.65, the specific sources do not need to be listed on the license as long as the licensee is authorized pursuant to 10 CFR 35.11 for medical use of byproduct material.

**Shielding Material/Depleted Uranium:** Some high activity radionuclide generators used to produce byproduct materials for 35.200 and 35.300 uses (e.g., Tc-99m generators) may include depleted uranium (i.e., uranium depleted in uranium-235 (U-235)) as shielding material. If a generator has depleted uranium shielding, an applicant should request authorization to possess depleted uranium as shielding material. Applicants receiving large therapy sources and devices also should determine if depleted uranium is used to shield the therapy sources and devices. If applicable, the applicant should request authorization to possess depleted uranium (i.e., uranium depleted in uranium-235 (U-235)) in quantities sufficient to include shielding material in both the device(s) and source containers used for source exchange and shielding for other devices. The applicant should review the manufacturer's specifications for each device specified in the license request to determine: (1) if depleted uranium is used to shield the source(s) within the device; and (2) the total quantity of depleted uranium present in the device (in kilograms). The applicant should also consult the manufacturer's specifications or the source supplier to determine if depleted uranium is contained in shielding source containers used during source exchange, as well as the total quantity of depleted uranium in such containers (in kilograms). The following format may be used:

Byproduct Material	Chemical/Physical Form	Maximum Amount
Depleted Uranium	Metal	999 kilograms

**Other Material:** The applicant should make a separate entry for other items that need to be listed (e.g., more byproduct material for *in vitro* testing than is allowed under 10 CFR 31.11, survey meter calibration source, dosimetry system constancy check source, material for *in vitro*, animal, or human research studies). The following format may be used:

Byproduct Material	Chemical/Physical Form	Maximum Amount
Any byproduct material permitted by 10 CFR 31.11	Prepackaged kits	50 millicuries

Sources that are authorized by 10 CFR 35.65, "Authorization for calibration, transmission, and references sources," should *not* be listed.

Applicants should number each line entry consecutively, following the 10 CFR Part 35 material.

**Blood Irradiators:** If the use of a device to irradiate blood is anticipated, the applicant should review NUREG-1556, Vol. 5, "Program-Specific Guidance About Self-Shielded Irradiator Licensees."

When determining both individual radionuclide and total quantities, all materials to be possessed at any one time under the license should be included [i.e., materials received awaiting use (new teletherapy or brachytherapy sources for exchange), materials in use or possessed, material used for shielding, and materials classified as waste awaiting disposal or held for decay-in-storage].

**Response from Applicant:** The applicant should submit the information as described above.

**8.6 ITEM 5: SEALED SOURCES AND DEVICES**

**Regulations:** 10 CFR 30.32(g); 10 CFR 30.33(a)(2); 10 CFR 32.210.

**Criteria:** In accordance with 10 CFR 30.32(g), applicants must provide the manufacturer’s name and model number for each requested sealed source and device (except for calibration, transmission, and reference sources authorized by 10 CFR 35.65). Licensees will be authorized to possess and use only those sealed sources and devices specifically approved or registered by NRC or an Agreement State.

Part 35	Applicability
100	
200	
300	
400	✓
500	✓
600	✓
1000	✓

**Discussion:** NRC or an Agreement State performs a safety evaluation of sealed sources and devices before authorizing a manufacturer to distribute the sources or devices to specific licensees. The safety evaluation is documented in an SSDR Certificate. Applicants must provide the manufacturer’s name and model number for each requested sealed source and device so that NRC can verify that they have been evaluated in an SSDR Certificate or specifically approved on a license. Applicants should include all possible new sources they might use, in order to minimize the need for license amendments if they change model or vendor. If such a review has not been conducted for the specific source/device model(s), licensees should request a copy of the latest version of NUREG-1556, Vol. 3, “Consolidated Guidance about Materials Licensees: Applications for Sealed Source and Device Evaluation and Registration” from NRC Regional Office and submit the information requested therein to NRC for review

An applicant may consult with the proposed supplier or manufacturer to ensure that requested sources and devices are compatible with each other and that they conform to the SSDR designations registered with NRC or an Agreement State. Licensees may not make any changes to the sealed source, device, or source-device combination that would alter the description or specifications from those indicated in the respective SSDR certificates without obtaining NRC’s prior permission in a license amendment. To ensure that sealed sources and devices are used in ways that comply with the SSDR Registry and registration certificates, applicants may want to obtain copies of the appropriate sections of the Registry certificates and review or discuss them with the manufacturer. The SSD Registry compilation of these registration certificates may be found at <<http://www.hsrdo.nrc.gov/nrc/sources/index.cfm>>.

**Response from Applicant:** If possession of sealed source(s) or device(s) is requested, the applicant shall submit the information described above.

**Reference:** See the Notice of Availability on the inside front cover of this report to obtain a copy of NUREG-1556, Vol. 3, “Consolidated Guidance about Materials Licensees: Applications for Sealed Source and Device Evaluation and Registration,” and NUREG-1556, Vol. 11, “Program-Specific Guidance About Licenses of Broad Scope.”

*Note:* Information on SSD registration certificates is also available on the Internet at <<http://www.nrc.gov/materials/miau/ssd/obtain-certs.html>> or by calling NRC's Registration Assistant toll-free at (800) 368-5642, extension 415-7217.

**8.7 ITEM 5: RECORDKEEPING FOR DECOMMISSIONING AND FINANCIAL ASSURANCE**

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓

**Regulations:** 10 CFR 30.34(b); 10 CFR 30.35.

**Criteria:** All licensees are required to maintain records important to decommissioning. Licensees authorized to possess licensed material in excess of the limits specified in 10 CFR 30.35 must provide evidence of financial assurance for decommissioning.

**Discussion:** All licensees are required, under 10 CFR 30.35(g), to maintain records important to decommissioning in an identified location. These records must, in part, identify all areas where licensed material is (or was) used or stored and any information relevant to spills (e.g., where contamination remains after cleanup procedures or when there is a reasonable likelihood that contaminants may have spread) and leaking sealed sources. As an alternative to the potential need for site characterizations, some licensees prefer to maintain information on surveys and leak tests on an ongoing basis and as a low-cost means of providing evidence and assurance of an appropriate decommissioning status upon the termination of licensed activities and/or release of a site for non-licensed use. Pursuant to 10 CFR 30.35(g), licensees must transfer the records important to decommissioning either to the new licensee before licensed activities are transferred or assigned in accordance with 10 CFR 30.34(b), and must transfer records to the appropriate NRC Regional Office before the license is terminated (see 30.51(b)).

Licensees using sealed sources authorized by 10 CFR 35 generally use licensed material in a manner that would preclude releases into the environment, would not cause the activation of adjacent materials, or would not contaminate work areas. The licensee's most recent leak test should demonstrate that there has been no leakage from the sealed sources while the sealed sources were in the licensee's possession. However, any leakage of the sealed source in excess of the regulatory limits would warrant further NRC review of decommissioning procedures on a case-by-case basis.

Licensees authorized to possess byproduct material in excess of the limits specified in 10 CFR 30.35 must also provide evidence of financial assurance for decommissioning. The requirements for financial assurance are specific to the types and quantities of byproduct material authorized on a license. Some medical use applicants and licensees may not need to take any action to comply with the financial assurance requirements because their total inventory of licensed material does not exceed the limits in 10 CFR 30.35 or because the half-life of the unsealed byproduct material used does not exceed 120 days. Applicants requesting licensed material with a half-life in excess of 120 days should determine whether financial assurance is necessary. In addition, applicants requesting more than one radionuclide must use the sum-of-the-ratios method to determine if financial assurance is needed.

## CONTENTS OF AN APPLICATION

Applications for authorization to possess and use unsealed byproduct material with a half-life exceeding 120 days must be accompanied by a decommissioning funding plan or certification of financial assurance when the trigger quantities given in 30.35(a) are exceeded. Acceptable methods of providing financial assurance include trust funds, escrow accounts, government funds, certificates of deposit, deposits of government securities, surety bonds, letters of credit, lines of credit, insurance policies, parent company guarantees, self guarantees, external sinking funds, statements of intent, special arrangements with government entities, and standby trust funds. NUREG-1727, "NMSS Decommissioning Standard Review Plan, Appendix F," dated September 2000 contains acceptable wording for each mechanism authorized by the regulation to guarantee or secure funds.

NRC will authorize sealed source possession exceeding the limits given in 10 CFR 30.35(d) without requiring decommissioning financial assurance, for the purpose of normal sealed source exchange, for no more than 30 days.

### Determining Need for Financial Assurance for Decommissioning

The half-lives of unsealed byproduct material used by medical licensees have traditionally been less than 120 days. Therefore, most medical use applicants need only consider licensed material in sealed sources to evaluate the need for financial assurance. Use Table 8.1 to determine if financial assurance is required for the sealed sources listed. If requesting sealed sources other than those listed or any other unsealed byproduct material with a half-life greater than 120 days, refer to 10 CFR 30.35 and Appendix B to Part 30 for possession limits requiring financial assurance. The sum of the fractions procedure is also depicted in Table 8.1 and must be used to determine the need for financial assurance for both sealed and unsealed byproduct material.

**Table 8.1 Worksheet for Determining Need for Financial Assurance for Sealed Sources**

Step Number	Description	Cobalt-60	Cesium-137	Strontium-90
1	Activity possessed, in curies*			
2	Activity requiring financial assurance, in curies	10,000	100,000	1,000
3	Divide data in Step 1 by data in Step 2 = FRACTION			
4	Add the fractions determined in Step 3			

\* This table uses only conventional units. The conversion to the International System of units (SI) is:  
1 Curie = 37 gigabecquerel.

As 10 CFR 30.35 describes, if the sum of the fractions is greater than or equal to 1, the applicant will need to submit a decommissioning funding plan or financial assurance, as applicable.

**Response from Applicants:** No response is needed from most applicants. If financial assurance is required, applicants must submit evidence as described above and as provided for in NUREG-1727.

**Reference:** See the Notice of Availability on the inside front cover of this report to obtain copies of NUREG-1727, "NMSS Decommissioning Standard Review Plan," dated September 2000.

**8.8 ITEM 6: PURPOSE(S) FOR WHICH LICENSED MATERIAL WILL BE USED**

**Regulations:** 10 CFR 30.33(a)(1); 10 CFR 35.100; 10 CFR 35.200; 10 CFR 35.300; 10 CFR 35.400; 10 CFR 35.500; 10 CFR 35.600; 10 CFR 35.1000.

**Criteria:** 10 CFR Part 35 divides byproduct material for medical use into seven types of use as follows:

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓

10 CFR 35.100	Medical Use of Unsealed Byproduct Material for Uptake, Dilution, and Excretion Studies for Which a Written Directive is Not Required
10 CFR 35.200	Medical Use of Unsealed Byproduct Material for Imaging and Localization Studies for Which a Written Directive is Not Required
10 CFR 35.300	Medical Use of Unsealed Byproduct Material for Which a Written Directive is Required
10 CFR 35.400	Medical Use of Sources for Manual Brachytherapy
10 CFR 35.500	Medical Use of Sealed Sources for Diagnosis
10 CFR 35.600	Medical Use of a Sealed Source(s) in a Device for Therapy-Teletherapy Unit
	Medical Use of a Sealed Source(s) in a Device for Therapy-Remote Afterloader Unit
	Medical Use of a Sealed Source(s) in a Device for Therapy-Gamma Stereotactic Radiosurgery Unit
10 CFR 35.1000	Other Medical Uses of Byproduct Material or Radiation from Byproduct Material

**Discussion: 35.100, 35.200, and 35.300 Use:** For 35.100, 35.200, and 35.300 use, the applicant should define the purpose of use by stating the applicable section of 10 CFR Part 35 (e.g., 10 CFR 35.100, 10 CFR 35.200) and the description of the applicable modality (e.g., any uptake dilution and excretion procedure for which a written directive is not required).

The use of unsealed byproduct material in therapy (10 CFR 35.300) involves administering a byproduct material, either orally or by injection, to treat or palliate a particular disease. The most common form of use of unsealed byproduct material for therapy is the treatment of hyperthyroidism with iodine-131 (I-131) sodium iodide. Other therapeutic procedures include, for example, ablation of thyroid cancer metastasis, treatment of malignant effusions, treatment of polycythemia vera and leukemia, palliation of bone pain in cancer patients, and radiation synovectomy for rheumatoid arthritis patients. References to particular diagnostic or treatment modalities in this section are intended to be examples and are not intended to imply that licensees are limited to these uses.

## CONTENTS OF AN APPLICATION

If an applicant's request is limited to I-131 under 10 CFR 35.300, the license will be limited to that radionuclide.

**35.400 Use:** For 35.400 use, the applicant should define the purpose of use by stating the applicable section of 10 CFR Part 35 (i.e., 10 CFR 35.400). If a source is to be used in a device, applicants may need to define the purpose of use by describing the manufacturer's name and model number of the device. The licensee should relate the sealed sources listed in Item 5 to the devices described in this item.

In manual brachytherapy several types of treatments are available. These may include, for example:

- Interstitial Treatment of Cancer.
- Eye Plaque Implants. This is considered interstitial, not topical, treatment.
- Intracavitary Treatment of Cancer. For purposes of NRC's sealed source and device evaluation on radiation safety issues, intraluminal use is considered analogous to intracavitary use.
- Topical (Surface) Applications.

**35.500 Use:** For 10 CFR 35.500 use, the applicant should define the purpose of use by stating the applicable section of 10 CFR 35 (i.e., 10 CFR 35.500) and describing the manufacturer's name(s) and model number(s) of devices containing sealed sources (where applicable). The licensee should correlate the sealed sources listed in Item 5 with the devices described in this item. Typically, a licensee should use the sealed sources according to manufacturer's radiation safety and handling instructions and must use the sources as approved in the SSDR.

**35.600 Use:** For 10 CFR 35.600 use, the applicant should define the purpose of use by stating the applicable section of 10 CFR Part 35.600 (e.g., teletherapy, remote afterloading, GSR) and describing the manufacturer's name(s) and model number(s) of the device containing a sealed source(s) (e.g., for use in a Manufacturer's Name and Unit Type, Model xxxx radiation therapy unit for the treatment of humans). The applicant should correlate the sealed source(s) listed in Item 5 with the device described in this item. If applicable, the applicant should state that depleted uranium is used as shielding for the device and specify that an additional source is requested to be stored in its shipping container incident to source replacement.

**35.1000 Use:** Applicants must apply for authorization to use byproduct material, or radiation therefrom, in medical applications under §35.1000 when the type of use is not covered under §§ 35.100-35.600.

When applying for use under provisions of 10 CFR 35.1000, applicants should describe the purpose of use and submit the information required under Section 35.12(b) through (d), review regulatory requirements in other Subparts of 10 CFR Part 35, and use them as a guide on how to determine what should be included in an application that is required in §35.12,. It is anticipated that many of the uses of byproduct material under the provisions of §35.1000 may involve

research or product development; thus, applicants should ensure review and compliance with 10 CFR 35.6, "Provisions for the protection of human research subjects," and 10 CFR 35.7, "FDA, other Federal, and State requirements." Use of byproduct material in a source or device after approval by U.S. Food and Drug Administration, e.g., under an IDE (investigational device exemption) or an IND (investigational new drug exemption), does not relieve individuals of the responsibility to obtain a license to use the byproduct material in medicine under the provisions of 10 CFR Part 35.

If the source for the type of use sought under 35.1000 is a sealed source, Section 8.6 of this guide describes the information that must be provided at the time of application. Broad scope licensees are exempted under 35.15(a) from requirements of 35.12(d) (which relates to including certain information in an application about radiation safety aspects of medical use under 35.1000). However, broad scope licensees should make sure that the quantity needed for the proposed use is authorized on their license or apply for an increase if not. Applicants should refer to IN 99-024, "Broad-Scope Licensees' Responsibilities for Reviewing and Approving Unregistered Sealed Sources and Devices" for more information on sealed sources.

Applicants for uses under 35.1000 should consult with their Regional Office to discuss the contents of their application.

**Non-Medical Uses:** Applicants may also describe non-medical uses (e.g., survey meter calibrations with NIST traceable brachytherapy sources) and reference the applicable radioactive material provided in response to Item 5.

**Response from Applicant:** The applicant must submit information regarding the purpose for which the licensed material will be used. The applicant should consider including the information described above, as applicable to the type of use(s) proposed.

**8.9 ITEM 7: INDIVIDUAL(S) RESPONSIBLE FOR RADIATION SAFETY PROGRAM AND THEIR TRAINING AND EXPERIENCE**

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓

**Regulations:** 10 CFR 30.33(a)(3); 10 CFR 33.13; 10 CFR 35.24; 10 CFR 35.50; 10 CFR 35.51; 10 CFR 35.55; 10 CFR 35.57; 10 CFR 35.59; 10 CFR 35.190; 10 CFR 35.290; 10 CFR 35.390; 10 CFR 35.392; 10 CFR 35.394; 10 CFR 35.490; 10 CFR 35.491; 10 CFR 35.590; and 10 CFR 35.690.

**Criteria:** The RSO, AUs, AMPs, and ANPs must have adequate training and experience.

**Discussion:** 10 CFR 35.24 provides the requirements regarding the authority and responsibilities for the radiation protection program, including those of the licensee's management and the RSO appointed by licensee management. Other personnel who have a role in the radiation protection program are AUs, AMPs, ANPs, and members of the RSC (if the licensee is required to establish a RSC). In 10 CFR 30.33(a)(3), NRC requires that an applicant be qualified by training and experience to use licensed materials for the purposes requested in

## CONTENTS OF AN APPLICATION

such a manner as to protect health and minimize danger to life or property. Subparts B, D, E, F, G, H, and J of 10 CFR Part 35 give specific criteria for acceptable training and experience for AUs for medical use, ANPs, the RSO, and AMPs.

A résumé or a curriculum vitae is likely to be insufficient because such documents usually do not supply all the information needed to evaluate an individual's training and experience for NRC purposes. Applicants should ensure that they submit the specific training information required by NRC regulations in Part 35. NRC Form 313A provides a convenient format for submitting this information.

Licensees are responsible for their radiation protection programs; it is essential that strong management control and oversight exist to ensure that licensed activities are conducted properly. The licensee's management must appoint an RSO, who agrees in writing to be responsible for implementing the radiation protection program, and must provide the RSO sufficient authority, organizational freedom, time, resources, and management prerogative to communicate with personnel and direct personnel regarding NRC regulations and license provisions, including: identifying radiation safety problems; initiating, recommending, or providing corrective actions; stopping unsafe operations; and verifying the implementation of corrective actions. Nevertheless, the licensee retains the ultimate responsibility for the conduct of licensed activities.

Licensees that are authorized for two or more different types of uses of byproduct material under Subparts E, F, and H, or two or more types of units under Subpart H are required under 10 CFR 35.24(f) to establish a Radiation Safety Committee (RSC) to oversee all uses of byproduct material permitted by the license. Membership of the committee must include an authorized user of each type of use permitted by the license, the Radiation Safety Officer, a representative of the nursing service, and a representative of management who is neither an authorized user nor the Radiation Safety Officer. The committee may include other members the licensee considers appropriate.

Licensees may contract for medical use services, including those involving patient services. However, the licensee should not assume that by hiring a contractor to provide certain services it has satisfied all regulatory requirements or that it has transferred responsibility for the licensed program to the contractor. Licensee management should ensure that adequate mechanisms for oversight are in place to determine that the radiation protection program, including training of contractor staff, is effectively implemented by the appropriate individuals.

**Response from Applicant:** Refer to the subsequent sections specific to the individuals described above.

**8.10 ITEM 7: RADIATION SAFETY OFFICER (RSO)**

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓

**Regulations:** 10 CFR 30.33(a)(3); 10 CFR 35.2; 10 CFR 35.14; 10 CFR 35.24; 10 CFR 35.50; 10 CFR 35.57; 10 CFR 35.59; 10 CFR 35.900; 10 CFR 35.2024.

**Criteria:** RSOs must have adequate training and experience. The training and experience requirements for the RSO are described in 10 CFR 35.50 or 35.900 and allow for the following four training pathways:

- Certification as provided in 10 CFR 35.50(a) (or 35.900(a) by one of the professional boards recognized by NRC in 10 CFR 35.900);
- Didactic training (200 hours) and 1 year of work experience as described in 10 CFR 35.900;
- Didactic training (200 hours), 1 year of work experience and preceptor statement as described in 10 CFR 35.50(b);
- Identification on the license as an AU, AMP, or ANP with experience in the radiation safety aspects of similar types of byproduct material use for which the individual has RSO responsibilities.

The licensee must also establish, in writing, the authority, duties, and responsibilities of the RSO.

**Discussion:** The RSO is responsible for day-to-day oversight of the radiation protection program. In accordance with 10 CFR 35.24, the licensee must provide the RSO sufficient authority, organizational freedom, time, and resources to perform his or her duties. Additionally, the RSO must have a sufficient commitment from management to fulfill the duties and responsibilities specified in 10 CFR 35.24 to ensure that radioactive materials are used in a safe manner. NRC requires the name of the RSO on the license, and an agreement in writing from the RSO, to ensure that licensee management has identified a responsible, qualified person and that the named individual knows of his or her designation and assumes the responsibilities of an RSO.

Usually, the RSO is a full-time employee of the licensed facility. NRC has authorized individuals that are not employed by the licensee, such as a consultant, to fill the role of RSO or to provide support to the facility RSO. In order to fulfill the duties and responsibilities, the RSO should be on site periodically to conduct meaningful, person-to-person interactions with licensee staff, commensurate with the scope of licensed activities, to satisfy requirements of 10 CFR 35.24. Appendix I contains a model RSO Delegation of Authority. Appendix B contains NRC Form NRC 313A, which can be used to document the RSO's training and experience.

## CONTENTS OF AN APPLICATION

**RSO Responsibilities:** Some of the typical duties and responsibilities of RSOs include ensuring the following:

- Unsafe activities involving licensed materials are stopped;
- Radiation exposures are ALARA;
- Material accountability and disposal;
- Interaction with NRC;
- Timely and accurate reporting and maintenance of appropriate records;
- Annual program audits;
- Proper use and routine maintenance;
- Personnel training; and
- Investigation of incidents involving byproduct material (e.g., medical events).

Appendix I contains a detailed list of typical duties and responsibilities of the RSO.

Applicants are reminded of recentness of training requirements described in 10 CFR 35.59. Specifically, RSO applicants must have successfully completed the applicable training and experience criteria described in 10 CFR Part 35 within 7 years preceding the date of the application. Alternatively, RSO applicants must have had related continuing education and experience since completing the required training and experience. This time provision applies to board certification as well as to other recognized training pathways.

**Response from Applicant:** Provide the following:

- Name of the proposed RSO.

**AND**

- Previous license number (if issued by NRC) or a copy of the license (if issued by an Agreement State) that authorized the uses requested and on which the individual was specifically named as the RSO.

**OR**

- Copy of the certification(s) for the board(s) recognized by NRC and as applicable to the types of use for which he or she has RSO responsibilities.

**OR**

- Description of the training and experience specified in 10 CFR 35.900(b).

**OR**

- Description of the training and experience specified in 10 CFR 35.50(b) demonstrating that the proposed RSO is qualified by training and experience as applicable to the types of use for which he or she has RSO responsibilities.

AND

- Written certification, signed by a preceptor RSO, that the above training and experience has been satisfactorily completed and that a level of radiation safety knowledge sufficient to function independently as an RSO for a medical use licensee has been achieved.

AND

- If applicable, description of recent related continuing education and experience as required by 10 CFR 35.59.

*Notes:*

- NRC Form 313A may be used to document training and experience (see Appendix B; note that former NRC Form 313B, used for preceptor statements, was incorporated in NRC Form 313A in 2002, and use of NRC Form 313B has been discontinued).
- The licensee must notify NRC within 30 days if an RSO permanently discontinues his or her duties under the license or has a name change under 10 CFR 35.14 and to request an amendment to change an RSO under 10 CFR 35.13.
- An AU, AMP, or ANP may be designated as the RSO on the license if the individual has experience with the radiation safety aspects of similar types of byproduct material use for which he or she has RSO responsibilities (see 10 CFR 35.50(c)) and, as required by 10 CFR 35.24(g), has sufficient time, authority, organizational freedom, resources, and management prerogative to perform the duties.
- Subpart J will be retained in Part 35 until October 24, 2004, and, until then, licensees may follow this provision of the rule to meet training and experience requirements.
- Descriptions of training and experience will be reviewed using the criteria listed above. NRC will review the documentation to determine if the applicable criteria in Subpart B or J are met. If the training and experience do not appear to meet the criteria in either Subpart B or J, NRC may request additional information from the applicant or may request the assistance of its Advisory Committee on the Medical Uses of Isotopes (ACMUI) in evaluating such training and experience.
- The training and experience for the RSO of a medical use broad scope license will be reviewed using the above criteria as well as criteria in 10 CFR Part 33.

**8.11 ITEM 7: AUTHORIZED USERS (AUs)**

**Regulations:** 10 CFR 30.33(a)(3); 10 CFR 35.2; 10 CFR 35.11; 10 CFR 35.14; 10 CFR 35.27; 10 CFR 35.57; 10 CFR 35.59; 10 CFR 35.190; 10 CFR 35.290; 10 CFR 35.390; 10 CFR 35.392; 10 CFR 35.394; 10 CFR 35.490; 10 CFR 35.491; 10 CFR 35.590; 10 CFR 35.690; 10 CFR 35 Subpart J.

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓

**Criteria:** Training and experience requirements for AUs are described in 10 CFR 35.190, 10 CFR 35.290, 10 CFR 35.390, 10 CFR 35.392, 10 CFR 35.394, 10 CFR 35.490, 10 CFR 35.491, 10 CFR 35.590, 10 CFR 35.690, or 10 CFR 35 Subpart J.

**Discussion:** The responsibilities of AUs involved in medical use include the following:

- Radiation safety commensurate with use of byproduct material;
- Administration of a radiation dose or dosage and how it is prescribed;
- Direction of individuals under the AU’s supervision in the preparation of byproduct material for medical use and in the medical use of byproduct material;
- Preparation of WDs, if required.

Applicants must meet recentness of training requirements described in 10 CFR 35.59. AU applicants must have successfully completed the applicable training and experience criteria described in 10 CFR Part 35 within 7 years preceding the date of the application. Alternatively, applicants must have had related continuing education and experience since completing the required training and experience. This time provision applies to board certification as well as to other recognized training pathways.

Technologists, therapists, or other personnel may use byproduct material for medical use under an AU’s supervision in accordance with 10 CFR 35.27, “Supervision,” and in compliance with applicable FDA, other Federal, and State requirements (10 CFR 35.7). Examples include FDA requirements for conduct of certain types of clinical research after submission of applications for INDs (Investigational New Drugs) and under the auspices of a Radioactive Drug Research Committee (21 CFR 361.1).

There is no NRC requirement that an AU must render an interpretation of a diagnostic image or results of a therapeutic procedure. NRC recognizes that the AU may or may not be the physician who interprets such studies. Additionally, NRC regulations do not restrict who can read and interpret diagnostic scans or the results of therapeutic procedures involving the administration of byproduct material to individuals.

**AU's for Non-Medical Uses:** For *in vitro* studies, animal research, calibration of survey instruments, and other uses that do not involve the intentional exposure of humans, the list of proposed AUs should include the individuals who will actually be responsible for the safe use of the byproduct material for the requested use.

An applicant should note which user will be involved with a particular use by referring to Items 5 and 6 of the application and providing information about the user's training and experience.

Authorized non-medical use or uses that do not involve the intentional exposure of humans (e.g., *in vitro* and animal research, calibration, dosimetry research) will be reviewed on a case-by-case basis.

**Response from Applicant: Provide the following:**

- Name of the proposed AU and uses requested.

**AND**

- Previous license number (if issued by NRC) or a copy of the license (if issued by an Agreement State) on which the physician was specifically named as an AU for the uses requested.

**OR**

- Copy of the certification(s) for the board(s) recognized by NRC under 10 CFR Part 35, Subparts D, E, F, G, H, and as applicable to the use requested.

**OR**

- Description of the training and experience identified in 10 CFR Part 35 Subpart J demonstrating that the proposed AU is qualified by training and experience for the use requested.

**OR**

- A description of the training and experience identified in 10 CFR Part 35 Subparts D, E, F, G, and H demonstrating that the proposed AU is qualified by training and experience for the use requested;

**AND**

- Written certification, signed by a preceptor physician AU, that the above training and experience has been satisfactorily completed and that a level of competency sufficient to function independently as an AU for the medical uses authorized has been achieved.

AND

- If applicable, description of recent related continuing education and experience as required by 10 CFR 35.59.

*Notes:*

- NRC Form 313A may be used to document training and experience (see Appendix B; note that former NRC Form 313B, used for preceptor statements, was incorporated in NRC Form 313A in 2002, and use of NRC Form 313B has been discontinued).
- Licensees must notify NRC within 30 days if an AU permanently discontinues his or her duties under the license or has a name change under 10 CFR 35.14.
- Subpart J will be retained in Part 35 until October 24, 2004, and, until then, licensees may follow this provision of the rule to meet training and experience requirements.
- Descriptions of training and experience will be reviewed using the criteria listed above. NRC will review the documentation to determine if the applicable criteria in 10 CFR Part 35 are met. If the training and experience do not appear to meet the 10 CFR Part 35 criteria, NRC may request additional information from the applicant or may request the assistance of its ACMUI in evaluating such training and experience.

*Note to reviewers:* Licenses will reflect any limitations on use for listed authorized users (e.g., whether administrations in excess of 33 mCi of iodine-131 are allowed and specific modalities under 10 CFR 35.600, etc.).

**8.12 ITEM 7: AUTHORIZED NUCLEAR PHARMACIST (ANP)**

**Regulations:** 10 CFR 30.33(a)(3); 10 CFR 35.2; 10 CFR 35.11; 10 CFR 35.14; 10 CFR 35.27; 10 CFR 35.55; 10 CFR 35.57; 10 CFR 35.59; 10 CFR 35.980.

**Criteria:** Training and experience requirements for ANPs are described in 10 CFR 35.55.

Part 35	Applicability
100	✓
200	✓
300	✓
400	
500	
600	
1000	✓

**Discussion:** At many licensed medical facilities, an ANP is directly involved with the preparation and administration of radiopharmaceuticals.

Technologists, or other personnel, may prepare byproduct material for medical use under an ANP's supervision in accordance with 10 CFR 35.27, "Supervision," and in compliance with applicable FDA, other Federal, and State requirements (10 CFR 35.7). (Preparation of byproduct material for medical use may also be performed under the supervision of a physician who is an authorized user.)

Applicants are reminded of recentness of training requirements described in 10 CFR 35.59. Specifically, nuclear pharmacist applicants must have successfully completed the applicable training and experience criteria described in 10 CFR Part 35 within 7 years preceding the date of the application. Alternatively, nuclear pharmacist applicants must have had related continuing education and experience since initially completing the required training and experience. This time provision applies to board certification as well as to other recognized training pathways.

**Response from Applicant:** Provide the following:

- Name of the proposed ANP.

**AND**

- Previous license number (if issued by NRC) or a copy of the license (if issued by an Agreement State) on which the individual was specifically named ANP.

**OR**

- Copy of the certification(s) for the radiopharmacy board(s) recognized by NRC under 10 CFR 35.55(a) or 10 CFR 35.980(a).

**OR**

- Description of the training and experience demonstrating that the proposed ANP is qualified by training and experience.

**AND**

- Written certification, signed by a preceptor ANP, that the above training and experience has been satisfactorily completed and that a level of competency
  - sufficient to function independently as an ANP has been achieved (10 CFR 35.55), or
  - sufficient to independently operate a nuclear pharmacy (10 CFR 35.980).

**AND**

- If applicable, description of recent related continuing education and experience as required by 10 CFR 35.59.

**Notes:**

- NRC Form 313A may be used to document training and experience (see Appendix B; note that former NRC Form 313B, used for preceptor statements, was incorporated in NRC Form 313A in 2002, and use of NRC Form 313B has been discontinued).
- Licensees must notify NRC within 30 days if an ANP permanently discontinues his or her duties under the license or has a name change under 10 CFR 35.14.

CONTENTS OF AN APPLICATION

- Subpart J will be retained in Part 35 until October 24, 2004, and, until then, licensees may follow this provision of the rule to meet training and experience requirements.
- Descriptions of training and experience will be reviewed using the criteria listed above. NRC will review the documentation to determine if the applicable criteria in Subparts B and J are met. If the training and experience do not appear to meet the criteria in Subparts B and J, NRC may request additional information from the applicant or may request the assistance of its ACMUI in evaluating such training and experience.

**8.13 ITEM 7: AUTHORIZED MEDICAL PHYSICIST (AMP)**

**Regulations:** 10 CFR 30.33(a)(3); 10 CFR 35.2; 10 CFR 35.14; 10 CFR 35.51; 10 CFR 35.57; 10 CFR 35.59; 10 CFR 35.433; 10 CFR 35.961.

**Criteria:** Training and experience requirements for AMPs are described in 10 CFR 35.51.

Part 35	Applicability
100	
200	
300	
400	✓
500	
600	✓
1000	✓

**Discussion:** At many licensed medical facilities conducting radiation therapy treatments, an AMP is directly involved with the calculation and administration of the radiation dose. The American Association of Physicists in Medicine (AAPM) suggests that a medical physicist limit his or her involvement in radiation therapy to areas for which he or she has established competency.

Applicants are reminded of recentness of training requirements described in 10 CFR 35.59. Specifically, medical physicist applicants must have successfully completed the applicable training and experience criteria described in 10 CFR Part 35 within 7 years preceding the date of the application. Alternatively, medical physicist applicants must have had related continuing education and experience since completing the required training and experience. This time provision applies to board certification as well as to other recognized training pathways.

**Response from Applicant:** Provide the following:

- Name of the proposed AMP.

**AND**

- Previous license number (if issued by NRC) or a copy of the license (if issued by an Agreement State) on which the individual was specifically named as an AMP for the units requested.

**OR**

- Copy of the certification(s) for the board(s) recognized by NRC in 10 CFR 35.51(a) or 10 CFR 35.961(a) or (b).

OR

- Description of the training and experience demonstrating that the proposed AMP is qualified by training and experience identified in 10 CFR 35.961(c) for the units requested.

OR

- Description of the training and experience demonstrating that the proposed AMP is qualified by training and experience identified in 10 CFR 35.51(b) for the units requested.

AND

- Written certification, signed by a preceptor AMP, that the above training and experience has been satisfactorily completed and that a level of competency sufficient to function independently as an AMP has been achieved.

AND

- If applicable, description of recent related continuing education and experience as required by 10 CFR 35.59.

*Notes:*

- NRC Form 313A may be used to document training and experience (see Appendix B; note that former NRC Form 313B, used for preceptor statements, was incorporated in NRC Form 313A in 2002, and use of NRC Form 313B has been discontinued).
- Licensees must notify NRC within 30 days if an AMP permanently discontinues his or her duties under the license or has a name change under 10 CFR 35.14.
- Subpart J will be retained in Part 35 until October 24, 2004, and, until then, licensees may follow this provision of the rule to meet training and experience requirements.
- Descriptions of training and experience will be reviewed using the criteria listed above. NRC will review the documentation to determine if the applicable criteria in Subparts B and J are met. If the training and experience do not appear to meet the criteria in Subparts B and J, NRC may request additional information from the applicant or may request the assistance of its ACMUI in evaluating such training and experience.

## 8.14 ITEM 9: FACILITIES AND EQUIPMENT

**Regulations:** 10 CFR 30.33(a)(2); 10 CFR 35.12(b)(1); 10 CFR 35.18(a).

**Criteria:** Facilities and equipment must be adequate to protect health and minimize danger to life or property.

**Discussion:** Requirements to provide information about the design and construction of facilities and safety equipment are contained in 10 CFR 30.33(a)(2), 35.12(b)(1), and 35.18(a).

Applications will be approved if, among other things, “the applicant’s proposed equipment and facilities are adequate to protect health and minimize danger to life or property.” Facility and equipment requirements depend on the scope of the applicant’s operations (e.g., planned use of the material, the types of radioactive emissions, the quantity and form of radioactive materials possessed, etc.). Applicants should focus particularly on operations using large quantities of radioactive materials; preparation steps involving liquids, gases, and volatile radioactive materials; and the use of alpha-emitters, high-energy photon-emitters, and high-energy beta-emitters.

**Response from Applicant:** Refer to Sections 8.15 through 8.19 for guidance.

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓

## 8.15 ITEM 9: FACILITY DIAGRAM

**Regulations:** 10 CFR 20.1003; 10 CFR 20.1101; 10 CFR 20.1201; 10 CFR 20.1301; 10 CFR 20.1302; 10 CFR 20.1601; 10 CFR 20.1602; 10 CFR 20.1901; 10 CFR 20.1902; 10 CFR 20.2102; 10 CFR 30.33(a)(2); 10 CFR 35.12; 10 CFR 35.14; 10 CFR 35.18(a)(3); 10 CFR 35.75; 10 CFR 35.315(a); 10 CFR 35.415; 10 CFR 35.615.

**Criteria:** In order to issue a license, the Commission must find that facilities and equipment must be adequate to protect health and minimize danger to life or property as required under 10 CFR 30.33(a) and/or 35.18(a).

**Discussion:** Applicants must describe the proposed facilities and equipment as required by 10 CFR 35.12. The facility diagram should include the room or rooms and adjacent areas where byproduct material is prepared, used, administered, and stored that is sufficient to demonstrate that the facilities and equipment are adequate to protect health and minimize danger to life or property.

For types of use permitted by 10 CFR 35.100 and 35.200, applicants should provide room numbers for areas in which byproduct materials are used or prepared for use (i.e., “hot labs”). When information regarding an area or room is provided, adjacent areas and rooms, including those above and below, should be described. For types of use permitted by 10 CFR 35.300 and 35.400, applicants should provide the above information and in addition they should provide the locations where sources are stored. Describe the rooms where patients will be housed if they

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓



## CONTENTS OF AN APPLICATION

The applicant should demonstrate that the limits specified in 10 CFR 20.1301(a) will not be exceeded. If the calculations demonstrate that these limits cannot be met, indicate any further steps that will be taken to limit exposure to individual members of the public. The applicant may consider the following options:

- Adding shielding to the barrier in question, with corresponding modification of the facility description if necessary.
- Requesting prior NRC authorization to operate up to an annual dose limit for an individual member of the public of 5 mSv (0.5 rem) and demonstrating that the requirements of 10 CFR 20.1301 will be met. The applicant must demonstrate the need for and the expected duration of operations that will result in an individual dose in excess of the limits specified in 10 CFR 20.1301(a). A program to assess and control dose within the 5 mSv (0.5 rem) annual limit and procedures to be followed to maintain the dose ALARA (10 CFR 20.1101) must be developed (see 10 CFR 20.1301(d)).

If applicants are proposing to use portable shielding to protect health and minimize danger to life or property, they should describe the alternative equipment and administrative procedures they propose to use for evaluation and approval by NRC. If applicants elect to use portable shielding they should commit to having administrative procedures to control configuration management to maintain dose within regulatory limits.

If radiopharmaceutical therapy and brachytherapy patient rooms are added after the initial license is issued, additional room diagrams should be submitted if the room design (including shielding) and the occupancy of adjacent areas are significantly different from the original diagrams provided. A written description should be submitted for simple changes.

For teletherapy units, it may be necessary to restrict use of the unit's primary beam if the treatment room's walls, ceiling, or floor will not adequately shield adjacent areas from direct or scattered radiation. Electrical, mechanical, or other physical means (rather than administrative controls) must be used to limit movement or rotation of the unit (e.g., electrical or mechanical stops). Some applicants have found it helpful to have a sample response for guidance. The following is an example of an acceptable response on the use of a rotational unit with an integral beam absorber (also called a beam catcher).

- "For the primary beam directed toward the integral beam absorber, electrical or mechanical stops are set so that the primary beam must be centered (within plus or minus 2 degrees) on the integral beam absorber and, in that configuration, the attenuated primary beam may be rotated 360 degrees pointing toward the floor, east wall, ceiling, and west wall."
- "For the primary beam directed away from the integral beam absorber, electrical or mechanical stops permit the unattenuated primary beam to be directed in a 95-degree arc from 5 degrees toward the west wall to vertically down toward the floor to 90 degrees toward the east wall."

Experience has shown that, given this type of example, many applicants can make changes to accommodate their own situations (e.g., use of a vertical unit, use of a rotational unit without an integral beam absorber).

**Response from Applicant:** Provide the following on the facility diagrams:

- Drawings should be to scale, and indicate the scale used.
- Location, room numbers, and principal use of each room or area where byproduct material is prepared, used or stored, as provided above under the heading "Discussion";
- Location, room numbers, and principal use of each adjacent room (e.g., office, file, toilet, closet, hallway), including areas above, beside, and below therapy treatment rooms; indicate whether the room is a restricted or unrestricted area as defined in 10 CFR 20.1003; and
- Provide shielding calculations and include information about the type, thickness, and density of any necessary shielding to enable independent verification of shielding calculations, including a description of any portable shields used (e.g., shielding of proposed patient rooms used for implant therapy including the dimensions of any portable shield, if one is used; source storage safe, etc.).

In addition to the above, for teletherapy and GSR facilities, applicants should provide the directions of primary beam usage for teletherapy units and, in the case of an isocentric unit, the plane of beam rotation.

**References:** National Council on Radiation Protection and Measurements (NCRP) Report 49, "Structural Shielding Design and Evaluation for Medical Use of X-Rays and Gamma Rays of Energies up to 10 MeV"; Report 102, "Medical X-Ray, Electron Beam and Gamma Ray Protection for Energies up to 50 MeV (Equipment Design, Performance and Use)"; and Report 40, "Protection Against Radiation from Brachytherapy Sources" may be helpful in responding to the items above. In addition, NUREG/CR-6276, "Quality Management in Remote Afterloading Brachytherapy" and NUREG/CR-6324, "Quality Assurance for Gamma Knives" may also be helpful in responding to the items above. However, please note that references to 10 CFR Part 35 in the NUREGs may be outdated because the rule was amended after these documents were published.

**8.16 ITEM 9: RADIATION MONITORING INSTRUMENTS**

**Regulations:** 10 CFR 20.1101; 10 CFR 20.1501; 10 CFR 20.2102; 10 CFR 20.2103(a); 10 CFR 30.3; 10 CFR 30.33(a)(2); 10 CFR 35.27; 10 CFR 35.61; 10 CFR 35.2061.

**Criteria:** All licensees shall possess calibrated radiation detection and measuring instruments that will be used for radiation protection, including survey and monitoring

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓

## CONTENTS OF AN APPLICATION

instruments and quantitative measuring instruments needed to monitor the adequacy of radioactive materials containment and contamination control.

**Discussion:** The radiation protection program that licensees are required to develop, document, and implement in accordance with 10 CFR 20.1101 must include provisions for survey instrument calibration (10 CFR 20.1501). Licensees shall possess instruments used to measure radiation levels, radioactive contamination, and radioactivity, as applicable. Instruments used for quantitative radiation measurements must be calibrated for the radiation measured. The instruments should be available for use at all times when byproduct material is in use. The licensee should possess survey instruments sufficiently sensitive to measure the type and energy of radiation used, including survey instruments used to locate low energy or low activity seeds (e.g., I-125, Pd-103) if they become dislodged in the operating room or patient's room.

Usually, it is not necessary for a licensee to possess a survey meter solely for use during sealed source diagnostic procedures, since it is not expected that a survey be performed each time such a procedure is performed. In these cases, it is acceptable for the meter to be available on short notice in the event of an accident or malfunction that could reduce the shielding of the sealed source(s). Surveys may be required to verify source integrity of the diagnostic sealed source and to ensure that dose rates in unrestricted areas and public and occupational doses are within regulatory limits.

Survey meter calibrations must be performed by persons, including licensed personnel, who are qualified to perform calibrations. One method a licensee may use to determine if the service is qualified to perform these activities is to determine that it has an NRC (or an equivalent Agreement State) license. Alternatively, an applicant may choose to develop, implement, and maintain procedures to ensure instruments are calibrated, or propose an alternate method for calibration

Appendix K provides guidance regarding appropriate instrumentation and model survey instrument calibration procedures to meet the requirements detailed in 10 CFR 35.61.

**Response from Applicant:** Provide the following:

- A statement that: "Radiation monitoring instruments will be calibrated by a person qualified to perform survey meter calibrations."

**AND/OR**

- A statement that: "We have developed and will implement and maintain written survey meter calibration procedures in accordance with the requirements in 10 CFR 20.1501 and that meet the requirements of 10 CFR 35.61."

**AND**

- A description of the instrumentation (e.g. gamma counter, solid state detector, portable or stationary count rate meter, portable or stationary dose rate or exposure rate meter, single or

multichannel analyzer, liquid scintillation counter, proportional counter) that will be used to perform required surveys.

AND

- A statement that: "We reserve the right to upgrade our survey instruments as necessary as long as they are adequate to measure the type and level of radiation for which they are used."

**Note:** If calibrations will not be performed by the licensee or by a person qualified to perform survey meter calibration, the applicant should propose an alternate method of calibration for review by NRC.

**References:** See the Notice of Availability on the inside front cover of this report to obtain a copy of NUREG-1556, Vol. 18, "Program-Specific Guidance About Service Provider Licenses," dated November 2000.

**8.17 ITEM 9: DOSE CALIBRATOR AND OTHER EQUIPMENT USED TO MEASURE DOSAGES OF UNSEALED BYPRODUCT MATERIAL**

Part 35	Applicability
100	✓*
200	✓*
300	✓*
400	
500	
600	
1000	✓*

**Regulations:** 10 CFR 30.3; 10 CFR 30.33; 10 CFR 35.27; 10 CFR 35.41; 10 CFR 35.60; 10 CFR 35.63; 10 CFR 35.2060; 10 CFR 35.2063.

\*If applicant will measure patient dosages or use other than unit dosages

**Criteria:** In 10 CFR 35.60 and 10 CFR 35.63, NRC describes requirements for the use, possession, calibration, and check of instruments (e.g., dose calibrators) used to measure patient dosages.

**Discussion:** As described in 10 CFR 35.63, dosage measurement is required for licensees who prepare patient dosages.

- If the licensee uses only unit dosages made by a manufacturer or preparer licensed under 10 CFR 32.72, (and does not split, combine, or otherwise modify unit dosages) the licensee is not required to possess an instrument to measure the dosage. Furthermore, licensees may rely on the provider's dose label for the measurement of the dosage and decay-correct the dosage to the time of administration.
- If the licensee performs direct measurements of dosages in accordance with 10 CFR 35.63 (e.g., prepares its own dosages, breaks up unit dosages for patient administration, or decides to measure unit dosages) the licensee is required to possess and calibrate all instruments used for measuring patient dosages.

Currently, no NRC-regulated alpha-emitting nuclides are used in unsealed form in medicine. This document, therefore, does not provide guidance on the measurement of these radionuclides.

## CONTENTS OF AN APPLICATION

Equipment used to measure dosages must be calibrated in accordance with nationally recognized standards (e.g., ANSI) or the manufacturer's instructions. The measurement equipment may be a well ion chamber, a liquid scintillation counter, etc., as long as the instrument can be calibrated appropriately and is both accurate and reliable.

For other than unit dosages, the activity must be determined by direct measurement, by a combination of radioactivity measurement and mathematical calculation, or by a combination of volumetric measurement and mathematical calculation. However, there are inherent technical difficulties to overcome. For beta-emitting radionuclides, these difficulties include dependence on geometry, lack of an industry standard for materials used in the manufacture of vials and syringes, and lack of a NIST-traceable standard for some radionuclides used. For instance, when determining the dosage of P-32, assays with a dose calibrator may result in inaccuracies caused by inherent variations in geometry; therefore, a volumetric measurement and mathematical calculation may be more accurate. Licensees must assay patient dosages in the same type of vial and geometry as used to determine the correct dose calibrator settings. Using different vials or syringes may result in measurement errors due, for example, to the variation of bremsstrahlung created by interaction between beta particles and the differing dosage containers. Licensees are reminded that beta emitters should be shielded using a low-atomic-numbered material to minimize the production of bremsstrahlung. When a high activity source is involved, consideration should be given to adding an outer shield made from material with a high atomic number to attenuate bremsstrahlung.

**Response from Applicant:** If applicable, provide the following:

- A statement that: "Equipment used to measure dosages will be calibrated in accordance with nationally recognized standards or the manufacturer's instructions."

### 8.18 ITEM 9: THERAPY UNIT — CALIBRATION AND USE

**Regulations:** 10 CFR 30.33(a)(2); 10 CFR 35.27; 10 CFR 35.432; 10 CFR 35.630; 10 CFR 35.632; 10 CFR 35.633; 10 CFR 35.635; 10 CFR 35.642; 10 CFR 35.643; 10 CFR 35.645; 10 CFR 35.2432; 10 CFR 35.2630; 10 CFR 35.2632; 10 CFR 35.2642; 10 CFR 35.2643; 10 CFR 35.2645.

**Criteria:** The above regulations contain NRC requirements, including recordkeeping requirements, for verification and periodic spot-checks of source activity or output. To perform these measurements, the applicant must possess appropriately calibrated dosimetry equipment. For manual brachtherapy sources and LDR remote afterloader sources licensees may use source activity or output determined by the manufacturer, provided that the manufacturer's measurements meet applicable requirements.

**Discussion:** Except for manual brachtherapy sources and low dose-rate remote afterloader sources where the source output or activity is determined by the manufacturer in accordance with 10 CFR Part 35, the applicant must possess a calibrated dosimetry system (e.g., Farmer chamber,

Part 35	Applicability
100	
200	
300	
400	✓*
500	
600	✓*
1000	✓

\* Special requirements re brachtherapy and LDR afterloader sources and Sr-90 sources

electrometer, well-type ionization chamber) that will be used to perform calibration measurements of sealed sources to be used for patient therapy. Dosimetry systems and/or sealed sources used to calibrate the licensee's dosimetry systems must be traceable to NIST or to a laboratory accredited by AAPM, pursuant to 10 CFR 35.630. The licensee must maintain records of calibrations of dosimetry equipment for the duration of the license.

The licensee's AMP must perform full calibrations of sealed sources and devices used for therapy in accordance with published protocols currently accepted by nationally recognized bodies (e.g., AAPM, ACR, ANSI). (Note: Calibration by an AMP is not required for manual brachytherapy sources, except for calculating the activity of Strontium-90 sources.) The licensee's AMP must calculate the activity of each strontium-90 source that is used to determine the treatment times for ophthalmic treatments. In addition, the licensee must perform spot-check measurements of sealed sources and devices used for therapy in accordance with written procedures established by the AMP (10 CFR 35.642, 10 CFR 35.643, and 10 CFR 35.645). Calibration procedures described by the AAPM or any published protocol approved by a nationally recognized body, as applicable, may be used. The calibration procedures should address, in part:

- The method used to determine the exposure rate (or activity) under specific criteria (i.e., distances used for the measurement, whether the measurement is an "in air" measurement or done using a phantom configuration of the chamber with respect to the source(s) and device, scatter factors used to compute the exposure rate, etc.).

Full calibrations must be performed before first medical use<sup>2</sup>, whenever spot-check measurements (if required) indicate that the output differs by more than 5% from the output obtained at the last full calibration corrected mathematically for decay, following replacement of the sources or reinstallation of the unit in a new location not previously described in the license, following any repairs of the unit that include removal of sealed sources or major repair of the components associated with the source exposure assembly, and at intervals as defined in 10 CFR 35.632, 10 CFR 35.633, and 10 CFR 35.635. Manual brachytherapy sources must be calibrated only initially, prior to use.

For sealed sources used in therapy, and in particular, for new types of use, licensees should select dosimetry equipment that will accurately measure the output or the activity of the source. Contact a Regional licensing specialist for additional assistance.

**Response from Applicant:** Provide the following:

- The applicant must provide the procedures required by 10 CFR 35.642, 10 CFR 35.643, and 10 CFR 35.645, if applicable to the license application.

---

<sup>2</sup>For brachytherapy sources, "first medical use" is defined as the first use following the effective date of the revised 10 CFR Part 35, October 24, 2002.

## CONTENTS OF AN APPLICATION

### References:

- AAPM Task Group No. 21, "A Protocol for the Determination of Absorbed Dose from High-Energy Photon and Electron Beams;"
- AAPM Task Group No. 40, "Comprehensive QA for Radiation Oncology," AAPM Report No. 54, "Stereotactic Radiosurgery;"
- AAPM Task Group No. 56, "Code of Practice for Brachytherapy Physics."

Copies of these documents and many other documents from AAPM referenced in this guide may be obtained from Medical Physics Publishing (MPP), 4513 Vernon Boulevard, Madison, WI 53705-4964 or ordered electronically from <<http://www.medicalphysics.org>>.

### 8.19 ITEM 9: OTHER EQUIPMENT AND FACILITIES

**Regulations:** 10 CFR 20.1101; 10 CFR 20.1801; 10 CFR 30.33(a)(2); 10 CFR 30.34; 10 CFR 35.12; 10 CFR 35.315; 10 CFR 35.415; 10 CFR 35.457; 10 CFR 35.615; 10 CFR 35.647; 10 CFR 35.657.

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓

**Criteria:** Facilities and equipment must be adequate to protect health and minimize danger to life or property.

**Discussion:** The applicant should describe, in Item 9 of the application, other equipment and facilities available for safe use and storage of byproduct material listed in Item 5 of this application. This description should be identified as Attachment 9.4.

The applicant must describe additional facilities and equipment for the radiopharmaceutical therapy program to safely receive, use, store, and dispose of radioactive material. The applicant should focus on facilities to be used for radioactive drug therapy administration and patient accommodations (i.e., private room with private bath). I-131 sodium iodide is the most widely used source of radiopharmaceutical therapy. If the radionuclide is administered in volatile liquid form, it is important to place the patient dosage in a closed environment (i.e., a fume hood). Also note there are hazards associated with volatile iodine in pill form; applicants should consider this in establishing their radiological controls. When patients are treated with I-131 sodium radioidide, sources of contamination include airborne I-131, urine, perspiration, saliva, and other secretions.

For **teletherapy, GSR, and HDR facilities**, the licensee shall require any individual entering the treatment room to ensure, through the use of appropriate radiation monitors, that radiation levels have returned to ambient levels. One method of meeting the requirements of 10 CFR 35.615(c) is a beam-on radiation monitor permanently mounted in each therapy treatment room that is equipped with an emergency power supply separate from the power supply for the therapy unit. Such beam-on monitors can provide a visible indication (e.g., flashing light) of an exposed or

partially exposed source. Applicants may propose an alternative to a permanently mounted monitor.

10 CFR 35.615(d) requires that, except for LDR units, each licensee shall construct or equip each treatment room so as to permit continuous observation of the patient while the patient is in the treatment room. If a shielded viewing window will be used, the thickness, density, and type of material used should be specified. If a closed-circuit television system (or some other electronic system) will be used to view the patient, the backup system or procedure to be used in case the electronic system malfunctions should be specified, or the applicant must commit to suspending all treatments until the electronic system is repaired and functioning again. The communication system should allow the patient to communicate with the unit operator in the event of medical difficulties. An open microphone system can be used to allow communication without requiring a patient to move to activate controls.

The regulations require adequate equipment and controls to maintain exposures of radiation to workers ALARA and within regulatory limits. 10 CFR 35.615(b), in part, requires that each door leading into the treatment room be provided with an electrical interlock system to control the on-off mechanism of the therapy unit. The interlock system must cause the source(s) to be shielded if the door to the treatment room is opened when the source is exposed. The interlock system must also prevent the operator from initiating a treatment cycle unless the treatment room entrance door is closed. Further, the interlock must be wired so that the source(s) cannot be exposed after interlock interruption until the treatment room door is closed and the source(s) on-off control is reset at the console.

Due to the unique characteristics of **PDR remote afterloaders** and the lack of constant surveillance of their operation, a more sophisticated alarm system is essential to ensure the patient is protected during treatment. In addition to the above, consider the following:

- The PDR device control console is *not* accessible to unauthorized personnel during treatment;
- A primary care provider checks the patient to ensure that the patient's device has not been moved, kinked, dislodged, or disconnected;
- A more sophisticated interlock/warning system is normally installed for PDR devices. This system should perform the following functions or possess the following characteristics:
  - The signal from the PDR device and the signal from the room radiation monitor should be connected in such a manner that an audible alarm sounds if the room monitor indicates the presence of radiation and the device indicates a "safe" or retracted position;
  - The alarm circuit should also be wired in such a manner that an audible alarm is generated for any device internal error condition that could indicate the unintended extension of the source. This would constitute a circuit that generates the audible alarm when either the "source retracted and radiation present" or appropriate internal error condition(s) exist;

## CONTENTS OF AN APPLICATION

- The "source safe and radiation present" signal should also be self-testing. If a "source not safe" input is received without a corresponding "radiation present" signal, the circuit should generate an interlock/warning circuit failure signal that will cause the source to retract. Reset this circuit manually before attempting to continue treatment;
- The audible alarm should be sufficiently loud to be clearly heard by the facility's responsible device/patient monitoring staff at all times; and
- No provisions for bypassing this alarm circuit or for permanently silencing the alarm should be made to the circuit as long as the room radiation monitor is indicating the presence of radiation. If any circuitry is provided to mute the audible alarm, such circuitry should not mute the alarm for a period of more than 1 minute. Controls that disable this alarm circuit or provide for silencing the alarm for periods in excess of 1 minute should be prohibited.

If the alarm circuit is inoperative for any reason, licensees should prohibit further treatment of patients with the device until the circuit has been repaired and tested. If the alarm circuit fails during the course of a patient treatment, the treatment in progress may continue as long as continuous surveillance of the device is provided during each treatment cycle or fraction.

Applicants may submit information on alternatives to fixed shielding as part of their facility description. This information must demonstrate that the shielding will remain in place during the course of patient treatment.

For patient rooms where **LDR remote afterloader** use is planned, neither a viewing nor an intercom system is required. However, the applicant should describe how the patient and device will be monitored during treatment to ensure that the sources and catheter guide tube are not disturbed during treatment and to provide for prompt detection of any operational problems with the LDR device during treatment.

**Response from Applicant:** For manual brachytherapy facilities, provide a description of the emergency response equipment. For teletherapy, GSR, and remote afterloader facilities, provide a description of the following:

- Warning systems and restricted area controls (e.g., locks, signs, warning lights and alarms, interlock systems) for each therapy treatment room;
- Area radiation monitoring equipment;
- Viewing and intercom systems (except for LDR units);
- Steps that will be taken to ensure that no two units can be operated simultaneously, if other radiation-producing equipment (e.g., linear accelerator, X-ray machine) are in the treatment room;

- Methods to ensure that whenever the device is not in use or is unattended, the console keys will be inaccessible to unauthorized persons; and
- Emergency response equipment.

**8.20 ITEM 10: RADIATION PROTECTION PROGRAM**

**Regulations:** 10 CFR 20.1101; 10 CFR 20.2102; 10 CFR 30.33; 10 CFR 30.34(e); 10 CFR 35.24; 10 CFR 35.26; 10 CFR 35.610; 10 CFR 35.2024; 10 CFR 35.2026.

**Criteria:** 10 CFR 20.1101 states that each licensee must develop, document, and implement a radiation protection program commensurate with the scope of the licensed activity. The program must be sufficient to ensure compliance with the provisions of Part 20 regulations. The licensee is responsible for the conduct of all licensed activities and the acts and omissions of individuals handling licensed material. 10 CFR 30.34(e) provides that NRC may incorporate into byproduct material licenses, at the time of issuance or thereafter, additional requirements and conditions that it deems appropriate or necessary to, in part, protect health or to minimize danger to life and property. 10 CFR 35.24 describes the licensee management's authorities and responsibilities for the radiation protection program. 10 CFR 35.26 sets forth four circumstances in which the licensee may revise its radiation protection program without NRC approval. For example, no NRC approval is required when the revision does not require a license amendment.

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓

**Discussion:** Applicants/licensees must abide by all applicable regulations, develop, implement, and maintain procedures when required, and/or provide requested information about the proposed radiation protection program during the licensing process. Tables C.1 and C.2 in Appendix C may be helpful in determining what information should be provided when requesting a license.

**Response from Applicant:** Respond to subsequent sections of this document regarding Item 10 of the application.

**8.21 ITEM 10: SAFETY PROCEDURES AND INSTRUCTIONS**

**Regulations:** 10 CFR 35.12(c)(2); 10 CFR 35.610; 10 CFR 35.642; 10 CFR 643; 10 CFR 35.645.

**Criteria:** Before using materials under 35.600, the applicant must develop, document, submit, and implement written safety procedures for emergency response. 10 CFR 35.610 requires, in part, that written procedures be developed, implemented, and maintained for responding to an abnormal situation involving a

Part 35	Applicability
100	
200	
300	
400	
500	
600	✓
1000	✓

## CONTENTS OF AN APPLICATION

remote afterloader unit, a teletherapy unit, or a gamma stereotactic radiosurgery unit. The procedures needed to meet 10 CFR 35.610 must include:

- Instructions for responding to equipment failures and the names of the individuals responsible for implementing corrective actions;
- The process for restricting access to and posting of the treatment area to minimize the risk of inadvertent exposure; and .
- The names and telephone numbers of AUs, AMPs, and the RSO to be contacted if the unit or console operates abnormally.

A copy of these procedures must be physically located at the therapy unit console. The instructions must inform the operator of procedures to be followed if the operator is unable to place the source(s) in the shielded position, or remove the patient from the radiation field with controls from outside the treatment room.

**Discussion:** The applicant must establish and follow written procedures for emergencies that may occur (e.g., a therapy source fails to retract or return to the shielded position, or a GSR couch fails to retract). A copy of the manufacturer's recommendations and instructions should be given to each individual performing therapy treatments or operating the therapy device. Practice drills, using nonradioactive (dummy) sources (when possible), must be practiced annually or more frequently, as needed. The drills should include dry runs of emergency procedures that cover stuck or dislodged sources and applicators (if applicable), and emergency procedures for removing the patient from the radiation field. Team practice may also be important for adequate emergency coordination for such maneuvers as removing a patient from a malfunctioning GSR unit and manual movement of the patient treatment table. These procedures, designed to minimize radiation exposure to patients, workers, and the general public should address the following points, as applicable to the type of medical use:

- When the procedures are to be implemented, such as any circumstance in which the source becomes dislodged, cannot be retracted to a fully shielded position, or the patient cannot be removed from the beam of radiation.
- The actions specified for emergency source recovery or shielding that primarily consider minimizing exposure to the patient and health care personnel while maximizing safety of the patient.
- The step-by-step actions for single or multiple failures that specify the individual(s) responsible for implementing the actions. The procedures should clearly specify which steps are to be taken under different scenarios. The procedure should specify situations in which surgical intervention may be necessary and the steps that should be taken in that event.
- Location of emergency source recovery equipment and specification of what equipment may be necessary for various scenarios. Emergency equipment should include shielded storage containers, remote handling tools, and if appropriate, supplies necessary to surgically remove

applicators or sources from the patient and tools necessary for removal of the patient from the device.

- Giving first consideration to minimizing exposure to the patient, usually by removing the patient from the room (rather than using tools to attempt to return the source to the off position). *Note:* If the first step of the emergency procedures for teletherapy units specifies pressing the emergency bar on the teletherapy unit console, the applicant is advised that this action may cause the source to return to the off position but may also cut power to the entire teletherapy unit or to the gantry or the couch.
- Instructing the staff to act quickly and calmly, and to avoid the primary beam of radiation.
- Specifying who is to be notified.
- Requirements to restrict (lock, as necessary) and post the treatment area with appropriate warning signs as soon as the patient and staff are out of the treatment room.

**Response from Applicant:** Provide procedures required by 10 CFR 35.610.

**References:** None.


### 8.22 ITEM 10: OCCUPATIONAL DOSE

**Regulations:** 10 CFR 20.1003, 10 CFR 20.1101; 10 CFR 20.1201; 10 CFR 20.1202; 10 CFR 20.1204; 10 CFR 20.1207; 10 CFR 20.1208; 10 CFR 20.1501; 10 CFR 20.1502; 10 CFR 20.2102; 10 CFR 20.2106.

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓

**Criteria:** Applicants must do either of the following:

- Demonstrate that unmonitored individuals are not likely to receive, in 1 year, a radiation dose in excess of 10 % of the allowable limits as shown in Figure 8.2.



**Figure 8.2 Annual Occupational Dose Limits for Adults**

OR

- Monitor external and/or internal occupational radiation exposure, if required by 10 CFR Part 20.1502.

**Discussion:** The radiation protection program that licensees are required to develop, document, and implement in accordance with 10 CFR 20.1101, must include provisions for monitoring occupational dose. The licensee must evaluate the exposure of all occupational workers (e.g., nurses, technologists) to determine if monitoring is required to demonstrate compliance with Subpart F of 10 CFR Part 20. Licensees must consider the internal and external dose and the occupational workers' assigned duties when evaluating the need to monitor occupational radiation exposure. Review of dosimetry histories for workers previously engaged in similar duties may be helpful in assessing potential doses.

When evaluating external dose from xenon gas, the licensee may take credit for the reduction of dose resulting from the use of xenon traps. Additionally, periodic checks of the trap effluent may be used to ensure proper operation of the xenon trap. Licensees may vent xenon gas directly to the atmosphere as long as the effluent concentration is within 10 CFR Part 20 limits.

When evaluating dose from aerosols, licensees may take credit for the reduction of dose resulting from the use of aerosol traps. Licensees may vent aerosols directly to the atmosphere as long as the effluent concentration is within 10 CFR Part 20 limits.

Appendix M provides a model procedure for monitoring external occupational exposure.

If external dose monitoring is necessary, the applicant should describe the type of personnel dosimetry, such as film badges, optically stimulated luminescence dosimeters (OSL), and thermoluminescent dosimeters (TLDs), that personnel will use. If occupational workers handle licensed material, the licensee should evaluate the need to provide extremity monitors, which are required if workers are likely to receive a dose in excess of 0.05 Sv (5 rems) shallow-dose equivalent (SDE), in addition to whole-body badges. Additionally, applicants should ensure that their personnel dosimetry program contains provisions that personnel monitoring devices be worn so that the part of the body likely to receive the greatest dose will be monitored.

Some licensees use self-reading dosimeters in lieu of processed dosimetry. This is acceptable if the regulatory requirements are met. See American National Standards Institute (ANSI) N322, "Inspection and Test Specifications for Direct and Indirect Reading Quartz Fiber Pocket Dosimeters," for more information. If pocket dosimeters are used to monitor personnel exposures, applicants should state the useful range of the dosimeters, along with the procedures and frequency for their calibration (10 CFR 20.1501(b)).

When personnel monitoring is needed, most licensees use either film badges or TLDs that are supplied by a processor holding current personnel dosimetry accreditation from the National Voluntary Laboratory Accreditation Program (NVLAP). Under 10 CFR 20.1501, licensees must verify that the processor is accredited by NVLAP for the type of radiation for which monitoring will be performed. Consult the NVLAP-accredited processor for its recommendations for exchange frequency and proper use.

It may be necessary to assess the intake of radioactivity for occupationally exposed individuals in accordance with 10 CFR 20.1204 and 20.1502. If internal dose assessment is necessary, the applicant shall measure the following:

- Concentrations of radioactive material in air in work areas; or
- Quantities of radionuclides in the body; or
- Quantities of radionuclides excreted from the body; or
- Combinations of these measurements.

The applicant should describe in its procedures the criteria used to determine the type of bioassay and the frequencies at which bioassay (both *in vivo* and *in vitro*) will be performed to evaluate intakes. The criteria also should describe how tables of investigational levels are derived, including the methodology used by the evaluated internal dose assessments, i.e., the empirical models used to interpret the raw bioassay data. The bioassay procedures should provide for baseline, routine, emergency, and follow-up bioassays. If a commercial bioassay service will be used, the applicant should ensure that the service is licensed by an NRC (or an equivalent Agreement State) license or provide another alternative for NRC to review.

RG 8.9, Revision 1, "Acceptable Concepts, Models, Equations, and Assumptions for a Bioassay Program," and NUREG/CR-4884, "Interpretation of Bioassay Measurements," outline acceptable criteria that applicants may use in developing their bioassay programs.

Regulatory Issue Summary (RIS) 2002-06, "Evaluating Occupational Dose for Individuals Exposed to NRC-Licensed Material and Medical X-Rays," provides guidance for evaluation of occupational dose when some exposure is due to x-rays and dosimeters are used to measure exposure behind lead aprons and elsewhere.

*Note:* The definition of "shallow-dose equivalent" in 10 CFR 20.1003 was revised, effective June 4, 2002<sup>3</sup> to change the area for averaging dose to skin from 1 square centimeter to 10 square centimeters (see NRC Regulatory Issue Summary 2002-10, "Revision of the Skin Dose Limit in 10 CFR Part 20").

**Response from Applicant:** If personnel monitoring is required, provide the following:

- A statement that: "Either we will perform a prospective evaluation demonstrating that unmonitored individuals are not likely to receive, in one year, a radiation dose in excess of 10% of the allowable limits in 10 CFR Part 20 or we will provide dosimetry that meets the requirements listed under "Criteria" in NUREG-1556, Vol. 9, "Consolidated Guidance About Materials Licenses: Program-Specific Guidance About Medical Use Licensees," dated October 2002."

**OR**

- A description of an alternative method for demonstrating compliance with the referenced regulations.

---

<sup>3</sup> 67 FR 16298

## CONTENTS OF AN APPLICATION

### References:

- National Institute of Standards and Technology (NIST) Publication 810, "National Voluntary Laboratory Accreditation Program Directory," is published annually and is available for purchase from GPO and on the Internet at <<http://ts.nist.gov/ts/htdocs/210/214/scopes/programs.htm>>.
- Copies of ANSI N322 may be obtained from the American National Standards Institute, 1430 Broadway, New York, NY 10018, or ordered electronically from <<http://www.ansi.org>>.
- NUREG/CR-4884, "Interpretation of Bioassay Measurements;"
- RG 8.9, Revision 1, "Acceptable Concepts, Models, Equations, and Assumptions for a Bioassay Program;" Regulatory Issue Summary 2002-06;
- "Evaluating Occupational Dose for Individuals Exposed to NRC-Licensed Material and Medical X-Rays;"
- NRC Regulatory Issue Summary 2002-06, "Evaluating Occupational Dose for Individuals Exposed to NRC-Licensed Material and Medical X-Rays;"
- NRC Regulatory Issue Summary 2002-10, "Revision of the Skin Dose Limit in 10 CFR Part 20."

See the Notice of Availability on the inside front cover of this report to obtain copies of these NRC documents. Copies of Regulatory Issue Summaries are also available on the NRC's web site in the Electronic Reading Room at <<http://www.nrc.gov/reading-rm/doc-collections/gen-comm/reg-issues/>>.

### 8.23 ITEM 10: AREA SURVEYS

**Regulations:** 10 CFR 20.1003; 10 CFR 20.1101; 10 CFR 20.1201; 10 CFR 20.1301; 10 CFR 20.1302; 10 CFR 20.1501; 10 CFR 20.1801; 10 CFR 20.1802; 10 CFR 20.2102; 10 CFR 20.2103; 10 CFR 20.2107; 10 CFR 35.70; 10 CFR 35.315; 10 CFR 35.404; 10 CFR 35.604; 10 CFR 35.2070.

**Criteria:** Licensees are required to make surveys of potential radiological hazards in their workplace. For example, licensees must perform surveys to:

- Ensure that licensed material will be used, transported, and stored in such a way that doses to members of the public do not exceed 1 mSv per year (100 millirem/year) and that the dose in any unrestricted area will not exceed 0.02 mSv (2 mrem) in any 1 hour from licensed operations;

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓

- Ensure that licensed material will be used, transported, and stored in such a way that occupational doses to individuals will not exceed the limits specified in 10 CFR 20.1201; and
- Control and maintain constant surveillance over licensed material that is not in storage and secure licensed material from unauthorized access or removal.
- Ensure that licensed material will be used, transported, and stored in such a way that the air emissions do not exceed the constraint value in 10 CFR 20.1101.

**Discussion:** The radiation protection program that licensees are required to develop, document, and implement in accordance with 10 CFR 20.1101 must include provisions for area surveys. Surveys are evaluations of radiological conditions and potential hazards. These evaluations may be measurements (e.g., radiation levels measured with survey instrument or results of wipe tests for contamination), calculations, or a combination of measurements and calculations. The selection and proper use of appropriate instruments is one of the most important factors in ensuring that surveys accurately assess radiological conditions.

There are many different kinds of surveys performed by licensees:

- Contamination:
  - Fixed;
  - Removable.
- Air Effluent;
- Water Effluent;
- Leak Test;
- Bioassays;
- Air Sample;
- Restricted Areas;
- Unrestricted Areas; and
- Personnel (during use, transfer, or disposal of licensed material).

Surveys are required when it is reasonable under the circumstances to evaluate a radiological hazard and when necessary for the licensee to comply with the appropriate regulations. The most important types of surveys are as follows:

- Surveys for radioactive contamination that could be present on surfaces of floors, walls, laboratory furniture, and equipment;
- Measurements of radioactive material concentrations in air for areas where radiopharmaceuticals are handled or processed in unsealed form and where operations could expose workers to the inhalation of radioactive material (e.g., radioiodine) or where licensed material is or could be released to unrestricted areas;

## CONTENTS OF AN APPLICATION

- Bioassays to determine the kinds, quantities, or concentrations, and in some cases, the location of radioactive material in the human body. Radioiodine uptake in a worker's thyroid gland is commonly measured by external counting using a specialized thyroid detection probe;
- Surveys of external radiation exposure levels in both restricted and unrestricted areas; and
- Surveys of radiopharmaceutical packages entering (e.g., from suppliers) and departing (e.g., returned radiopharmaceuticals to the supplier).

The frequency of routine surveys depends on the nature, quantity, and use of radioactive materials, as well as the specific protective facilities, equipment, and procedures that are designed to protect workers and the public from external and internal exposure. Also, the frequency of the survey depends on the type of survey. Appendix R contains model procedures that represent one acceptable method of establishing survey frequencies for ambient radiation level and contamination surveys. For example, licensees are required to perform daily surveys in all areas used for the preparation and administration of radiopharmaceuticals for which a written directive is required (diagnostic activities exceeding 30  $\mu\text{Ci}$  of I-131 and all therapy treatments); when the licensee administers radiopharmaceuticals requiring a WD in a patient's room, the licensee is not required to perform a survey of the patient's room. Licensees should perform surveys after the patient's release. Licensees must perform surveys prior to the release of the room for unrestricted use. Licensees should be cognizant of the requirement to perform surveys to demonstrate the public dose limits are not exceeded.

Because therapy sealed sources (including applicators and catheters) may become dislodged during implantation or after surgery, and inadvertently lost or removed, the following surveys shall be performed:

- Immediately after implanting sources in a patient or a human research subject, the licensee shall make a survey to locate and account for all sources that have not been implanted; and
- Immediately after removing the last temporary implant source from a patient or human research subject, the licensee shall make a survey of the patient or human research subject with a radiation detection survey instrument to confirm that all sources have been removed.

In addition, licensees should also consider the following:

- The therapy patient's bed linens before removing them from the patient's room;
- The operating room and the patient's room after source implantation (e.g., radiation level and/or visual check);
- All trash exiting the patient's room; and
- Areas of public access in and around the patient's room.

**Response from Applicant:** Provide the following statement:

“We have developed and will implement and maintain written procedures for area surveys in accordance with 10 CFR 20.1101 that meet the requirements of 10 CFR 20.1501 and 10 CFR 35.70.”

**8.24 ITEM 10: SAFE USE OF UNSEALED LICENSED MATERIAL**

Part 35	Applicability
100	✓
200	✓
300	✓
400	
500	
600	
1000	✓

**Regulations:** 10 CFR 20.1101; 10 CFR 20.1301; 10 CFR 20.1302; 10 CFR 20.2102; 10 CFR 20.2103; 10 CFR 30.33(a)(2); 10 CFR 30.34(e); 10 CFR 35.27; 10 CFR 35.69; 10 CFR 35.70; 10 CFR 35.310.

**Criteria:** Before using licensed material, the licensee must develop and implement a radiation protection program that includes safe use of unsealed licensed material.

**Discussion:** The radiation protection program that licensees are required to develop, document, and implement in accordance with 10 CFR 20.1101 must include provisions for safe use of licensed material. Licensees are responsible for developing, documenting, and implementing procedures to ensure the security and safe use of all licensed material from the time it arrives at their facilities until it is used, transferred, and/or disposed. The written procedures should provide reasonable assurance that only appropriately trained personnel will handle and use licensed material without undue hazard to themselves, other workers, or members of the public.

In addition, licensees must develop, implement, and maintain procedures for protective measures to be taken by occupational workers to maintain their doses ALARA. Protective measures may include:

- Use of syringe shields and/or vial shields;
- Wearing laboratory coats and gloves when handling unsealed byproduct material; and
- Monitoring hands after handling unsealed byproduct material.

Appendix T contains model procedures that provide one method for safe use of unsealed licensed material.

**Response from Applicant:** Provide the following statement:

“We have developed and will implement and maintain procedures for safe use of unsealed byproduct material that meet the requirements of 10 CFR 20.1101 and 10 CFR 20.1301.”

### 8.25 ITEM 10: SPILL PROCEDURES

**Regulations:** 10 CFR 19.11(a)(3); 10 CFR 20.1101; 10 CFR 20.1406; 10 CFR 20.2202; 10 CFR 20.2203; 10 CFR 30.32; 10 CFR 30.35(g); 10 CFR 30.50; 10 CFR 30.51; 10 CFR 35.27.

**Criteria:** Before using licensed material, the licensee must develop, document, and implement a radiation protection program that includes proper response to spills of licensed material.

**Discussion:** The radiation protection program that licensees are required to develop, document, and implement in accordance with 10 CFR 20.1101 must include provisions for responding to spills or other contamination events in order to prevent the spread of radioactive material. Appendix N contains model emergency response procedures, including model spill procedures. Spill procedures should address all types and forms of licensed material used and should be posted in restricted areas where licensed materials are used or stored. The instructions should specifically state the names and telephone numbers of persons to be notified (e.g., RSO, staff, state and local authorities, and NRC, when applicable). Additionally, the instructions should contain procedures for evacuation of the area, containment of spills and other releases, appropriate methods for reentering, and for decontaminating facilities (when necessary).

**Response from Applicant:** Provide the following statement:

“We have developed and will implement and maintain written procedures for safe response to spills of licensed material in accordance with 10 CFR 20.1101.”

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓*
500	✓*
600	
1000	✓

\*If source does not meet sealed source definition in 10 CFR Part 35

### 8.26 ITEM 10: INSTALLATION, MAINTENANCE, ADJUSTMENT, REPAIR, AND INSPECTION OF THERAPY DEVICES CONTAINING SEALED SOURCES

**Regulations:** 10 CFR 20.1101; 10 CFR 30.32; 10 CFR 30.34; 10 CFR 35.605; 10 CFR 35.655; 10 CFR 35.2605; 10 CFR 35.2655.

Part 35	Applicability
100	
200	
300	
400	
500	
600	✓
1000	✓

**Criteria:** In accordance with 10 CFR 35.605 and 10 CFR 35.655, licensees must ensure that therapy devices containing sealed sources are installed, maintained, adjusted, repaired, and inspected by persons specifically licensed to conduct these activities. The above activities should be conducted according to the manufacturers' written recommendations and instructions and according to the SSDR. In addition, 10 CFR 35.655 requires that teletherapy and GSR units be fully inspected and serviced during source replacement or at intervals not to exceed 5 years, whichever comes first, to ensure that the source exposure mechanism functions properly. Maintenance is necessary to ensure that the device functions as designed and source integrity is not compromised.

**Discussion:** Maintenance and repair includes installation, replacement, and relocation or removal of the sealed source(s) or therapy unit that contains a sealed source(s). Maintenance and repair also includes any adjustment involving any mechanism on the therapy device, treatment console, or interlocks that could expose the source(s), reduce the shielding around the source(s), affect the source drive controls, or compromise the radiation safety of the unit or the source(s).

NRC requires that maintenance and repair (as defined above) be performed only by persons specifically licensed by NRC or an Agreement State to perform such services. Most licensee employees do not perform maintenance and repair because they do not have the specialized equipment and technical expertise to perform these activities. Applicants requesting authorization to possess and use LDR remote afterloaders should review 10 CFR 35.605 before responding to this item. 10 CFR 35.605 allows for an AMP to perform certain service activities with regard to LDR remote afterloader units.

**Response from Applicant:** No response is necessary if the licensee contracts with personnel who are licensed by NRC or an Agreement State to install, maintain, adjust, repair, and inspect the specific therapy device possessed by the licensee. However, if the applicant requests that an employee who is trained by the manufacturer be authorized to perform the aforementioned activities, the applicant must provide sufficient information to allow the NRC to evaluate and approve such authorization (see CFR 35.605 and 10 CFR 35.655). This should include the following:

- Name of the proposed employee and types of activities requested;

**AND**

- Description of the training and experience demonstrating that the proposed employee is qualified by training and experience for the use requested;

**AND**

- Copy of the manufacturer's training certification and an outline of the training in procedures to be followed.

**Note:** The applicant should specify only those installation, maintenance, inspection, adjustment, and repair functions described in a certificate or letter from the manufacturer of the device that documents the employee's training in the requested function(s).

## 8.27 ITEM 10: MINIMIZATION OF CONTAMINATION

**Regulations:** 10 CFR 20.1406; 10 CFR 35.67.

**Criteria:** Applicants for new licenses must describe in the application how facility design and procedures for operation will minimize, to the extent practicable, contamination of the facility and the environment, facilitate eventual decommissioning, and minimize, to the extent practicable, the generation of radioactive waste.

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓

**Discussion:** All applicants for new licenses need to consider the importance of designing and operating their facilities to minimize the amount of radioactive contamination generated at the site during its operating lifetime and to minimize the generation of radioactive waste during decontamination. This is especially important for licensed activities involving unsealed byproduct material. As described in Item 8.25, "Spill Procedures," cleanup procedures should be implemented for contamination events. Recommended limits for acceptable levels of surface contamination in restricted and unrestricted areas are provided in Appendix R, Tables R.2 and R.3.

Sealed sources and devices that are approved by NRC or an Agreement State and located and used according to their SSDR Certificates usually pose little risk of contamination. Leak tests performed as specified in the SSDR Certificate should identify defective sources. Leaking sources must be immediately withdrawn from use and stored, repaired, or disposed of according to NRC requirements. These steps minimize the spread of contamination and reduce radioactive waste associated with decontamination efforts.

**Response from Applicant:** A response from applicants is not required under the following condition: the NRC will consider that the above criteria have been met if the information provided in applicant's responses satisfy the criteria in Sections 8.14, 8.15, 8.20, 8.24, 8.26, and 8.28, on the topics: Facility and Equipment; Facility Diagram; Radiation Protection Program; Safety Program; and Waste Management.

## 8.28 ITEM 11: WASTE MANAGEMENT

**Regulations:** 10 CFR 20.1101; 10 CFR 20.1301; 10 CFR 20.1302; 10 CFR 20.1501; 10 CFR 20.1904; 10 CFR 20.2001-2007; 10 CFR 20.2102; 10 CFR 20.2103; 10 CFR 20.2107; 10 CFR 20.2108; 10 CFR 30.33(a)(2); 10 CFR 30.41; 10 CFR 30.51; 10 CFR 31.11; 10 CFR 35.92; 10 CFR 35.2092; 10 CFR 61.3; 10 CFR 71.5.

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓

**Criteria:** Licensed materials must be disposed of in accordance with NRC requirements by:

- Transfer to an authorized recipient (10 CFR 30.41(b));
- Decay-in-storage;
- Release in effluents within the limits in 10 CFR 20.1301; or
- As authorized under 10 CFR 20.2002 through 20.2005.

**Discussion:** The radiation protection program that licensees are required to develop, document, and implement in accordance with 10 CFR 20.1101 must include provisions for waste disposal of licensed material. Appendix W contains model procedures that represent one way to provide for decay-in-storage and generator or other licensed material return. Applicants are reminded to take into account the following information when they develop procedures (as applicable):

- Except for material suitable for decay-in-storage and some animal carcasses handled by the licensee, solids are transferred to an authorized recipient licensed to receive such waste in accordance with 10 CFR 20.2001(b), 10 CFR 20.2006, or in applicable regulations in 10 CFR Parts 30 or 61. Follow the packaging instructions received from the transfer agent and the burial site operator. Keep the consignment sheet from the transfer agent as the record of disposal.
- When setting up a program for decay-in-storage, consider short-term and long-term storage. Consider designing long-term storage to allow for segregation of wastes with different half-lives (e.g., the use of multiple shielded containers) and use of containers with shielded covers to maintain occupational exposure at ALARA levels. Storage areas must be in a secure location.
- Waste from *in vitro* kits (except mock iodine-125) that are generally licensed under 10 CFR 31.11 is exempt from waste disposal regulations in 10 CFR Part 20, as set forth in 10 CFR 31.11(f). Radioactive labels should be defaced or removed. There is no need to keep any record of release or make any measurement.
- Consider the monitoring and control mechanisms in place to ensure compliance with the appropriate requirements regarding the release of material into air and water under 10 CFR 20.1302 and 20.2003, respectively.
  - Regulations for disposal in the sanitary sewer appear in 10 CFR 20.2003. Material must be readily soluble or dispersible in the water. There are also monthly and annual limits, based on the total sanitary sewerage release of the facility. (Excreta from patients undergoing medical diagnosis or therapy are not subject to these limitations; see 10 CFR 20.2003(b)).
  - Limits on permissible concentrations in effluents to unrestricted areas are enumerated in Table II of Appendix B to 10 CFR Part 20. These limits apply at the boundary of the restricted area.
  - Liquid scintillation-counting media containing 1.85 kBq (0.05  $\mu$ Ci) per gram of H-3 or C-14 may be disposed of without regard to its radioactivity (10 CFR 20.2005(a)(1)).

## CONTENTS OF AN APPLICATION

- If applicants/licensees propose to treat or dispose of licensed material by incineration, they must comply with 10 CFR 20.2004. Contact the appropriate NRC Regional Office for guidance on treatment or disposal of material by incineration.
- Applicants that wish to use waste volume reduction operations (e.g., compactors) should provide a detailed description (as outlined below), along with their response to Item 8.16 (Facility Diagram):
  - A description of the compactor to demonstrate that it is designed to safely compact the waste generated (e.g., manufacturer's specifications, annotated sketches, photographs);
  - The types, quantities, and concentrations of the waste to be compacted;
  - An analysis of the potential for airborne release of radioactive material during compaction activities;
  - The location of the compactors in the waste processing area(s), as well as a description of the ventilation and filtering systems used in conjunction with the compactors, and procedures for monitoring filter blockage and exchange;
  - Methods used to monitor worker breathing zones and/or exhaust systems;
  - The types and frequencies of surveys that will be performed for contamination control in the compactor area;
  - The instructions provided to compactor operators, including instructions for protective clothing, checks for proper functioning of equipment, method of handling uncompacted waste, and examining containers for defects.

**Nuclear pacemakers:** Medical licensees are often the first to come into contact with plutonium-powered pacemakers or the first to be contacted by nursing homes and funeral homes when a patient with an implanted pacemaker dies. In such cases and when the licensee is not responsible for control or disposal of the pacemaker, notify the NRC and attempt to contact the hospital where the pacemaker was implanted to arrange for explantation. The licensee which implanted the device is responsible for the follow-up, explantation, and return of the pacemaker to the manufacturer for proper disposal. NRC Information Notice 98-12, "Licensees' Responsibilities Regarding Reporting and Follow-up Requirements for Nuclear-Powered Pacemakers," provides additional information.

**Response from Applicant:** Provide the following statement:

"We have developed and will implement and maintain written waste disposal procedures for licensed material in accordance with 10 CFR 20.1101, that also meet the requirements of the applicable section of Subpart K to 10 CFR Part 20 and 10 CFR 35.92."

**8.29 ITEM 12: FEES**

**Regulations:** 10 CFR 170.31.

On NRC Form 313, enter the appropriate fee category from 10 CFR 170.31 and the amount of the fee enclosed with the application.

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓

**8.30 ITEM 13: CERTIFICATION**

Individuals acting in a private capacity are required to date and sign NRC Form 313. Otherwise, representatives of the corporation or legal entity filing the application should date and sign NRC Form 313. These representatives must be authorized to make binding commitments and to sign official documents on behalf of the applicant. An application for licensing a medical facility must be signed by the applicant's or licensee's management. The individual who signs the application should be identified by title of the office held. As discussed previously in "Management Responsibility," signing the application acknowledges management's commitment and responsibilities for the radiation protection program. Management includes the chief executive officer or other individual having the authority to manage, direct, or administer the licensee's activities, or those persons' delegate or delegates. NRC will return all unsigned applications for proper signature.

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓

*Note:* It is a criminal offense to make a willful false statement or representation on applications or correspondence (18 U.S.C. 1001).

**PROGRAM-RELATED GUIDANCE -- NO RESPONSE REQUIRED FROM  
APPLICANTS ON NRC FORM 313**

The information provided in the following sections is included because this topic is a key element of a licensee's program and the information is provided as guidance to applicants in setting up their programs to satisfy regulatory requirements.

**8.31 SAFETY INSTRUCTION FOR INDIVIDUALS WORKING IN OR FREQUENTING RESTRICTED AREAS**

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓

**Regulations:** 10 CFR 19.12; 10 CFR 35.27; 10 CFR 35.310; 10 CFR 35.410; 10 CFR 35.610; 10 CFR 35.2310.

**Criteria:** Individuals working with or in the vicinity of licensed material must have adequate safety instruction as required by 10 CFR Parts 19 and 35. For individuals who, in the course of employment, are likely to receive in a year an occupational dose of radiation over 1 millisievert (mSv) [100 millirem (mrem)], the licensee must provide safety instructions as required by 10 CFR 19.12. Additional requirements for training in radiation safety for individuals involved with therapeutic treatment of patients are described in 10 CFR 35.310, 10 CFR 35.410, and 10 CFR 35.610. 10 CFR 35.27 requires the licensee's AUs and ANPs to provide safety instruction to all personnel using byproduct material under their supervision.

**Discussion:** AUs, ANPs, AMPs, RSOs, and their supervised employees are most likely to receive doses in excess of 1 mSv (100 mrem) in a year. However, licensees also must evaluate potential radiation doses received by any individual working in or frequenting restricted areas. All individuals working with or around licensed materials should receive safety instruction commensurate with their assigned duties, and if it is likely that they could receive doses over 1 mSv (100 mrem) in a year, they must receive instruction as specified by 10 CFR 19.12. For example, a licensee might determine that housekeeping staff, while not likely to receive doses over 1 mSv (100 mrem), should be informed of the nature of the licensed material and the meaning of the radiation symbol, and instructed not to touch the licensed material and to remain out of the room if the door to the licensed material storage location is open. Providing minimal instruction to ancillary staff (e.g., housekeeping, security, etc.) may assist in controlling abnormal events, such as loss of radioactive material.

In addition to safety instruction required by 10 CFR 19.12 and in accordance with 10 CFR 35.310, 10 CFR 35.410, and 10 CFR 35.610, the licensee must provide radiation safety instruction to personnel (e.g., nurses) caring for patients undergoing radiopharmaceutical therapy and/or implant therapy who cannot be released in accordance with 10 CFR 35.75. This safety instruction should be commensurate with the duties of the personnel and include safe handling, patient control, visitor control, contamination control, waste control, and notification of the RSO and the AU if the patient has a medical emergency or dies.

In accordance with 10 CFR 35.27(a), individuals working with licensed material under the supervision of an AU must receive instruction on the licensee's written radiation protection procedures, written directive procedures, and NRC regulations and license conditions with respect to the use of byproduct material.

In accordance with 10 CFR 35.27(b), a licensee that permits the preparation of byproduct material for medical use by an individual under the supervision of an ANP or an AU, as allowed by 10 CFR 35.11(b)(2), shall instruct supervised individuals in the preparation of byproduct

## PROGRAM-RELATED GUIDANCE

material for medical use and require the individuals to follow their instructions, the licensee's written radiation protection procedures, the license conditions, and NRC regulations. 10 CFR 35.27(c) states that a licensee that permits supervised activities, under paragraph 10 CFR 35.27(a) and (b), is responsible for the acts and omissions of the supervised individuals.

Appendix J provides a model training program that provides one way to satisfy the requirements referenced above.

**Response from Applicant:** No response is necessary.

### 8.32 PUBLIC DOSE

**Regulations:** 10 CFR 20.1003; 10 CFR 20.1101, 10 CFR 20.1301; 10 CFR 20.1302; 10 CFR 20.1801; 10 CFR 20.1802; 10 CFR 20.2107.

**Criteria:** Licensees must do the following:

- Ensure that licensed material will be used, transported, and stored in such a way that members of the public will not receive more than 1 mSv (100 mrem) in 1 year, and the dose in any unrestricted area will not exceed 0.02 mSv (2 mrem) in any one hour from licensed operations.
- Ensure air emissions of radioactive materials to the environment will not result in exposures to individual members of the public in excess of 0.1 mSv (10 mrem) (TEDE) in one year from these emissions.
- Control and maintain constant surveillance of licensed material that is not in storage and secure stored licensed material from unauthorized access, removal, or use.

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓

**Discussion:** Members of the public include persons who are not radiation workers. This includes workers who live, work or may be near locations where licensed material is used or stored and employees whose assigned duties do not include the use of licensed materials and who work in the vicinity where it is used or stored. Public dose is controlled, in part, by ensuring that licensed material is secure (e.g., located in a locked area) to prevent unauthorized access or use by individuals coming into the area. Some medical use devices containing licensed material are usually restricted by controlling access to the keys needed to operate the devices and/or to keys to the locked storage area. Only AUs and personnel using byproduct material under their supervision should have access to these keys.

Typical unrestricted areas may include offices, shops, laboratories, areas outside buildings, property, and nonradioactive equipment storage areas. The licensee does not control access to these areas for purposes of controlling exposure to radiation or radioactive materials; however, the licensee may control access to these areas for other reasons, such as security.

For areas adjacent to facilities where licensed material is used or stored, calculations or a combination of calculations and measurements (e.g., using an environmental TLD) are often used to show compliance.

The definition of "public dose" in 10 CFR 20.1003 does not include doses received due to exposure to patients released in accordance with 10 CFR 35.75. Dose to members of the public in waiting rooms was addressed in Informational Notice (IN) 94-09.<sup>4</sup> The provisions of 10 CFR 20.1301(a) should not be applied to radiation received by a member of the general public from patients released under 10 CFR 35.75. If a patient is released pursuant to 10 CFR 35.75, licensees are not required to limit the radiation dose to members of the public (e.g., visitor in a waiting room) from a patient to 0.02mSv (2mrem) in any one hour. Patient waiting rooms need only be controlled for those patients not meeting the release criteria in 10 CFR 35.75.

10 CFR 20.1301(c) allows licensees to permit visitors to a patient who cannot be released under 10 CFR 35.75 to receive a dose greater than 0.1 rem (1 mSv) provided the dose does not exceed 0.5 rem (5 mSv) and the authorized user has determined before the visit that it is appropriate.

In assessing adequacy of facilities to control public dose, licensees should consider the design factors discussed under "Facility Diagram" in Section 8.15 and may find confirmatory surveys to be useful in assuring compliance with 10 CFR 20.1301.

The licensee must control emissions of byproduct material to air such that the individual member of the public likely to receive the highest total effective dose equivalent (TEDE) does not exceed the constraint level of 0.10 mSv (10 mrem) per year from those emissions. If exceeded, the licensee must report this in accordance with 10 CFR 20.2203, and take prompt actions to ensure against recurrence.

**Response from Applicant:** No response required.

### 8.33 OPENING PACKAGES

**Regulations:** 10 CFR 20.1906; 10 CFR 20.2103.

**Criteria:** Licensees must ensure that packages are opened safely and that the requirements of 10 CFR 20.1906 are met. Licensees must retain records of package surveys in accordance with 10 CFR 20.2103.

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓

**Discussion:** Licensees must establish, maintain, and retain written procedures for safely opening packages to ensure that the monitoring requirements of 10 CFR 20.1906 are met and that radiation exposure to personnel coming near or in contact with the packages containing radioactive material are ALARA.

<sup>4</sup>IN 94-09 - Release of Patients with Residual Radioactivity from Medical Treatment and Control of Areas Due to Presence of Patients Containing Radioactivity Following Implementation, February 1994.

PROGRAM-RELATED GUIDANCE

Appendix P contains model procedures that represent one method for safely opening packages containing radioactive materials. Applicants are reminded that 10 CFR 20.1906(b) requires, in part, that licensees monitor the external surfaces of a labeled package for radioactive contamination within 3 hours of receipt if it is received during normal working hours, or not later than 3 hours from the beginning of the next working day if it is received after working hours.

**Response from Applicant:** No response required.

### 8.34 PROCEDURES FOR ADMINISTRATIONS WHEN A WRITTEN DIRECTIVE IS REQUIRED

**Regulations:** 10 CFR 35.27; 10 CFR 35.40; 10 CFR 35.41; 10 CFR 35.2040; 10 CFR 35.2041.

**Criteria:** 10 CFR 35.40 sets forth the requirements for WDs. 10 CFR 35.41 requires medical use licensees to develop, maintain, and implement written procedures to provide high confidence that licensed material is administered as directed by authorized users.

**Discussion:** The procedures do not need to be submitted to NRC. This gives licensees the flexibility to revise the procedures to enhance effectiveness without obtaining NRC approval. Appendix S provides guidance on developing the procedures.

**Response from Applicant:** No response required.

Part 35	Applicability
100	
200	
300	✓
400	✓
500	
600	✓
1000	✓

### 8.35 RELEASE OF PATIENTS OR HUMAN RESEARCH SUBJECTS

**Regulations:** 10 CFR 35.75; 10 CFR 35.2075.

**Criteria:** Licensees may release from confinement patients or human research subjects (patients) who have been administered licensed material if the TEDE to any other individual from exposure to the released patient is not likely to exceed 5 mSv (0.5 rem). Licensees must provide radiation safety instructions to patients released (or their parent or guardian) in accordance with 10 CFR 35.75(b).

**Discussion:** 10 CFR 35.75 requires that the licensee provide the released individual (patient) with instructions, including written instructions, on actions recommended to maintain doses to other individuals ALARA if the TEDE to any other individual is likely to exceed 1 mSv (0.1 rem). If the dose to a breast-feeding infant or a child could exceed 1 mSv (0.1 rem), assuming there was no interruption of breast-feeding, the instructions also shall include:

Part 35	Applicability
100	
200	
300	✓
400	✓
500	
600	
1000	✓

- Guidance on the interruption or discontinuation of breast-feeding; and
- Information on the potential consequences of failure to follow the guidance.

Appendix U provides guidance to the applicant on one way for determining when:

- The licensee may authorize the release of a patient who has been administered radiopharmaceuticals or who has been treated with implants containing radioactive material (Section 1), and
- Instructions to the patient are required by 10 CFR 35.75(b) (Section 2).
- Appendix U lists activities for commonly used radionuclides and the corresponding dose rates with which a patient may be released in compliance with the dose limits in 10 CFR 35.75.

**Response from Applicant:** No response required.

### 8.36. MOBILE MEDICAL SERVICE

**Regulations:** 10 CFR 35.2; 10 CFR 35.12; 10 CFR 35.18; 10 CFR 35.80; 10 CFR 35.647; 10 CFR 35.2080; 10 CFR 35.2647; 10 CFR 71.5; 10 CFR 71.12; 10 CFR 71.13; 10 CFR 71.14; 10 CFR 71.37; 10 CFR 71.38; Subpart H of 10 CFR Part 71; 10 CFR 150.20; 49 CFR Parts 171-178.

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓

**Criteria:** In addition to the requirements in 10 CFR 35.80, and 35.647 as applicable, mobile medical service licensees must comply with all other applicable regulations.

**Discussion:** Applicants for licensure of mobile medical services should review Sections 8.1 through 8.30 of this NUREG for information to be submitted as part of their applications; many of the requirements in these sections are relevant to use of byproduct material by mobile medical service providers with details being dependent upon the scope of such programs. "Temporary job site" means a location, other than specific location(s) of use authorized on the license, where mobile medical services are conducted. Mobile medical service licensees may transport licensed material and equipment into a client's building, or may bring patients into the transport (e.g., van). In either case, the van should be located on the client's property that is under the client's control.

Self-contained mobile medical service involves a mobile treatment or administration facility that provides ready-to-deliver mobile medical services on arrival at a client's site. Companies providing transportation only will not be licensed for medical use under 10 CFR Part 35. Before using a remote afterloader for this type of service, the device should be installed in an appropriately shielded treatment room.

## PROGRAM-RELATED GUIDANCE

The general types of services provided as mobile medical services are:

- Mobile medical services (byproduct material, trained personnel, and facility) that provide the device/facility (e.g., in-van use) and treatment of (or administration to) patients at the client site. These mobile medical service providers are responsible for all aspects of byproduct material use and authorized patient treatments (or administrations).
- Mobile medical service providers (byproduct material and trained personnel) that provide the transportation to and use of the byproduct material within the client's facility. These mobile medical service providers are also responsible for all aspects of byproduct material use and authorized patient treatments (or administrations).

Mobile medical service licensees must ensure that the criteria in 10 CFR 35.75 are met before releasing patients treated in their facilities.

Refer to Appendix V for additional guidance on information to provide in applications.

*Note:* Agreement State licensees that request reciprocity for activities conducted in non-Agreement States are subject to the general license provisions described in 10 CFR 150.20. This general license authorizes persons holding a specific license from an Agreement State to conduct the same activity in non-Agreement States if the specific license issued by the Agreement State does not limit the authorized activity to specific locations or installations. NRC licensees who wish to conduct operations at temporary job sites in an Agreement State should contact that state's Radiation Control Program Office for information about state regulations, including notification requirements, and to determine if mobile medical services are allowed within the Agreement State through reciprocity. Therefore, to ensure compliance with Agreement State reciprocity requirements, an NRC licensee shall request authorization well in advance of scheduled work. In addition to the requirements specified in 10 CFR 150.20, applicants requesting a mobile medical service license should contact all states where they plan to conduct mobile medical services, to clarify requirements associated with an authorization to practice medicine within the state's jurisdiction.

**Response from Applicant:** No response required.

### 8.37 AUDIT PROGRAM

**Regulations:** 10 CFR 20.1101; 10 CFR 20.2102.

**Criteria:** Under 10 CFR 20.1101, all licensees must annually review the content and implementation of the radiation protection program. The review should ensure the following:

- Compliance with NRC and applicable DOT regulations and the terms and conditions of the license; and
- Occupational doses and doses to members of the public are ALARA (10 CFR 20.1101).

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓

**Discussion:** The applicant should develop and implement procedures for the required review or audit of the radiation protection program's content and implementation. Appendix L contains model procedures that are only a suggested guide and are one way to meet this requirement. Some sections of Appendix L may not be pertinent to every licensee or to each review or audit. For example, licensees do not need to address areas that do not apply to their activities, and activities that have not occurred since the last review or audit need not be reviewed at the next review or audit. Reviews or audits of the content and implementation of the radiation protection program must be conducted at least annually.

NRC encourages licensee management to conduct performance-based reviews by observing work in progress, interviewing staff about the radiation protection program, and spot-checking required records. As part of their review programs, licensees should consider performing unannounced audits of authorized and supervised users to determine if, for example, Operating and Emergency Procedures are available and are being followed.

It is essential that once identified, violations and radiation safety concerns are corrected comprehensively and in a timely manner. The following three-step corrective action process has proven effective:

- Conduct a complete and thorough review of the circumstances that led to the violation.
- Identify the root cause of the violation.
- Take prompt and comprehensive corrective actions that will address the immediate concerns and prevent recurrence of the violation.

NRC's goal is to encourage prompt identification and prompt, comprehensive correction of violations and deficiencies.

**Response from Applicant:** No response is necessary.

**References:** See the Notice of Availability on the inside front cover of this report to obtain copies of: NUREG-1600, "General Statement of Policy and Procedures on NRC Enforcement Actions," and IN 96-28, "Suggested Guidance Relating to Development and Implementation of Corrective Action," dated May 1, 1996. NUREG-1600 is also available on the Internet at the NRC's web site, <<http://www.nrc.gov/reading-rm/doc-collections/nuregs/staff/sr1600/>>.

### 8.38 OPERATING AND EMERGENCY PROCEDURES

**Regulations:** 10 CFR 19.11(a)(3); 10 CFR 20.1101; 10 CFR 20.1601; 10 CFR 20.1602; 10 CFR 20.1801; 10 CFR 20.1802; 10 CFR 20.1906; 10 CFR 20.2102; 10 CFR 20.2201-2203; 10 CFR 21.21; 10 CFR 30.50; 10 CFR 35.12; 10 CFR 35.41; 10 CFR 35.75; 10 CFR 35.310; 10 CFR 35.315; 10 CFR 35.404; 10 CFR 35.406; 10 CFR 35.410; 10 CFR 35.415; 10 CFR 35.610; 10 CFR 35.615; 10 CFR 35.3045; 10 CFR 35.3047; 10 CFR 35.3067.

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓

**Criteria:** This section summarizes operating and emergency procedures. Many of these procedures are covered in greater detail in other sections of this document.

- Develop, implement, and maintain specific operating and emergency procedures containing the following elements:
  - Instructions for opening packages containing licensed material (see Section 8.33);
  - Using licensed material, operating therapy treatment devices, and performing routine maintenance on devices containing sealed sources, according to the manufacturer's written recommendations and instructions and in accordance with regulatory requirements (see Section 8.26);
  - Instructions for conducting area radiation level and contamination surveys (see Section 8.23);
  - Instructions for administering licensed material in accordance with the WD (see Section 8.34);
  - Steps to ensure that patient release is in accordance with 10 CFR 35.75 (see Section 8.35);
  - Instructions for calibration of survey and dosage measuring instruments (see Sections 8.16 and 8.17);
  - Periodic spot checks of therapy device units, sources, and treatment facilities (see Section 8.18);
  - Instructions for radioactive waste management (see Section 8.28);
  - Steps to take, and whom to contact (e.g., RSO, local officials), when the following has occurred: (a) leaking or damaged source, (b) device malfunction and/or damage, (c) licensed material spills, (d) theft or loss of licensed material, or (e) any other incidents involving licensed material (see Sections 8.25, 8.44);

- Steps for source retrieval and access control of damaged sealed source(s) and/or malfunctioning devices containing sealed source(s) (see Section 8.21);
- Steps to take if a therapy patient undergoes emergency surgery or dies.

**AND**

The licensee should consider the following:

- Make operating procedures, including emergency procedures, available to all users (e.g., post the procedures or the location of procedure storage);
- Maintain a current copy of the procedures at each location of use (or, if this is not practicable, post a notice describing the procedures and stating where they may be examined).
- When developing the procedures described above, the licensee is reminded that 10 CFR 20.1101(b) requires that the licensee use, to the extent practical, procedures and engineering controls based on sound radiation protection principles to achieve occupational doses and doses to members of the public that are ALARA.
- When receiving and using byproduct material, the licensee is reminded that it must be licensed to possess the byproduct material and that the radioactive material must be secured (or controlled) and accounted for at all times.

**Discussion:** Sealed sources and unsealed byproduct material used for therapy can deliver significant doses in a short time. 10 CFR 20.1601, 10 CFR 20.1602, 10 CFR 20.1801, and 10 CFR 20.1802 describe access control to high and very high radiation areas and the security of licensed material. Unauthorized access to licensed material by untrained individuals could lead to a significant radiological hazard. Many licensees achieve access control by permitting only trained individuals to have access to licensed material (e.g., keys, lock combinations, security badges). Accountability of licensed material may be ensured by conducting physical inventories, controlling receipt and disposal, and maintaining use records.

If a therapy patient undergoes emergency surgery or dies, it is necessary to ensure the safety of others attending the patient. As long as the patient's body remains unopened, the radiation received by anyone near it is due almost entirely to gamma rays. The change in emphasis when an operation or autopsy is to be performed is due to the possible exposure of the hands and face to relatively intense beta radiation. Procedures for emergency surgery or autopsy can be found in Section 5.3 of NCRP Report No. 37, "Precautions In The Management of Patients Who Have Received Therapeutic Amounts of Radionuclides."

Applicants should develop emergency procedures that address a spectrum of incidents (e.g., major spills, leaking source, medical events, interlock failure, stuck source, etc.).

After its occurrence becomes known to the licensee, NRC must be notified when an incident involving licensed material occurs. Refer to the regulations (10 CFR 20.2201-20.2203, 10 CFR

PROGRAM-RELATED GUIDANCE

30.50, 10 CFR 21.21, 10 CFR 35.3045, 10 CFR 35.3047, and 10 CFR 35.3067) for a description of when notifications are required.

Appendix N provides model procedures that are one method for responding to some types of emergencies.

**Response from Applicant:** No response is necessary.

**Reference:** Copies of NCRP Report No. 37, "Precautions In The Management of Patients Who Have Received Therapeutic Amounts of Radionuclides," NCRP Report No. 105, "Radiation Protection for Medical and Allied Health Personnel," 1989, and NCRP Report No. 107, "Implementation of the Principle of As Low As Reasonably Achievable (ALARA) for Medical and Dental Personnel," 1990," may be obtained from the National Council on Radiation Protection and Measurements, 7910 Woodmont Avenue, Suite 800, Bethesda, MD 20814-3095, or ordered electronically at <<http://www.ncrp.com>>.

### 8.39 MATERIAL RECEIPT AND ACCOUNTABILITY

**Regulations:** 10 CFR 20.1801; 10 CFR 20.1802; 10 CFR 20.1906; 10 CFR 20.2201; 10 CFR 30.35(g)(2); 10 CFR 30.41; 10 CFR 30.51; 10 CFR 35.67; 10 CFR 35.406.

**Criteria:** To maintain accountability of licensed material, licensees must do the following:

- Secure licensed material;
- Maintain records of receipt, transfer, and disposal of licensed material; and
- Conduct physical inventories at required frequencies to account for licensed material.

**Discussion:** Licensed materials must be tracked from "cradle to grave" to ensure accountability, identify when licensed material could be lost, stolen, or misplaced, and ensure that possession limits listed on the license are not exceeded.

**Response from Applicant:** No response is necessary.

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓

### 8.40 ORDERING AND RECEIVING

**Regulations:** 10 CFR 20.1801; 10 CFR 20.1802; 10 CFR 20.1906; 10 CFR 30.51.

**Criteria:** 10 CFR 20.1906 contains the requirements for receiving packages containing licensed material. Additionally, the security of licensed material, required by 10 CFR 20.1801 and 10 CFR 20.1802, must be considered for all receiving areas. 10 CFR 30.51 requires licensees, in part, to maintain records showing the receipt of byproduct material.

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓

**Discussion:** Licensees must ensure that the type and quantity of licensed material possessed is in accordance with the license. Additionally, licensees must ensure that packages are secured and radiation exposure from packages is minimized.

Appendix O contains model procedures that are one method for ordering and receiving licensed material.

**Response from Applicant:** No response is necessary.

### 8.41 SEALED SOURCE INVENTORY

**Regulations:** 10 CFR 20.1801; 10 CFR 20.1802; 10 CFR 30.51; 10 CFR 35.67; 10 CFR 35.406; 10 CFR 35.2067; 10 CFR 35.2406.

**Criteria:** NRC requires the licensee in possession of a sealed source or brachytherapy source to conduct a semi-annual physical inventory of all such sources in its possession.

**Discussion:** According to 10 CFR 35.67, the licensee must conduct a semi-annual physical inventory of all sealed sources and brachytherapy sources in its possession. Individual GSR sources are exempt from this physical inventory requirement, as stated in 10 CFR 35.67(g). However, the licensee must maintain records of GSR source receipt, transfer, and disposal, under 10 CFR 30.51, to indicate the current inventory of sources at the licensee's facility.

**Response from Applicant:** No response is necessary.

Part 35	Applicability
100	✓*
200	✓*
300	✓*
400	✓
500	✓
600	✓
1000	✓

\* Sealed sources for calibration, transmission, and reference use (35.65)

### 8.42 RECORDS OF DOSAGES AND USE OF BRACHYTHERAPY SOURCE

**Regulations:** 10 CFR 30.51; 10 CFR 35.63; 10 CFR 35.2063; 10 CFR 35.2204; 10 CFR 35.2406.

**Criteria:** Licensees must record the use of licensed material to reflect proper use and accountability. Records of use must be maintained for 3 years.

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	
600	
1000	✓

**Discussion:** Licensees are required to make and maintain records of each dosage and administration prior to medical use. The records must include:

- Radiopharmaceutical;
- Patient's or human research subject's name or identification number (if one has been assigned);

#### PROGRAM-RELATED GUIDANCE

- Prescribed dosage, determined dosage, or a notation that the total activity is less than 1.1 MBq (30  $\mu$ Ci);
- Date and time of dosage determination; and
- Name of the individual who determined the dosage.

Dosage determination for unit dosages may be made either by direct measurement or by a decay correction based on the determination (e.g., measurement) made by the manufacturer or preparer licensed under 10 CFR 32.72 or equivalent Agreement State requirements.

If molybdenum concentration is measured under 10 CFR 35.204, records of molybdenum concentration must be made under 10 CFR 35.2204 and must include, for each measured elution of technetium-99m:

- Ratio of the measurements expressed as kBq ( $\mu$ Ci) of molybdenum-99 per MBq (mCi) of technetium-99m;
- Date and time of the measurement; and
- Name of the individual who made the measurement.

If the licensee uses manual brachytherapy sources, the following records of use must be kept:

- When temporary implant brachytherapy sources are removed from storage, a record will include the number and activity of sources removed, the time and date they were removed from storage, the location of use, and the name of the individual who removed them from storage.
- When temporary implant brachytherapy sources are returned to storage, a record will include the number and activity of sources returned, the time and date they were returned to storage, and the name of the individual who returned them to storage.
- For permanent implants, a record will be made and will include the number and activity of sources removed from storage, the date they were removed from storage, the name of the individual who removed them from storage, the number and activity of sources not implanted, the date they were returned to storage, the name of the individual who returned them to storage, and the number and activity of sources permanently implanted in the patient or human research subject.

**Response from Applicant:** No response is necessary.

### 8.43 RECORDKEEPING

**Regulations:** 10 CFR Part 20, Subpart L; 10 CFR 30.51; 10 CFR Part 35 Subpart L.

**Criteria:** Licensees must maintain records as provided in 10 CFR Part 20, Subpart L; 10 CFR 30.51; and 10 CFR Part 35 Subpart L.

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓

**Discussion:** The licensee must maintain certain records to comply with NRC regulations, the conditions of the license, and commitments made in the license application and correspondence with NRC. Operating procedures should identify which individuals in the organization are responsible for maintaining which records.

A table of recordkeeping requirements appears in Appendix X.

**Response from Applicant:** No response is necessary.

### 8.44 REPORTING

**Regulations:** 10 CFR Part 20, Subpart M; 10 CFR 21.21; 10 CFR 30.50; 10 CFR Part 35, Subpart M.

**Criteria:** Licensees are required to report to NRC via telephone, written report, or both in the event that the safety or security of byproduct material may be compromised. The specific events that require reporting are explained in Subpart M of Part 35, Subpart M of Part 20; and in 10 CFR CFR 21.21 and 30.50. The timing and type of report are specified within these parts.

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓

**Discussion:** The NRC requires licensees to report incidents that might compromise the health and safety of patients, health care providers, or the public. Therefore, Parts 20, 21, 30, and 35 include provisions that describe reporting requirements associated with the medical use of byproduct material.

A table of reporting requirements appears in Appendix Y.

**Response from Applicant:** No response is necessary.

## 8.45 LEAK TESTS

**Regulations:** 10 CFR 20.1501; 10 CFR 20.2103; 10 CFR 30.53; 10 CFR 35.67; 10 CFR 35.2067; 10 CFR 35.3067.

**Criteria:** NRC requires testing to determine if there is any radioactive leakage from sealed sources.

**Discussion:** Licensees must perform leak testing of sealed sources, e.g., calibration, transmission, and reference sources, or brachytherapy sources in accordance with 10 CFR 35.67.

Appendix Q provides model procedures that are one way to perform leak testing. 10 CFR 35.67 requires licensees to perform leak tests at six-month intervals or at other intervals approved by NRC or an Agreement State and specified in the SSDR certificate and before first use unless accompanied by a certificate indicating that the test was performed within the past 6 months. The measurement of the leak test sample is a quantitative analysis requiring that instrumentation used to analyze the sample be capable of detecting 185 Bq (0.005  $\mu$ Ci) of radioactivity on the sample. Leak test samples should be collected at the most accessible area where contamination would accumulate if the sealed source were leaking.

The leak test may be performed in-house or by a contractor who is authorized by NRC or an Agreement State to perform leak tests as a service to other licensees.

The licensee or contractor does not need to leak-test sources if:

- Sources contain only byproduct material with a half-life of less than 30 days;
- Sources contain only byproduct material as a gas;
- Sources contain 3.7 MBq (100  $\mu$ Ci) or less of beta-emitting or gamma-emitting material, or 0.37 MBq (10  $\mu$ Ci) or less of alpha-emitting material;
- Sources contain Ir-192 seeds in nylon ribbon; or
- Sources are stored and not being used. The licensee, shall, however, test each such source for leakage before any use or transfer unless it has been leak-tested within 6 months before the date of use or transfer.

**Response from Applicant:** No response is necessary.

**References:** See the Notice of Availability on the inside front cover of this report to obtain a copy of NUREG-1556, Vol. 18, "Program-Specific Guidance About Service Provider Licenses," dated November 2000.

Part 35	Applicability
100	✓*
200	✓*
300	✓*
400	✓
500	✓
600	✓
1000	✓

\*If possess sealed sources under 35.65

## 8.46 SAFETY PROCEDURES FOR TREATMENTS WHEN PATIENTS ARE HOSPITALIZED

**Regulations:** 10 CFR 20.1101; 10 CFR 20.1301; 10 CFR 20.1501; 10 CFR 20.1801; 10 CFR 20.2103; 10 CFR 35.310; 10 CFR 35.315; 10 CFR 35.404; 10 CFR 35.410; 10 CFR 35.415; 10 CFR 35.604; 10 CFR 35.610; 10 CFR 35.615; 10 CFR 35.2404.

Part 35	Applicability
100	
200	
300	✓
400	✓
500	
600	✓
1000	✓

**Criteria:** Applicants must develop and implement procedures to ensure that access to therapy treatment rooms, and exposure rates from therapy treatments, are limited to maintain doses to occupational workers and members of the public within regulatory limits.

**Discussion:** 10 CFR 35.315, 10 CFR 35.415, and 10 CFR 35.615 require licensees to take certain safety precautions for uses of byproduct material involving radiopharmaceutical therapy, manual brachytherapy, or remote afterloader brachytherapy involving patients who cannot be released in accordance with 10 CFR 35.75. This section of the guidance does not include guidance on this subject for teletherapy or GSR outpatient treatments. The precautions described below are provided to help ensure compliance with the exposure limits in 10 CFR Part 20.

10 CFR 35.404(b) and 10 CFR 35.604(a) require licensees to perform a radiation survey of the patient (and the remote afterloader unit) immediately after removing the last temporary implant source from the patient and prior to releasing the patient from licensee control. This is done to confirm that all sources have been removed and accounted for. 10 CFR 35.615(e) requires that when sources are placed within the patient's body, licensed activities be limited to treatments that allow for expeditious removal of a decoupled or jammed source.

In addition, applicants must take the following steps for patients who cannot be released under 10 CFR 35.75:

- Provide a room with a private sanitary facility for patients treated with a radiopharmaceutical therapy dosage (Note: 10 CFR 35.315(a) allows for a room shared with another radiopharmaceutical therapy patient);
- Provide a private room for patients implanted with brachytherapy sources (Note: 10 CFR 35.415 allows for a room shared with another brachytherapy patient);
- Visibly post a "Radioactive Materials" sign on the patient's room and note on the door or in the patient's chart where and how long visitors may stay in the patient's room (10 CFR 35.315 and 10 CFR 35.415);
- Either monitor material and items removed from the patient's room (e.g., patient linens, surgical dressings) with a radiation detection survey instrument set on its most sensitive scale with no interposed shielding to determine that their radioactivity cannot be distinguished

## PROGRAM-RELATED GUIDANCE

from the natural background radiation level or handle them as radioactive waste (10 CFR 35.315 and 10 CFR 20.1501); and

- Notify the RSO, or his/her designee, and AU as soon as possible if the patient has a medical emergency or dies (10 CFR 35.315, 10 CFR 35.415, and 10 CFR 35.615).

10 CFR 20.1501 requires licensees to perform adequate surveys to evaluate the extent of radiation levels. Therefore, licensees must evaluate the exposure rates around patients who are hospitalized in accordance with 10 CFR 35.75 following the dosage administration or implant (e.g., measured exposure rates, combination of measured and calculated exposure rates).

10 CFR 20.1801 requires licensees to secure licensed material in storage from unauthorized access or removal. Access control and appropriate training of authorized personnel may prevent unauthorized removal of licensed material temporarily stored in the patient's room and unnecessary personnel exposures.

In order to control exposures to individuals in accordance with 10 CFR Part 20, the licensee should consider briefing patients on radiation safety procedures for confinement to bed, visitor control, identification of potential problems, notification of medical staff in the event of problems, and other items as applicable and consistent with good medical care.

**Response from Applicant:** No response is necessary.

## 8.47 TRANSPORTATION

**Regulations:** 10 CFR 20.1101; 10 CFR 30.41; 10 CFR 30.51; 10 CFR 71.5; 10 CFR 71.9; 10 CFR 71.12; 10 CFR 71.13; 10 CFR 71.14; 10 CFR 71.37; 10 CFR 71.38; Subpart H of 10 CFR Part 71; 49 CFR Parts 171-178.

**Criteria:** Applicants who will prepare for shipment, ship, or transport radioactive materials, including radioactive waste, must develop, implement, and maintain safety programs for the transport of radioactive material to ensure compliance with NRC and DOT regulations.

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓

**Discussion:** Most packages of licensed material for medical use contain quantities of radioactive material that require use of Type A packages. Additionally, many packages shipped by medical licensees (e.g., unused radiopharmaceutical dosages) frequently meet the "Limited Quantity" criteria described in 49 CFR 173.421 and are therefore excepted from certain DOT requirements, provided certain other less restrictive requirements are met [e.g., activity in the package is less than the limited quantity and the radiation level on the surface of the package does not exceed 0.005 mSv per hour (0.5 mrem per hour)].

The general license in 10 CFR 71.12, "General license: NRC-approved package," provides the authorization used by most licensees to transport, or to deliver to a carrier for transport, licensed material in a package for which a license, certificate of compliance, or other approval has been

issued by NRC. This general license is subject to certain conditions. 10 CFR 71.5 sets forth the requirements for transportation of licensed material. 10 CFR 71.9 exempts any physician licensed by a state to dispense drugs in the practice of medicine, who is also licensed under 10 CFR Part 35 or the equivalent Agreement State regulations from the requirements in 10 CFR 71.5. This exemption applies to transport by the physician of licensed material for use in the practice of medicine.

Some medical use licensees (e.g., teletherapy or gamma stereotactic radiosurgery) may need to ship licensed material in Type B packages. 10 CFR 71.12-71.14 sets forth the Type B package requirements for transporting or delivering the package to a carrier for transport. These include registration as a user of the package and having an NRC-approved quality assurance (QA) plan. For information about these QA plans, see Revision 1 of RG 7.10, "Establishing Quality Assurance Programs for Packaging Used in the Transport of Radioactive Material," dated June 1986. For further information about registering as a user of a package or submitting a QA program for review, contact NRC's Spent Fuel Project Office by calling NRC toll-free at (800) 368-5642, extension 415-8500. For information about associated fees, contact NRC's OCFO by calling NRC toll-free at (800) 368-5642, extension 415-7544.

Some medical use licensees that ship radioactive material have chosen to transfer possession of radioactive materials to a manufacturer (or service licensee) with an NRC or Agreement State license, who then acts as the shipper. The manufacturer (or service licensee), who is subject to the provisions of 10 CFR 71.12 or 10 CFR 71.14, as appropriate, then becomes responsible for proper packaging of the radioactive materials and compliance with NRC and DOT regulations. Licensees who do this must ensure that the manufacturer (or service licensee):

- Is authorized to possess the licensed material (see 10 CFR 30.41).
- Actually takes possession of the licensed material under its license.

Licensees should also ensure that the manufacturer (or service licensee) is authorized to possess the material at temporary job sites (e.g., the licensee's facilities).

During an inspection, NRC uses the provisions of 10 CFR 71.5 and a Memorandum of Understanding with DOT on the Transportation of Radioactive Material (signed June 6, 1979) to examine and enforce various DOT requirements applicable to medical use licensees. Appendix Z lists major DOT regulations that apply to medical licensees.

**Response from Applicant:** No response is needed from applicants during the licensing phase. However, before making shipments of licensed materials on its own in a Type B package, a licensee must have registered with NRC as a user of the package and obtained NRC's approval of its QA program. Transportation issues will be reviewed during inspection.

#### References:

- "A Review of Department of Transportation Regulations for Transportation of Radioactive Materials" can be obtained by calling DOT's Office of Hazardous Material Initiatives and Training at (202) 366-4425.

#### PROGRAM-RELATED GUIDANCE

- See the Notice of Availability on the inside front cover of this report to obtain a copy of the Memorandum of Understanding with DOT on the Transportation of Radioactive Material, signed June 6, 1979; Revision 1 of RG 7.10, "Establishing Quality Assurance Programs for Packaging Used in the Transport of Radioactive Material," dated June 1986; and NUREG-1556 Vol. 18, Program-Specific Guidance About Service Provider Licenses."

## 9 AMENDMENTS AND RENEWALS TO A LICENSE

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓

**Regulations:** 10 CFR 35.13.

Licenses are responsible for applying for amendments to licenses and for keeping them up-to-date. Furthermore, to continue a license after its expiration date, the licensee must submit an application for a license renewal at least 30 days before the expiration date (10 CFR 2.109, 10 CFR 30.36(a)).

10 CFR 35.13 requires a licensee to apply for and receive a license amendment before several activities can occur, including:

- Receipt or use of byproduct material for a type of use permitted by Part 35, but not authorized on the licensee's current Part 35 license;
- Permitting anyone to work as an AU, AMP, or ANP, unless the individual meets one of the exceptions listed in 10 CFR 35.13(b) (Supply information required to document training and experience on NRC Form 313A for change or addition of AU, AMP, ANP, or RSO);
- Changing the RSO;
- Receiving byproduct material in excess of the amount, or receiving radionuclides or forms different than currently authorized on the NRC license;
- Changing an area or address of use identified in the application or on the license, except for areas of use where byproduct material is used only in accordance with either 10 CFR 35.100 or 10 CFR 35.200; and
- Revising procedures required by 10 CFR 35.610, 35.642, 35.643, and 35.645, when the revision reduces the level of radiation safety.

In case of a medical emergency requiring an expedited license amendment, contact the NRC regional materials licensing staff.

For both renewal and amendment requests, applicants should do the following:

- Use the most recent guidance in preparing an amendment or renewal request;
- Submit in duplicate either an NRC Form 313 or a letter requesting an amendment or renewal; and
- Provide the license number.

## 10 APPLICATIONS FOR EXEMPTIONS

**Regulations:** 10 CFR 19.31; 10 CFR 20.2301; 10 CFR 30.11; 10 CFR 35.15; 10 CFR 35.19.

**Criteria:** Licensees may request exemptions to regulations. The licensee must demonstrate that the exemption is authorized by law, will not endanger life or property or the common defense and security, and is otherwise in the public interest.

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓

**Discussion:** Various sections of NRC's regulations address requests for exemptions (e.g., 10 CFR 19.31, 10 CFR 20.2301, 10 CFR 30.11(a)). These regulations state that NRC may grant an exemption, acting on its own initiative or on an application from an interested person.

Exemptions are not intended to revise regulations, are not intended for large classes of licenses, and are generally limited to unique situations. Exemption requests should be accompanied by descriptions of the following:

- Exemption and justification of why it is needed.
- Proposed compensatory safety measures intended to provide a level of health and safety equivalent to the regulation for which the exemption is being requested.
- Alternative methods for complying with the regulation and why compliance with the existing regulations is not feasible.

Until NRC has granted an exemption in writing, NRC expects strict compliance with all applicable regulations.

Type A broad scope licensees are granted certain exemptions as described in 10 CFR 35.15.

# 11 TERMINATION OF ACTIVITIES

Part 35	Applicability
100	✓
200	✓
300	✓
400	✓
500	✓
600	✓
1000	✓

**Regulations:** 10 CFR 20.1401; 10 CFR 20.1402; 10 CFR 20.1403; 10 CFR 20.1404; 10 CFR 20.1405; 10 CFR 20.1406; 10 CFR 30.34(b); 10 CFR 30.35(g); 10 CFR 30.36; 10 CFR 30.51(f).

**Criteria:** Pursuant to the regulations described above, the licensee must do the following:

- Notify NRC, in writing, within 60 days of:
  - the expiration of its license;
  - a decision to permanently cease licensed activities at the entire site (regardless of contamination levels);
  - a decision to permanently cease licensed activities in any separate building or outdoor area if it contains residual radioactivity making it unsuitable for release according to NRC requirements;
  - no principal activities having been conducted at the entire site under the license for a period of 24 months; and
  - no principal activities having been conducted for a period of 24 months in any separate building or outdoor area if it contains residual radioactivity making it unsuitable for release according to NRC requirements.
- Submit a decommissioning plan, if required by 10 CFR 30.36(g);
- Conduct decommissioning, as required by 10 CFR 30.36(h) and (j); and
- Submit, to the appropriate NRC Regional Office, a completed NRC Form 314, "Certificate of Disposition of Materials," (or equivalent information) and demonstrate that the premises are suitable for release for unrestricted use (e.g., results of final survey).
- Before a license is terminated, send the records important to decommissioning to the appropriate NRC Regional Office. If licensed activities are transferred or assigned in accordance with 10 CFR 30.34(b), transfer records important to decommissioning to the new licensee.

## TERMINATION OF ACTIVITIES

**Discussion:** Useful guidance and other aids related to decommissioning are:

- NUREG-1727, "NMSS Decommissioning Standard Review Plan," dated September 2000.
- NUREG/BR-0241, "NMSS Handbook for Decommissioning Fuel Cycle and Materials Licenses," dated March 1997, contain the current regulatory guidance concerning decommissioning of facilities and termination of licenses.
- Appendix B of NUREG/BR-0241 contains a comprehensive list of NRC's decommissioning regulations and guidance. NUREG-1727 contains a list of superceded guidance; however, due to ongoing revisions, applicants are encouraged to consult with NRC staff regarding updates of decommissioning guidance.
- NUREG-1575, "Multi-Agency Radiation Survey and Site Investigation Manual (MARSSIM)," dated December 1997, should be reviewed by licensees who have large facilities to decommission.
- An acceptable computer code for calculating screening values to demonstrate compliance with the unrestricted dose limits is DandD, Version 2.1.0, (McFadden and others, 2001).
- NUREG-1727 includes a table (Table C2.2) of acceptable license termination screening values of common beta/gamma radionuclides for building surface contamination. NUREG-1727 also contains methods for conducting site-specific dose assessments for facilities with contamination levels above those in the table.

**Response from Applicant:** The applicant is not required to submit a response to NRC during the initial application. The licensee's obligations are to undertake the necessary decommissioning activities, to submit NRC Form 314 or equivalent information, and to perform any other actions as summarized in the "Criteria."

### References:

- Copies of NRC Form 314, "Certificate of Disposition of Materials," are available upon request from NRC's Regional Offices. (See Figure 2.1 for addresses and telephone numbers.)
- McFadden, K., D.A. Brosseau, W.A. Beyeler, and C.D. Updegraff, "Residual Radioactive Contamination from Decommissioning - User's Manual DandD Version 2.1," NUREG/CR-5512, Volume 2, U.S. Nuclear Regulatory Commission, Washington, DC, April 2001.

# APPENDICES

**APPENDICES A-H**  
**FORMS AND SAMPLES**

**APPENDIX A**

**NRC Form 313  
“Application for Material License”**

<b>NRC FORM 313</b> (8-1999) 10 CFR 30.32-33 34.35.36.39 and 40	<b>U. S. NUCLEAR REGULATORY COMMISSION</b>	APPROVED BY OMB NO 3150-0120 EXPIRES 10/10/2005	Estimated burden per response to comply with this mandatory information collection request: 7.4 hours. Submittal of the application is necessary to determine that the applicant is qualified and that adequate procedures exist to protect the public health and safety. Send comments regarding burden estimate to the Records Management Branch (T 8 E6), U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001, or by internet e-mail to bjs1@nrc.gov and to the Desk Officer, Office of Information and Regulatory Affairs, NE08-10202 (3150-0120), Office of Management and Budget, Washington, DC 20503. If a means used to impose an information collection does not display a currently valid OMB control number, NRC may not conduct or sponsor, and a person is not required to respond to the information collection.		
APPLICATION FOR MATERIAL LICENSE					
<b>INSTRUCTIONS: SEE THE APPROPRIATE LICENSE APPLICATION GUIDE FOR DETAILED INSTRUCTIONS FOR COMPLETING APPLICATION. SEND TWO COPIES OF THE ENTIRE COMPLETED APPLICATION TO THE NRC OFFICE SPECIFIED BELOW.</b>					
APPLICATION FOR DISTRIBUTION OF EXEMPT PRODUCTS FILE APPLICATIONS WITH DIVISION OF INDUSTRIAL AND MEDICAL NUCLEAR SAFETY OFFICE OF NUCLEAR MATERIALS SAFETY AND SAFEGUARDS U.S. NUCLEAR REGULATORY COMMISSION WASHINGTON, DC 20555-0001  ALL OTHER PERSONS FILE APPLICATIONS AS FOLLOWS IF YOU ARE LOCATED IN  CONNECTICUT, DELAWARE, DISTRICT OF COLUMBIA, MAINE, MARYLAND, MASSACHUSETTS, NEW HAMPSHIRE, NEW JERSEY, NEW YORK, PENNSYLVANIA, RHODE ISLAND, OR VERMONT, SEND APPLICATIONS TO  LICENSING ASSISTANT SECTION NUCLEAR MATERIALS SAFETY BRANCH U.S. NUCLEAR REGULATORY COMMISSION REGION I 475 ALLENDALE ROAD KING OF PRUSSIA, PA 19406-1415  ALABAMA, FLORIDA, GEORGIA, KENTUCKY, MISSISSIPPI, NORTH CAROLINA, PUERTO RICO, SOUTH CAROLINA, TENNESSEE, VIRGINIA, VIRGIN ISLANDS, OR WEST VIRGINIA SEND APPLICATIONS TO  SAM NUNN ATLANTA FEDERAL CENTER U.S. NUCLEAR REGULATORY COMMISSION REGION II 61 FORSYTH STREET, S.W. SUITE 23785 ATLANTA, GEORGIA 30303-8931		IF YOU ARE LOCATED IN  ILLINOIS, INDIANA, IOWA, MICHIGAN, MINNESOTA, MISSOURI, OHIO, OR WISCONSIN SEND APPLICATIONS TO  MATERIALS LICENSING SECTION U.S. NUCLEAR REGULATORY COMMISSION, REGION III 801 WARRENVILLE RD LISLE, IL 60532-4351  ALASKA, ARIZONA, ARKANSAS, CALIFORNIA, COLORADO, HAWAII, IDAHO, KANSAS, LOUISIANA, MONTANA, NEBRASKA, NEVADA, NEW MEXICO, NORTH DAKOTA, OKLAHOMA, OREGON, PACIFIC TRUST TERRITORIES, SOUTH DAKOTA, TEXAS, UTAH, WASHINGTON, OR WYOMING, SEND APPLICATIONS TO  NUCLEAR MATERIALS LICENSING SECTION U.S. NUCLEAR REGULATORY COMMISSION REGION IV 611 RYAN PLAZA DRIVE, SUITE 400 ARLINGTON, TX 76011-8064			
PERSONS LOCATED IN AGREEMENT STATES SEND APPLICATIONS TO THE U.S. NUCLEAR REGULATORY COMMISSION ONLY IF THEY WISH TO POSSESS AND USE LICENSED MATERIAL IN STATES SUBJECT TO U.S. NUCLEAR REGULATORY COMMISSION JURISDICTIONS					
1 THIS IS AN APPLICATION FOR (Check appropriate item) <input type="checkbox"/> A NEW LICENSE <input type="checkbox"/> B AMENDMENT TO LICENSE NUMBER _____ <input type="checkbox"/> C RENEWAL OF LICENSE NUMBER _____		2 NAME AND MAILING ADDRESS OF APPLICANT (include Zip code)			
3 ADDRESS(ES) WHERE LICENSED MATERIAL WILL BE USED OR POSSESSED		4 NAME OF PERSON TO BE CONTACTED ABOUT THIS APPLICATION  TELEPHONE NUMBER			
SUBMIT ITEMS 5 THROUGH 11 ON 8-1/2 X 11" PAPER. THE TYPE AND SCOPE OF INFORMATION TO BE PROVIDED IS DESCRIBED IN THE LICENSE APPLICATION GUIDE.					
5 RADIOACTIVE MATERIAL a. Element and mass number b. chemical and/or physical form and c. maximum amount which will be possessed at any one time		6 PURPOSE(S) FOR WHICH LICENSED MATERIAL WILL BE USED			
7 INDIVIDUAL(S) RESPONSIBLE FOR RADIATION SAFETY PROGRAM AND THEIR TRAINING EXPERIENCE		8 TRAINING FOR INDIVIDUALS WORKING IN OR FREQUENTING RESTRICTED AREAS			
9 FACILITIES AND EQUIPMENT		10 RADIATION SAFETY PROGRAM			
11 WASTE MANAGEMENT		12 LICENSEE FEES (See 10 CFR 170 and Section 170.31) FEE CATEGORY: _____ AMOUNT ENCLOSED \$ _____			
13 CERTIFICATION (Must be completed by applicant) THE APPLICANT UNDERSTANDS THAT ALL STATEMENTS AND REPRESENTATIONS MADE IN THIS APPLICATION ARE BINDING UPON THE APPLICANT.  THE APPLICANT AND ANY OFFICIAL EXECUTING THIS CERTIFICATION ON BEHALF OF THE APPLICANT, NAMED IN ITEM 2, CERTIFY THAT THIS APPLICATION IS PREPARED IN CONFORMITY WITH TITLE 10 CODE OF FEDERAL REGULATIONS PARTS 30.32, 33, 34, 35, 36, 39, AND 40, AND THAT ALL INFORMATION CONTAINED HEREIN IS TRUE AND CORRECT TO THE BEST OF THEIR KNOWLEDGE AND BELIEF.  WARNING: 18 U.S.C. SECTION 1001, ACT OF JUNE 25, 1948, 62 STAT. 749, MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.					
CERTIFYING OFFICER - TYPED/PRINTED NAME AND TITLE		SIGNATURE	DATE		
FOR NRC USE ONLY					
TYPE OF FEE	FEE LOG	FEE CATEGORY	AMOUNT RECEIVED \$	CHECK NUMBER	COMMENTS
APPROVED BY				DATE	

**APPENDIX B**

**NRC Form 313A**

**“Training and Experience and Preceptor Statement”**

NRC FORM 313A  
(10-2002)

U.S. NUCLEAR REGULATORY COMMISSION

APPROVED BY OMB: NO. 3150-0120  
EXPIRES: 10/31/2005

**TRAINING AND EXPERIENCE AND PRECEPTOR STATEMENT**

**PART I -- TRAINING AND EXPERIENCE**

**Note:** Descriptions of training and experience must contain sufficient detail to match the training and experience criteria in the applicable regulations.

1. Name of Individual, Proposed Authorization (e.g., Radiation Safety Officer), and Applicable Training Requirements (e.g., 10 CFR 35.50)

2. For Physicians, Podiatrists, Dentists, Pharmacists -- State or Territory Where Licensed

**3. CERTIFICATION**

Specialty Board	Category	Month and Year Certified

*Stop here when using Board Certification to meet 10 CFR Part 35 training and experience requirements.*

**4. DIDACTIC OR CLASSROOM AND LABORATORY TRAINING (optional for Medical Physicists)**

Description of Training	Location	Clock Hours	Dates of Training
Radiation Physics and Instrumentation			
Radiation Protection			
Mathematics Pertaining to the Use and Measurement of Radioactivity			
Radiation Biology			
Chemistry of Byproduct Material for Medical Use			
OTHER			

TRAINING AND EXPERIENCE AND PRECEPTOR STATEMENT (continued)

5a. WORK EXPERIENCE WITH RADIATION

Description of Experience	Name of Supervising Individual(s)	Location and Corresponding Materials License Number	Dates and Clock Hours of Experience

5b. SUPERVISED CLINICAL CASE EXPERIENCE

Radionuclide	Type of Use	No. of Cases Involving Personal Participation	Name of Supervising Individual	Location and Corresponding Materials License Number	Dates and Clock Hours of Experience

NRC FORM 313A  
(10-2002)

U.S. NUCLEAR REGULATORY COMMISSION

**TRAINING AND EXPERIENCE AND PRECEPTOR STATEMENT (continued)**

**6. FORMAL TRAINING (applies to Medical Physicists and Therapy Physicists)**

Degree, Area of Study or Residency Program	Name of Program and Location with Corresponding Materials License Numbers	Dates	Name of Organization that Approved the Program (e.g., Accreditation Council for Graduate Medical Education) and the Applicable Regulation (e.g., 10 CFR 35.490)

**7. RADIATION SAFETY OFFICER -- ONE-YEAR FULL-TIME TRAINING**

- YES Completed 1-year of full-time radiation safety experience (in areas identified in item 5a) under supervision
- N/A of \_\_\_\_\_ the RSO for License No. \_\_\_\_\_

**8. MEDICAL PHYSICIST -- ONE YEAR FULL-TIME TRAINING/WORK EXPERIENCE**

- YES Completed 1-year of full-time training in therapeutic radiological physics under the supervision of
- N/A \_\_\_\_\_ who meets requirements for Authorized Medical Physicists; and
- YES Completed 1-year of full-time work experience (for areas identified in item 5a) for \_\_\_\_\_
- N/A modality(ies) under the supervision of \_\_\_\_\_ who meets requirements for Authorized Medical Physicists for \_\_\_\_\_ modality(ies).

**9. SUPERVISING INDIVIDUAL -- IDENTIFICATION AND QUALIFICATIONS**

The training and experience indicated above was obtained under the supervision of (if more than one supervising individual is needed to meet requirements in 10 CFR 35, provide the following information for each):

- A. Name of Supervisor \_\_\_\_\_
- B. Supervisor is:
  - Authorized User
  - Radiation Safety Officer
  - Authorized Medical Physicists
  - Authorized Nuclear Pharmacists
- C. Supervisor meets requirements of Part 35, Section(s) \_\_\_\_\_ for medical uses in Part 35, Section(s) \_\_\_\_\_
- D. Address \_\_\_\_\_
- E. Materials License Number \_\_\_\_\_

NRC FORM 313A  
(10-2002)

U.S. NUCLEAR REGULATORY COMMISSION

**TRAINING AND EXPERIENCE AND PRECEPTOR STATEMENT (continued)**

**PART II -- PRECEPTOR STATEMENT**

**Note:** *This part must be completed by the individual's preceptor. If more than one preceptor is necessary to document experience, obtain a separate preceptor statement from each. This part is not required to meet the training requirements in 10 CFR 35.590.*

Item 10 must be completed for Nuclear Pharmacists meeting the requirements of 10 CFR Part 35, Subpart J. Preceptors do not have to complete items 11a, 11b, or the certifying statements for other individuals meeting the requirements of 10 CFR Part 35, Subpart J.

YES 10. The individual named in item 1 has satisfactorily completed the training requirements in  
 N/A 10 CFR 35.980 and is competent to independently operate a nuclear pharmacy.

YES 11a. The individual named in Item 1 has satisfactorily completed the requirements in Part 35, Section(s)  
 N/A and Paragraph(s) \_\_\_\_\_.

YES 11b. The individual named in Item 1. is competent to independently function as an authorized  
 N/A \_\_\_\_\_ for \_\_\_\_\_ uses.

**12. PRECEPTOR APPROVAL AND CERTIFICATION**

I certify the approval of item 10 and certify I am an Authorized Nuclear Pharmacist;

**or**

I certify the approval of items 11a and 11b and certify I am an Authorized Nuclear Pharmacist;

**or**

I certify the approval of Items 11a and 11b; and I certify that I meet the requirements of \_\_\_\_\_  
 or equivalent Agreement State requirements to be a preceptor authorized \_\_\_\_\_  
 for the following uses of byproduct material: \_\_\_\_\_

A. Address

B. Materials License Number

C. NAME OF PRECEPTOR (*print clearly*)

D. SIGNATURE -- PRECEPTOR

E. DATE

## **APPENDIX C**

### **License Application Checklists**

## License Application Checklists

This Appendix contains checklists that may be used to assist in organizing an application.

Table C.1, Applicability Table, may be used to determine if particular information must be provided or if "N/A" may be the response to each item that follows. To determine those items to which you must respond, "highlight" the columns under the categories of materials you requested in Item 5 (e.g., 35.300, 35.400, etc.). If any "Y" beside an item is highlighted, you must provide detailed information in response to that item. If the letters "N/A" (not applicable) are highlighted, you may respond "N/A" on your application. If any "N" beside an item is highlighted, no information in response is required, but NRC regulations that apply to the given category apply to your type of license. If any "P" beside an item is highlighted, you should provide a commitment as described in the section referenced in the body of this document. If any "G" beside an item is highlighted, see subsequent sections for required responses. "APP" indicates that an appendix found in this document addresses the item.

**Table C.1 Applicability Table**

Item #	Topic	35.100/200	35.300	35.400	35.500	35.600	35.1000	APP
8.5	Unsealed Byproduct Material – Uptake, Dilution, Excretion, Imaging, and Localization Studies	Y						
8.5	Unsealed Byproduct Material – Written Directive Required		Y					
8.5	Manual Brachytherapy			Y				
8.5	Sealed Sources for Diagnosis				Y			
8.5	Teletherapy Units					Y		
8.5	Remote Afterloader Units					Y		
8.5	Gamma Stereotactic Radiosurgery Units					Y		
8.5	Other Medical Uses						Y	
8.6	Sealed Sources and Devices	N	N	Y	Y	Y	Y	
8.7	Financial Assurance Determination	Y	Y	Y	Y	Y	Y	
8.8	Purpose(s) for Which Licensed Material Will be Used	Y	Y	Y	Y	Y	Y	
8.9	Training and Experience	G	G	G	G	G	G	
8.10	Radiation Safety Officer	Y	Y	Y	Y	Y	Y	I, D

Table C.1 (continued)

Item #	Topic	35.100/200	35.300	35.400	35.500	35.600	35.1000	APP
8.11	Authorized User(s) (AUs)	Y	Y	Y	Y	Y	Y	D
8.12	Authorized Nuclear Pharmacist (ANP)	Y	Y	N/A	N/A	N/A	Y	D
8.13	Authorized Medical Physicist (AMP)	N/A	N/A	Y*	N/A	Y	Y	D
8.14	Facilities and Equipment	G	G	G	G	G	G	
8.15	Facility Diagram	Y	Y	Y	Y	Y	Y	
8.16	Radiation Monitoring Instruments	Y, P	Y, P	Y, P	Y, P	Y, P	Y, P	K
8.17	Dose Calibrator and Other Equipment	P	P	N/A	N/A	N/A	P	
8.18	Therapy Unit - Calibration and Use	N/A	N/A	N	N/A	Y	N	
8.19	Other Equipment and Facilities	N	N	N	N	Y	N	
8.20	Radiation Protection Program	G	G	G	G	G	G	
8.21	Safety Procedures and Instructions	N/A	N/A	N/A	N/A	Y	N/A	
8.22	Occupational Dose	P	P	P	P	P	P	M
8.23	Area Surveys	P	P	P	P	P	P	R
8.24	Safe Use of Unsealed Licensed Material	P	P	N/A	N/A	N/A	P	T
8.25	Spill Procedures	P	P	P	N/A	N/A	P	N
8.26	Service of Therapy Devices Containing Sealed Sources	N/A	N/A	N/A	N/A	Y	Y	
8.27	Minimization of Contamination	N	N	N	N	N	N	
8.28	Waste Management	P	P	P	P	P	P	W
8.29	Fees	Y	Y	Y	Y	Y	Y	
8.30	Certification	Y	Y	Y	Y	Y	Y	
8.31	Safety Instruction for Individuals in Restricted Areas	N	N	N	N	N	N	J
8.32	Public Dose	N	N	N	N	N	N	
8.33	Opening Packages	N	N	N	N	N	N	

Table C.1 (continued)

Item #	Topic	35.100/200	35.300	35.400	35.500	35.600	35.1000	APP
8.34	Written Directive Procedures	N/A	N	N	N/A	N	N	S
8.35	Release of Patients or Human Research Subjects	N	N	N	N/A	N/A	N	U
8.36	Mobile Medical Service	N	N	N	N	N	N	V
8.37	Audit Program	N	N	N	N	N	N	W
8.38	Operating and Emergency Procedure	N	N	N	N	N	N	N
8.39	Material Receipt and Accountability	N	N	N	N	N	N	
8.40	Ordering and Receiving	N	N	N	N	N	N	O
8.41	Sealed Source Inventory	N	N	N	N	N	N	
8.42	Records of Dosages and Use of Brachytherapy Source	N	N	N	N	N	N	
8.43	Recordkeeping	N	N	N	N	N	N	X
8.44	Reporting	N	N	N	N	N	N	Y
8.45	Leak Tests	N	N	N	N	N	N	Q
8.46	Safety Procedures for Treatments when Patients are Hospitalized	N/A	N	N	N/A	N**	N	
8.47	Transportation	N	N	N	N	N	N	Z

\* Y beside item 8.13 for use under 35.400 applies to Sr-90 only.

\*\* N/A for teletherapy and gamma stereotactic radiosurgery outpatient treatments.

APPENDIX C

Table C.2 outlines the detailed responses that may be made to Items 5 and 6 on Form 313 for type of radioactive material requested and purposes for which it will be used. For example, if the applicant is seeking a license for unsealed byproduct material under 10 CFR 35.100 or 35.200, then the applicant should check the "yes" column next to 10 CFR 35.100 and 35.200 in Table C.2. The table then indicates appropriate responses for that type of use. An applicant may copy the checklist and include it in the license application.

**Table C.2 Items 5 and 6 on NRC Form 313: Radioactive Material And Use**  
*(If using this checklist, check applicable rows and fill in details, and attach copy of checklist to the application.)*

Yes	Radionuclide	Form or Manufacturer/ Model No.	Maximum Quantity	Purpose of Use
	Any byproduct material permitted by 10 CFR 35.100	Any	As needed	Any uptake, dilution, and excretion study permitted by 10 CFR 35.100.
	Any byproduct permitted by 10 CFR 35.200	Any	As needed	Any imaging and localization study permitted by 10 CFR 35.200.
	Any byproduct material permitted by 10 CFR 35.300	Any	___ millicuries	Any radiopharmaceutical therapy procedure permitted by 10 CFR 35.300.
	Iodine-131	Any	___ millicuries	Administration of I-131 sodium iodide.
	Byproduct material permitted by 10 CFR 35.400 (Radionuclide _____)	Sealed source or device (Manufacturer _____, Model No. _____)	___ millicuries	Any brachytherapy procedure permitted by 10 CFR 35.400.
	Byproduct material permitted by 10 CFR 35.400 (Radionuclide _____)	Sealed source or device (Manufacturer _____, Model No. _____)	___ millicuries	Any brachytherapy procedure permitted by 10 CFR 35.400.
	Byproduct material permitted by 10 CFR 35.400 (Radionuclide _____)	Sealed source or device (Manufacturer _____, Model No. _____)	___ millicuries	Any brachytherapy procedure permitted by 10 CFR 35.400.
	Byproduct material permitted by 10 CFR 35.400 (Radionuclide _____)	Sealed source or device (Manufacturer _____, Model No. _____)	___ millicuries	Any brachytherapy procedure permitted by 10 CFR 35.400.

Table C.2 (continued)

Yes	Radionuclide	Form or Manufacturer/ Model No.	Maximum Quantity	Purpose of Use
	Strontium-90	Sealed source or device (Manufacturer _____, Model No. _____)	___ millicuries	Treatment of superficial eye conditions using an applicator distributed pursuant to 10 CFR 32.74 and permitted by 10 CFR 35.400.
	Byproduct material permitted by 10 CFR 35.500 Check all that apply: <input type="checkbox"/> Gd-153; <input type="checkbox"/> I-125; <input type="checkbox"/> Other, describe	Sealed source or device (Manufacturer _____, Model No. _____)	___ curies per source and ___ curies total	Diagnostic medical use of sealed sources permitted by 10 CFR 35.500 in compatible devices registered pursuant to 10 CFR 30.32(g).
	Iridium-192	Sealed source or device (Manufacturer _____, Model No. _____)	___ curies per source and ___ curies total	One source for medical use permitted by 10 CFR 35.600, in a Manufacturer _____ Model No. _____ remote afterloading brachytherapy device. One source in its shipping container as necessary for replacement of the source in the remote afterloader device.
	Cobalt-60	Sealed source or device (Manufacturer _____, Model No. _____)	___ curies per source and ___ curies total	One source for medical use permitted by 10 CFR 35.600, in a Manufacturer _____ Model No. _____ teletherapy unit. One source in its shipping container as necessary for replacement of the source in the teletherapy unit.

Table C.2 (continued)

Yes	Radionuclide	Form or Manufacturer/ Model No.	Maximum Quantity	Purpose of Use
	Cobalt-60	Sealed source or device (Manufacturer _____, Model No. _____)	___ curies per source and ___ curies total	For medical use permitted by 10 CFR 35.600, in a Manufacturer _____ Model No. _____ stereotactic radiosurgery device. Sources in the shipping container as necessary for replacement of the sources in the stereotactic radiosurgery device.
	Any byproduct material under 10 CFR 31.11	Prepackaged kits	___ millicuries	<i>In vitro</i> studies.
	Depleted uranium	Metal	___ kilograms	Shielding in a teletherapy unit.
	Depleted uranium	Metal	___ kilograms	Shielding in a linear accelerator.
	Any radionuclide in excess of 30 millicuries for use in calibration, transmission, and reference sources. (List radionuclide: _____)	Sealed source or device (Manufacturer _____, Model No. _____)	___ millicuries	For use in a Manufacturer _____ Model No. _____ for calibration and checking of licensee's survey instruments.
	Americium-241	Sealed source or device (Manufacturer _____, Model No. _____)	___ millicuries per source and ___ millicuries total	Use as an anatomical marker.
	Plutonium (principal radionuclide Pu-238)	Sealed sources	___ millicuries per source and ___ grams total	As a component of Manufacturer _____ Model No. _____, nuclear-powered cardiac pacemakers for clinical evaluation in accordance with manufacturer's protocol dated _____. This authorization includes: follow-up, explantation, recovery, disposal, and implantation.
	Other	Form or Manufacturer/ Model No.	___ millicuries	Purpose of use

Table C.2 (continued)

[The table content is extremely faint and illegible due to low contrast and scan quality. It appears to be a multi-column table with several rows of data.]

APPENDIX C

Table C.3 is a checklist that may be used to identify the attached documents that the applicant is supplying for items for which a response is required. For example, an applicant may fill in the name(s) of Radiation Safety Officer in Table C.3 and then check the boxes indicating which documents pertaining to the RSO are being included in the license application. An applicant may copy the checklist and include it in the license application.

**Table C.3 Items 7 through 11 on NRC Form 313: Training & Experience, Facilities & Equipment, Radiation Protection Program, and Waste Disposal**

*(Check all applicable rows and fill in details and attach a copy of the checklist to the application or provide information separately.)*

Item Number and Title	Suggested Response	Check box to indicate material included in application
<p>Item 7: Radiation Safety Officer Name: _____</p>	<p>Previous license number (if issued by NRC) or a copy of the license (if issued by an Agreement State) that authorized the uses requested and on which the individual was specifically named as the RSO.</p> <p style="text-align: center;"><b>OR</b></p> <p>Copy of the certification(s) for the board(s) recognized by NRC and as applicable to the types of use for which he or she has RSO responsibilities.</p> <p style="text-align: center;"><b>OR</b></p> <p>Description of the training and experience specified in 10 CFR 35.900(b).</p> <p style="text-align: center;"><b>OR</b></p> <p>Description of the training and experience specified in 10 CFR 35.50(b) demonstrating that the proposed RSO is qualified by training and experience as applicable to the types of use for which he or she has RSO responsibilities.</p> <p style="text-align: center;"><b>AND</b></p> <p>Written certification, signed by a preceptor RSO, that the above training and experience has been satisfactorily completed and that a level of radiation safety knowledge sufficient to function independently as an RSO for a medical use licensee has been achieved.</p> <p style="text-align: center;"><b>AND</b></p> <p>If applicable, description of recent related continuing education and experience as required by 10 CFR 35.59.</p>	<p><input type="checkbox"/></p>

Table C.3 (continued)

Item Number and Title	Suggested Response	Check box to indicate material included in application
<p>Item 7: Authorized Users Names and Requested Uses for Each Individual</p>	<p>Previous license number (if issued by NRC) or a copy of the license (if issued by an Agreement State) on which the physician was specifically named as an AU for the uses requested.</p> <p style="text-align: center;"><b>OR</b></p> <p>Copy of the certification(s) for the board(s) recognized by NRC under 10 CFR Part 35, Subparts D, E, F, G, H, and as applicable to the use requested.</p> <p style="text-align: center;"><b>OR</b></p> <p>Description of the training and experience identified in 10 CFR Part 35 Subpart J demonstrating that the proposed AU is qualified by training and experience for the use requested.</p> <p style="text-align: center;"><b>OR</b></p> <p>A description of the training and experience identified in 10 CFR Part 35 Subparts D, E, F, G, and H demonstrating that the proposed AU is qualified by training and experience for the use requested;</p> <p style="text-align: center;"><b>AND</b></p> <p>Written certification, signed by a preceptor physician AU, that the above training and experience has been satisfactorily completed and that a level of competency sufficient to function independently as an AU for the medical uses authorized has been achieved.</p> <p style="text-align: center;"><b>AND</b></p> <p>If applicable, description of recent related continuing education and experience as required by 10 CFR 35.59.</p>	<p><input type="checkbox"/></p> <p><input type="checkbox"/></p> <p><input type="checkbox"/></p> <p><input type="checkbox"/></p> <p><input type="checkbox"/></p> <p><input type="checkbox"/></p>
<p>Item 7: Authorized Nuclear Pharmacists</p> <p>Names: _____</p>	<p>Previous license number (if issued by NRC) or a copy of the license (if issued by an Agreement State) on which the individual was specifically named ANP.</p> <p style="text-align: center;"><b>OR</b></p> <p>Copy of the certification(s) for the radiopharmacy board(s) recognized by NRC under 10 CFR 35.55(a) or 10 CFR 35.980(a).</p> <p style="text-align: center;"><b>OR</b></p> <p>Description of the training and experience demonstrating that the proposed ANP is qualified by training and experience.</p> <p style="text-align: center;"><b>AND</b></p>	<p><input type="checkbox"/></p> <p><input type="checkbox"/></p> <p><input type="checkbox"/></p>



Table C.3 (continued)

Item Number and Title	Suggested Response	Check box to indicate material included in application
<p>Item 7: Authorized Medical Physicists</p> <p>Names: _____</p>	<p>Previous license number (if issued by NRC) or a copy of the license (if issued by an Agreement State) on which the individual was specifically named as an AMP for the units requested.</p> <p style="text-align: center;">OR</p> <p>Copy of the certification(s) for the board(s) recognized by NRC in 10 CFR 35.51(a) or 10 CFR 35.961(a) or (b).</p> <p style="text-align: center;">OR</p> <p>Description of the training and experience demonstrating that the proposed AMP is qualified by training and experience identified in 10 CFR 35.961(c) for the units requested.</p> <p style="text-align: center;">OR</p> <p>Description of the training and experience demonstrating that the proposed AMP is qualified by training and experience identified in 10 CFR 35.51(b) for the units requested.</p> <p style="text-align: center;">AND</p> <p>Written certification, signed by a preceptor AMP, that the above training and experience has been satisfactorily completed and that a level of competency sufficient to function independently as an AMP has been achieved.</p> <p style="text-align: center;">AND</p> <p>If applicable, description of recent related continuing education and experience as required by 10 CFR 35.59.</p>	<p><input type="checkbox"/></p> <p><input type="checkbox"/></p> <p><input type="checkbox"/></p> <p><input type="checkbox"/></p> <p><input type="checkbox"/></p>
<p>Item 9: Facility Diagram</p>	<p>A diagram is enclosed that describes the facilities and identifies activities conducted in all contiguous areas surrounding the area(s) of use. The following information is included:</p> <ul style="list-style-type: none"> <li>• Drawings should be to scale, and indicate the scale used.</li> <li>• Location, room numbers, and principal use of each room or area where byproduct material is prepared, used or stored, as provided above under the heading "Discussion";</li> <li>• Location, room numbers, and principal use of each adjacent room (e.g., office, file, toilet, closet, hallway), including areas above, beside, and below therapy treatment rooms; indicate whether the room is a restricted or unrestricted area as defined in 10 CFR 20.1003; and</li> <li>• Provide shielding calculations and include information about the type, thickness, and density of any necessary shielding to enable independent verification of shielding calculations, including a description of any portable shields used (e.g., shielding of proposed patient rooms used for implant therapy including the dimensions of any portable shield, if one is used; source storage safe, etc.).</li> </ul>	<p><input type="checkbox"/></p> <p><input type="checkbox"/></p> <p><input type="checkbox"/></p> <p><input type="checkbox"/></p>

Table C.3 (continued)

Item Number and Title	Suggested Response	Check box to indicate material included in application
	In addition to the above, for teletherapy and GSR facilities, applicants should provide the directions of primary beam usage for teletherapy units and, in the case of an isocentric unit, the plane of beam rotation.	<input type="checkbox"/>



Table C.3 (continued)

Item Number and Title	Suggested Response	Check box to indicate material included in application
Item 10: Safety Procedures and Instructions	Attached procedures required by 10 CFR 35.610	<input type="checkbox"/>
Item 10: Occupational Dose	<p>A statement that: "Either we will perform a prospective evaluation demonstrating that unmonitored individuals are not likely to receive, in one year, a radiation dose in excess of 10% of the allowable limits in 10 CFR Part 20 or we will provide dosimetry that meets the requirements listed under "Criteria" in NUREG-1556, Vol. 9, "Consolidated Guidance About Materials Licenses: Program-Specific Guidance About Medical Use Licensees," dated October 2002."</p> <p style="text-align: center;"><b>OR</b></p> <p>A description of an alternative method for demonstrating compliance with the referenced regulations.</p>	<p><input type="checkbox"/></p> <p><input type="checkbox"/></p>
Item 10: Area Surveys	A statement that: "We have developed and will implement and maintain written procedures for area surveys in accordance with 10 CFR 20.1101 that meet the requirements of 10 CFR 20.1501 and 10 CFR 35.70."	<input type="checkbox"/>
Item 10: Safe Use of Unsealed Licensed Material	A statement that: "We have developed and will implement and maintain procedures for safe use of unsealed byproduct material that meet the requirements of 10 CFR 20.1101 and 10 CFR 20.1301."	<input type="checkbox"/>
Item 10: Spill Procedures	A statement that: "We have developed and will implement and maintain written procedures for safe response to spills of licensed material in accordance with 10 CFR 20.1101."	<input type="checkbox"/>
Item 10: Installation, Maintenance, Adjustment, Repair, and Inspection of Therapy Devices Containing Sealed Sources	<p>Name of the proposed employee and types of activities requested:</p> <hr/> <p style="text-align: center;"><b>AND</b></p> <p>Description of the training and experience demonstrating that the proposed employee is qualified by training and experience for the use requested.</p> <p style="text-align: center;"><b>AND</b></p> <p>Copy of the manufacturer's training certification and an outline of the training in procedures to be followed.</p>	<p><input type="checkbox"/></p> <p><input type="checkbox"/></p> <p><input type="checkbox"/></p>
Item 10: Minimization of Contamination	A response is not required under the following condition: the NRC will consider that the above criteria have been met if the information provided in applicant's responses satisfy the criteria in Sections 8.14; 8.15; 8.20; 8.24; 8.26; and 8.28, on the topics: Facility and Equipment; Facility Diagram; Radiation Protection Program; Safety Program; and Waste Management.	N/A

Table C.3 (continued)

<b>Item Number and Title</b>	<b>Suggested Response</b>	<b>Check box to indicate material included in application</b>
Item 11: Waste Management	A statement that: "We have developed and will implement and maintain written waste disposal procedures for licensed material in accordance with 10 CFR 20.1101, that also meet the requirements of the applicable section of Subpart K to 10 CFR Part 20 and 10 CFR 35.92."	<input type="checkbox"/>

## **APPENDIX D**

**Documentation of Training and Experience  
To Identify Individuals on a License as  
Authorized User, Radiation Safety Officer,  
Authorized Nuclear Pharmacist, or  
Authorized Medical Physicist**

## **Documentation of Training and Experience to Identify Individuals on a License as Authorized User, Radiation Safety Officer, Authorized Nuclear Pharmacist, or Authorized Medical Physicist**

### **I. Experienced Authorized Users, Authorized Medical Physicists, Authorized Nuclear Pharmacists, or Radiation Safety Officer**

An applicant or licensee that is adding an experienced authorized user, authorized medical physicist, authorized nuclear pharmacist, or radiation safety officer to its medical use license only needs to provide evidence that the individual is listed on a medical use license issued by the Commission or Agreement State, a permit issued by a Commission master material licensee, a permit issued by a Commission or Agreement State broad scope licensee, or a permit issued by a Commission master material broad scope permittee before October 24, provided that the individual is authorized for the same types of use(s) requested in the application under review, and the individual meets the recentness of training criteria described in 10 CFR 35.59. When adding an experienced authorized nuclear pharmacist to the license, the applicant may also provide evidence that the individual is listed on an NRC or Agreement State commercial nuclear pharmacy license or identified as an authorized nuclear pharmacist by a commercial nuclear pharmacy authorized to identify authorized nuclear pharmacists. For individuals who have been previously authorized by, but not listed on, the commercial nuclear pharmacy license, medical broad scope license, or master materials license medical broad scope permit, the applicant should submit either verification of previous authorizations granted or evidence of acceptable training and experience.

### **II. Applicants that Include Individuals for New Authorized User, Authorized Medical Physicist, Authorized Nuclear Pharmacist or Radiation Safety Officer Recognition by NRC**

Applicants should complete the appropriate sections on NRC Form 313A to show that the individuals meet the appropriate training and experience criteria in 10 CFR Part 35 subparts B, D, E, F, G, H, or J. NRC Form 313A was developed to provide a single location where six different professional groups (physicians, dentist, podiatrist, medical physicist, pharmacist, and radiation safety officer) and ten different medical sub-specialties could document completion of appropriate training and experience requirements. Therefore, some of the sections will not be applicable for each group.

There are two different training and experience routes recognized to qualify an individual as an authorized user, authorized medical physicist, authorized nuclear pharmacist, or radiation safety officer. The first is by means of certification by a professional board recognized by NRC as provided in 10 CFR 35.50(a), 35.51(a), 35.55(a), 35.190(a), 35.290(a), 35.390(a), 35.392(a), 35.394(a), 35.490(a), 35.590(a), or 35.690(a), or until October 25, 2004, a professional board listed in 10 CFR Part 35 Subpart J.

The second route is by meeting the structured educational program, supervised work experience, and preceptor certification requirements in 10 CFR Part 35, subparts B, D, E, F, G, and H. Until

## APPENDIX D

October 25, 2004 this route also includes the classroom and laboratory training and supervised clinical or work experience requirements in 10 CFR Part 35, Subpart J.

### III. Recentness of Training

The required training and experience described in 10 CFR Part 35 must be obtained within the 7 years preceding the date of the application, or the individual must document having had related continuing education, retraining, and experience since obtaining the required training and experience. Examples of acceptable continuing education and experience include the following:

- Successful completion of classroom and laboratory review courses that include radiation safety practices relative to the proposed type of authorized medical use;
- Practical and laboratory experience with patient procedures using radioactive material for the same use(s) for which the applicant is requesting authorization;
- Practical and laboratory experience under the supervision of an AU at the same or another licensed facility that is authorized for the same use(s) for which the applicant is requesting authorization; and
- For therapy devices, experience with the therapy unit and/or comparable linear accelerator experience and completion of an in-service review of operating and emergency procedures relative to the therapy unit to be used by the applicant.

### IV. Instructions and Guidance for Filling Out NRC Form 313A

*Note: Individuals who have been certified by boards recognized by the NRC need only complete items 1, 2 and 3 of NRC Form 313A. Information for all other individuals to be listed on the license as an authorized user, authorized medical physicist, authorized nuclear pharmacist or Radiation Safety Officer must be provided in subsequent sections of NRC Form 313A.*

#### Part I. Training and Experience

Provide information for each individual for whom authorization is sought.

##### **Item 1. Name of individual, proposed authorization, and applicable training requirements.**

Provide the individual's complete name so that NRC can distinguish the training and experience received from that received by others with a similar name, specify the type authorization being requested (Radiation Safety Officer, authorized user, authorized medical physicist, and authorized nuclear pharmacist), and applicable training requirements.

*Note: Do not include personal or private information (e.g., date of birth, social security number) as part of your qualification documentation.*

##### **Item 2. State or territory where licensed**

NRC requires physicians, dentists, podiatrists, and pharmacists to be licensed by a state or territory of the United States, the District of Columbia, or the Commonwealth of Puerto Rico to prescribe drugs in the practice of medicine, practice dentistry, practice podiatry, or practice pharmacy, respectively (see 10 CFR 35.2, "Definition").

### **Item 3. Certification**

The applicant should provide the complete name of the specialty board and the category (or subspecialty) if the board recognizes more than one certification specialty. The month and year certified is used to establish recentness of training, to confirm that NRC recognizes that board's certifications, and to verify that the applicant meets the training requirements.

**If an individual to be listed on the license as an authorized user, authorized medical physicist, authorized nuclear pharmacist, or Radiation Safety Officer cannot meet the Specialty Board certification route, the applicant must fill out the appropriate remaining sections of NRC Form 313A.**

### **Item 4. "Didactic or Classroom and Laboratory Training"; 5a, "Work Experience with Radiation," and 5b. "Supervised Clinical Case Experience."**

Applicants with individuals needing authorized user and authorized nuclear pharmacist recognition should consider the completion of Items 4 and 5 as a unit since the hours required for training and supervised work experience are usually added together to meet the requirement. Because the applicant is not required to receive the training described in Item 4 at one location or at one time, space is provided to identify each location and date of training. The clock hours must be indicated for those individuals that must meet a minimum number of training and work experience hours. The specific number of hours needed for each training element will depend upon that individual's needs. Item 4 is optional for the medical physicist applicant because the regulations do not specify the specific training elements needed to be an authorized medical physicist.

Most applicants will only complete Item 5a because most individuals are required to have supervised work experience but are not required to have specific clinical case experience. Those individuals that have to document clinical case experiences (e.g., the Strontium-90 eye applicator users) may also have specific work experience elements that need to be documented in Item 5b, as part of the clinical case experience.

*Note:* Didactic or Classroom and Laboratory Training may be provided at medical teaching/university institutions. In some cases, a course may be specifically provided for that particular need and taught in consecutive days; in others, the period may be a semester or quarter as part of the formal curriculum. The required "structural educational programs" or "training" may be obtained in any number of settings, locations, and educational situations. There are no specific requirements on who can provide the training or set hours for each of the training elements. If the applicant is seeking authorization under the requirements of 10 CFR Part 35 subparts B, D, E, F, G, and H, the preceptor is responsible for the initial determination of the adequacy of the training (and work experience) to permit the individual to function independently.

APPENDIX D

**Item 6. Formal Training**

This item is completed for individuals qualified to be medical physicists, therapy physicians meeting the requirements in 10 CFR 35.490, 35.690, 35.940 and 35.960, as well as radiation safety officers.

**Item 7. Radiation Safety Officer – One Year Full-Time Work Experience**

This item is used to document that the applicant meets the regulatory requirement of one full year of full time work experience in the areas which are listed in Item 5a.

**Item 8. Medical Physicist – One-Year Full-Time Training and Experience**

This section is used to document that the Medical Physicist has received one full year of full time training and one full year of work experience. Both are required to be under the supervision of an authorized medical physicist but they do not have to be under the same medical physicist.

**Item 9. Supervising Individual**

Item 9 need only be completed by an applicant seeking to have an individual listed on the license as an AU, RSO, or AMP under Part 35 Subparts B, D, E, F, or H. If an applicant is following the training and experience requirements in Subpart J, it is sufficient to identify the supervising individual and licensed facility in Items 5a and 5b. In addition, the use of Item 9 is also dependent on whether information on the identity, qualifications and location (license number) of the supervising individual has already been provided elsewhere on NRC Form 313A (e.g., in Items 7, 8 or 12).

*Note: If the individual had more than one supervisor, all supervising individual names must be listed in Items 5a and 5b and Item 9 filled out for each.*

*Note: The use of Item 9 is dependent on whether the information on the identity, qualifications, and locations (license number) are already provided elsewhere on NRC Form 313A (e.g., in Items 7, 8 or 12).*

*Note: The authorized nuclear pharmacist applicant is required to have supervised practical experience in a nuclear pharmacy but the individual(s) providing the supervision are not specified. Therefore the authorized nuclear pharmacist applicant does not need to identify a supervising individual in Item 5a or complete Item 9.*

**Part II Preceptor Statement**

NRC defines the term "preceptor" in 10 CFR 35.2, "Definitions," to mean "an individual who provides or directs the training and experience required for an individual to become an authorized user, an authorized medical physicist, an authorized nuclear pharmacist, or a radiation Safety Officer." While the supervising individual for the work experience may also be the preceptor, the preceptor does not have to be the supervising individual as long as the preceptor directs the training and experience required. The preceptor has to certify in writing that the

## APPENDIX D

individual applying for recognition as an authorized user, authorized nuclear pharmacist, authorized medical physicist or radiation safety officer (pursuant to 10 CFR Part 35, Subparts B, C, D, E, F, or H and 10 CFR 35.980) has satisfactorily completed the appropriate training and experience criteria and can function independently. This preceptor also has to meet specific requirements.

NRC recognizes supervised work experience, such as that described in 10 CFR 35.290(c), conducted under a preceptor in a licensed material use program. A preceptor is an AU who provides frequent direction, instruction, and direct oversight of the student as the student completes the required work experience in the use of byproduct material. Preceptorships may occur at various licensed facilities, from a large teaching university hospital to a small private practice. However, work experience for sealed source therapy, as described in 10 CFR 35.490(b)(1), 10 CFR 35.491(b)(2), and 10 CFR 35.690(b)(1) must have been gained at a medical institution. When the supervised work experience is complete, the applicant should submit either the preceptor forms, NRC Form 313A as attachments to NRC Form 313, "Application for Material License," or a letter from the preceptor that indicates that the applicant has obtained all required experience elements.

### **Items 10, 11 and 12**

An applicant requesting authorized nuclear pharmacist status for a pharmacist under 10 CFR 35.980(b) is required to provide a different certification statement than a pharmacist under Subpart B, 10 CFR 35.55(b). Item 10 confirms the certification needed to meet the requirements under 10 CFR Part 35, Subpart J and is applicable only for the pharmacist meeting the requirements in 10 CFR 35.980(b). Item 12 is to provide the preceptor authorized nuclear pharmacist's signature.

### **Items 11 and 12**

Item 11 has two components: 11a. certifies that the applicant has satisfactorily completed the training and supervised work experience requirements and 11b. certifies that the applicant has the competency to function independently.

*Note: 10 CFR Part 35 Subpart G does not require a preceptor statement.*

Item 12 requires the preceptor's signature, and certification. The preceptor must indicate whether the preceptor is an Authorized Nuclear Pharmacist certifying the applicant under Subpart B (10 CFR 35.980(b) or Subpart B (10 CFR 35.55(b)), or whether the preceptor has another authorized position.

## **APPENDIX E**

### **Sample License Application**

# APPLICATION FOR MATERIAL LICENSE

Estimated burden per response to comply with this mandatory information collection request, 7 4 hours. Submittal of the application is necessary to determine that the applicant is qualified and that adequate procedures exist to protect the public health and safety. Send comments regarding burden estimate to the Records Management Branch (T-8 E6), U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001, or by internet e-mail to bjs1@nrc.gov, and to the Desk Officer, Office of Information and Regulatory Affairs, NEOB-10202, (3150-0120), Office of Management and Budget, Washington, DC 20503. If a means used to impose an information collection does not display a currently valid OMB control number, NRC may not conduct or sponsor, and a person is not required to respond to, the information collection.

INSTRUCTIONS: SEE THE APPROPRIATE LICENSE APPLICATION GUIDE FOR DETAILED INSTRUCTIONS FOR COMPLETING APPLICATION  
 SEND TWO COPIES OF THE ENTIRE COMPLETED APPLICATION TO THE NRC OFFICE SPECIFIED BELOW.

**APPLICATION FOR DISTRIBUTION OF EXEMPT PRODUCTS FILE APPLICATIONS WITH:**

DIVISION OF INDUSTRIAL AND MEDICAL NUCLEAR SAFETY  
 OFFICE OF NUCLEAR MATERIALS SAFETY AND SAFEGUARDS  
 U.S. NUCLEAR REGULATORY COMMISSION  
 WASHINGTON, DC 20555-001

**ALL OTHER PERSONS FILE APPLICATIONS AS FOLLOWS:**

**IF YOU ARE LOCATED IN:**  
 CONNECTICUT, DELAWARE, DISTRICT OF COLUMBIA, MAINE, MARYLAND,  
 MASSACHUSETTS, NEW HAMPSHIRE, NEW JERSEY, NEW YORK, PENNSYLVANIA,  
 RHODE ISLAND, OR VERMONT, SEND APPLICATIONS TO:

LICENSING ASSISTANT SECTION  
 NUCLEAR MATERIALS SAFETY BRANCH  
 U.S. NUCLEAR REGULATORY COMMISSION, REGION I  
 475 ALLENDALE ROAD  
 KING OF PRUSSIA, PA 19406-1415

**ALABAMA, FLORIDA, GEORGIA, KENTUCKY, MISSISSIPPI, NORTH CAROLINA,  
 PUERTO RICO, SOUTH CAROLINA, TENNESSEE, VIRGINIA, VIRGIN ISLANDS, OR**

**IF YOU ARE LOCATED IN**

**ILLINOIS, INDIANA, IOWA, MICHIGAN, MINNESOTA, MISSOURI, OHIO, OR WISCONSIN,  
 SEND APPLICATIONS TO:**

MATERIALS LICENSING SECTION  
 U.S. NUCLEAR REGULATORY COMMISSION, REGION III  
 801 WARRENVILLE RD  
 LISLE, IL 60532-4351

**ALASKA, ARIZONA, ARKANSAS, CALIFORNIA, COLORADO, HAWAII, IDAHO, KANSAS,  
 LOUISIANA, MONTANA, NEBRASKA, NEVADA, NEW MEXICO, NORTH DAKOTA,  
 OKLAHOMA, OREGON, PACIFIC TRUST TERRITORIES, SOUTH DAKOTA, TEXAS, UTAH,  
 WASHINGTON, OR WYOMING, SEND APPLICATIONS TO:**

NUCLEAR MATERIALS LICENSING SECTION  
 U.S. NUCLEAR REGULATORY COMMISSION, REGION IV  
 611 RYAN PLAZA DRIVE, SUITE 400  
 ARLINGTON, TX 76011-8084

PERSONS LOCATED IN AGREEMENT STATES SEND APPLICATIONS TO THE U.S. NUCLEAR REGULATORY COMMISSION ONLY IF THEY WISH TO POSSESS AND USE LICENSED MATERIAL IN STATES SUBJECT TO U.S. NUCLEAR REGULATORY COMMISSION JURISDICTIONS.

<p>1. THIS IS AN APPLICATION FOR (Check appropriate item)</p> <p><input checked="" type="checkbox"/> A. NEW LICENSE</p> <p><input type="checkbox"/> B. AMENDMENT TO LICENSE NUMBER _____</p> <p><input type="checkbox"/> C. RENEWAL OF LICENSE NUMBER _____</p>	<p>2. NAME AND MAILING ADDRESS OF APPLICANT (Include Zip code)</p> <p><b>Dr. Noe Directive</b>  <b>Suite 112</b>  <b>2 Physician Circle Parkway</b>  <b>Anytown, PA 02201</b></p>
---	---

<p>3. ADDRESS(ES) WHERE LICENSED MATERIAL WILL BE USED OR POSSESSED</p> <p><b>Suite 112</b>  <b>2 Physician Circle Parkway</b>  <b>Anytown, PA 02201</b></p>	<p>4. NAME OF PERSON TO BE CONTACTED ABOUT THIS APPLICATION</p> <p><b>Noe Directive, MD</b></p> <p>TELEPHONE NUMBER</p> <p><b>(123) 456-7890</b></p>
--	--

SUBMIT ITEMS 5 THROUGH 11 ON 8-1/2 X 11" PAPER THE TYPE AND SCOPE OF INFORMATION TO BE PROVIDED IS DESCRIBED IN THE LICENSE APPLICATION GUIDE

<p>5. RADIOACTIVE MATERIAL</p> <p>a. Element and mass number; b. chemical and/or physical form; and c. maximum amount which will be possessed at any one time.</p>	<p>6. PURPOSE(S) FOR WHICH LICENSED MATERIAL WILL BE USED</p>			
<p>7. INDIVIDUAL(S) RESPONSIBLE FOR RADIATION SAFETY PROGRAM AND THEIR TRAINING EXPERIENCE</p>	<p>8. TRAINING FOR INDIVIDUALS WORKING IN OR FREQUENTING RESTRICTED AREAS</p>			
<p>9. FACILITIES AND EQUIPMENT</p>	<p>10. RADIATION SAFETY PROGRAM</p>			
<p>11. WASTE MANAGEMENT.</p>	<p>12. LICENSEE FEES (See 10 CFR 170 and Section 170.31)</p> <table border="1"> <tr> <td>FEE CATEGORY</td> <td>7C</td> <td>AMOUNT ENCLOSED \$D, DDD CC</td> </tr> </table>	FEE CATEGORY	7C	AMOUNT ENCLOSED \$D, DDD CC
FEE CATEGORY	7C	AMOUNT ENCLOSED \$D, DDD CC		

13. CERTIFICATION (Must be completed by applicant) THE APPLICANT UNDERSTANDS THAT ALL STATEMENTS AND REPRESENTATIONS MADE IN THIS APPLICATION ARE BINDING UPON THE APPLICANT.

THE APPLICANT AND ANY OFFICIAL EXECUTING THIS CERTIFICATION ON BEHALF OF THE APPLICANT, NAMED IN ITEM 2, CERTIFY THAT THIS APPLICATION IS PREPARED IN CONFORMITY WITH TITLE 10, CODE OF FEDERAL REGULATIONS, PARTS 30, 32, 33, 34, 35, 36, 39, AND 40, AND THAT ALL INFORMATION CONTAINED HEREIN IS TRUE AND CORRECT TO THE BEST OF THEIR KNOWLEDGE AND BELIEF.

WARNING 18 U.S.C. SECTION 1001 ACT OF JUNE 25, 1948 62 STAT. 749 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO

<p>CERTIFYING OFFICER - TYPED/PRINTED NAME AND TITLE</p> <p><b>Noe Directive, MD - President</b></p>	<p>SIGNATURE</p> <p><i>Noe Directive</i></p>	<p>DATE</p> <p><b>April 11, 2003</b></p>
--	--	--

FOR NRC USE ONLY					
TYPE OF FEE	FEE LOG	FEE CATEGORY	AMOUNT RECEIVED \$	CHECK NUMBER	COMMENTS
APPROVED BY				DATE	

**TRAINING AND EXPERIENCE AND PRECEPTOR STATEMENT**

**PART I -- TRAINING AND EXPERIENCE**

**Note:** Descriptions of training and experience must contain sufficient detail to match the training and experience criteria in the applicable regulations.

1. Name of Individual, Proposed Authorization (e.g., Radiation Safety Officer), and Applicable Training Requirements (e.g., 10 CFR 35.50)

**Noe Directive, MD Authorized user, 10 CFR 35.190, 10 CFR 35.290**

2. For Physicians, Podiatrists, Dentists, Pharmacists -- State or Territory Where Licensed

**Pennsylvania Medical License PA-MD-XXYYY**

**3. CERTIFICATION**

Specialty Board	Category	Month and Year Certified
N/A		

*Stop here when using Board Certification to meet 10 CFR Part 35 training and experience requirements.*

**4. DIDACTIC OR CLASSROOM AND LABORATORY TRAINING (optional for Medical Physicists)**

Description of Training	Location	Clock Hours	Dates of Training
Radiation Physics and Instrumentation	Radiation 200 for Diagnostic Physicians Sample Medical School 4321 Broad Street	50	July 1 to August 15, 2002
Radiation Protection	Anytown, PA	50	"
Mathematics Pertaining to the Use and Measurement of Radioactivity	"	50	"
Radiation Biology	"	50	"
Chemistry of Byproduct Material for Medical Use	"	50	"
OTHER			

TRAINING AND EXPERIENCE AND PRECEPTOR STATEMENT (continued)

5a. WORK EXPERIENCE WITH RADIATION

Description of Experience	Name of Supervising Individual(s)	Location and Corresponding Materials License Number	Dates and Clock Hours of Experience
Ordering, receiving, and unpacking radioactive material safely and performing the related radiation surveys.	Thomas Group, D.O.	See Item 9	August 2002 to March 2003 100
Calibrating instruments used to determine the activity of dosages and performing checks for proper operation of survey meters.	"	"	100
Calculating, measuring, and safely preparing patient or human research subject dosages	"	"	100
Using administrative controls to prevent a medical event involving the use of unsealed byproduct material	"	"	50
Using procedures to safely contain spilled radioactive material and using proper decontamination procedures	"	"	50
Administering dosages of radioactive drugs to patients or human research subjects	"	"	100
Eluting generators, measuring and testing the eluate, and processing the eluate with reagent kits to prepare labeled radioactive drugs.	Jane Diagnostic, MD	See Item 12	50
			Total > 700 hours

5b. SUPERVISED CLINICAL CASE EXPERIENCE - N/A

Radionuclide	Type of Use	No. of Cases Involving Personal Participation	Name of Supervising Individual	Location and Corresponding Materials License Number	Dates and Clock Hours of Experience
N/A					

NRC FORM 313A  
(10-2002)

U.S. NUCLEAR REGULATORY COMMISSION

## TRAINING AND EXPERIENCE AND PRECEPTOR STATEMENT (continued)

## 6. FORMAL TRAINING (applies to Medical Physicists and Therapy Physicists)

Degree, Area of Study or Residency Program	Name of Program and Location with Corresponding Materials License Numbers	Dates	Name of Organization that Approved the Program (e.g., Accreditation Council for Graduate Medical Education) and the Applicable Regulation (e.g., 10 CFR 35.490)
N/A			

## 7. RADIATION SAFETY OFFICER -- ONE-YEAR FULL-TIME TRAINING

- YES Completed 1-year of full-time radiation safety experience (in areas identified in item 5a) under supervision of
- N/A \_\_\_\_\_ the RSO for License No. \_\_\_\_\_

## 8. MEDICAL PHYSICIST -- ONE YEAR FULL-TIME TRAINING/WORK EXPERIENCE

- YES Completed 1-year of full-time training in therapeutic radiological physics under the supervision of
- N/A \_\_\_\_\_ who meets requirements for Authorized Medical Physicists; and
- YES Completed 1-year of full-time work experience (for areas identified in item 5a) for \_\_\_\_\_
- N/A modality(ies) under the supervision of \_\_\_\_\_ who meets requirements for Authorized Medical Physicists for \_\_\_\_\_ modality(ies).

## 9. SUPERVISING INDIVIDUAL -- IDENTIFICATION AND QUALIFICATIONS

The training and experience indicated above was obtained under the supervision of (if more than one supervising individual is needed to meet requirements in 10 CFR 35, provide the following information for each):

A. Name of Supervisor

 Supervisor is:**Thomas Group, D.O.**

- Authorized User
- Radiation Safety Officer

- Authorized Medical Physicists
- Authorized Nuclear Pharmacists

C. Supervisor meets requirements of Part 35, Section(s) **35.290**for medical uses in Part 35, Section(s) **35.100 and 35.200**

D. Address

**Sample Medical-Institution Limited**  
**1234 Main Street**  
**Anytown, PA 02120**

E. Materials License Number

**99-02120-01**

TRAINING AND EXPERIENCE AND PRECEPTOR STATEMENT (continued)

PART II - PRECEPTOR STATEMENT

Note: This part must be completed by the individual's preceptor. If more than one preceptor is necessary to document experience, obtain a separate preceptor statement from each. This part is not required to meet the training requirements in 10 CFR 35.590.

Item 10 must be completed for Nuclear Pharmacists meeting the requirements of 10 CFR Part 35, Subpart J. Preceptors do not have to complete items 11a, 11b, or the certifying statements for other individuals meeting the requirements of 10 CFR Part 35, Subpart J.

YES 10. The individual named in item 1 has satisfactorily completed the training requirements in  
 N/A 10 CFR 35.980 and is competent to independently operate a nuclear pharmacy.

YES 11a. The individual named in Item 1 has satisfactorily completed the requirements in Part 35, Section(s) and  
 N/A Paragraph(s) 35.190(c) and 35.290(c)

YES 11b. The individual named in Item 1. is competent to independently function as an authorized  
 N/A User for 10 CFR 35.100 and 200 uses.

12. PRECEPTOR APPROVAL AND CERTIFICATION

I certify the approval of item 10 and certify I am an Authorized Nuclear Pharmacist;

or

I certify the approval of items 11a and 11b, and certify I am an Authorized Nuclear Pharmacist;

or

I certify the approval of items 11a and 11b, and I certify that I meet the requirements of 35.290  
or equivalent Agreement State requirements to be a preceptor authorized User  
for the following uses of byproduct material: 35.100 and 35.200

A. Address

**Sample Medical Institution Limited**  
**1234 Main Street**  
**Anytown, PA 02120**

B. Materials License Number

99-02120-01

C. NAME OF PRECEPTOR (print clearly)

**Jane Diagnostic, MD**

D. SIGNATURE -- PRECEPTOR

*Jane Diagnostic*

E. DATE

**4-11-03**

Noe Directive, M.D. Attachment 1 of 6

Table C.2 outlines the detailed responses that may be made to Items 5 and 6 on Form 313 for type of radioactive material requested and purposes for which it will be used. For example, if the applicant is seeking a license for unsealed byproduct material under 10 CFR 35.100 or 35.200, then the applicant should check the "yes" column next to 10 CFR 35.100 and 35.200 in Table C.2. The table then indicates appropriate responses for that type of use. An applicant may copy the checklist and include it in the license application.

**Table C.2 Items 5 and 6 on NRC Form 313: Radioactive Material And Use**

*(If using this checklist, check applicable rows and fill in details, and attach copy of checklist to the application.)*

Yes	Radionuclide	Form or Manufacturer/ Model No.	Maximum Quantity	Purpose of Use
<input checked="" type="checkbox"/>	Any byproduct material permitted by 10 CFR 35.100	Any	As needed	Any uptake, dilution, and excretion study permitted by 10 CFR 35.100.
<input checked="" type="checkbox"/>	Any byproduct permitted by 10 CFR 35.200	Any	As needed	Any imaging and localization study permitted by 10 CFR 35.200.
	Any byproduct material permitted by 10 CFR 35.300	Any	___ millicuries	Any radiopharmaceutical therapy procedure permitted by 10 CFR 35.300.
	Iodine-131	Any	___ millicuries	Administration of I-131 sodium iodide.
	Byproduct material permitted by 10 CFR 35.400 (Radionuclide _____)	Sealed source or device (Manufacturer _____, Model No. _____)	___ millicuries	Any brachytherapy procedure permitted by 10 CFR 35.400.
	Byproduct material permitted by 10 CFR 35.400 (Radionuclide _____)	Sealed source or device (Manufacturer _____, Model No. _____)	___ millicuries	Any brachytherapy procedure permitted by 10 CFR 35.400.
	Byproduct material permitted by 10 CFR 35.400 (Radionuclide _____)	Sealed source or device (Manufacturer _____, Model No. _____)	___ millicuries	Any brachytherapy procedure permitted by 10 CFR 35.400.
	Byproduct material permitted by 10 CFR 35.400 (Radionuclide _____)	Sealed source or device (Manufacturer _____, Model No. _____)	___ millicuries	Any brachytherapy procedure permitted by 10 CFR 35.400.

Noe Directive, M.D. Attachment 2 of 6

Yes	Radionuclide	Form or Manufacturer/ Model No.	Maximum Quantity	Purpose of Use
	Strontium-90	Sealed source or device (Manufacturer _____, Model No. _____)	_____ millicuries	Treatment of superficial eye conditions using an applicator distributed pursuant to 10 CFR 32.74 and permitted by 10 CFR 35.400.

Table C.3 is a checklist that may be used to identify the attached documents that the applicant is supplying for items for which a response is required. For example, an applicant may fill in the name(s) of Radiation Safety Officer in Table C.3 and then check the boxes indicating which documents pertaining to the RSO are being included in the license application. An applicant may copy the checklist and include it in the license application.

**Table C.3 Items 7 through 11 on NRC Form 313: Training & Experience, Facilities & Equipment, Radiation Protection Program, and Waste Disposal**

*(Check all applicable rows and fill in details and attach a copy of the checklist to the application or provide information separately.)*

Item Number and Title	Suggested Response	Check box to indicate material included in application
Item 7: Radiation Safety Officer Name: <u>Noe Directive, M.D.</u>	Previous license number (if issued by NRC) or a copy of the license (if issued by an Agreement State) that authorized the uses requested and on which the individual was specifically named as the RSO.	<input type="checkbox"/>
	OR	
	Copy of the certification(s) for the board(s) recognized by NRC and as applicable to the types of use for which he or she has RSO responsibilities.	<input type="checkbox"/>
	OR	
	Description of the training and experience specified in 10 CFR 35.900(b).	<input type="checkbox"/>
	OR	
	Description of the training and experience specified in 10 CFR 35.50(b) demonstrating that the proposed RSO is qualified by training and experience as applicable to the types of use for which he or she has RSO responsibilities.	<input checked="" type="checkbox"/>
	AND	



Noe Directive, M.D. Attachment 4 of 6  
Table C.3 (continued)

Item Number and Title	Suggested Response	Check box to indicate material included in application
<p>Item 7: Authorized Users Names and Requested Uses for Each Individual Noe Directive, M.D., 35.100; 35.200</p>	<p>Previous license number (if issued by NRC) or a copy of the license (if issued by an Agreement State) on which the physician was specifically named as an AU for the uses requested.</p> <p style="text-align: center;">OR</p> <p>Copy of the certification(s) for the board(s) recognized by NRC under 10 CFR Part 35, Subparts D, E, F, G, H, and as applicable to the use requested.</p> <p style="text-align: center;">OR</p> <p>Description of the training and experience identified in 10 CFR Part 35 Subpart J demonstrating that the proposed AU is qualified by training and experience for the use requested.</p> <p style="text-align: center;">OR</p> <p>A description of the training and experience identified in 10 CFR Part 35 Subparts D, E, F, G, and H demonstrating that the proposed AU is qualified by training and experience for the use requested;</p> <p style="text-align: center;">AND</p> <p>Written certification, signed by a preceptor physician AU, that the above training and experience has been satisfactorily completed and that a level of competency sufficient to function independently as an AU for the medical uses authorized has been achieved.</p> <p style="text-align: center;">AND</p> <p>If applicable, description of recent related continuing education and experience as required by 10 CFR 35.59.</p>	<p><input type="checkbox"/></p> <p><input type="checkbox"/></p> <p><input type="checkbox"/></p> <p><input checked="" type="checkbox"/></p> <p><input checked="" type="checkbox"/></p> <p><input type="checkbox"/></p>
<p>Item 7: Authorized Nuclear Pharmacists Names: N/A</p>	<p>Previous license number (if issued by NRC) or a copy of the license (if issued by an Agreement State) on which the individual was specifically named ANP.</p> <p style="text-align: center;">OR</p> <p>Copy of the certification(s) for the radiopharmacy board(s) recognized by NRC under 10 CFR 35.55(a) or 10 CFR 35.980(a).</p> <p style="text-align: center;">OR</p> <p>Description of the training and experience demonstrating that the proposed ANP is qualified by training and experience.</p> <p style="text-align: center;">AND</p> <p>Written certification, signed by a preceptor ANP, that the above training and experience has been satisfactorily completed and that a level of competency</p> <ul style="list-style-type: none"> <li>• sufficient to function independently as an ANP has been achieved (10 CFR 35.55), or</li> <li>• sufficient to independently operate a nuclear pharmacy (10 CFR 35.980).</li> </ul> <p style="text-align: center;">AND</p>	<p><input type="checkbox"/></p> <p><input type="checkbox"/></p> <p><input type="checkbox"/></p> <p><input type="checkbox"/></p>

Noe Directive, M.D. Attachment 3 of 6  
**Table C.3 (continued)**

Item Number and Title	Suggested Response	Check box to indicate material included in application
	If applicable, description of recent related continuing education and experience as required by 10 CFR 35.59.	<input type="checkbox"/>

Noe Directive, M.D. Attachment 4 of 6  
Table C.3 (continued)

Item Number and Title	Suggested Response	Check box to indicate material included in application
<p>Item 7: Authorized Medical Physicists</p> <p>Names: <u>N/A</u></p>	<p>Previous license number (if issued by NRC) or a copy of the license (if issued by an Agreement State) on which the individual was specifically named as an AMP for the units requested.</p> <p style="text-align: center;"><b>OR</b></p> <p>Copy of the certification(s) for the board(s) recognized by NRC in 10 CFR 35.51(a) or 10 CFR 35.961(a) or (b).</p> <p style="text-align: center;"><b>OR</b></p> <p>Description of the training and experience demonstrating that the proposed AMP is qualified by training and experience identified in 10 CFR 35.961(c) for the units requested.</p> <p style="text-align: center;"><b>OR</b></p> <p>Description of the training and experience demonstrating that the proposed AMP is qualified by training and experience identified in 10 CFR 35.51(b) for the units requested.</p> <p style="text-align: center;"><b>AND</b></p> <p>Written certification, signed by a preceptor AMP, that the above training and experience has been satisfactorily completed and that a level of competency sufficient to function independently as an AMP has been achieved.</p> <p style="text-align: center;"><b>AND</b></p> <p>If applicable, description of recent related continuing education and experience as required by 10 CFR 35.59.</p>	<p><input type="checkbox"/></p> <p><input type="checkbox"/></p> <p><input type="checkbox"/></p> <p><input type="checkbox"/></p> <p><input type="checkbox"/></p> <p><input type="checkbox"/></p>
<p>Item 9: Facility Diagram</p>	<p>A diagram is enclosed that describes the facilities and identifies activities conducted in all contiguous areas surrounding the area(s) of use. The following information is included:</p> <ul style="list-style-type: none"> <li>• Drawings should be to scale, and indicate the scale used.</li> <li>• Location, room numbers, and principal use of each room or area where byproduct material is prepared, used or stored, as provided above under the heading "Discussion";</li> <li>• Location, room numbers, and principal use of each adjacent room (e.g., office, file, toilet, closet, hallway), including areas above, beside, and below therapy treatment rooms; indicate whether the room is a restricted or unrestricted area as defined in 10 CFR 20.1003; and</li> <li>• Provide shielding calculations and include information about the type, thickness, and density of any necessary shielding to enable independent verification of shielding calculations, including a description of any portable shields used (e.g., shielding of proposed patient rooms used for implant therapy including the dimensions of any portable shield, if one is used; source storage safe, etc.).</li> </ul>	<p><input checked="" type="checkbox"/></p>

Noe Directive, M.D. Attachment 3 of 6  
Table C.3 (continued)

Item Number and Title	Suggested Response	Check box to indicate material included in application
	In addition to the above, for teletherapy and GSR facilities, applicants should provide the directions of primary beam usage for teletherapy units and, in the case of an isocentric unit, the plane of beam rotation.	<input type="checkbox"/>
<p>Item 9: Radiation Monitoring Instruments</p> <p>Description of the instrumentation that will be used to perform required surveys: <u>Geiger Mueller-Range 1-1000 mR/hr and NaI Well Counter - sensitivity 22 dpm/cm<sup>2</sup></u></p>	<p>A statement that: "Radiation monitoring instruments will be calibrated by a person qualified to perform survey meter calibrations."</p> <p>AND/OR</p> <p>A statement that: "We have developed and will implement and maintain written survey meter calibration procedures in accordance with the requirements in 10 CFR 20.1501 and that meet the requirements of 10 CFR 35.61."</p> <p>AND</p> <p>A description of the instrumentation (e.g., gamma counter, solid state detector, portable or stationary count rate meter, portable or stationary dose rate or exposure rate meter, single or multichannel analyzer, liquid scintillation counter, proportional counter) that will be used to perform required surveys.</p> <p>AND</p> <p>A statement that: "We reserve the right to upgrade our survey instruments as necessary as long as they are adequate to measure the type and level of radiation for which they are used."</p>	<p><input checked="" type="checkbox"/></p> <p><input type="checkbox"/></p> <p><input checked="" type="checkbox"/></p> <p><input checked="" type="checkbox"/></p>
Item 9: Dose Calibrator and Other Dosage Measuring Equipment N/A	A statement that: "Equipment used to measure dosages will be calibrated in accordance with nationally recognized standards or the manufacturer's instructions."	<input type="checkbox"/>
Item 9: Therapy Unit - Calibration and Use N/A	We are providing the procedures required by 10 CFR 35.642, 10 CFR 35.643, and 10 CFR 35.645, if applicable to the license application.	<input type="checkbox"/>
Item 9: Other Equipment and Facilities N/A	<p>Attached is a description identified as Attachment 9.4, of additional facilities and equipment.</p> <p>For manual brachytherapy facilities, we are providing a description of the emergency response equipment.</p> <p>For teletherapy, GSR, and remote afterloader facilities, we are providing a description of the following:</p> <ul style="list-style-type: none"> <li>• Warning systems and restricted area controls (e.g., locks, signs, warning lights and alarms, interlock systems) for each therapy treatment room;</li> <li>• Area radiation monitoring equipment;</li> <li>• Viewing and intercom systems (except for LDR units);</li> </ul>	<p><input type="checkbox"/></p> <p><input type="checkbox"/></p> <p><input type="checkbox"/></p> <p><input type="checkbox"/></p> <p><input type="checkbox"/></p>

Noe Directive, M.D. Attachment 4 of 6  
 Table C.3 (continued)

Item Number and Title	Suggested Response	Check box to indicate material included in application
	<ul style="list-style-type: none"> <li>• Steps that will be taken to ensure that no two units can be operated simultaneously, if other radiation-producing equipment (e.g., linear accelerator, X-ray machine) are in the treatment room;</li> <li>• Methods to ensure that whenever the device is not in use or is unattended, the console keys will be inaccessible to unauthorized persons; and</li> <li>• Emergency response equipment.</li> </ul>	<input type="checkbox"/>   <input type="checkbox"/>   <input type="checkbox"/>
Item 10. Safety Procedures and Instructions N/A	Attached procedures required by 10 CFR 35.610	<input type="checkbox"/>


Noe Directive, M.D. Attachment 3 of 6  
Table C.3 (continued)

Item Number and Title	Suggested Response	Check box to indicate material included in application
Item 10: Occupational Dose	<p>A statement that: "Either we will perform a prospective evaluation demonstrating that unmonitored individuals are not likely to receive, in one year, a radiation dose in excess of 10% of the allowable limits in 10 CFR Part 20 or we will provide dosimetry that meets the requirements listed under "Criteria" in NUREG-1556, Vol. 9, "Consolidated Guidance About Materials Licenses: Program-Specific Guidance About Medical Use Licensees," dated October 2002."</p> <p style="text-align: center;"><b>OR</b></p> <p>A description of an alternative method for demonstrating compliance with the referenced regulations.</p>	<p><input checked="" type="checkbox"/></p> <p><input type="checkbox"/></p>
Item 10: Area Surveys	A statement that: "We have developed and will implement and maintain written procedures for area surveys in accordance with 10 CFR 20.1101 that meet the requirements of 10 CFR 20.1501 and 10 CFR 35.70."	<input checked="" type="checkbox"/>
Item 10: Safe Use of Unsealed Licensed Material	A statement that: "We have developed and will implement and maintain procedures for safe use of unsealed byproduct material that meet the requirements of 10 CFR 20.1101 and 10 CFR 20.1301."	<input checked="" type="checkbox"/>
Item 10: Spill Procedures	A statement that: "We have developed and will implement and maintain written procedures for safe response to spills of licensed material in accordance with 10 CFR 20.1101."	<input checked="" type="checkbox"/>
<p>Item 10: Installation, Maintenance, Adjustment, Repair, and Inspection of Therapy Devices Containing Sealed Sources</p> <p>N/A</p>	<p>Name of the proposed employee and types of activities requested:</p> <hr/> <p style="text-align: center;"><b>AND</b></p> <p>Description of the training and experience demonstrating that the proposed employee is qualified by training and experience for the use requested.</p> <p style="text-align: center;"><b>AND</b></p> <p>Copy of the manufacturer's training certification and an outline of the training in procedures to be followed.</p>	<p><input type="checkbox"/></p> <p><input type="checkbox"/></p> <p><input type="checkbox"/></p>
<p>Item 10: Minimization of Contamination</p> <p>N/A</p>	A response is not required under the following condition: the NRC will consider that the above criteria have been met if the information provided in applicant's responses satisfy the criteria in Sections 8.14, 8.15, 8.20, 8.24, 8.26, and 8.28, on the topics: Facility and Equipment; Facility Diagram; Radiation Protection Program; Safety Program; and Waste Management.	N/A

**Noe Directive, M.D. Attachment 4 of 6  
Table C.3 (continued)**

<b>Item Number and Title</b>	<b>Suggested Response</b>	<b>Check box to indicate material included in application</b>
Item 11: Waste Management	A statement that: "We have developed and will implement and maintain written waste disposal procedures for licensed material in accordance with 10 CFR 20.1101, that also meet the requirements of the applicable section of Subpart K to 10 CFR Part 20 and 10 CFR 35.92."	<input checked="" type="checkbox"/>

Dr. Noe Directive



*Notes:*

- 1) *Radioactive material delivered to hot lab.*
- 2) *Counter surfaces are stainless steel and floors are seamless vinyl to facilitate cleanup and minimize permanent contamination.*
- 3) *Unoccupied basement located underneath facility and Suite 212 (a doctor's office) located above facility.*

**APPENDIX F**

**Sample Licenses**

## Sample Licenses

The license conditions listed in the example licenses come from the standard conditions in NUREG 1556 Volume 20, "Guidance About Administrative Licensing Procedures," with some modifications to reflect provisions of 10 CFR Part 35. The modified conditions are as follows:

- Standard tie-down condition (standard condition 38) modified to reflect 10 CFR 35.26
- Decay-in-storage condition (standard condition 140) modified to reflect 10 CFR 35.92
- Sealed sources leak test condition (standard condition 165) modified to reflect 10 CFR 35.67

When preparing licenses, please refer to the latest revision of NUREG 1556 Volume 20 for the most current versions of the license conditions.

**MATERIALS LICENSE**

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 36, 39, 40, and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations, and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

Licensee	
1. Norma L. Vision, M.D.	3. License number 52-02210-01
2. Suite 201 1234 Bright Sun Drive Sun City, Puerto Rico 02210	4. Expiration date March 31, 2015
	5. Docket No. 030-02210 Reference No.

6. Byproduct, source, and/or special nuclear material	7. Chemical and/or physical form	8. Maximum amount that licensee may possess at any one time under this license
A. Strontium 90 permitted by 10 CFR 35.400	A. Sealed Source (DuPont Merck Pharmaceutical Co. Model NB-1)	A. 120 millicuries

9. Authorized use:  
A. Strontium-90 for ophthalmic radiotherapy permitted by 10 CFR 35.400

★ **CONDITIONS** ★

- 10. Licensed material may be used or stored only at the licensee's facilities located at Suite 201, 1234 Bright Sun Drive, Sun City, Puerto Rico.
- 11. The Radiation Safety Officer for this license is Cecil Source, Ph.D.
- 12. Licensed material is only authorized for use by, or under the supervision of:
  - A. Individuals permitted to work as an authorized user and/or authorized medical physicist in accordance with 10 CFR 35.13 and 35.14.
  - B. Authorized user and use: Norma L. Vision, M.D. - Strontium 90 for ophthalmic radiotherapy.
  - C. Authorized medical physicist: Cecil Source, Ph.D.

**MATERIALS LICENSE  
SUPPLEMENTARY SHEET**


License Number  
52-02210-01

Docket or Reference Number  
030-02210

13. The licensee is authorized to transport licensed material in accordance with the provisions of 10 CFR Part 71, "Packaging and Transportation of Radioactive Material."

14. Except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures, listed below. This license condition applies only to those procedures that are required to be submitted in accordance with the regulations. Additionally, this license condition does not limit the licensee's ability to make changes to the radiation protection program as provided for in 10 CFR 35.26. The U.S. Nuclear Regulatory Commission's regulations shall govern unless the statements, representations, and procedures in the licensee's application and correspondence are more restrictive than the regulations.

A. Application dated March 15, 2005



For the U.S. Nuclear Regulatory Commission

Date \_\_\_\_\_

By \_\_\_\_\_

Division of Nuclear Materials Safety

**MATERIALS LICENSE**

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 36, 39, 40, and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations, and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

Licensee

1. Sample Medical Broad Scope

3. License number 99-02110-01

2. 300 Main Street

4. Expiration date March 31, 2013

Anytown, Pennsylvania 02110

5. Docket No. 030-02110

Reference No.

6. Byproduct, source, and/or special nuclear material

7. Chemical and/or physical form

8. Maximum amount that licensee may possess at any one time under this license

A. Any byproduct material with atomic numbers 1 through 83

A. Any

A. 200 millicuries per radionuclide and 15 curies total

B. Any byproduct material with atomic numbers 3 through 83

B. Sealed Sources

B. 15 curies per radionuclide and 15 curies total

C. Hydrogen 3

C. Any

C. 2 curies

D. Carbon 14

D. Any

D. 1 curie

E. Phosphorus 32

E. Any

E. 2 curies

F. Sulfur 35

F. Any

F. 2 curies

G. Chromium 51

G. Any

G. 500 millicuries

H. Molybdenum 99

H. Any

H. 10 curies

I. Technetium 99m

I. Any

I. 10 curies

J. Iridium 192

J. Sealed Sources (US Atomic Model IR-192HDR)

J. 12 curies per source and 24 curies total

9. Authorized use:

A. through I. Medical diagnosis, therapy and research in humans in accordance with any applicable U.S. Food and Drug Administration (FDA) requirements. Research and development as

**MATERIALS LICENSE  
SUPPLEMENTARY SHEET**

License Number  
99-02110-01

Docket or Reference Number  
030-02110

defined in 10 CFR 30.4, including animal studies; instrument calibration; student instruction; and in-vitro studies.

- J. One source for medical use described in 10 CFR 35.600, in a US Atomic Model IR-192THER remote afterloader unit. The source activity may not exceed 10 curies at the time of installation. One source in its shipping container as necessary for replacement of the source in the remote afterloader unit.

CONDITIONS

10. Licensed material may be used or stored only at the licensee's facilities located at 300 Main Street, Anytown, Pennsylvania.
11. A. The Radiation Safety Officer for this license is Patty Melt, Ph.D.
- B. The use of licensed material in or on humans shall be by an authorized user as defined in 10 CFR 35.2.
- C. Individuals designated to work as authorized users, authorized nuclear pharmacists, or authorized medical physicists as defined in 10 CFR 35.2, shall meet the training, experience, and recency of training criteria established in 10 CFR Part 35, and shall be designated, in writing, by the licensee's Radiation Safety Committee.
- D. Licensed material for other than human use shall be used by, or under the supervision of, individuals designated by the Radiation Safety Committee.
12. The licensee shall not use licensed material in field applications where it is released except as provided otherwise by specific condition of this license.
13. Experimental animals, or the products from experimental animals, that have been administered licensed materials shall not be used for human consumption.
14. For sealed sources not associated with 10 CFR Part 35 use, the following conditions apply:
- A. Sealed sources shall be tested for leakage and/or contamination at intervals not to exceed the intervals specified in the certificate of registration issued by the U.S. Nuclear Regulatory Commission under 10 CFR 32.210 or under equivalent regulations of an Agreement State.
- B. Notwithstanding Paragraph A of this Condition, sealed sources designed to primarily emit alpha particles shall be tested for leakage and/or contamination at intervals not to exceed 3 months.

**MATERIALS LICENSE  
SUPPLEMENTARY SHEET**License Number  
99-02110-01Docket or Reference Number  
030-02110

- C. Each sealed source fabricated by the licensee shall be inspected and tested for construction defects, leakage, and contamination prior to any use or transfer as a sealed source.
- D. In the absence of a certificate from a transferor indicating that a leak test has been made within the intervals specified in the certificate of registration issued by the U.S. Nuclear Regulatory Commission under 10 CFR 32.210 or under equivalent regulations of an Agreement State, prior to the transfer, a sealed source received from another person shall not be put into use until tested and the test results received.
- E. Sealed sources need not be tested if they contain only hydrogen-3; or they contain only a radioactive gas; or the half-life of the isotope is 30 days or less; or they contain not more than 100 microcuries of beta- and/or gamma-emitting material or not more than 10 microcuries of alpha-emitting material.
- F. Sealed sources need not be tested if they are in storage and are not being used; however, when they are removed from storage for use or transferred to another person and have not been tested within the required leak test interval, they shall be tested before use or transfer. No sealed source shall be stored for a period of more than 10 years without being tested for leakage and/or contamination.
- G. The leak test shall be capable of detecting the presence of 0.005 microcurie (185 becquerels) of radioactive material on the test sample. If the test reveals the presence of 0.005 microcurie (185 becquerels) or more of removable contamination, a report shall be filed with the U.S. Nuclear Regulatory Commission in accordance with 10 CFR 30.50(c)(2), and the source shall be removed immediately from service and decontaminated, repaired, or disposed of in accordance with Commission regulations.
- H. Tests for leakage and/or contamination, including leak test sample collection and analysis, shall be performed by the licensee or by other persons specifically licensed by the U.S. Nuclear Regulatory Commission or an Agreement State to perform such services.
15. Sealed sources or detector cells containing licensed material shall not be opened or sources removed from source holders by the licensee.
16. The licensee shall conduct a physical inventory every six months, or at other intervals approved by the U.S. Nuclear Regulatory Commission, to account for all sources and/or devices received and possessed under the license.
17. A. Detector cells containing a titanium tritide foil or a scandium tritide foil shall only be used in conjunction with a properly operating temperature control mechanism that prevents the foil temperature from exceeding that specified in the certificate of registration issued by NRC pursuant to 10 CFR 32.210 or the equivalent regulations from an Agreement State.

**MATERIALS LICENSE  
SUPPLEMENTARY SHEET**

License Number  
99-02110-01

Docket or Reference Number  
030-02110

- B. When in use, detector cells containing a titanium tritide foil or a scandium tritide foil shall be vented to the outside.
18. For radioactive material held for decay in storage other than that held in accordance with 10 CFR 35.92, the licensee is authorized to hold radioactive material with a physical half-life of less than or equal to 120 days for decay-in-storage before disposal in ordinary trash, provided:
- A. Waste to be disposed of in this manner shall be held for decay a minimum of 10 half-lives.
  - B. Before disposal as ordinary trash, the waste shall be surveyed at the container surface with the appropriate survey instrument set on its most sensitive scale and with no interposed shielding to determine that its radioactivity cannot be distinguished from background. All radiation labels shall be removed or obliterated.
  - C. A record of each such disposal permitted under this license condition shall be retained for 3 years. The record must include the date of disposal, the date on which the byproduct material was placed in storage, the radionuclides disposed, the survey instrument used, the background dose rate, the dose rate measured at the surface of each waste container, and the name of the individual who performed the disposal.
19. The licensee is authorized to transport licensed material in accordance with the provisions of 10 CFR Part 71, "Packaging and Transportation of Radioactive Material."
20. Except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures, listed below. This license condition applies only to those procedures that are required to be submitted in accordance with the regulations. Additionally, this license condition does not limit the licensee's ability to make changes to the radiation protection program as provided for in 10 CFR 35.26. The U.S. Nuclear Regulatory Commission's regulations shall govern unless the statements, representations, and procedures in the licensee's application and correspondence are more restrictive than the regulations.
- A. Application dated December 20, 2002
  - B. Letter dated February 15, 2003
- For the U.S. Nuclear Regulatory Commission

Date \_\_\_\_\_

By \_\_\_\_\_

Division of Nuclear Materials Safety

**MATERIALS LICENSE**

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 36, 39, 40, and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations, and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

Licensee	
1. Sample Medical Institution Limited	3. License number 99-02120-01
2. 1234 Main Street Anytown, Pennsylvania 02120	4. Expiration date October 31, 2012
	5. Docket No. 030-02120 Reference No.

6. Byproduct, source, and/or special nuclear material	7. Chemical and/or physical form	8. Maximum amount that licensee may possess at any one time under this license
A. Any byproduct material permitted by 10 CFR 35.100	A. Any	A. As needed
B. Any byproduct material permitted by 10 CFR 35.200	B. Any	B. As needed
C. Any byproduct material permitted by 10 CFR 35.300	C. Any	C. 900 millicuries
D. Any byproduct material permitted by 10 CFR 35.400	D. Sealed Sources (US Atomic Models Ir-192L, Cs-137V, and I-125M)	D. 2 curies
E. Any byproduct material permitted by 10 CFR 35.500	E. Sealed Sources (US Atomic Model I-125P and GD-153A)	E. 0.3 curie per source and 2 curies total
F. Any byproduct material permitted by 10 CFR 31.11	F. Prepackaged Kits	F. 5 millicuries
G. Phosphorus 32	G. Sealed Sources (US Atomic Model P-32IVB)	G. 600 millicuries per source assembly and 1.2 curies total
H. Cesium 137	H. Sealed Source (US Atomic Model CS-137C)	H. 200 millicuries

**MATERIALS LICENSE  
SUPPLEMENTARY SHEET**

License Number  
99-02120-01

Docket or Reference Number  
030-02120

- |   |  |  |
|---|--|--|
| 6. Byproduct, source, and/or special nuclear material | 7. Chemical and/or physical form               | 8. Maximum amount that licensee may possess at any one time under this license |
| I. Iridium 192  | I. Sealed Sources (US Atomic Model IR-192HDR2) | I. 10 curies per source and 20 curies total                                    |
| J. Depleted Uranium                                   | J. Metal                                       | J. 999 kilograms   |

9. Authorized use:

- A. Any uptake, dilution and excretion study permitted by 10 CFR 35.100.
- B. Any imaging and localization study permitted by 10 CFR 35.200.
- C. Any therapy procedure permitted by 10 CFR 35.300.
- D. Any manual brachytherapy procedure permitted by 10 CFR 35.400.
- E. Diagnostic medical use of sealed sources permitted by 10 CFR 35.500 in compatible devices registered pursuant to 10 CFR 30.32(g).
- F. In vitro studies.
- G. One source assembly for medical use in a US Atomic Model INTRA intravascular brachytherapy remote afterloader unit. One source assembly in its shipping container, as necessary for replacement of the source assembly in the intravascular brachytherapy remote afterloader unit.
- H. For use in a US Atomic Model CS-137SC for calibrations and checking of licensee's survey instruments.
- I. One source for medical use described in 10 CFR 35.600, in a US Atomic Model IR-192THER remote afterloading brachytherapy unit. One source in its shipping container as necessary for replacement of the source in the remote afterloader unit.
- J. Shielding in a linear accelerator.

CONDITIONS

10. Licensed material may be used or stored only at the licensee's facilities located at 1234 Main Street, Anytown, Pennsylvania.
11. The Radiation Safety Officer for this license is Melba Physicist, M.S.
12. Licensed material is only authorized for use by, or under the supervision of:
- A. Individuals permitted to work as an authorized user, authorized nuclear pharmacist, and/or authorized medical physicist in accordance with 10 CFR 35.13 and 35.14.

**MATERIALS LICENSE  
SUPPLEMENTARY SHEET**

License Number  
99-02120-01

Docket or Reference Number  
030-02120

B. The following individuals are authorized users for medical use:

	<u>Material and Use</u>
Jane Diagnostic, M.D.	35.100; 35.200; 35.300; 35.500; <i>In vitro</i> studies; Gadolinium-153 for patient attenuation correction
Thomas Group, D.O.	35.100; 35.200; 35.300 except Iodine-131
Gilbert Lawrence, M.D.	35.100; 35.200; 35.500; Oral administration of sodium iodide I-131 in quantities less than or equal to 33 millicuries
John Therapy, M.D.	35.400; Phosphorus-32 for intravascular brachytherapy procedures; Iridium-192 for uses in a High Dose Rate Remote Afterloading Brachytherapy Unit; Depleted Uranium

C. The following individuals are authorized users for non-medical uses:

	<u>Material and Use</u>
James Pathology	<i>In vitro</i> studies
Cecil Source, Ph.D.	Cesium-137 for calibration of instruments

D. The following individual is an authorized medical physicist:

	<u>Material and Use</u>
Melba Physicist, M.S.	Iridium-192 in a High Dose Rate Remote Afterloader Unit for calibrations, spot-checks, and training

E. Intravascular brachytherapy procedures shall be conducted under the supervision of the authorized user, who will consult with the interventional cardiologist/physician and authorized medical physicist prior to initiating treatment. The procedures shall be conducted in the physical presence of the authorized user or the authorized medical physicist.

13. In addition to the possession limits in Item 8, the licensee shall further restrict the possession of licensed material to quantities below the minimum limit specified in 10 CFR 30.35(d) for establishing decommissioning financial assurance.

**MATERIALS LICENSE  
SUPPLEMENTARY SHEET**

License Number

99-02120-01

Docket or Reference Number

030-02120


14. The intravascular brachytherapy afterloader device shall be inspected and serviced at intervals recommended by the manufacturer, and maintenance and repair shall be performed only by the manufacturer or persons specifically licensed by NRC or an Agreement State to perform such services.
15. For sealed sources not associated with 10 CFR Part 35 use, the following conditions apply:
- Sealed sources shall be tested for leakage and/or contamination at intervals not to exceed the intervals specified in the certificate of registration issued by the U.S. Nuclear Regulatory Commission under 10 CFR 32.210 or under equivalent regulations of an Agreement State.
  - Notwithstanding Paragraph A of this Condition, sealed sources designed to primarily emit alpha particles shall be tested for leakage and/or contamination at intervals not to exceed 3 months.
  - In the absence of a certificate from a transferor indicating that a leak test has been made within the intervals specified in the certificate of registration issued by the U.S. Nuclear Regulatory Commission under 10 CFR 32.210 or under equivalent regulations of an Agreement State, prior to the transfer, a sealed source received from another person shall not be put into use until tested and the test results received.
  - Sealed sources need not be tested if they contain only hydrogen-3; or they contain only a radioactive gas; or the half-life of the isotope is 30 days or less; or they contain not more than 100 microcuries of beta- and/or gamma-emitting material or not more than 10 microcuries of alpha-emitting material.
  - Sealed sources need not be tested if they are in storage and are not being used; however, when they are removed from storage for use or transferred to another person and have not been tested within the required leak test interval, they shall be tested before use or transfer. No sealed source shall be stored for a period of more than 10 years without being tested for leakage and/or contamination.
  - The leak test shall be capable of detecting the presence of 0.005 microcurie (185 becquerels) of radioactive material on the test sample. If the test reveals the presence of 0.005 microcurie (185 becquerels) or more of removable contamination, a report shall be filed with the U.S. Nuclear Regulatory Commission in accordance with 10 CFR 30.50(c)(2), and the source shall be removed immediately from service and decontaminated, repaired, or disposed of in accordance with Commission regulations.
  - Tests for leakage and/or contamination, including leak test sample collection and analysis, shall be performed by the licensee or by other persons specifically licensed by the U.S. Nuclear Regulatory Commission or an Agreement State to perform such services.
16. The licensee shall conduct a physical inventory every six months, or at other intervals approved by the U.S. Nuclear Regulatory Commission, to account for all sources and/or devices received and possessed under the license.

**MATERIALS LICENSE  
SUPPLEMENTARY SHEET**

License Number  
99-02120-01

Docket or Reference Number  
030-02120

- 17. Sealed sources or detector cells containing licensed material shall not be opened or sources removed from source holders by the licensee.
- 18. The licensee is authorized to transport licensed material in accordance with the provisions of 10 CFR Part 71, "Packaging and Transportation of Radioactive Material."
- 19. Except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures, listed below. This license condition applies only to those procedures that are required to be submitted in accordance with the regulations. Additionally, this license condition does not limit the licensee's ability to make changes to the radiation protection program as provided for in 10 CFR 35.26. The U.S. Nuclear Regulatory Commission's regulations shall govern unless the statements, representations, and procedures in the licensee's application and correspondence are more restrictive than the regulations.
  - A. Application dated June 10, 2002
  - B. Letter dated September 30, 2002



For the U.S. Nuclear Regulatory Commission

Date \_\_\_\_\_

By \_\_\_\_\_

Division of Nuclear Materials Safety

**MATERIALS LICENSE**

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 36, 39, 40, and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations, and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

<p style="text-align: center;">Licensee</p> <p>1. Thomas I. Royed, M.D.</p> <p>2. Suite 301 2 Physician Circle Parkway Anytown, Pennsylvania 02200</p>	<p>3. License number 99-02200-01</p> <p>4. Expiration date December 31, 2012</p> <p>5. Docket No. 030-02200 Reference No.</p>
--	---

<p>6. Byproduct, source, and/or special nuclear material</p> <p>A. Iodine-131 permitted by 10 CFR 35.300</p>	<p>7. Chemical and/or physical form</p> <p>A. Any</p>	<p>8. Maximum amount that licensee may possess at any one time under this license</p> <p>A. 500 millicuries</p>
--	---	---

9. Authorized use:

A. Any iodine-131 procedure permitted by 10 CFR 35.300 for which the patient can be released under the provisions of 10 CFR 35.75.

★ ★ ★ ★ ★  
CONDITIONS

- 10. Licensed material may be used or stored only at the licensee's facilities located at Suite 301, 2 Physician Circle Parkway, Anytown, Pennsylvania.
- 11. The Radiation Safety Officer for this license is Roger O. Blation, M.D.
- 12. Licensed material is only authorized for use by, or under the supervision of:
  - A. Individuals permitted to work as an authorized user in accordance with 10 CFR 35.13 and 35.14.

**MATERIALS LICENSE  
SUPPLEMENTARY SHEET**

License Number  
99-02200-01

Docket or Reference Number  
030-02200

B. The following individuals are authorized users for the materials and uses indicated:

Authorized Users

Material and Use

Roger O. Blation, M.D.

Oral administration of sodium iodide I-131


Thomas I. Royed, M.D.

Oral administration of sodium iodide I-131 in quantities less than or equal to 33 millicuries

13. The licensee is authorized to transport licensed material in accordance with the provisions of 10 CFR Part 71, "Packaging and Transportation of Radioactive Material."

14. Except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures, listed below. This license condition applies only to those procedures that are required to be submitted in accordance with the regulations. Additionally, this license condition does not limit the licensee's ability to make changes to the radiation protection program as provided for in 10 CFR 35.26. The U.S. Nuclear Regulatory Commission's regulations shall govern unless the statements, representations, and procedures in the licensee's application and correspondence are more restrictive than the regulations.

A. Application dated October 30, 2002



For the U.S. Nuclear Regulatory Commission

Date \_\_\_\_\_

By \_\_\_\_\_

**MATERIALS LICENSE**

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 36, 39, 40, and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations, and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

Licensee	
1. Manuel U. Seeds	3. License number <b>99-02200-01</b>
2. Suite 106 3 Physician Circle Parkway Anytown, Pennsylvania 02200	4. Expiration date <b>September 30, 2014</b>
	5. Docket No. <b>030-02200</b> Reference No.

6. Byproduct, source, and/or special nuclear material	7. Chemical and/or physical form	8. Maximum amount that licensee may possess at any one time under this license
A. Any byproduct material permitted by 10 CFR 35.400	A. Sealed Sources (US Atomic Model US-I-125-10L)	A. 500 millicuries
9. Authorized use:		
A. Any manual brachytherapy procedure permitted by 10 CFR 35.400 for which the patient can be released under the provisions of 10 CFR 35.75.		

★ **CONDITIONS** ★

- 10. Licensed material may be used or stored only at the licensee's facilities located at Suite 106, 3 Physician Circle Parkway, Anytown, Pennsylvania.
- 11. The Radiation Safety Officer for this license is Manuel U. Seeds, M.D.
- 12. Licensed material is only authorized for use by, or under the supervision of:
  - A. Individuals permitted to work as an authorized user in accordance with 10 CFR 35.13 and 35.14.
  - B. The following individuals are authorized users for the materials and uses indicated:

Authorized Users

Manuel U. Seeds, M.D.

Material and Use

10 CFR 35.400


**MATERIALS LICENSE  
SUPPLEMENTARY SHEET**

License Number  
99-02200-01

Docket or Reference Number  
030-02200

- 13. The licensee is authorized to transport licensed material in accordance with the provisions of 10 CFR Part 71, "Packaging and Transportation of Radioactive Material."
- 14. Except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures, listed below. This license condition applies only to those procedures that are required to be submitted in accordance with the regulations. Additionally, this license condition does not limit the licensee's ability to make changes to the radiation protection program as provided for in 10 CFR 35.26. The U.S. Nuclear Regulatory Commission's regulations shall govern unless the statements, representations, and procedures in the licensee's application and correspondence are more restrictive than the regulations.

A. Application dated July 20, 2004



For the U.S. Nuclear Regulatory Commission

Date \_\_\_\_\_

By \_\_\_\_\_

Division of Nuclear Materials Safety

**MATERIALS LICENSE**

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 36, 39, 40, and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations, and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

Licensee		
1. Noe Directive, M.D.		3. License number 99-02201-01
2. Suite 112 2 Physician Circle Parkway Anytown, Pennsylvania 02201		4. Expiration date June 30, 2013
		5. Docket No. 030-02201 Reference No.
6. Byproduct, source, and/or special nuclear material	7. Chemical and/or physical form	8. Maximum amount that licensee may possess at any one time under this license
A. Any byproduct material permitted by 10 CFR 35.100	A. Any	A. As needed
B. Any byproduct material permitted by 10 CFR 35.200	B. Any, except generators	B. As needed
9. Authorized use:		
A. Any uptake, dilution and excretion study permitted by 10 CFR 35.100.		
B. Any imaging and localization study permitted by 10 CFR 35.200.		

**CONDITIONS**

- 10. Licensed material may be used or stored only at the licensee's facilities located at Suite 112, 2 Physician Circle Parkway, Anytown, Pennsylvania.
- 11. The Radiation Safety Officer for this license is Noe Directive, MD.
- 12. Licensed material is only authorized for use by, or under the supervision of:
  - A. Individuals permitted to work as an authorized user in accordance with 10 CFR 35.13 and 35.14.

**MATERIALS LICENSE  
SUPPLEMENTARY SHEET**

License Number  
99-02201-01

Docket or Reference Number  
030-02201

B. The following individuals are authorized users for the materials and uses indicated:

Authorized Users


Material and Use

Noe Directive, M.D.

35.100; 35.200

- 13. In addition to the possession limits in Item 8, the licensee shall further restrict the possession of licensed material to quantities below the minimum limit specified in 10 CFR 30.35(d) for establishing decommissioning financial assurance.
- 14. The licensee is authorized to transport licensed material in accordance with the provisions of 10 CFR Part 71, "Packaging and Transportation of Radioactive Material."
- 15. Except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures, listed below. This license condition applies only to those procedures that are required to be submitted in accordance with the regulations. Additionally, this license condition does not limit the licensee's ability to make changes to the radiation protection program as provided for in 10 CFR 35.26. The U.S. Nuclear Regulatory Commission's regulations shall govern unless the statements, representations, and procedures in the licensee's application and correspondence are more restrictive than the regulations.

A. Application dated April 11, 2003



For the U.S. Nuclear Regulatory Commission

Date \_\_\_\_\_

By \_\_\_\_\_

**MATERIALS LICENSE**

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 36, 39, 40, and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations, and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

Licensee	
1. Sample Mobile Nuclear Medicine  2. Suite 214 2 Physician Circle Parkway Anytown, Pennsylvania 02220	3. License number 99-02220-01  4. Expiration date December 31, 2012  5. Docket No. 030-02220 Reference No.
6. Byproduct, source, and/or special nuclear material  A. Any byproduct material permitted by 10 CFR 35.100  B. Any byproduct material permitted by 10 CFR 35.200	Chemical and/or physical form  A. Any  B. Any, except generators
9. Authorized use:  A. Any uptake, dilution and excretion study permitted by 10 CFR 35.100. B. Any imaging and localization study permitted by 10 CFR 35.200.	8. Maximum amount that licensee may possess at any one time under this license  A. As needed  B. As needed

**CONDITIONS**

10. Licensed material may be used or stored at the licensee's facilities located at Suite 214, 2 Physician Circle Parkway, Anytown, Pennsylvania and may be used at temporary job sites of the licensee anywhere in the United States where the U.S. Nuclear Regulatory Commission maintains jurisdiction for regulating the use of licensed material, including areas of exclusive Federal jurisdiction within Agreement States.

If the jurisdiction status of a Federal facility within an Agreement State is unknown, the licensee should contact the Federal agency controlling the job site in question to determine whether the proposed job site is an area of exclusive Federal jurisdiction. Authorization for use of radioactive materials at job sites in Agreement States not under exclusive Federal jurisdiction shall be obtained from the appropriate state regulatory agency.

**MATERIALS LICENSE  
SUPPLEMENTARY SHEET**

License Number  
99-02220-01

Docket or Reference Number  
030-02220

11. Licensed material is only authorized for use by, or under the supervision of:

- A. Individuals permitted to work as an authorized user in accordance with 10 CFR 35.13 and 35.14.
- B. The following individuals are authorized users for the materials and uses indicated:

Authorized Users

Material and Use

Thomas Group, D.O.

35.100; 35.200

12. The Radiation Safety Officer for this license is Thomas Group, D.O.

13. In addition to the possession limits in Item 8, the licensee shall further restrict the possession of licensed material to quantities below the minimum limit specified in 10 CFR 30.35(d) for establishing decommissioning financial assurance.

14. The licensee is authorized to transport licensed material in accordance with the provisions of 10 CFR Part 71, "Packaging and Transportation of Radioactive Material."

15. Except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures, listed below. This license condition applies only to those procedures that are required to be submitted in accordance with the regulations. Additionally, this license condition does not limit the licensee's ability to make changes to the radiation protection program as provided for in 10 CFR 35.26. The U.S. Nuclear Regulatory Commission's regulations shall govern unless the statements, representations, and procedures in the licensee's application and correspondence are more restrictive than the regulations.

A. Application dated November 15, 2002

For the U.S. Nuclear Regulatory Commission

Date \_\_\_\_\_

By \_\_\_\_\_

Division of Nuclear Materials Safety

**MATERIALS LICENSE**

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 36, 39, 40, and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations, and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

<p>Licensee</p> <p>1. Sample Teletherapy</p> <p>2. 200 Cobalt Street Anytown, Pennsylvania 02300</p>	<p>3. License number 99-02300-01</p> <p>4. Expiration date June 30, 2013</p> <p>5. Docket No. 030-02300 Reference No.</p>
--	---

<p>6. Byproduct, source, and/or special nuclear material</p> <p>A. Cobalt 60 permitted by 10 CFR 35.600</p> <p>B. Depleted Uranium</p>	<p>Chemical and/or physical form</p> <p>A. Sealed Sources (US Atomic Model US-CO-60TELE)</p> <p>B. Metal</p>	<p>8. Maximum amount that licensee may possess at any one time under this license</p> <p>A. 5,500 curies per source and 11,000 curies total</p> <p>B. 999 kilograms</p>
--	--	---

9. Authorized use:

A. One source for medical use permitted by 10 CFR 35.600, in a US Atomic Model TELE teletherapy unit. One source in its shipping container as necessary for replacement of the source in the teletherapy unit.

B. Shielding in a teletherapy unit.

**CONDITIONS**

- 10. Licensed material may be used or stored only at the licensee's facilities located at 200 Cobalt Street, Anytown, Pennsylvania.
- 11. The Radiation Safety Officer for this license is Sarah Smith, M.S.
- 12. Licensed material is only authorized for use by, or under the supervision of:
  - A. Individuals permitted to work as an authorized user, and/or authorized medical physicist in accordance with 10 CFR 35.13 and 35.14.

**MATERIALS LICENSE  
SUPPLEMENTARY SHEET**

License Number  
99-02300-01

Docket or Reference Number  
030-02300

B. The following individual is an authorized user for medical use:

David Jones, M.D.

Material and Use

Cobalt-60 for medical uses in a Teletherapy Unit;  
Depleted Uranium

C. The following individual is an authorized medical physicist:

Sarah Smith, M.S.

Material and Use

Cobalt-60 in a Teletherapy Unit for calibrations, spot-  
checks, and training

- 13. The licensee is exempt from decommissioning financial assurance requirements for possession of licensed material in sealed sources in quantities greater than the limits in 10 CFR 30.35(d) for the purpose of source changes only. This exemption is granted for no more than 30 days for any one source change.
- 14. The licensee is authorized to transport licensed material in accordance with the provisions of 10 CFR Part 71, "Packaging and Transportation of Radioactive Material."
- 15. Except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures, listed below. This license condition applies only to those procedures that are required to be submitted in accordance with the regulations. Additionally, this license condition does not limit the licensee's ability to make changes to the radiation protection program as provided for in 10 CFR 35.26. The U.S. Nuclear Regulatory Commission's regulations shall govern unless the statements, representations, and procedures in the licensee's application and correspondence are more restrictive than the regulations.

A. Application dated March 19, 2003

For the U.S. Nuclear Regulatory Commission

Date \_\_\_\_\_

By \_\_\_\_\_

Division of Nuclear Materials Safety

**MATERIALS LICENSE**

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 36, 39, 40, and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations, and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

Licensee	
1. Sample Gamma Stereotactic	3. License number 99-02310-01
2. 100 Main Street Anytown, Pennsylvania 02310	4. Expiration date March 31, 2013
	5. Docket No. 030-02310 Reference No.
6. Byproduct, source, and/or special nuclear material	7. Chemical and/or physical form
A. Cobalt 60 permitted by 10 CFR 35.600	A. Sealed Sources (US Atomic Model US-CO-60STER)
	8. Maximum amount that licensee may possess at any one time under this license
	A. 33 curies per source and 10,000 curies total
9. Authorized use:	
A. For medical use permitted by 10 CFR 35.600 in a US Atomic Model STEREO gamma stereotactic radiosurgery unit. Sources in the shipping container as necessary for replacement of the sources in the gamma stereotactic radiosurgery unit.	

CONDITIONS

- 10. Licensed material may be used or stored only at the licensee's facilities located at 100 Main Street, Anytown, Pennsylvania.
- 11. The Radiation Safety Officer for this license is Kimberly Therapy, Ph.D.
- 12. Licensed material is only authorized for use by, or under the supervision of:
  - A. Individuals permitted to work as an authorized user, and/or authorized medical physicist in accordance with 10 CFR 35.13 and 35.14.

**MATERIALS LICENSE  
SUPPLEMENTARY SHEET**

License Number  
99-02310-01

Docket or Reference Number  
030-02310

B. The following individuals are authorized users for medical uses:

Material and Use

John Smith, M.D.

Cobalt-60 for medical uses in a Gamma Stereotactic Radiosurgery Unit

Jessica Water, M.D.

Cobalt-60 for medical uses in a Gamma Stereotactic Radiosurgery Unit

C. The following individuals are authorized medical physicists:

Material and Use

Kimberly Therapy, Ph.D.

Cobalt-60 in a Gamma Stereotactic Radiosurgery Unit for calibrations, spot-checks, and training

Ronald Stereo, M.S.

Cobalt-60 in a Gamma Stereotactic Radiosurgery Unit for calibrations, spot-checks, and training

13. The licensee is authorized to transport licensed material in accordance with the provisions of 10 CFR Part 71, "Packaging and Transportation of Radioactive Material."

14. Except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures, listed below. This license condition applies only to those procedures that are required to be submitted in accordance with the regulations. Additionally, this license condition does not limit the licensee's ability to make changes to the radiation protection program as provided for in 10 CFR 35.26. The U.S. Nuclear Regulatory Commission's regulations shall govern unless the statements, representations, and procedures in the licensee's application and correspondence are more restrictive than the regulations.

A. Application dated December 15, 2002

B. Letter dated March 4, 2003

For the U.S. Nuclear Regulatory Commission

Date \_\_\_\_\_

By \_\_\_\_\_

Division of Nuclear Materials Safety

**MATERIALS LICENSE**

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 36, 39, 40, and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations, and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

Licensee		
1. Sample Pacemaker License	3. License number SNM-22160	
2. 100 Medical Center Drive Anytown, Pennsylvania 22160	4. Expiration date October 31, 2012	
	5. Docket No. 070-22160 Reference No.	
6. Byproduct, source, and/or special nuclear material	7. Chemical and/or physical form	8. Maximum amount that licensee may possess at any one time under this license
A. Plutonium (principal radionuclide Pu-238)	A. Sealed Sources (US Atomic Model US-PU-238)	A. 5 curies per source and 50 curies total
9. Authorized use:		
A. As a component of US Atomic Model PACE nuclear-powered pacemakers for clinical evaluation in accordance with manufacturer's protocol dated October 19, 1995. This authorization includes: follow-up, explantation, recovery, and disposal, but not implantation.		

**CONDITIONS**


10. Licensed material may be used or stored only at the licensee's facilities located at 100 Medical Center Drive, Anytown, Pennsylvania.
11. The Radiation Safety Officer for this license is Chief Radiologist, M.D.
12. The physicians responsible for follow-up, explantation, and return of nuclear-powered pacemakers to the manufacturer for proper disposal are Chief Cardiosurgeon, M.D.
13. The specified possession limit for nuclear-powered pacemakers includes all licensed material possessed by the licensee under this license whether in storage, implanted in patients, or otherwise in use.

**MATERIALS LICENSE  
SUPPLEMENTARY SHEET**

License Number  
SNM-22160

Docket or Reference Number  
070-22160

- 14. The licensee shall continue patient follow-up and replacement procedures for the nuclear-powered pacemaker during the life of the patient. Procedures for recovery and authorized disposal of the nuclear-powered pacemaker by return to the manufacturer shall be followed upon the death of the patient.
- 15. The licensee shall report to the U.S. Nuclear Regulatory Commission's Regional Office referenced in Appendix D of 10 CFR Part 20, within 10 days after discovery of loss of contact with a nuclear-powered pacemaker patient.
- 16. Sealed sources or detector cells containing licensed material shall not be opened or sources removed from source holders by the licensee.
- 17. The licensee is authorized to transport licensed material in accordance with the provisions of 10 CFR Part 71, "Packaging and Transportation of Radioactive Material."
- 18. Except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures, listed below. The U.S. Nuclear Regulatory Commission's regulations shall govern unless the statements, representations, and procedures in the licensee's application and correspondence are more restrictive than the regulations.
  - A. Application dated September 30, 2002
  - B. Letter dated October 15, 2002



For the U.S. Nuclear Regulatory Commission

Date \_\_\_\_\_

By \_\_\_\_\_

Division of Nuclear Materials Safety