

Region I NMSS Licensee Event Report

Licensee	AGILENT
Event Description	Notice of General License Device with High Removable Activity

License #	07-28762-016	Docket #	03002988	MLER-RI	2003-001
Event Date	JAN 2003	Report Date	JAN 10, 2003		

1. REPORTING REQUIREMENT

<input type="checkbox"/> 10 CFR 20.2201 Theft or Loss	<input type="checkbox"/> 10 CFR 35.33 Misadministration
<input type="checkbox"/> 10 CFR 20.2203 30 Day Report	<input type="checkbox"/> License Condition
<input type="checkbox"/> 10 CFR 30.50 Report	
<input checked="" type="checkbox"/> Other <u>10 CFR 31.5</u>	

2. REGION I RESPONSE

<input type="checkbox"/> Immediate Site Inspection	Inspector/Date	
<input type="checkbox"/> Special Inspection	Inspector/Date	
<input type="checkbox"/> Telephone Inquiry	Inspector/Date	
<input type="checkbox"/> Preliminary Report	<input type="checkbox"/> Daily Report	
<input checked="" type="checkbox"/> Information Entered in RI Log	<input checked="" type="checkbox"/> Review at next inspection	
Report referred to: _____		

3. ~~REPORTING REQUIREMENT~~

<input checked="" type="checkbox"/> Description of Event	<input checked="" type="checkbox"/> Corrective Actions
<input checked="" type="checkbox"/> Levels of RAM Involved	<input type="checkbox"/> Calculations Adequate
<input checked="" type="checkbox"/> Cause of Event	<input type="checkbox"/> Additional Information Requested from Licensee

4. ~~SPECIAL INSURCTIONS OR COMMENTS~~

<input checked="" type="checkbox"/>		
Completed by:	Betsy Ullrich	Date
Reviewed by:	[Signature]	Date
		1/16/03
		1/16/2003

 **Agilent Technologies**
2850 Centerville Road
Wilmington, DE 19808

RECEIVED
REGION 1

2003 JAN 10 PM 1:55

Division of Nuclear Materials Safety
U.S. Nuclear Regulatory Commission
Region I
475 Allendale Road,
King of Prussia, PA 19406

Re: Notice of General License device with high removable activity (NRC License 07-28762-02G)

January 8, 2003

Dear Sir or Madam,

Per the requirements of 10 CFR 31.5, I am notifying your office of the following devices returned to us by a customer from which we have obtained removable activity wipes in excess of 0.005 μCi . The cells were evaluated by support engineering and their disposition is identified below. I have contacted the customer for further information and will notify the cell supplier of the high wipes.

Serial #	Model #	Activity (μCi)	Disposition	Comments
L0026	19233	0.189	Scrapped	Cell 19 years old. Source was corroded.
L2242	19233	0.026	Scrapped	Cell 14 years old. Source dirty and corroded

Please contact me at 302-633-8262 if there are any questions.

Thank you,



David S. Bennett
Radiation Safety Officer