

**From:** <Tim.Daly@ahlstrom.com>  
**To:** "Stephanie Tullio" <SXT4@nrc.gov>  
**Date:** Thu, Jan 23, 2003 1:29 PM  
**Subject:** Re: Fwd: NRC License renewal info. request.

Stephanie,

I would like the license section you referenced to read as follows:

A. Installation, initial radiation surveys, relocation, removal from service, dismantling, alignment, and non-routine maintenance or repair of components related to the radiological safety of the gauge shall be performed only by, Timothy Daly or other individuals who have completed the training specified in our application dated 11/25/2002 and accompanied by cover letter dated Dec. 5, 2002 or by persons specifically licensed by the U.S. Nuclear Regulatory Commission or an Agreement State to perform such services.

B. The following services shall not be performed by the licensee: Replacement and disposal of the sealed sources. These services shall be performed only by persons specifically licensed by the U.S. Nuclear Regulatory Commission or an Agreement State to perform such services.

To address the points you brought up concerning Appendix N of NUREG 1556 I have modified the "Procedure for Non-Routine Maintenance of ABB/Measurex Basis Weight Gauges" that was sent with the application.

I have also completed a NUREG 1556 section J dosimetry evaluation to establish that extremity dosimetry is not required.

Please let me know if this is satisfactory.

Tim Daly  
Ahlstrom FiberComposites

Ahlstrom Windsor Locks LLC  
Two Elm Street  
Windsor Locks, Connecticut 06096  
Tel: +1 860 654 8560  
Fax : +1 860 654 8569

"Stephanie Tullio" <SXT4@nrc.gov>  
01/22/2003 02:16 PM

To: <Tim.Daly@Ahlstrom.com>  
cc:  
Subject: Fwd: NRC License renewal info. request.

----- Message from "Stephanie Tullio" <SXT4@nrc.gov> on Wed, 22 Jan 2003  
14:11:57 -0500 -----

To:  
<Timothy.Daly@ahlstrom.com>  
Subject:  
NRC License renewal info. request.  
Tim,

Below are non-routine maintenance applications, please identify those you  
will be performing.

A. Installation, initial radiation surveys, relocation,  
removal from service, dismantling, alignment, replacement, disposal of the  
sealed source and non-routine maintenance or repair of components related  
to the radiological safety of the gauge shall be performed only by,  
[insert names] or other individuals who have completed the training  
specified in [application/letter] dated [insert dates] or by persons  
specifically licensed by the U.S. Nuclear Regulatory Commission or an  
Agreement State to perform such services.

B. The following services shall not be performed by the  
licensee: [Insert the services NOT authorized in this license, delete  
remaining:] Installation, initial radiation surveys, relocation, removal  
from service, dismantling, alignment, replacement, disposal of the sealed

sources and non-routine maintenance or repair of components related to the radiological safety of the gauge. These services shall be performed only by persons specifically licensed by the U.S. Nuclear Regulatory Commission or an Agreement State to perform such services.

Any questions please call 610-337-5240.

Thank you,  
Stephanie Tullio, Health Physicist  
USNRC, RI