

**Region I
NMSS Licensee Event Report**

Licensee	AGILENT Technologies		
Event Description	Notice of General License Denial w/ High Removable Activity		
License #	07-28762- 02G	Docket #	03002988
Event Date	2003	Report Date	3/31/03
		MLER-RI	2003-021

1. REPORTING REQUIREMENT

<input type="checkbox"/>	10 CFR 20.2201 Theft or Loss	<input type="checkbox"/>	10 CFR 35.33 Misadministration
<input type="checkbox"/>	10 CFR 20.2203 30 Day Report	<input type="checkbox"/>	License Condition
<input type="checkbox"/>	10 CFR 30.50 Report		
<input checked="" type="checkbox"/>	Other <u>10 CFR 31.5</u>		

2. REGION I RESPONSE

<input type="checkbox"/>	Immediate Site Inspection	Inspector/Date	
<input type="checkbox"/>	Special Inspection	Inspector/Date	
<input type="checkbox"/>	Telephone Inquiry	Inspector/Date	
<input type="checkbox"/>	Preliminary Report	<input type="checkbox"/>	Daily Report
<input checked="" type="checkbox"/>	Information Entered in RI Log	<input checked="" type="checkbox"/>	Review at next inspection
<input type="checkbox"/>	Report referred to: _____		

3. REPORT EVALUATION

<input checked="" type="checkbox"/>	Description of Event	<input type="checkbox"/>	Corrective Actions
<input checked="" type="checkbox"/>	Levels of RAM Involved	<input type="checkbox"/>	Calculations Adequate
<input checked="" type="checkbox"/>	Cause of Event	<input type="checkbox"/>	Additional Information Requested from Licensee

4. SPECIAL INSTRUCTIONS OR COMMENTS

Public Non-Public

Completed by:	<u>Bob Ullrich</u> 4/2/03	Date	<u>BU</u> 4/2/03
Reviewed by:	<u>John D. Kern</u> 4/2/2003	Date	4/2/2003

 **Agilent Technologies**
2850 Centerville Road
Wilmington, DE 19808

RECEIVED
REGION I

2003 MAR 31 PM 1:05

Division of Nuclear Materials Safety
U.S. Nuclear Regulatory Commission
Region I
475 Allendale Road,
King of Prussia, PA 19406

Re: Notice of General License device with high removable activity (NRC License 07-28762-02G)

March 28, 2003

Dear Sir or Madam,

Per the requirements of 10 CFR 31.5, I am notifying your office of a general license device recently received from a source supplier from which we have obtained a removable activity wipe in excess of 0.005 μCi . The cell was evaluated by support engineering and its disposition is identified below. I will notify the cell supplier of the high wipe.

<u>Serial #</u>	<u>Model #</u>	<u>Activity (μCi)</u>	<u>Disposition</u>
F4458	19233	0.008	Scrapped

Comments:

Source was dirty and black. Cell was 11 years old.

Please contact me at 302-633-8262 if there are any questions.

Thank you,



David S. Bennett
Radiation Safety Officer