

Region I
NMSS Licensee Event Report

Licensee	Agilent		
Event Description	Notice of General License Device with High Removable		
	Activity		
License #	07-28762- ⁶³⁶	Docket #	03002988
Event Date	2003	Report Date	4-7-03
		MLER-RI	2003-023

1. **REPORTING REQUIREMENT**

<input type="checkbox"/> 10 CFR 20.2201 Theft or Loss <input type="checkbox"/> 10 CFR 20.2203 30 Day Report <input type="checkbox"/> 10 CFR 30.50 Report <input checked="" type="checkbox"/> Other <u>10 CFR 31.5</u>	<input type="checkbox"/> 10 CFR 35.33 Misadministration <input type="checkbox"/> License Condition
--	---

2. **REGION I RESPONSE**

<input type="checkbox"/> Immediate Site Inspection <input type="checkbox"/> Special Inspection <input type="checkbox"/> Telephone Inquiry <input type="checkbox"/> Preliminary Report <input checked="" type="checkbox"/> Information Entered in RI Log <input type="checkbox"/> Report referred to: _____	<table border="1" style="width: 100%;"><tr><td style="width: 50%;">Inspector/Date</td><td></td></tr><tr><td>Inspector/Date</td><td></td></tr><tr><td>Inspector/Date</td><td></td></tr><tr><td><input type="checkbox"/> Daily Report</td><td></td></tr><tr><td><input type="checkbox"/> Review at next inspection</td><td></td></tr></table>	Inspector/Date		Inspector/Date		Inspector/Date		<input type="checkbox"/> Daily Report		<input type="checkbox"/> Review at next inspection	
Inspector/Date											
Inspector/Date											
Inspector/Date											
<input type="checkbox"/> Daily Report											
<input type="checkbox"/> Review at next inspection											

3. **REPORT EVALUATION**

<input checked="" type="checkbox"/> Description of Event <input checked="" type="checkbox"/> Levels of RAM Involved <input checked="" type="checkbox"/> Cause of Event	<input checked="" type="checkbox"/> Corrective Actions <input type="checkbox"/> Calculations Adequate <input type="checkbox"/> Additional Information Requested from Licensee
--	---

4. **SPECIAL INSTRUCTIONS OR COMMENTS**


Public

X

Non-Public

	Initials/Date	
Completed by:	Bety Ullrich 4/10/03	Date
Reviewed by:	JLD/Km	Date

	Initials/Date
	BY 4/10/03
	4/10/2003

 **Agilent Technologies**
2850 Centerville Road
Wilmington, DE 19808

RECEIVED
REGION 1

2003 APR -7 PM 1:38

Division of Nuclear Materials Safety
U.S. Nuclear Regulatory Commission
Region I
475 Allendale Road,
King of Prussia, PA 19406

Re: Notice of General License device with high removable activity (NRC License 07-28762-02G)

April 4, 2003

Dear Sir or Madam,

Per the requirements of 10 CFR 31.5, I am notifying your office of a general license device recently received from a source supplier from which we have obtained a removable activity wipe in excess of $0.005 \mu\text{Ci}$. The cell was evaluated by support engineering and its disposition is identified below. I will notify the cell supplier of the high wipe.


Serial #	Model #	Activity (μCi)	Disposition
F5044	19233	0.604	Scrapped

Comments:

Source was very dirty. Cell was 10 years old.

Please contact me at 302-633-8262 if there are any questions.

Thank you,


David S. Bennett
Radiation Safety Officer