

UNITED STATES
NUCLEAR REGULATORY COMMISSION
OFFICE OF NUCLEAR MATERIAL SAFETY AND SAFEGUARDS
WASHINGTON, D.C. 20555

October 29, 1987

NRC INFORMATION NOTICE NO. 87-55: PORTABLE MOISTURE/DENSITY GAUGES: RECENT
INCIDENTS OF PORTABLE GAUGES BEING STOLEN
OR LOST

[Refer also to NRC Information Notice No. 86-67, August 15, 1986, Same subject]

Addressees:

All NRC licensees authorized to possess portable gauges.

Purpose:

This notice is intended to inform licensees of the recent increase in incidents of portable gauges being lost or stolen during use, transportation, and storage. It is expected that recipients will review the information for applicability to their licensed activities and consider actions, if appropriate, to preclude similar problems from occurring at their facilities. However, suggestions contained in this information notice do not constitute NRC requirements; therefore, no specific action or written response is required.

Description of Circumstances:

Several incidents have occurred recently where portable moisture and density gauges have been lost or stolen from licensees, either from job sites or from vehicles during transportation. Gauges lost or stolen were not secured or were stolen while left unattended by the users.

In one recent event, a gauge was stolen after the user failed to lock the case and failed to chain the case to the truck, which itself was left unattended. This event resulted in four violations of NRC requirements. In another case, a gauge was stolen after being left outdoors, unattended, over a weekend, at a construction site.

Discussion:

A primary contributing factor in these incidents was the failure of the gauge users to secure and maintain control over the gauges.

Title 10 CFR 20.207 requires that licensed material (e.g., portable gauges) must be under the constant surveillance and immediate control of the licensee, or must be secured in storage. At construction sites, the licensee must constantly control access to the gauge during use at the site, and keep it in locked storage when not in use.

8710230213

ZA

IDTR-11C

NRC licensees transporting portable gauges are subject to the regulations in 10 CFR Part 71. Section 71.5(a) incorporates certain regulations (49 CFR 170-189) of the Department of Transportation (DOT), to which these licensees are also subject. Licensees who transport gauges to and from temporary job sites in private vehicles are shippers acting as private carriers, and as such, must comply with the DOT regulations governing both shippers and carriers. Title 49 CFR 177.842(d) requires that packages containing radioactive material (i.e., the gauge in its case) must be blocked and braced to prevent movement of the package during transportation (see Information Notice 87-31, attached). For pickup trucks, this requirement is usually met when the gauge is secured within its case, and the case is secured and locked to the bed of the truck.

The NRC has also noted several incidents where gauges were damaged in transportation accidents or by being run over by construction vehicles. The most prevalent cause of such incidents is the failure by the licensee to secure and lock the cargo door, combined with a failure to properly secure the device/case to the bed of the vehicle. Licensees are reminded that they should use care to avoid these types of incidents. If a gauge is damaged, the area around the gauge should be restricted, and the licensee's radiation safety officer and NRC should be promptly notified.

No specific action or written response is required by this notice. If you have any questions regarding this information notice, please contact the appropriate NRC regional office or this office.



Richard E. Cunningham, Director
Division of Industrial and
Medical Nuclear Safety
Office of Nuclear Material Safety
and Safeguards

Technical Contact: Don Mackenzie, NMSS
(301) 427-4052

Attachments:

1. Information Notice No. 86-67
2. List of Recently Issued NRC Information Notices

UNITED STATES
NUCLEAR REGULATORY COMMISSION
OFFICE OF INSPECTION AND ENFORCEMENT
WASHINGTON, D.C. 20555

AUGUST 15, 1986

IE INFORMATION NOTICE NO. 86-67: PORTABLE MOISTURE/DENSITY GAUGES: RECENT INCIDENTS AND COMMON VIOLATIONS OF REQUIREMENTS FOR USE, TRANSPORTATION, AND STORAGE

Addressees:

All NRC licensees authorized to possess, use, transport, and store sealed sources contained in portable gauges used to measure the moisture content and/or density of construction materials.

Purpose:

This notice is intended to bring to the attention of licensees the recent increase in incidents involving the use, transportation, and storage of portable gauges and the number of common violations identified during NRC inspections. It is expected that recipients will review the information for applicability to their facilities and consider actions, if appropriate, to preclude similar problems from occurring at their facilities. However, suggestions contained in this information notice do not constitute NRC requirements; therefore, no specific action or written response is required.

Description of Circumstances:

An abnormally high number of incidents have occurred recently where portable moisture/density gauges have been damaged at temporary job sites by heavy construction equipment or where the gauges have been lost or stolen from licensee vehicles during transportation. Gauges damaged at construction sites were left unattended. Gauges lost or stolen from vehicles were not secured to the vehicle or were stolen while left unattended by the users.

Inspections initiated by these incidents and routine inspections that have been performed reveal common violations of NRC requirements. These violations include failure to:

- (1) have a shipping paper in the transport vehicle
- (2) transport gauges in authorized packages

- (3) maintain records of tests performed on transport cases and on sealed sources
- (4) use authorized and/or qualified users
- (5) use authorized storage locations
- (6) conduct leak tests and physical inventories to conduct those tests and inventories within the required time interval
- (7) wear film or TLD badges or estimate doses to personnel who had lost their badges or evaluate and report possible overexposures

Discussion:


The cause of these incidents, invariably, is the failure of the gauge users to secure and maintain control over the gauges.

10 CFR 20.207 requires that licensed material (in these cases the sealed sources in the gauges) must be under the constant surveillance and immediate control of the licensee or must be secured in storage when in an unrestricted area. An unrestricted area is defined in 10 CFR 20.3(a)(17) as any area to which access is not controlled by the licensee for purposes of protection of individuals from exposure to radiation and radioactive materials and any area used for residential quarters. A construction site is usually an unrestricted area; so the licensee must constantly control access to the gauge while in storage or during use at the site.

NRC licensees transporting portable gauges are subject to 10 CFR 71. 10 CFR 71.5(a) incorporates certain regulations (49 CFR 170-189) of the Department of Transportation (DOT) to which these licensees are also subject. Licensees who transport gauges to and from temporary job sites in licensee or private vehicles are acting as private shippers and, as such, must comply with the DOT regulations governing shippers. 49 CFR 177.842(d) requires that packages containing radioactive material (i.e., the gauge in its case and containing radioactive sealed sources) must be blocked and braced to prevent movement of the package during transportation. For pickup trucks, this requirement is usually met when the gauge, in its case, is chained or tied to the bed of the truck.

Licensees are reminded that they must use, transport, and store the gauges in accordance with the conditions of their NRC license, other commitments made to the NRC, and applicable regulations. A discussion of other requirements and license conditions commonly violated is attached.

No specific action or written response is required by this notice. If you have any questions regarding this information notice, please contact the Regional Administrator of the appropriate NRC regional office or this office.



James G. Partlow, Director
Division of Inspection Programs
Office of Inspection and Enforcement

Technical Contact: J. R. Metzger, IE
(301) 492-4947

Attachments:

1. Other Common Violations
2. List of Recently Issued IE Information Notices

OTHER COMMON VIOLATIONS

49 CFR 172.201, 172.202, 172.203(d), 172.204, and 177.817(e) specify the contents of the shipping paper and the location in a vehicle where the shipping paper must be stored. Shipping papers must not be stored in or on the case containing the gauge or in the glove compartment of the vehicle.

49 CFR 173.475 requires that before each shipment the shipper ensures by examination or appropriate test that the proper shipping case or box is used, that the case is in unimpaired physical condition, and that each closure device on the case is properly installed, secured, and free of defects.

49 CFR 172.301, 172.304, 172.308, 172.310, 172.403, and 178.350 specify the labeling and marking of the case used for shipping. Vehicles transporting one or more gauges usually do not require placards (49 CFR 172.504, Table 1).

49 CFR 173.415(a) and 173.476(a) require that shippers maintain on file results of tests conducted on shipping cases and on the sealed sources contained in the gauges. Licensees can usually obtain these test results from the manufacturer of the gauges.

Gauges must be used only by properly authorized and trained individuals. A common license condition usually names individuals who are authorized to use the gauges or who must be physically present to supervise their use. Occasionally, licensees are allowed by a license condition to name authorized users. Authorized users must have successfully completed an approved training course given by the manufacturer or a consulting firm. Certain licensees are authorized to provide in-house training. The Radiation Protection Officer (RPO) also may be named on the license, but is always specified in documents submitted to the NRC. When this individual is no longer employed as RPO, the licensee must request an amendment to the license to name a replacement.

A specific license condition or statement in a submitted document specifies the temporary and permanent storage facilities for the gauges. Private residences (including basements and garages) are usually not allowed as storage facilities. Vehicles used to temporarily store gauges overnight at private residences must not be used by individuals who are not authorized users for purposes other than those authorized on the license. For example, friends or relatives of an authorized user must not use a vehicle containing a gauge unless those individuals are performing a purpose authorized on the license as authorized users.

A specific license condition will specify the interval between leak tests. This interval is usually 6 months and may be greater on some licenses for gauges that are stored and not being used. Many licensees also are required by license condition to conduct a physical inventory at 6-month intervals.

Licensees are usually committed by a referenced document to wear film or TLD badges when using or transporting the gauges. When not worn by the users, the badges must not be stored near or on the gauges. For lost badges, the licensee is required by 10 CFR 20.201 to estimate the user's dose for the period for which the badge was lost. For film or TLD badge reports indicating that doses greater than the limit in 10 CFR 20.101 may have been received, the licensee must evaluate the reading (10 CFR 20.201) to determine if it has been caused by an actual exposure to the user. This evaluation includes, at a minimum, questioning the user about the use of the gauge and badge and having the film or TLD badge processor reexamine the badge. Although, true overexposures of gauge users are rare, the occurrences must be reported to the NRC as required by 10 CFR 20.405.

LIST OF RECENTLY ISSUED
 IE INFORMATION NOTICES

Information Notice No.	Subject	Date of Issue	Issued to
86-66	Potential For Failure Of Replacement AC Coils Supplied By The Westinghouse Electric Corporation For Use In Class 1E Motor Starters And Contractors	8/15/86	All power reactor facilities holding an OL or CP
86-65	Malfunctions Of ITT Barton Model 580 Series Switches During Requalification Testing	8/14/86	All power reactor facilities holding an OL or CP
86-64	Deficiencies In Upgrade Programs For Plant Emergency Operating Procedures	8/14/86	All power reactor facilities holding an OL or CP
86-63	Loss Of Safety Injection Capability	8/6/86	All PWR facilities holding an OL or CP
86-62	Potential Problems In Westinghouse Molded Case Circuit Breakers Equipped With A Shunt Trip	7/31/86	All power reactor facilities holding an OL or CP
86-61	Failure Of Auxiliary Feed-water Manual Isolated Valve	7/28/86	All power reactor facilities holding a CP
86-60	Unanalyzed Post-LOCA Release Paths	7/28/86	All power reactor facilities holding an OL or CP
86-31 Sup. 1	Unauthorized Transfer And Loss Of Control Of Industrial Nuclear Gauges	7/14/86	All NRC general licensees that possess and use industrial nuclear gauges
86-59	Increased Monitoring Of Certain Patients With Implanted Coratomic, Inc. Model C-100 and C-101 Nuclear-Powered Cardiac Pacemakers	7/14/86	All NRC licensees authorized to use nuclear-powered cardiac pacemakers

OL = Operating License
 CP = Construction Permit

LIST OF RECENTLY ISSUED
INFORMATION NOTICES 1987

Information Notice No.	Subject	Date of Issuance	Issued to
87-54	Emergency Response Exercises	10/23/87	All holders of OLs or CPs for nuclear power reactors.
87-53	Auxiliary Feedwater Pump Trips Resulting from Low Suction Pressure	10/20/87	All holders of OLs or CPs for nuclear power reactors.
87-52	Insulation Breakdown of Silicone Rubber-Insulated Single Conductor Cables During High Potential Testing	10/16/87	All holders of OLs or CPs for nuclear power reactors.
87-51	Failure of Low Pressure Safety Injection Pump Due to Seal Problems	10/13/87	All nuclear power reactor facilities holding an OL or CP.
87-50	Potential LOCA at High- and Low-Pressure Interfaces from Fire Damage	10/9/87	All nuclear power reactor facilities holding an OL or CP.
87-49	Deficiencies in Outside Containment Flooding Protection	10/9/87	All nuclear power reactor facilities holding an OL or CP.
87-48	Information Concerning the Use of Anaerobic Adhesive/ Sealants	10/9/87	All nuclear power reactor facilities holding an OL or CP.
87-47	Transportation of Radiography Devices	10/5/87	All NRC licensees authorized to manufacture, distribute and/or operate radiographic exposure devices and/or source changers.
87-46	Undetected Loss of Reactor Coolant	9/30/87	All PWR facilities holding on OL or CP.

OL = Operating License
CP = Construction Permit

NRC licensees transporting portable gauges are subject to the regulations in 10 CFR Part 71. Section 71.5(a) incorporates certain regulations (49 CFR 170-189) of the Department of Transportation (DOT), to which these licensees are also subject. Licensees who transport gauges to and from temporary job sites in private vehicles are shippers acting as private carriers, and as such, must comply with the DOT regulations governing both shippers and carriers. Title 49 CFR 177.842(d) requires that packages containing radioactive material (i.e., the gauge in its case) must be blocked and braced to prevent movement of the package during transportation (see Information Notice 87-31, attached). For pickup trucks, this requirement is usually met when the gauge is secured within its case, and the case is secured and locked to the bed of the truck.

The NRC has also noted several incidents where gauges were damaged in transportation accidents or by being run over by construction vehicles. The most prevalent cause of such incidents is the failure by the licensee to secure and lock the cargo door, combined with a failure to properly secure the device/case to the bed of the vehicle. Licensees are reminded that they should use care to avoid these types of incidents. If a gauge is damaged, the area around the gauge should be restricted, and the licensee's radiation safety officer and NRC should be promptly notified.

No specific action or written response is required by this notice. If you have any questions regarding this information notice, please contact the appropriate NRC regional office or this office.

Richard E. Cunningham, Director
 Division of Industrial and
 Medical Nuclear Safety
 Office of Nuclear Material Safety
 and Safeguards

Technical Contact: Don Mackenzie, NMSS
 (301) 427-4052

Attachments:

1. Information Notice No. 86-67
2. List of Recently Issued NRC Information Notices
 (*See previous concurrence)

OFC:	Editor*	IMOB*	IMAB*	:SGTB *	SGOB*	FCOB*	FCOB*
NAME:	EKraus	JRMetzger/jp	MLamastra	CEWilliams	AGrella	DACool	JWNHickey
DATE:	/ / 87	/ / 87	/ / 87	/ / 87	/ / 87	/ / 87	/ / 87
OFC:	NRR	IMNS	IMNS				
NAME:	Cross	GLSjoblom	RECunningham				
DATE:	10/23/87	187	10/22/87				

NRC licensees transporting portable gauges are subject to the regulations in 10 CFR Part 71. Section 71.5(a) incorporates certain regulations (49 CFR 170-189) of the Department of Transportation (DOT), to which these licensees are also subject. Licensees who transport gauges to and from temporary job sites in licensee or private vehicles are acting as private shippers, and as such, must comply with the DOT regulations governing shippers. Title 49 CFR 177.842(d) requires that packages containing radioactive material (i.e., the gauge in its case) must be blocked and braced to prevent movement of the package during transportation (see Information Notice 87-31, attached). For pickup trucks, this requirement is usually met when the gauge, in its case, is secured and locked to the bed of the truck.

The NRC has also noted several incidents where gauges were damaged in transportation accidents or by being run over by construction vehicles. Licensees are reminded that they should use care to avoid these types of incidents. If a gauge is damaged, the area around the gauge should be restricted, and the licensee's radiation safety officer and NRC should be promptly notified.

No specific action or written response is required by this notice. If you have any questions regarding this information notice, please contact the appropriate NRC regional office or this office.

Richard E. Cunningham, Director
 Division of Industrial and
 Medical Nuclear Safety
 Office of Nuclear Material Safety
 and Safeguards

Technical Contact: Don Mackenzie, NMSS
 (301) 427-4052

Attachments:

- Information Notice No. 87-31
- List of Recently Issued NRC Information Notices

*correct subject
 to changes on
 attached routing.*

OFC: Editor	IMOB	IMAB	SGTB	SGOB	FCOB	FCOB
NAME: EKraus	JRM/zger/jp	MLamastra	CEW/Items Mackenzie	AGrella	DACool	JWRHickey
DATE: 10/5/87	10/8/87	10/8/87	10/16/87	10/20/87	10/18/87	10/18/87
OFC: NRR	IMNS	IMNS				
NAME: CRossi	GLSjoblom	RECunningham				
DATE: / / 87	/ / 87	/ / 87				