

# Region I NMSS Licensee Event Report

Licensee	Agilent		
Event Description	Notice of General License Device with High Removable		
Activity	026		

License #	07-2872	Docket #	03002988	MLER-RI	2003-038
Event Date	May 03	Report Date	5-29-03		

**1. REPORTING REQUIREMENT**

<input type="checkbox"/> 10 CFR 20.2201 Theft or Loss	<input type="checkbox"/> 10 CFR 35.33 Misadministration
<input type="checkbox"/> 10 CFR 20.2203 30 Day Report	<input type="checkbox"/> License Condition
<input type="checkbox"/> 10 CFR 30.50 Report	
<input checked="" type="checkbox"/> Other <u>10 CFR 31.5</u>	

**2. REGION I RESPONSE**

<input type="checkbox"/> Immediate Site Inspection	Inspector/Date	
<input type="checkbox"/> Special Inspection	Inspector/Date	
<input type="checkbox"/> Telephone Inquiry	Inspector/Date	
<input type="checkbox"/> Preliminary Report	<input type="checkbox"/> Daily Report	
<input checked="" type="checkbox"/> Information Entered in RI Log	<input checked="" type="checkbox"/> Review at next inspection	
Report referred to: _____		


**3. REPORT EVALUATION**

<input checked="" type="checkbox"/> Description of Event	<input checked="" type="checkbox"/> Corrective Actions
<input checked="" type="checkbox"/> Levels of RAM Involved	<input type="checkbox"/> Calculations Adequate
<input checked="" type="checkbox"/> Cause of Event	<input type="checkbox"/> Additional Information Requested from Licensee

**4. SPECIAL INSTRUCTIONS OR COMMENTS**

Public  Non-Public

<b>Completed by:</b>	Initials/Date Betty Ullrich	<b>Date</b>	Initials/Date 6/12/03
<b>Reviewed by:</b>	Initials/Date [Signature]	<b>Date</b>	Initials/Date 6/17/2003

 **Agilent Technologies**  
2850 Centerville Road  
Wilmington, DE 19808

RECEIVED  
REGION 1

2003 MAY 29 PM 12: 21

Division of Nuclear Materials Safety  
U.S. Nuclear Regulatory Commission  
Region I  
475 Allendale Road,  
King of Prussia, PA 19406

Re: Notice of General License device with high removable activity (NRC License 07-28762-02G)

May 28, 2003

Dear Sir or Madam,

Per the requirements of 10 CFR 31.5, I am notifying your office of the following devices returned to us by customers from which we have obtained removable activity wipes in excess of 0.005  $\mu\text{Ci}$ . Cell serial # K3491 was from a U.S. customer and the others from overseas. The cells were evaluated by support engineering and their disposition is identified below. I have contacted the customer for further information and will notify the cell supplier of the high wipes.

Serial #	Model #	Activity ( $\mu\text{Ci}$ )	Disposition	Comments
K3491	19233	0.056	Scrapped	Cell was 6 years old, source dark and very dirty
K3546	G1533A	0.011	Scrapped	Cell was 7 years old and rusted
K3797	G1533A	0.037	Scrapped	Cell was 6 years old and the seal was broken
L0870	19233	0.018	Scrapped	Cell was 18 years old, source dark and dirty
F1196	19233	0.013	Scrapped	Cell was 12 years old and rusted

Please contact me at 302-633-8262 if there are any questions.

Thank you,



David S. Bennett  
Radiation Safety Officer