

Region I NMSS Licensee Event Report

Licensee	Agilent
Event Description	Notice of General License Device with High Removable
Activity	

License #	07-28762- ^{02G}	Docket #	03002988	MLER-RI	2003-046
Event Date	JUNE 03	Report Date	06/25/03		

1. REPORTING REQUIREMENT

<input type="checkbox"/> 10 CFR 20.2201 Theft or Loss <input type="checkbox"/> 10 CFR 20.2203 30 Day Report <input type="checkbox"/> 10 CFR 30.50 Report <input checked="" type="checkbox"/> Other <u>31.5</u>	<input type="checkbox"/> 10 CFR 35.33 Misadministration <input type="checkbox"/> License Condition
---	---

2. REGION I RESPONSE

<input type="checkbox"/> Immediate Site Inspection <input type="checkbox"/> Special Inspection <input type="checkbox"/> Telephone Inquiry <input type="checkbox"/> Preliminary Report <input checked="" type="checkbox"/> Information Entered in RI Log <input type="checkbox"/> Report referred to: _____	<table border="1" style="width: 100%; border-collapse: collapse;"> <tr><td style="width: 50%;">Inspector/Date</td><td></td></tr> <tr><td>Inspector/Date</td><td></td></tr> <tr><td>Inspector/Date</td><td></td></tr> </table> <input type="checkbox"/> Daily Report <input checked="" type="checkbox"/> Review at next inspection	Inspector/Date		Inspector/Date		Inspector/Date	
Inspector/Date							
Inspector/Date							
Inspector/Date							


3. REPORT EVALUATION

<input checked="" type="checkbox"/> Description of Event <input checked="" type="checkbox"/> Levels of RAM Involved <input type="checkbox"/> Cause of Event	<input checked="" type="checkbox"/> Corrective Actions <input type="checkbox"/> Calculations Adequate <input type="checkbox"/> Additional Information Requested from Licensee
---	---

4. SPECIAL INSTRUCTIONS OR COMMENTS

Public for 7/3/03 Non-Public

Completed by:	<table style="width: 100%;"> <tr><td style="width: 50%;">Initials/Date</td><td style="width: 50%;">Date</td></tr> <tr><td style="text-align: center;">[Signature]</td><td style="text-align: center;">7/3/03</td></tr> </table>	Initials/Date	Date	[Signature]	7/3/03	Date	<table style="width: 100%;"> <tr><td style="width: 50%;">Initials/Date</td><td style="width: 50%;">Date</td></tr> <tr><td></td><td style="text-align: center;">7/11/2003</td></tr> </table>	Initials/Date	Date		7/11/2003
Initials/Date	Date										
[Signature]	7/3/03										
Initials/Date	Date										
	7/11/2003										
Reviewed by:	<table style="width: 100%;"> <tr><td style="width: 50%;">Initials/Date</td><td style="width: 50%;">Date</td></tr> <tr><td style="text-align: center;">[Signature]</td><td></td></tr> </table>	Initials/Date	Date	[Signature]		Date					
Initials/Date	Date										
[Signature]											

 **Agilent Technologies**
2850 Centerville Road
Wilmington, DE 19808

RECEIVED
REGION 1

2003 JUN 26 PM 1:02

Division of Nuclear Materials Safety
U.S. Nuclear Regulatory Commission
Region I
475 Allendale Road,
King of Prussia, PA 19406

Re: Notice of General License device with high removable activity (NRC License 07-28762-02G)

June 25, 2003


Dear Sir or Madam,

Per the requirements of 10 CFR 31.5, I am notifying your office of a general license device returned to our site by a customer from which we have obtained a removable activity wipe in excess of 0.005 μCi . The cell was evaluated by support engineering and its disposition is identified below. I will notify the cell supplier of the high wipe.

<u>Serial #</u>	<u>Model #</u>	<u>Activity (μCi)</u>	<u>Disposition</u>	<u>Comment</u>
U0771	G22397A	0.035	Scrapped	Cell was 5 years old

Please contact me at 302-633-8262 if there are any questions.

Thank you,


David S. Bennett
Radiation Safety Officer