

**Region I  
NMSS Licensee Event Report**

Licensee	<i>Agilent</i>		
Event Description	Notice of General License Device with High Removable		
	Activity		
License #	07-28762-6	Docket #	03002988
Event Date	JUNE 03	Report Date	06/25/03
		MLER-RI	2003-046

**1. REPORTING REQUIREMENT**

<input type="checkbox"/> 10 CFR 20.2201 Theft or Loss <input type="checkbox"/> 10 CFR 20.2203 30 Day Report <input type="checkbox"/> 10 CFR 30.50 Report <input checked="" type="checkbox"/> Other <u>31.5</u>	<input type="checkbox"/> 10 CFR 35.33 Misadministration <input type="checkbox"/> License Condition
---	---

**2. REGION I RESPONSE**

<input type="checkbox"/> Immediate Site Inspection <input type="checkbox"/> Special Inspection <input type="checkbox"/> Telephone Inquiry <input type="checkbox"/> Preliminary Report <input checked="" type="checkbox"/> Information Entered in RI Log <input type="checkbox"/> Report referred to: _____	<table border="1" style="width: 100%; border-collapse: collapse;"><tr><td style="width: 50%;">Inspector/Date</td><td></td></tr><tr><td>Inspector/Date</td><td></td></tr><tr><td>Inspector/Date</td><td></td></tr></table> <input type="checkbox"/> Daily Report <input checked="" type="checkbox"/> Review at next inspection	Inspector/Date		Inspector/Date		Inspector/Date	
Inspector/Date							
Inspector/Date							
Inspector/Date							

**3. REPORT EVALUATION**

<input checked="" type="checkbox"/> Description of Event <input checked="" type="checkbox"/> Levels of RAM Involved <input type="checkbox"/> Cause of Event	<input checked="" type="checkbox"/> Corrective Actions <input type="checkbox"/> Calculations Adequate <input type="checkbox"/> Additional Information Requested from Licensee
---	---

**4. SPECIAL INSTRUCTIONS OR COMMENTS**


Public

☒ *for 7/3/03*

Non-Public

Completed by:

Initials/Date

*Boyle 7/3/03*

Date


Initials/Date

Reviewed by:

*Joe D / km*

Date

*7/11/2003*

 **Agilent Technologies**  
2850 Centerville Road  
Wilmington, DE 19808

RECEIVED  
REGION 1

2003 JUN 26 PM 1:02

Division of Nuclear Materials Safety  
U.S. Nuclear Regulatory Commission  
Region I  
475 Allendale Road,  
King of Prussia, PA 19406

Re: Notice of General License device with high removable activity (NRC License 07-28762-02G)

June 25, 2003


Dear Sir or Madam,

Per the requirements of 10 CFR 31.5, I am notifying your office of a general license device returned to our site by a customer from which we have obtained a removable activity wipe in excess of 0.005  $\mu\text{Ci}$ . The cell was evaluated by support engineering and its disposition is identified below. I will notify the cell supplier of the high wipe.

Serial #	Model #	Activity ( $\mu\text{Ci}$ )	Disposition	Comment
U0771	G22397A	0.035	Scrapped	Cell was 5 years old

Please contact me at 302-633-8262 if there are any questions.

Thank you,

  
David S. Bennett  
Radiation Safety Officer