

Region I
NMSS Licensee Event Report

Licensee	AGILENT		
Event Description	NOTICE OF GEN LICENSE DEVICE WITH HIGH REMOVABLE ACTIVITY		
	026		
License #	07-28762-	Docket #	03002988
Event Date	JULY 03	Report Date	7-11-03
		MLER-RI	2003-049

1. REPORTING REQUIREMENT

<input type="checkbox"/>	10 CFR 20.2201 Theft or Loss	<input type="checkbox"/>	10 CFR 35.33 Misadministration
<input type="checkbox"/>	10 CFR 20.2203 30 Day Report	<input type="checkbox"/>	License Condition
<input type="checkbox"/>	10 CFR 30.50 Report		
<input checked="" type="checkbox"/>	Other 10 CFR 31.5		

2. REGION I RESPONSE

<input type="checkbox"/>	Immediate Site Inspection	Inspector/Date	
<input type="checkbox"/>	Special Inspection	Inspector/Date	
<input type="checkbox"/>	Telephone Inquiry	Inspector/Date	
<input type="checkbox"/>	Preliminary Report	<input type="checkbox"/>	Daily Report
<input checked="" type="checkbox"/>	Information Entered in RI Log	<input checked="" type="checkbox"/>	Review at next inspection
<input type="checkbox"/>	Report referred to: _____		

3. REPORT EVALUATION

<input checked="" type="checkbox"/>	Description of Event	<input checked="" type="checkbox"/>	Corrective Actions
<input checked="" type="checkbox"/>	Levels of RAM Involved	<input type="checkbox"/>	Calculations Adequate
<input checked="" type="checkbox"/>	Cause of Event	<input type="checkbox"/>	Additional Information Requested from Licensee

4. SPECIAL INSTRUCTIONS OR COMMENTS


Public

☒

Non-Public

☐

Completed by:	Initials/Date Beth Ullrich	Date 9/10/03
Reviewed by:	Initials/Date [Signature]	Date 9.10.2003

 **Agilent Technologies**
2850 Centerville Road
Wilmington, DE 19808

RECEIVED
REGION I

2003 JUL 14 PM 1:49

Division of Nuclear Materials Safety
U.S. Nuclear Regulatory Commission
Region I
475 Allendale Road,
King of Prussia, PA 19406

Re: Notice of General License device with high removable activity (NRC License 07-28762-02G)

July 11, 2003


Dear Sir or Madam,

Per the requirements of 10 CFR 31.5, I am notifying your office of the following devices returned to us by customers from which we have obtained removable activity wipes in excess of 0.005 μCi . Cell serial #F5728 was from a U.S. customer and N0282 was returned from overseas. The cells were evaluated by support engineering and their disposition is identified below. I have contacted the customer for further information and will notify the cell supplier of the high wipes.

Serial #	Model #	Activity (μCi)	Disposition	Comments
N0282	G1533A	0.121	Scrapped	Outside body rusted, source was dark and dirty Cell was 5 years old
F5728	19233	0.006	Scrapped	Source was unusual color inside. Cell was 7 years old.

Please contact me at 302-633-8262 if there are any questions.

Thank you,


David S. Bennett
Radiation Safety Officer