

S O U T H W E S T R E S E A R C H I N S T I T U T E

Department of Quality Assurance
Calibration Laboratory

CERTIFICATE OF CALIBRATION

Issued to: DIV20 B51 KRISTI MEYER

Device No: 2552

Manufacturer: BRAINARD KILMAN

Model: E400

Nomenclature: PRESSURE INDICATOR

Serial Number: E400-341

SwRI No: NONE

Cal interval 12 Mo.

Remarks

Transducer E124 S/N 206

Accuracy: +/-1 PSI

Procedure: MFGR

ENVIRONMENT

Temperature: 0 Humidity: 0 Location: SWRI BLDG. 145

CONCLUSION

Tolerance/Remarks: Received into the system, introduced or reactivated

Standard Reading:	Unit Under Test:
150.0 psi	150.0 psi
100.0 psi	99.6 psi
50.0 psi	49.8 psi
30.0 psi	30.0 psi
0.0 psi	0.0 psi

Calibration was in accord with requirements of MIL-STD-45662A. Measurements are traceable to the National Institute of Standards and Technology. Inspection and test data are on file and available for inspection.

Signed _____

Calibration Date: 08/02/93

Cal interval: 12 Months

Record Number: 00011881

Next Calibration Due: 08/02/94

S O U T H W E S T R E S E A R C H I N S T I T U T E

Department of Quality Assurance
Calibration Laboratory

Device Serial No: E400-341

Calibration Date: 08/02/93

STANDARDS

Standard No: 1554 Manufacturer: AMETEK

Model: RK-300

Nomenclature: DEADWEIGHT TESTER

Serial No: 82167

Cal.Due: 03/26/94

Cal.Rec.No: 00010846