

SOUTHWEST RESEARCH INSTITUTE
Department of Quality Assurance
Calibration Laboratory • 522-5215

WORK ORDER

CERTIFICATE # 25032 ASSET # 002512 DATE 07/11/97

ITEM DATA:
Manufacturer Kossow Model ASTM 82C
Description 1/10/1000/10 Serial # 924500
Accessories _____

ACTION REQUESTED Cal

CUSTODIAN Dr. P. David Dun

Turned in by: _____ Phone 1070

CHARGE # 20.5700561 Date Required _____

INSTRUMENT USED ON: NUCLEAR DOD NASA GLP SPPE
 OTHER _____

COPY OF CALIBRATION CERTIFICATE Yes No

CONDITION RECEIVED: _____ Out of tolerance, repaired to specifications
_____ In tolerance, minor adjustments/repairs made
 In tolerance, no adjustments/repairs
_____ Out of tolerance, adjusted to specifications
_____ Received into system, introduced or reactivated
_____ Calibration interval
_____ Reliability code

ACTION TAKEN: (Calibration/Repair/Parts) Calibrated by Lawrence
445
- 0.02°C 0.0°C

CAL ENVIRONMENT:
Temperature 77°F Humidity 27%RH

CALIBRATED/REPAIRED:
By R. [Signature] Cal Procedure WI-9-30-7103
Date 15 Apr 97 Accuracy +/- 1°C
Cal Interval 12 mos Time to complete:
Next Cal due 15 Apr 98 Cal 2.0 Repair _____
Standards used (Asset#) 106, 5174

DATE COMPLETED 15 Apr 97
DATE PICKED UP 11/7/97 PICKED UP BY [Signature]

25032