

SOUTHWEST RESEARCH INSTITUTE  
Department of Quality Assurance  
Calibration Laboratory • 522-5215

WORK ORDER

CERTIFICATE # 25032 ASSET # 002512 DATE 07/11/97

ITEM DATA:  
Manufacturer Kossow Model ASTM 82C  
Description 1/10/1000tor Serial # 924500  
Accessories \_\_\_\_\_

ACTION REQUESTED Cal

CUSTODIAN Dr. P. David Dun

Turned in by: \_\_\_\_\_ Phone 1070

CHARGE # 20.5700561 Date Required \_\_\_\_\_

INSTRUMENT USED ON:  NUCLEAR  DOD  NASA  GLP  SPPE  
 OTHER \_\_\_\_\_

COPY OF CALIBRATION CERTIFICATE  Yes  No

CONDITION RECEIVED: \_\_\_\_\_ Out of tolerance, repaired to specifications  
\_\_\_\_\_ In tolerance, minor adjustments/repairs made  
 In tolerance, no adjustments/repairs  
\_\_\_\_\_ Out of tolerance, adjusted to specifications  
\_\_\_\_\_ Received into system, introduced or reactivated  
\_\_\_\_\_ Calibration interval  
\_\_\_\_\_ Reliability code

ACTION TAKEN: (Calibration/Repair/Parts) Calibrated by Lawrence  
445  
- 0.02°C 0.0°C

CAL ENVIRONMENT:  
Temperature 77°F Humidity 27%RH

CALIBRATED/REPAIRED:  
By R. [Signature] Cal Procedure WI-9-30-7103  
Date 15 Apr 97 Accuracy +/- 1°C  
Cal Interval 12 mos Time to complete:  
Next Cal due 15 Apr 98 Cal 2.0 Repair \_\_\_\_\_  
Standards used (Asset#) 106, 5174

DATE COMPLETED 15 Apr 97  
DATE PICKED UP 11/7/97 PICKED UP BY [Signature]

25032