

SOUTHWEST RESEARCH INSTITUTE

Department of Quality Assurance
Calibration Laboratory

CERTIFICATE OF CALIBRATION
08/04/95

Issued to: DARRELL DUNN DIV20 ,B57
Manufacturer: PARR
Nomenclature: TEMPERATURE CONTROLLER
Serial Number: 0690A
Notes:

Asset Number: 002026
Model Number: 4841
SwRI/Div. I.D. #:

ENVIRONMENTAL CONDITIONS

Temperature: 70.0F

Relative Humidity: 38%

CALIBRATION INFORMATION

Procedure Number: MFGR
Remarks: CAL WITH "J" THERMOCOUPLE TO +/-2DEG C
LIMITED CALIBRATION - TEMP. ONLY

Accuracy: +/-2DEG.C
Received IN Tolerance

Calibration was in accordance with requirements of MIL-STD-45662A.
Measurements are traceable to the National Institute of Standards and Technology. Inspection and test data are on file and available for inspection.

STANDARDS USED FOR CERTIFICATION

Asset #	Serial #	Mfg	Model #	Nomenclature	Cal Date	Int.	Cal Due
000182	5200003	FLUKE	5700A	CALIBRATOR	06/07/95	3	09/07/95
002868	311210	KAYE	X0251	ICE POINT REFERENCE	01/03/95	12	01/03/96

Certified by : 

Certificate#: 17596

Calibration Date: 08/04/95
Interval: 6 months
Next Calibration Due: 02/04/96