

SOUTHWEST RESEARCH INSTITUTE  
Department of Quality Assurance  
Calibration Laboratory • 522-5215

**WORK ORDER**

CERTIFICATE # 25036 ASSET # 107046 DATE 07 Apr 97

ITEM DATA:  
Manufacturer Fisher Model 14-795-100  
Description TEMPERATURE Serial # 0505002  
Accessories \_\_\_\_\_

ACTION REQUESTED 11

CUSTODIAN John, Dept. Chem

Turned in by: \_\_\_\_\_ Phone 7050

CHARGE # 705100501 Date Required \_\_\_\_\_

INSTRUMENT USED ON:  NUCLEAR  DOD  NASA  GLP  SPPE  
 OTHER \_\_\_\_\_

COPY OF CALIBRATION CERTIFICATE  Yes  No

CONDITION RECEIVED: \_\_\_\_\_ Out of tolerance, repaired to specifications  
\_\_\_\_\_ In tolerance, minor adjustments/repairs made  
 In tolerance, no adjustments/repairs  
\_\_\_\_\_ Out of tolerance, adjusted to specifications  
\_\_\_\_\_ Received into system, introduced or reactivated  
\_\_\_\_\_ Calibration interval  
\_\_\_\_\_ Reliability code

ACTION TAKEN: (Calibration/Repair/Parts) Calibrated per procedure  
std 405  
-0.03°C 0.02

**25036**

CAL ENVIRONMENT:  
Temperature 77 °F Humidity 27 %RH

CALIBRATED/REPAIRED:  
By [Signature] Cal Procedure WI-9-30-7163  
Date 15 Apr 97 Accuracy ±1°C  
Cal Interval 12 mos Time to complete:  
Next Cal due 15 Apr 98 Cal 2.0 Repair \_\_\_\_\_  
Standards used (Asset#) 106, 5174

DATE COMPLETED 15 Apr 97  
DATE PICKED UP 4/17/97 PICKED UP BY [Signature]