

SOUTHWEST RESEARCH INSTITUTE
Department of Quality Assurance
Calibration Laboratory • 522-5215

WORK ORDER

CERTIFICATE # 25036 ASSET # 107046 DATE 07 Apr 97

ITEM DATA:
Manufacturer Fisher Model 14-795-100
Description TEMPERATURE Serial # 0585002
Accessories _____

ACTION REQUESTED 11

CUSTODIAN John, Dept. Chem

Turned in by: _____ Phone 7050

CHARGE # 705100501 Date Required _____

INSTRUMENT USED ON: NUCLEAR DOD NASA GLP SPPE
 OTHER _____

COPY OF CALIBRATION CERTIFICATE Yes No

CONDITION RECEIVED: _____ Out of tolerance, repaired to specifications
_____ In tolerance, minor adjustments/repairs made
 In tolerance, no adjustments/repairs
_____ Out of tolerance, adjusted to specifications
_____ Received into system, introduced or reactivated
_____ Calibration interval
_____ Reliability code

ACTION TAKEN: (Calibration/Repair/Parts) Calibrated per procedure
std 405
-0.03°C 0.02

25036

CAL ENVIRONMENT:
Temperature 77 °F Humidity 27 %RH

CALIBRATED/REPAIRED:
By [Signature] Cal Procedure WI-9-30-7163
Date 15 Apr 97 Accuracy ±1°C
Cal Interval 12 mos Time to complete:
Next Cal due 15 Apr 98 Cal 2.0 Repair _____
Standards used (Asset#) 106, 5174

DATE COMPLETED 15 Apr 97
DATE PICKED UP 4/17/97 PICKED UP BY [Signature]