

SOUTHWEST RESEARCH INSTITUTE  
Department of Quality Assurance  
Calibration Laboratory • 522-5215

WORK ORDER

CERTIFICATE # 25035 ASSET # 002046 DATE 01 Apr 97

ITEM DATA:  
Manufacturer Fisherbrand Model 141-903-103  
Description Flow Indicator Serial # 032707  
Accessories \_\_\_\_\_

ACTION REQUESTED Cal

CUSTODIAN Dr. W. D. Damm

Turned in by: \_\_\_\_\_ Phone 1000

CHARGE # 701 5700561 Date Required \_\_\_\_\_

INSTRUMENT USED ON:  NUCLEAR  DOD  NASA  GLP  SPPE  
 OTHER \_\_\_\_\_

COPY OF CALIBRATION CERTIFICATE  Yes  No

CONDITION RECEIVED: \_\_\_\_\_ Out of tolerance, repaired to specifications  
\_\_\_\_\_ In tolerance, minor adjustments/repairs made  
 In tolerance, no adjustments/repairs  
\_\_\_\_\_ Out of tolerance, adjusted to specifications  
\_\_\_\_\_ Received into system, introduced or reactivated  
\_\_\_\_\_ Calibration interval  
\_\_\_\_\_ Reliability code

ACTION TAKEN: (Calibration/Repair/Parts) Calibrated Per Researcher  
std 447  
0.00 °C 0.0 °C

CAL ENVIRONMENT:  
Temperature 77 °F Humidity 27 %RH

CALIBRATED/REPAIRED:  
By [Signature] Cal Procedure WI-9-30-7H03  
Date 15 Apr 97 Accuracy ± 1%  
Cal Interval 12 mths Time to complete:  
Next Cal due 15 Apr 98 Cal 2.0 Repair \_\_\_\_\_  
Standards used (Asset#) 106, 5174

DATE COMPLETED 15 Apr 97  
DATE PICKED UP 4/17/97 PICKED UP BY [Signature]

25035