

SOUTHWEST RESEARCH INSTITUTE  
Department of Quality Assurance  
Calibration Laboratory • 522-5215

WORK ORDER

CERTIFICATE # 24193 ASSET # 002105 DATE 03 Feb 97

ITEM DATA:

Manufacturer FOWLER Model S2008-008  
Description 8" CALIPER Serial # 20-8C-1  
Accessories CASE

ACTION REQUESTED CAL

CUSTODIAN D. DUNN

Turned in by: D. DUNN Phone \_\_\_\_\_

CHARGE # 20-5708-561 Date Required \_\_\_\_\_

INSTRUMENT USED ON:  NUCLEAR  DOD  NASA  GLP  SPPE  
 OTHER \_\_\_\_\_

COPY OF CALIBRATION CERTIFICATE  Yes  No

CONDITION RECEIVED: \_\_\_\_\_ Out of tolerance, repaired to specifications  
\_\_\_\_\_ In tolerance, minor adjustments/repairs made  
 In tolerance, no adjustments/repairs  
\_\_\_\_\_ Out of tolerance, adjusted to specifications  
\_\_\_\_\_ Received into system, introduced or reactivated  
 Calibration internal  
 Reliability code

24193

ACTION TAKEN: (Calibration/Repair/Parts) \_\_\_\_\_

CAL ENVIRONMENT:  
Temperature 70 °F Humidity 30 %RH

CALIBRATED/REPAIRED:  
By 8114 Cal Procedure W1930MS07  
Date 2-11-97 Accuracy ±.001  
Cal Interval 6m Time to complete:  
Next Cal due 8-11-97 Cal 8 Repair \_\_\_\_\_  
Standards used (Asset#) 002171

DATE COMPLETED 2-11-97  
DATE PICKED UP 2/17/97 PICKED UP BY [Signature]

**CALIBRATION CHECK FORM**

Date Calibrated 2-11-97 Work Order 24193  
 Technician 8114  
 Unit Under Test \_\_\_\_\_  
 Manufacturer Fowler Model 32-068-008 SN 20-86-1 ASN \_\_\_\_\_

WI-9-30- MS07  
 Rev \_\_\_\_\_ Chg \_\_\_\_\_  
 Page \_\_\_\_\_ of \_\_\_\_\_

STEP	FUNCTION OR RANGE		APPLIED	TOLERANCE		MEASURED VALUES		P/F
				MIN	MAX	AS FOUND	RELEASED	
1	1.000	O.D.		.9995	1.0005	1.000		P
2	4.000	O.D.		3.9995	4.0005	4.000		P
3	6.000	O.D.		5.9995	6.0005	6.000		P
4	8.000	O.D.		7.9995	8.0005	8.000		P
5	10.000	O.D.		9.9995	10.0005			
6	12.000	O.D.		11.9995	12.0005			
7	1.000	I.D.		.9995	1.0005	1.000		P
8	4.000	I.D.		3.9995	4.0005	4.000		P
9	6.000	I.D.		5.9995	6.0005	6.000		P
10	8.000	I.D.		7.9995	8.0005	8.000		P
11	10.000	I.D.		9.9995	10.0005			
12	12.000	I.D.		11.9995	12.0005			
13	1.000	DEPTH		.9995	1.0005	1.000		P
14	4.000	DEPTH		3.9995	4.0005	4.000		P
15	6.000	DEPTH		5.9995	6.0005	6.000		P
16	8.000	DEPTH		7.9995	8.0005	8.000		P
17	10.000	DEPTH		9.9995	10.0005			
18	12.000	DEPTH		11.9995	12.0005			

SOUTHWEST RESEARCH INSTITUTE  
Department of Quality Assurance  
Calibration Laboratory • 522-5215

**WORK ORDER**

CERTIFICATE # 26597 ASSET # 002185 DATE 25 Aug 97

ITEM DATA:

Manufacturer Fowler Model S2-08-008  
Description 0" dia caliper Serial # 20-80-1  
Accessories \_\_\_\_\_

ACTION REQUESTED MI

CUSTODIAN DWRO, Small, Dean

Turned in by: \_\_\_\_\_ Phone 6090

CHARGE # 20-5708-561 Date Required \_\_\_\_\_

INSTRUMENT USED ON:  DOD/NASA  NUCLEAR  GLP  SPPE  ISO  
 OTHER \_\_\_\_\_

COPY OF CALIBRATION CERTIFICATE  Yes  No

NEW WORK  Yes  No If yes, an evaluation shall be made to verify capabilities.

By \_\_\_\_\_ Date \_\_\_\_\_

CONDITION RECEIVED: \_\_\_\_\_ (F) Out of tolerance, repaired to specifications  
\_\_\_\_\_ (G) In tolerance, minor adjustments/repairs made  
 (J) In tolerance, no adjustments/repairs  
\_\_\_\_\_ (K) Out of tolerance, adjusted to specifications  
\_\_\_\_\_ (S) Received into system, introduced or reactivated

ACTION TAKEN: (Calibration/Repair/Parts) \_\_\_\_\_

Calibrated

CAL ENVIRONMENT: Temperature 70 °F Humidity 62 %RH

CALIBRATED/REPAIRED:  
By 8114 Cal Procedure W1930MS08  
Date 9-9-97 Accuracy .001  
Cal Interval 6M Reliability Code: 7  
Next Cal due 3-9-97 Cal Time 8 Repair Time \_\_\_\_\_  
Standards used (Asset#) 002171

DATE COMPLETED \_\_\_\_\_  
DATE PICKED UP 9/30/97 PICKED UP BY [Signature]

26597

# CALIBRATION CHECK FORM

Date Calibrated 9-9-97 Work Order 26577  
 Technician 814  
 Unit Under Test \_\_\_\_\_  
 Manufacturer FOWLER Model 520B-008 SN 20-8C-1 ASN 002185

STEP	FUNCTION OR RANGE	APPLIED	TOLERANCE		MEASURED VALUES		P/F
			MIN	MAX	AS FOUND	RELEASED	
1	0.000 OD		.0005	.0005	.000	—	P
2	2.125				2.125	—	P
3	4.100				4.100	—	P
4	6.500				6.500	—	P
5	8.000				8.000	—	P
6	1.000 ID				1.000	—	P
7	2.125				2.125	—	P
8	4.100				4.100	—	P
9	6.500				6.500	—	P
10	8.000				8.000	—	P
11	0.000 Depth						0.000
12	2.125		2.125	—			P
13	4.100		4.100	—			P
14	6.500		6.500	—			P
15	8.000		8.000	—			P
16	0.000 Depth				0.000	—	P
17	2.125				2.125	—	P
18	4.100				4.100	—	P
19	6.500				6.500	—	P
20	8.000				8.000	—	P

SOUTHWEST RESEARCH INSTITUTE  
Department of Quality Assurance  
Calibration Laboratory • 522-5215

**WORK ORDER**

CERTIFICATE # 28699 ASSET # 002185 DATE 19 Feb 98

ITEM DATA:

Manufacturer Fowler Model S2-008-008  
Description 6" dial caliper Serial # 20-80-1  
Accessories OSP

ACTION REQUESTED Cal MTG du 1906829

CUSTODIAN Dr. W. Howell Dunn

Turned in by: \_\_\_\_\_ Phone 6090

CHARGE # 20-1402-571 Date Required \_\_\_\_\_

INSTRUMENT USED ON: DOD/NASA  NUCLEAR  GLP  SPPE  ISO  
OTHER \_\_\_\_\_

COPY OF CALIBRATION CERTIFICATE  Yes  No

NEW WORK  Yes  No If yes, an evaluation shall be made to verify capabilities.  
By \_\_\_\_\_ Date \_\_\_\_\_

CONDITION RECEIVED: \_\_\_\_\_ (F) Out of tolerance, repaired to specifications  
\_\_\_\_\_ (G) In tolerance, minor adjustments/repairs made  
 (J) In tolerance, no adjustments/repairs  
\_\_\_\_\_ (K) Out of tolerance, adjusted to specifications  
\_\_\_\_\_ (S) Received into system, introduced or reactivated

ACTION TAKEN: (Calibration/Repair/Parts) Calibrated Per Procedure

CAL ENVIRONMENT:  
Temperature 69 °F Humidity 45 %RH

CALIBRATED/REPAIRED:  
By [Signature] Cal Procedure CAL-MS-007  
Date 23 Feb 98 Accuracy 1/100  
Cal Interval 6 mos Reliability Code: 8  
Next Cal due 23 Aug 98 Cal Time 1.5 Repair Time \_\_\_\_\_  
Standards used (Asset#) 2171, 1741

DATE COMPLETED 23 Feb 98  
DATE PICKED UP 2/25/98 PICKED UP BY [Signature]

28699

# CALIBRATION CHECK FORM

Date Calibrated 23 FEB 85 Work Order 28699  
 Technician 8216 CAL. PROCEDURE: CLCP-M3-007  
 Unit Under Test DIAL CALIPER  
 Manufacturer Fowler Model 52-008-008 SN 20-85-1 ASN 2/85

STEP	FUNCTION OR RANGE	APPLIED	TOLERANCE		MEASURED VALUES		P/F
			MIN	MAX	AS FOUND	RELEASED	
	O.D		± 0.001				
	25% RANGE	2.000"			2.000"		P
	50% RANGE	4.000"			4.000"		P
	75% RANGE	6.000"			6.000"		P
	100% RANGE	8.000"			8.000"		P
	I.D						
	25% RANGE	2.000"			2.000"		P
	50% RANGE	4.000"			4.000"		P
	75% RANGE	6.000"			6.000"		P
	100% RANGE	8.000"			8.000"		P
	DEPTH						
	25% RANGE	2.000"			2.000"		P
	50% RANGE	4.000"			4.000"		P
	75% RANGE	6.000"			6.000"		P
	100% RANGE	7.500"			7.500"		P
	OFFSET DEPTH						
	25% RANGE	2.000"			2.000"		P
	50% RANGE	4.000"			4.000"		P
	75% RANGE	6.000"			6.000"		P
	100% RANGE	8.000"			8.000"		P

SOUTHWEST RESEARCH INSTITUTE  
Department of Quality Assurance  
Calibration Laboratory • 522-5215

WORK ORDER

CERTIFICATE # 31064 ASSET # 002105 DATE 3/1/98

ITEM DATA:

Manufacturer Fowler Model 52-008-008  
Description 8" dial caliper Serial # 20-80-1  
Accessories MSA spv 4926829

ACTION REQUESTED cal

CUSTODIAN Div. 80, Darrell Dunn

Turned in by: WAE Phone 6090

CHARGE # 20-1102-571 Date Required \_\_\_\_\_

INSTRUMENT USED ON:  (DOD/NASA)  (NUCLEAR)  (GLP)  (SPPE)  (ISO)  
 OTHER

COPY OF CALIBRATION CERTIFICATE  (Yes)  (No)

NEW WORK  Yes  No If yes, an evaluation shall be made to verify capabilities.

By WAE Date 03/31/98

Work involves proprietary/confidential information or equipment  (Yes)  (No)

CONDITION RECEIVED: \_\_\_\_\_ (F) Out of tolerance, repaired to specifications  
\_\_\_\_\_ (G) In tolerance, minor adjustments/repairs made  
 (J) In tolerance, no adjustments/repairs  
\_\_\_\_\_ (K) Out of tolerance, adjusted to specifications  
\_\_\_\_\_ (S) Received into system, introduced or reactivated

ACTION TAKEN: (Calibration/Repair/Parts) Calibrated Per Specifications

CAL ENVIRONMENT:  
Temperature 68 °F Humidity 47 %RH

CALIBRATED/REPAIRED:  
By K. [Signature] Cal Procedure CLCP-45-007  
Date 3/5/98 Accuracy 4/29 [Signature]  
Cal Interval 6 mths Reliability Code: 9  
Next Cal due 3/5/99 Cal Time 1.5 Repair Time \_\_\_\_\_  
Standards used (Asset#) 0465

DATE COMPLETED 3/5/98  
DATE PICKED UP 2/22/98 PICKED UP BY [Signature]

31064

# CALIBRATION CHECK FORM

Date Calibrated 3/5/98 Work Order 31064  
 Technician 8216 Cal. Procedure: CLCP-115-007  
 Unit Under Test Digital Caliper Dual Caliper 8"  
 Manufacturer Fowler Model 52-008-008 SN 20-8C-1 ASN 2185

STEP	FUNCTION OR RANGE	APPLIED	TOLERANCE		MEASURED VALUES		P/F
			MIN	MAX	AS FOUND	RELEASED	
	<u>O.D.</u>		<u>± 0.001"</u>				
	<u>25% RANGE</u>	<u>2.000"</u>			<u>2.000"</u>		<u>P</u>
	<u>50% RANGE</u>	<u>4.000"</u>			<u>4.000"</u>		<u>P</u>
	<u>75% RANGE</u>	<u>6.000"</u>			<u>6.000"</u>		<u>P</u>
	<u>100% RANGE</u>	<u>8.000"</u>			<u>8.000"</u>		<u>P</u>
	<u>I.D.</u>						
	<u>25% RANGE</u>	<u>2.000"</u>			<u>2.000"</u>		<u>P</u>
	<u>50% RANGE</u>	<u>4.000"</u>			<u>4.000"</u>		<u>P</u>
	<u>75% RANGE</u>	<u>6.000"</u>			<u>6.000"</u>		<u>P</u>
	<u>100% RANGE</u>	<u>8.000"</u>			<u>8.000"</u>		<u>P</u>
	<u>DEPTH</u>						
	<u>25% RANGE</u>	<u>2.000"</u>			<u>2.000"</u>		<u>P</u>
	<u>50% RANGE</u>	<u>4.000"</u>			<u>4.000"</u>		<u>P</u>
	<u>75% RANGE</u>	<u>6.000"</u>			<u>6.000"</u>		<u>P</u>
	<u>100% RANGE</u>	<u>8.000"</u>			<u>8.000"</u>		<u>P</u>
	<u>STEP</u>						
	<u>25% RANGE</u>	<u>2.000"</u>			<u>2.000"</u>		<u>P</u>
	<u>50% RANGE</u>	<u>4.000"</u>			<u>4.000"</u>		<u>P</u>
	<u>75% RANGE</u>	<u>6.000"</u>			<u>6.000"</u>		<u>P</u>
	<u>100% RANGE</u>	<u>8.000"</u>			<u>8.000"</u>		<u>P</u>

*Handwritten signature*



SOUTHWEST RESEARCH INSTITUTE  
Department of Quality Assurance  
Calibration Laboratory • 522-5215

WORK ORDER

CERTIFICATE # 33570 ASSET # 002185 DATE 05 MARCH 99

ITEM DATA:

Manufacturer Foxy Model 52-008-008  
Description dial caliper Serial # 20-80-1  
Accessories OSP NEB Shim 9920829

ACTION REQUESTED MI

CUSTODIAN Div 20, Darrell Dunn

Turned in by: \_\_\_\_\_ Phone 6090

CHARGE # 20-442-571 OH 03/2 Date Required \_\_\_\_\_

INSTRUMENT USED ON:  (DOD/NASA)  (NUCLEAR)  (GLP)  (SPPE)  (ISO)  
 OTHER \_\_\_\_\_

COPY OF CALIBRATION CERTIFICATE  (Yes)  (No)

NEW WORK  Yes  No If yes, an evaluation shall be made to verify capabilities.

By MDI Date 03-25-99

Work involves proprietary/confidential information or equipment  (Yes)  (No)

CONDITION RECEIVED: \_\_\_\_\_ (F) Out of tolerance, repaired to specifications  
\_\_\_\_\_ (G) In tolerance, minor adjustments/repairs made  
 (J) In tolerance, no adjustments/repairs  
\_\_\_\_\_ (K) Out of tolerance, adjusted to specifications  
\_\_\_\_\_ (S) Received into system, introduced or reactivated

ACTION TAKEN: (Calibration/Repair/Parts) Calibrated Per Request

CAL ENVIRONMENT:  
Temperature 69 °F Humidity 42 %RH

CALIBRATED/REPAIRED:  
By 76-01-01-99 K. K. K. Cal Procedure CLCP-MS-207, DEC 97  
Date 16 APR 99 Accuracy 1/2  
Cal Interval 6 mos Reliability Code: 10  
Next Cal due 16 SEP 1999 Cal Time 1.5 Repair Time \_\_\_\_\_  
Standards used (Asset#) E465

DATE COMPLETED 16 APR 99  
DATE PICKED UP 3/18/99 PICKED UP BY [Signature]

33570

# CALIBRATION CHECK FORM

Date Calibrated 16 MAR 99 Work Order 33570  
 Technician R.P.B. Cal. Procedure: CLCA-M5-007, DEC 97  
 Unit Under Test DIAL CALYPER, 8"  
 Manufacturer FOWLER + NICHOLSON Model 52-008-008 SN 20-8C-1 ASN 3/85

STEP	FUNCTION OR RANGE	APPLIED	TOLERANCE		MEASURED VALUES		P/F
			MIN	MAX	AS FOUND	RELEASED	
	ZERO		0		0.000"		P
	O.D.		±0.001"				
	25% RANGE	2.000"			2.000"		P
	50% RANGE	4.000"			4.000"		P
	75% RANGE	6.000"			6.000"		P
	100% RANGE	8.000"			8.000"		P
	I.D.						
	25% RANGE	2.000"			2.000"		P
	50% RANGE	4.000"			4.000"		P
	75% RANGE	6.000"			6.000"		P
	100% RANGE	8.000"			8.000"		P
	DEPTH						
	25% RANGE	2.000"			2.000"		P
	50% RANGE	4.000"			4.000"		P
	75% RANGE	6.000"			6.000"		P
	100% RANGE	8.000"			8.000"		P
	OFFSET DEPTH						
	25% RANGE	2.000"			2.000"		P
	50% RANGE	4.000"			3.999"		P
	75% RANGE	6.000"			6.000"		P
	100% RANGE	8.000"			8.000"		P

*R.P.B.*

SOUTHWEST RESEARCH INSTITUTE  
Department of Quality Assurance  
Calibration Laboratory • 522-5215

WORK ORDER

CERTIFICATE # 35948 ASSET # 002185 DATE 09 Sept. 99

ITEM DATA:

Manufacturer Fowler Model 52-008-008  
Description dial caliper Serial # 20-80-1  
Accessories case

ACTION REQUESTED cal

CUSTODIAN Div. 20, Darrell Dunn

Turned in by: \_\_\_\_\_ Phone 6090

CHARGE # 20-04 Date Required \_\_\_\_\_

INSTRUMENT USED ON:  (DOD/NASA)  (NUCLEAR)  (GLP)  (SPPE)  (ISO)  
 OTHER \_\_\_\_\_

COPY OF CALIBRATION CERTIFICATE  (Yes)  (No)

NEW WORK  Yes  No If yes, an evaluation shall be made to verify capabilities.

By MAD Date 09-09-99

Work involves proprietary/confidential information or equipment  (Yes)  (No)

- CONDITION RECEIVED:
- (F) Out of tolerance, repaired to specifications
  - (G) In tolerance, minor adjustments/repairs made
  - (J) In tolerance, no adjustments/repairs
  - (K) Out of tolerance, adjusted to specifications
  - (S) Received into system, introduced or reactivated

ACTION TAKEN: (Calibration/Repair/Parts) calibrated per procedure

CAL ENVIRONMENT:

Temperature 68 °F Humidity 42%RH

CALIBRATED/REPAIRED:

By [Signature] Cal Procedure CL-11, 11/24/99

Date 14 Sep 99 Accuracy [Signature]

Cal Interval 6 mos Reliability Code: 11

Next Cal due 14 Mar 00 Cal Time 1.5 Repair Time \_\_\_\_\_

Standards used (Asset#) 7308, 6465

DATE COMPLETED 14 Sep 99

DATE PICKED UP 9/23/99 PICKED UP BY [Signature]

35948



# WORK ORDER 38598

Date Received 4/10/00

Asset No. 002185 Manufacturer FOWLER & NSK Model 52-008-008  
Description DIAL CALIPER Serial Number 20-8C-1  
Accessory Received/Required CASE  
Div/CC ID NONE Accessory to Asset No. N/A Accuracy +/-0.001"  
Div/CC DIV20 Location B57 Custodian DARRELL DUNN Tel. 6090  
Charge/Project No. 20.00751.006 Proprietary/Confidential N Date Required ROUTINE  
Work Requested CALIBRATION  
Receiving Inspection O.K.  
Delivered By DARRELL DUNN Tel. 6090

### WORK HISTORY

Date	Start Time	Stop Time	Notes

### PARTS

Part Name	Part Number	Cost	Failure Description

### WORK SUMMARY

Failure Description \_\_\_\_\_  
Repair Action \_\_\_\_\_  
Cal Procedure CL-15, 5/99 Temp 68 F Hum 44 %  
Tech 8216 Cal Hrs. 1.5 Repair Hrs. \_\_\_\_\_ Part Cost \_\_\_\_\_  
Action Taken Calibrated Per Procedure  
Standards Used 7705, 6465  
Date Cal 12 Apr 00 Int. 6 Mo. Date Due 12 Oct 00 Reliability Code 12  
Date Picked Up 5/2/00 Picked Up By [Signature]

38598





Southwest Research Institute  
6220 Culebra Road  
San Antonio, TX 78238  
(210) 522-5215  
Department of Quality Assurance  
Calibration Laboratory



Certificate #  
0972-01

## Certificate of Calibration

19 April 2000

**Issued to:** DARRELL DUNN DIV20 B57  
**Manufacturer/Model:** FOWLER & NSK 52-008-008  
**Description:** DIAL CALIPER  
**Serial Number:** 20-8C-1  
**Asset Number:** 002185

This certifies the above item was calibrated in compliance with MIL-STD-45662A and ANSI/NC SL Z540-1-1994. The results of this calibration relate only to the individual item as described above. Standards used in this calibration, described in the referenced calibration procedure with associated uncertainties or tolerances, are traceable to the National Institute of Standards and Technology (NIST). Supporting documentation relative to traceability is on file and available for examination upon request. This certificate is not to be reproduced, except in full, without the written approval of the Southwest Research Institute Department of Quality Assurance Calibration Laboratory.

This laboratory is accredited by the American Association for Laboratory Accreditation (A2LA) and the results of this calibration certificate were determined in accordance with the terms of accreditation unless stated otherwise below.

The uncertainty of the calibration was sufficient to determine that the item met the manufacturer's published specifications unless stated otherwise below.

**Ambient Conditions:** Temperature: 68.0 Degrees Fahrenheit Humidity: 44 % RH

**Calibration Date:** 19 Apr 00 **Calibration Procedure:** CL-11, MAY 99

**Condition as Received:** IN TOLERANCE

**Condition as Released:** IN TOLERANCE

**Remarks:**

**Approved by:**

  
\_\_\_\_\_  
Jim Patterson, Supervisor, or Walt Hill, Metrologist

**Certificate #** 38598

m:\a2la.rpt Rev date 14 Dec 99

**Measurements performed by:**

  
\_\_\_\_\_  
Ken Harp, Technician

Page 1 of 1

# SOUTHWEST RESEARCH INSTITUTE

## Calibration Laboratory

### WORK ORDER

Processed by RCRUZ at 8:44:40AM on 10/24/00

I4213

ASSET NO. 002185

Work Order No 444041096

Arrived 10/24/00

Manufacturer FOWLER & NSK

Model 52-008-008

Instrument Type/Class DIAL CALIPER

Serial No. 20-8C-1

Accessory No. Calibration Procedure CL-11, MAY 99

Location B57

Div/Client DIV20

Custodian DARRELL DUNN

Mail Stop B57

Tel. 6090

Charge/Project No. 20.00751.006

Delivered By / Telephone DARRELL DUNN

**IN4CAL**

Special Instructions \_\_\_\_\_

### WORK NOTES

Date	Hours	Remarks/Notes
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____

### REPAIR PARTS

Date	Hours	Part Name	Part Number	Failure Description	Cost
_____	_____	_____	_____	_____	_____
_____	_____	_____	_____	_____	_____
_____	_____	_____	_____	_____	_____
_____	_____	_____	_____	_____	_____
_____	_____	_____	_____	_____	_____

**44096**

### WORK SUMMARY

Failure Description \_\_\_\_\_

Repair Action \_\_\_\_\_

Calibration Procedure CL-11, May 99 Temp 88F Hum. 44%

Tech 8216 Totals Cal Hours 1 Repair Hours \_\_\_\_\_ Parts Cost \_\_\_\_\_

Standards Used 6445

Date Picked Up 11/6/2000

Picked Up By Darrell Dunn







Southwest Research Institute  
6220 Culebra Road  
San Antonio, TX 78238  
(210) 522-5215  
Department of Quality Assurance  
Calibration Laboratory



Certificate #  
0972-01

## Certificate of Calibration

2 November 2000

**Issued to:** DARRELL DUNN DIV20 B57  
**Manufacturer/Model:** FOWLER & NSK 52-008-008  
**Description:** CALIPER  
**Serial Number:** 20-8C-1  
**Asset Number:** 002185

This certifies the above item was calibrated in compliance with MIL-STD-45662A and ANSI/NCSL Z540-1-1994. The results of this calibration relate only to the individual item as described above. Standards used in this calibration, described in the referenced calibration procedure with associated uncertainties or tolerances, are traceable to the National Institute of Standards and Technology (NIST). Supporting documentation relative to traceability is on file and available for examination upon request. This certificate is not to be reproduced, except in full, without the written approval of the Southwest Research Institute Department of Quality Assurance Calibration Laboratory.

This laboratory is accredited by the American Association for Laboratory Accreditation (A2LA) and the results of this calibration certificate were determined in accordance with the terms of accreditation unless stated otherwise below.

The uncertainty of the calibration was sufficient to determine that the item met the manufacturer's published specifications unless stated otherwise below.

**Ambient Conditions:** Temperature: 68.0 Degrees Fahrenheit Humidity: 44 % RH

**Calibration Date:** 2 Nov 00 **Calibration Procedure:** CL-11, MAY 99

**Condition as Received:** IN TOLERANCE

**Condition as Released:** IN TOLERANCE

**Remarks:**

**Approved by:**

Jim Patterson, Supervisor, or Walt Hill, Metrologist

Certificate # 444041096

m:\a2la.rpt Rev date 22 May 00

**Measurements performed by:**

Ken Harp, Technician







Southwest Research Institute  
6220 Culebra Road  
San Antonio, TX 78238  
(210) 522-5215  
Department of Quality Assurance  
Calibration Laboratory



Certificate #  
0972-01

## Certificate of Calibration

30 April 2001

**Issued to:** DARRELL DUNN DIV20 B57  
**Manufacturer/Model:** FOWLER & NSK 52-008-008  
**Description:** CALIPER  
**Serial Number:** 20-8C-1  
**Asset Number:** 002185  
**Work Order Number:** 444043358

This certifies the above item was calibrated in compliance with MIL-STD-45662A and ANSI/NCSL Z540-1-1994. The results of this calibration relate only to the individual item as described above. Standards used in this calibration, described in the referenced calibration procedure with associated uncertainties or tolerances, are traceable to the National Institute of Standards and Technology (NIST). Supporting documentation relative to traceability is on file and available for examination upon request. This certificate is not to be reproduced, except in full, without the written approval of the Southwest Research Institute Department of Quality Assurance Calibration Laboratory.

This laboratory is accredited by the American Association for Laboratory Accreditation (A2LA) and the results of this calibration certificate were determined in accordance with the terms of accreditation unless stated otherwise below.

The uncertainty of the calibration was sufficient to determine that the item met the manufacturer's published specifications unless stated otherwise below.

**Ambient Conditions:** Temperature: 68.0 Degrees Fahrenheit Humidity: 42 % RH

**Calibration Date:** 30 Apr 01 **Calibration Procedure:** CL-11, MAY 99

**Condition as Received:** IN TOLERANCE

**Condition as Released:** IN TOLERANCE

**Remarks:**

**Approved by:**

Walt Hill, Supervisor  
Institute Calibration Laboratory

**Measurements performed by:**

Ken Harp, Technician

# SOUTHWEST RESEARCH INSTITUTE

## Calibration Laboratory

### WORK ORDER

Received by RCRUZ, 12/6/01 2:25:38PM

|||||

Arrived 12/6/01

Work Order **444046374**

Asset No. 002185 Manufacturer FOWLER & NSK

Model 52-008-008

Equipment Type CALIPER

Serial No. 20-8C-1

Accessory No.

Interval 6 M

Calibration Procedure CL-11, MAY 99

Location B57

Div/Client DIV20

Custodian DARRELL DUNN

Mail Stop B57

Tel 6090

**IN4CAL**

Special Instructions 20.00751.006

Notify before adjustments or repairs. (  Provide data with certificate (  Certificate Typ. \_\_\_\_\_

Charge/Project No. 00751.006 1.20

Requester / Telephone \_\_\_\_\_

This information is correct for the work requested. 

### WORK NOTES

Date	Hours	Remarks/Notes
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____

Date	Hours	Part Name	Part Number	Failure Description	Cost
_____	_____	_____	_____	_____	_____
_____	_____	_____	_____	_____	_____
_____	_____	_____	_____	_____	_____
_____	_____	_____	_____	_____	_____
_____	_____	_____	_____	_____	_____

### WORK SUMMARY

Failure Description \_\_\_\_\_

Repair Action \_\_\_\_\_

Tech 8216 Cal Hrs. 1 Repair Hrs \_\_\_\_\_ Parts Cost \_\_\_\_\_ Temp 68 F Hum. 33 %

Standards Used 6465

Date Picked Up 1/7/2002

Picked Up By 

**444046374**

# CALIBRATION CHECK FORM

Date Calibrated 31 Dec 01 Work Order 444076374

Technician 8216 Calibration Procedure: CL-11. May99

Unit Under Test DIAL CALIPERS, 1"

Manufacturer Fowler & NSK Model 52-008-008 SN 20-80-1 ASN 2185

STEP	FUNCTION OR RANGE	APPLIED	TOLERANCE		MEASURED VALUES		P/F
			MIN	MAX	AS FOUND	RELEASED	
4.3.2	SEPARATION OF JAWS (CLOSED)						
	Visible Light	NA			None		P
4.3.3	JAW PARALLELISM						
	(Nearest point to TI beam)	NA			/		
	(Opposite end of Jaws)	NA	< +/- 1 div difference				
4.3.5	OUTSIDE SCALE CALIBRATION		+/- 0.001"				
	25% Range	2.0000"		"	2.0002"		P
	50% Range	4.0000"		"	4.0005"		P
	75% Range	6.0000"		"	6.0000"		P
	100% Range	8.0000"		"	8.0005"		P
4.4.8	INSIDE SCALE CALIBRATION		+/- 0.001"				
		0.5000"		"	0.4995"		P
	50% Range	4.0000"		"	3.9993"		P
4.4.11	DEPTH SCALE CALIBRATION		+/- 0.001"				
		1.0000"		"	1.0000" None		P
---	STEP SCALE CALIBRATION		+/- 0.001"				
		1.0000"		"	0.9995"		P







Southwest Research Institute  
6220 Culebra Road  
San Antonio, TX 78238  
(210) 522-5215  
Department of Quality Assurance  
Calibration Laboratory

## Certificate of Calibration

31 December 2001

**Issued to:** DARRELL DUNN DIV20 B57  
**Manufacturer/Model:** FOWLER & NSK 52-008-008  
**Description:** CALIPER  
**Serial Number:** 20-8C-1  
**Asset Number:** 002185  
**Work Order Number:** 444046374

This certifies the above item was calibrated in compliance with MIL-STD-45662A and ANSI/NCSL Z540-1-1994. Standards used in this calibration, described in the referenced calibration procedure with associated uncertainties or tolerances, are traceable to the National Institute of Standards and Technology (NIST). Supporting documentation relative to traceability is on file and is available for examination upon request. This certificate is not to be reproduced, except in full, without the written approval of the Southwest Research Institute Department of Quality Assurance Calibration Laboratory.

The uncertainty of the calibration was sufficient to determine that the item met the manufacturer's published specifications unless stated otherwise below.

**Ambient Conditions:** Temperature: 68.0 Degrees Fahrenheit Humidity: 33 % RH


**Calibration Date:** 31 Dec 01 **Calibration Procedure:** CL-11, MAY 99

**Condition as Received:** IN TOLERANCE


**Condition as Returned:** IN TOLERANCE

**Remarks:**

**Approved by:**

  
Walt Hill, Supervisor  
Institute Calibration Laboratory

**Measurements performed by:**

  
Ken Harp, Technician

# SOUTHWEST RESEARCH INSTITUTE

## Calibration Laboratory

### WORK ORDER

Received by RCRUZ, 6/14/02 9:00:15AM

||||| ||||| ||||| ||||| ||||| ||||| |||||

Arrived 6/14/02

Work Order **444049039**

Asset No. 002185 Manufacturer FOWLER & NSK

Model 52-008-008

Equipment Type CALIPER

Serial No. 20-8C-1

Accessory No.

Interval 6 M

Calibration Procedure CL-11, MAY 99

Location B57

Div/Client DIV20

Custodian DARRELL DUNN

Mail Stop B57

Tel 6090

**IN LINE**

Special Instructions 20.00751.006

Notify before adjustments or repairs. ( ) Provide data with certificate ( ) Certificate Typ\_\_\_\_\_

Charge/Project No. 00751.006 1.20

Requester / Telephone \_\_\_\_\_

This information is correct for the work requested. *[Signature]*

### WORK NOTES

Date Hours Remarks/Notes

6/14/02 20 cal, limited calibration

Date Hours Part Name Part Number Failure Description Cost

Date	Hours	Part Name	Part Number	Failure Description	Cost

### WORK SUMMARY

Failure Description with Inside function does not meet tolerance.

Repair Action n/a

Tech RJDK Cal Hrs. 20 Repair Hrs \_\_\_\_\_ Parts Cost \_\_\_\_\_ Temp 68 F Hum. 38 %

Standards Used GAGS

Date Picked Up 6/14/02

Picked Up By *[Signature]*

**49039**



Southwest Research Institute  
6220 Culebra Road  
San Antonio, TX 78238  
(210) 522-5215  
Department of Quality Assurance  
Calibration Laboratory

## Certificate of Calibration

17 June 2002

**Issued to:** DARRELL DUNN DIV20 B57  
**Manufacturer/Model:** FOWLER & NSK 52-008-008  
**Description:** CALIPER  
**Serial Number:** 20-8C-1  
**Asset Number:** 002185  
**Work Order Number:** 444049039

This certifies the above item was calibrated in compliance with MIL-STD-45662A and ANSI/NCSL Z540-1-1994. Standards used in this calibration, described in the referenced calibration procedure with associated uncertainties or tolerances, are traceable to the National Institute of Standards and Technology (NIST). Supporting documentation relative to traceability is on file and is available for examination upon request. This certificate is not to be reproduced, except in full, without the written approval of the Southwest Research Institute Department of Quality Assurance Calibration Laboratory.

The uncertainty of the calibration was sufficient to determine that the item met the manufacturer's published specifications unless stated otherwise below.

**Ambient Conditions:** Temperature: 68.0 Degrees Fahrenheit Humidity: 36 % RH

**Calibration Date:** 17 Jun 02 **Calibration Procedure:** CL-11, MAY 99

**Condition as Received:** OUT OF TOLERANCE

**Condition as Returned:** LIMITED CALIBRATION

**Remarks:** LIMITED CALIBRATION, THE INSIDE FUNCTION OF CALIPER DOES NOT MEET TOLERANCE AND SHOULD NOT BE USED.

Approved by:

Walt Hill, Metrology Group Leader  
Institute Calibration Laboratory

Measurements performed by:

Roger Dykstra, Technician

Southwest Research Institute  
Calibration Laboratory  
Calibration Data sheet.

As left data

<b>Work Order</b> 444049039	<b>Mfr</b> Fowler & NSK	<b>Tech:</b> R Dykstra	
<b>Asset No.</b> 002185	<b>Model</b> 52-008-008	<b>Procedure:</b> CL-11, May99	
<b>Serial No.</b> 20-8C-1	<b>Type</b> Dial Caliper	<b>Cal Date:</b> 6/14/02	

**Remarks:**

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor (k=2) providing a level of confidence of approximately 95%.  
The Difference is equal to TI reading - Std reading.  
The results can be Pass, Fail, or if blank "not determinable". If "not determinable" it is up to the end user to determine if results meet their needs. Limited calibration Do Not use the inside function, does not meet tolerance. The accuracy of the 0.001 Inch is in the procedure.

**Parameter: Length**

	Applied Value(in.)	Indicated Value(in.)	Diff. (in.)	TI Tolerance(in.)	Uncertainty (in)	Results
--	--------------------	----------------------	-------------	-------------------	------------------	---------

**SEPARATION OF JAWS (CLOSED)**

**Visible Light**                      NA      No

**JAW PARALLELISM**

(Nearest point to TI beam)      NA  
(Opposite end of Jaws)              NA              0.000000      -1 Minor Div.

**OUTSIDE SCALE CALIBRATION**

2.0000	2.0000	0.0000	0.001	0.00058	Pass
4.0000	4.0000	0.0000	0.001	0.00058	Pass
6.0000	6.0002	0.0002	0.001	0.00058	Pass
8.0000	8.0002	0.0002	0.001	0.00058	Pass

**INSIDE SCALE CALIBRATION**

0.5000	0.4990	-0.0010	0.001	0.0006	
4.0000	3.9980	-0.0020	0.001	0.0003	Fail

**DEPTH SCALE CALIBRATION**

1.0000	0.9996	-0.0004	0.001	0.00029	Pass
--------	--------	---------	-------	---------	------

**STEP SCALE CALIBRATION**

1.0000	0.9996	-0.0004	0.001	0.00029	Pass
--------	--------	---------	-------	---------	------

Southwest Research Institute  
Calibration Laboratory  
Calibration Data sheet.

As found data

<b>Work Order</b> 444049039	<b>Mfr</b> Fowler & NSK	<b>Tech:</b> R Dykstra
<b>Asset No.</b> 002185	<b>Model</b> 52-008-008	<b>Procedure:</b> CL-11, May99
<b>Serial No.</b> 20-8C-1	<b>Type</b> Dial Caliper	<b>Cal Date:</b> 6/14/02

**Remarks:**

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor (k=2) providing a level of confidence of approximately 95%.  
The Difference is equal to TI reading - Std reading.  
The results can be Pass, Fail, or if blank "not determinable". If "not determinable" it is up to the end user to determine if results meet their needs.  
The accuracy of the 0.001 Inch is in manual.

**Parameter: Length**

	Applied Value(in.)	Indicated Value(in.)	Diff. (in.)	TI Tolerance(in.)	Uncertainty (in)	Results
--	--------------------	----------------------	-------------	-------------------	------------------	---------

**SEPARATION OF JAWS (CLOSED)**

Visible Light                      NA      No

**JAW PARALLELISM**

(Nearest point to TI beam)      NA  
(Opposite end of Jaws)              NA              0.000000      -1 Minor Div.

**OUTSIDE SCALE CALIBRATION**

	2.0000	2.0000	0.0000	0.001	0.00058	Pass
	4.0000	4.0000	0.0000	0.001	0.00058	Pass
	6.0000	6.0002	0.0002	0.001	0.00058	Pass
	8.0000	8.0002	0.0002	0.001	0.00058	Pass

**INSIDE SCALE CALIBRATION**

	0.5000	0.4990	-0.0010	0.001	0.0006	
	4.0000	3.9980	-0.0020	0.001	0.0003	Fail

**DEPTH SCALE CALIBRATION**

	1.0000	0.9996	-0.0004	0.001	0.00029	Pass
--	--------	--------	---------	-------	---------	------

**STEP SCALE CALIBRATION**

	1.0000	0.9996	-0.0004	0.001	0.00029	Pass
--	--------	--------	---------	-------	---------	------