

SOUTHWEST RESEARCH INSTITUTE
Department of Quality Assurance
Calibration Laboratory • 522-5215

WORK ORDER

CERTIFICATE # 33134 ASSET # 2089 DATE 02 FEB 99

ITEM DATA:

Manufacturer KEITHLEY Model 614
Description ELECTROMETER Serial # 0704936
Accessories _____

ACTION REQUESTED _____

CUSTODIAN DARRELL DUNN Div 20

Turned in by: _____ Phone _____

CHARGE # 20-1402-571 Date Required _____

INSTRUMENT USED ON: (DOD/NASA) (NUCLEAR) (GLP) (SPPE) (ISO)
 OTHER _____

COPY OF CALIBRATION CERTIFICATE (Yes) (No)

NEW WORK Yes No If yes, an evaluation shall be made to verify capabilities.

By [Signature] Date 02/02/99

Work involves proprietary/confidential information or equipment (Yes) (No)

CONDITION RECEIVED: _____ (F) Out of tolerance, repaired to specifications
_____ (G) In tolerance, minor adjustments/repairs made
_____ (J) In tolerance, no adjustments/repairs
 (K) Out of tolerance, adjusted to specifications
 (S) Received into system, introduced or reactivated

ACTION TAKEN: (Calibration/Repair/Parts) ref to Radio Dev. ST# 341685

CAL ENVIRONMENT:
Temperature 73 °F Humidity 32 %RH

CALIBRATED/REPAIRED:
By [Signature] San Antonio, TX Cal Procedure MPG
Date 03/02/99 Accuracy MPG
Cal Interval 12mos Reliability Code: 1
Next Cal due 03/02/00 Cal Time _____ Repair Time _____
Standards used (Asset#) Vendor

DATE COMPLETED 08 March 99
DATE PICKED UP 3/4/99 PICKED UP BY [Signature]

33134

ROTHE DEVELOPMENT METROLOGY SERVICES

CALIBRATION DATA: KEITHLEY 614 ELECTROMETER

CUSTOMER: Southwest Research
 WO NUMBER: 70172
 SERIAL: 0704936
 CUST ID: _____

DATE: 2 Mar 99
 TECH: YMH
 INST NO: 28151

CALIBRATION DATA TAKEN

INCOMING ✓
 OUTGOING ✓

CONDITION OF EQUIPMENT

IN TOLERANCE ✓
 OUT OF TOLERANCE _____

INPUT CURRENT VERIFICATION

RANGE	INPUT	MIN	READING	MAX
20 pA	NONE	-00.06	<u>.00</u>	+00.06

DC VOLTMETER ACCURACY

RANGE	INPUT	MIN	READING	MAX
.2 V	0.0 mV	-.00002	<u>-.00001</u>	+.00002
	+190 mV	+.18983	<u>.18998</u>	+.19017
	-190 mV	-.18983	<u>.19000</u>	-.19017
2 V	+1.9 V	+1.8984	<u>1.9000</u>	+1.9016
	-1.9 V	-1.8984	<u>1.8998</u>	-1.9016
20 V	+19 V	+18.984	<u>18.999</u>	+19.016
	-19 V	-18.984	<u>18.998</u>	-19.016

DC CURRENT ACCURACY


RANGE	INPUT	MIN	READING	MAX
200 uA	+190 uA	+189.3	<u>189.9</u>	+190.7
200 nA	+190 nA	+189.0	<u>190.1</u>	+191.0
2000 pA	+1900 pA	+1871	<u>1900</u>	+1929

RESISTANCE ACCURACY

RANGE		ACTUAL INPUT	TOLERANCE	READING
20 K Ω	(19)	<u>19.00</u>	± 11 C	<u>19.00</u>
200 K Ω		<u>190.0</u>	± 11 C	<u>189.8</u>
20 M Ω	(10)	<u>10.00</u>	± 10 C	<u>9.99</u>
20 G Ω	(10)	<u>9.98</u>	± 2 %	<u>9.96</u>

CHARGE ACCURACY

RANGE	INPUT	MIN	READING	MAX
2 nC	1.0 VDC INTO 1000 pF	0.9500	<u>.9923</u>	1.0500



CERT. NUMBER: 56841

ROTHE DEVELOPMENT, INC.
METROLOGY SERVICES DIVISION
4614 SINCLAIR RD., SAN ANTONIO, TEXAS 78222 PH:210-648-3131

CERTIFICATE OF CALIBRATION

ISSUED TO: Southwest Research Institute
6220 Culebra, Bldg. 64/Division 30
San Antonio, TX 78284
522-5460

(30) MFG: Keithley
MODEL: 614
NOMEN: Electrometer
S/N: 0704936
CUST. ID:

CAL DATE: 03/02/1999
DUE DATE: 03/02/2000

CONTROL NO.: 103 - 28151
TECHNICIAN: 9
SPECIFICATIONS: MFG
PROCEDURE: MFG
WORK ORDER: 70172
CUSTOMER P.O.: 02127/ST341685/20-1402571
RECEIVED CONDITION: IN TOLERANCE
RETURNED CONDITION: IN TOLERANCE

CALIBRATION PERFORMED AT: RDMSD
CALIBRATION INTERVAL: 12mos.
TEMPERATURE: 73 °F
RELATIVE HUMIDITY: 32 %

DATE RECEIVED: 02/11/1999

COMMENTS:

ATTACHMENTS: CALIBRATION DATA

All Calibrations performed at Rothe Development, Inc. Metrology Services Division meet the requirements of ANSI/NC SL Z540-1-1994, ISO/IEC GUIDE 25, and ISO 10012-1, and are traceable to the National Institute of Standards and Technology. The collective uncertainty of the measurement(s) does not exceed 25% (TUR \geq 4:1) of the instrument specification(s) unless noted in the COMMENTS section.

TR#	MFG	MODEL	SERIAL NO.	DUE DATE
235	GUILDLINE	6500A	59660	06/23/1999
20	FLUKE	5700A/EP	4605002	03/28/1999
258	FLUKE	5725A	6585002	03/28/1999

APPROVED BY: *Detlef Stemmer*

CMS

QCO

DATE: 03/02/1999

WORK ORDER 38128

Date Received 3/8/00

Asset No. 007089 Manufacturer KEITHLEY Model 614
 Description ELECTROMETER Serial Number 0704936
 Accessory Received/Required TRIAx CABLE
 Div/CC ID NONE Accessory to Asset No. N/A Accuracy MFG SPECS
 Div/CC DIV20 Location B57 Custodian DARRELL DUNN Tel. 6090
 Charge/Project No. 20.00751.006 Proprietary/Confidential N Date Required ROUTINE
 Work Requested CALIBRATION.
 Receiving Inspection O.K.
 Delivered By DARRELL DUNN Tel. 6090

WORK HISTORY

Date	Start Time	Stop Time	Notes
<u>03/0</u>			<u>Roche Dev. ST# 366546</u>

PARTS

Part Name	Part Number	Cost	Failure Description


38128

WORK SUMMARY

Failure Description _____
 Repair Action _____
 Cal Procedure MFG Temp 73 F Hum 32 %
 Tech Roche Dev. Cal Hrs. _____ Repair Hrs. _____ Part Cost _____
 Action Taken Calibrated
 Standards Used Vonder
 Date Cal 03-18-00 Int. 12 Mo. Date Due 03-18-01 Reliability Code 7
 Date Picked Up 3/22/2000 Picked Up By [Signature]

**SOUTHWEST RESEARCH INSTITUTE
CALIBRATION LABORATORY
MEMORANDUM**

March 22, 2000


To: DARRELL DUNN DIV20 B57
From: Jim Patterson 
Subject: Calibration by an approved supplier

Manufacturer/Model: KEITHLEY 614
Description: ELECTROMETER
Serial Number: 0704936
Asset Number: 007089
Calibration Due: March 16, 2001
Remarks: ROTHE CERTIFICATE NO.: 28151:853363078.

Supplier: ROTHE DEVELOPMENT, INC., SAN ANTONIO, TX

The above item was sent to an approved supplier who is listed on the SwRI Approved Suppliers List (ASL). This supplier is qualified to supply a product or service in support of project activities that require the use of approved suppliers. Please notify Mark Romero, extension 5215, of any discrepancies with the item or calibration documentation.

Attachment(s) 1



CERT. NUMBER:
28151:953363078

ROTHE DEVELOPMENT, INC.
METROLOGY SERVICES DIVISION
4614 SINCLAIR RD., SAN ANTONIO, TEXAS 78222 PH:210-648-3131

CERTIFICATE OF CALIBRATION

ISSUED TO: Southwest Research Institute (30)
6220 Culebra
San Antonio, TX 78284

MFG: Keithley
MODEL: 614
NOMEN: Electrometer
S/N: 0704936
CUST. ID:

CAL DATE: 3/18/2000
DUE DATE: 3/18/2001

CONTROL NO.: 103 - 28151
TECHNICIAN: 4
SPECIFICATIONS: MFG
PROCEDURE: MFG
WORK ORDER: 002002625
CUSTOMER P.O.: 02127/ST366546/20.00760.
RECEIVED CONDITION: IN TOLERANCE
RETURNED CONDITION: IN TOLERANCE

CALIBRATION PERFORMED AT: RDMSD
CALIBRATION INTERVAL: 12Mo.
TEMPERATURE: 73.0 °F
RELATIVE HUMIDITY: 32 %

DATE RECEIVED: 3/13/2000

COMMENTS:

ATTACHMENTS:

All Calibrations performed at Rothe Development, Inc. Metrology Services Division meet the requirements of ANSI/NCSL Z540-1-1994, ISO/IEC GUIDE 25, and ISO 10012-1, and are traceable to the National Institute of Standards and Technology. The collective uncertainty of the measurement(s) does not exceed 25% (TUR_>4:1) of the instrument specification(s) unless noted in the COMMENTS section.

TR#	MFG	MODEL	SERIAL NO.	DUE DATE
20	FLUKE	5700A/EP	4605002	3/23/2000
258	FLUKE	5725A	6585002	3/23/2000
235	GUILDLINE	6500A	59660	10/26/2000
75	JRL	NMN	75	11/15/2000

APPROVED BY: *Arthur J. Stemmer*

CMS

QCO

DATE:

3/20/2000

SOUTHWEST RESEARCH INSTITUTE

Calibration Laboratory

WORK ORDER

Processed by RCRUZ at 9:33:28AM on 5/8/01

||||| ||||| ||||| ||||| ||||| ||||| ||||| |||||

Work Order 444043488

Arrived 5/8/01

Asset No. 007089 Manufacturer KEITHLEY

Model 614

Instrument Type/Class ELECTROMETER

Serial No. 0704936

Accessory No. Calibration Procedure VENDOR

Location B57

Div/Client DIV20

Custodian DARRELL DUNN

Mail Stop B57

Tel. 6090

Charge/Project No. 20.00751.006

Delivered By / Telephone

IN4CAL

Special Instructions _____

WORK NOTES

Date	Hours	Remarks/Notes
<u>5-9-01</u>		<u>Rothe POC. Jose Cal. # 110⁰⁰ S.7 # 380600 shipped</u>
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____

ROTHE (210) 646-3151
ISO 9002
INST NO. 28151
CAL ON MAY 14 2001
CAL ON MAY 14 2002
TRACEABLE TO NIST

REPAIR PARTS

Date	Hours	Part Name	Part Number	Failure Description	Cost
_____	_____	_____	_____	_____	_____
_____	_____	_____	_____	_____	_____
_____	_____	_____	_____	_____	_____
_____	_____	_____	_____	_____	_____

WORK SUMMARY

Failure Description MF

Repair Action _____

Calibration Procedure MF6 Temp 73 Hum. 36 %

Tech Rothe Totals Cal Hours _____ Repair Hours _____ Parts Cost _____

Standards Used Vendors

Date Picked Up 5/18/2001 Picked Up By Darrell Dunn

444043488

**SOUTHWEST RESEARCH INSTITUTE
CALIBRATION LABORATORY
MEMORANDUM**

May 17, 2001

To: DARRELL DUNN DIV20 B57

From: Walt Hill, Supervisor 
Institute Calibration Laboratory

Subject: Calibration by an approved supplier

Manufacturer/Model: KEITHLEY 614
Description: ELECTROMETER
Serial Number: 0704936
Asset Number: 007089
Work Order Number: 444043488
Date Calibrated: May 14, 2001
Remarks: ROTHE CERT # 28151:989851053

Supplier: ROTHE DEVELOPMENT, INC., SAN ANTONIO, TX

The above item was sent to an approved supplier who is listed on the SwRI Approved Suppliers List (ASL). This supplier is qualified to supply a product or service in support of project activities that require the use of approved suppliers. Please notify the Institute Calibration Laboratory, extension 5215, of any discrepancies with the item or calibration documentation.

Attachment(s) 1



CERT. NUMBER:
28151:989851053

ROTHE DEVELOPMENT, INC.
METROLOGY SERVICES DIVISION

4614 SINCLAIR RD., SAN ANTONIO, TEXAS 78222 PH:210-648-3131

CERTIFICATE OF CALIBRATION

ISSUED TO:	Southwest Research Institute	MFG:	Keithley	CAL DATE:	5/14/2001
	(30)	MODEL:	614	DUE DATE:	5/14/2002
	6220 Culebra	NOMEN:	Electrometer		
	San Antonio, TX 78284	S/N:	0704936		
		CUST. ID:			

CONTROL NO.: 103 - 28151
 TECHNICIAN: 7
 SPECIFICATIONS: MFG
 PROCEDURE: MFG
 WORK ORDER: 002012362
 CUSTOMER P.O.: 01571R/ST380600/20.00760
 RECEIVED CONDITION: IN TOLERANCE
 RETURNED CONDITION: IN TOLERANCE

CALIBRATION PERFORMED AT: RDMSD
 CALIBRATION INTERVAL: 12Mo.
 TEMPERATURE: 73.0 °F
 RELATIVE HUMIDITY: 36 %
 DATE RECEIVED: 5/10/2001

COMMENTS:

ATTACHMENTS:

All Calibrations performed at Rothe Development, Inc. Metrology Services Division meet the requirements of ANSI/NCSL Z540-1-1994, ISO/IEC GUIDE 25, and ISO 10012-1, and are traceable to the National Institute of Standards and Technology. The collective uncertainty of the measurement(s) does not exceed 25% (TUR \geq 4:1) of the instrument specification(s) unless noted in the COMMENTS section.

TR#	MFG	MODEL	SERIAL NO.	DUE DATE
20	FLUKE	5700A/EP	4605002	6/21/2001
258	FLUKE	5725A	6585002	6/21/2001
235	GUILDLINE	6500A	59660	11/1/2001
75	JRL	NMN	75	1/22/2002

APPROVED BY: Peter J. Stemann

CMS · QCO

DATE:

5/15/2001

SOUTHWEST RESEARCH INSTITUTE

Calibration Laboratory

WORK ORDER

Received by RCRUZ, 5/14/02 8:12:36AM

Arrived 5/14/02

Work Order **444048529**

Asset No. 007089 Manufacturer KEITHLEY

Model 614

Equipment Type ELECTROMETER

Serial No. 0704936

Accessory No.

Interval 12 M

Calibration Procedure ROTHER

Location B57

Div/Client DIV20

Custodian DARRELL DUNN

Mail Stop B57

Tel 6090

IN LINE

Special Instructions _____

Notify before adjustments or repairs. () Provide data with certificate () Certificate Typ _____

Charge/Project No. 00751.006 1.20

Requester / Telephone DARRELL DUNN X6090

This information is correct for the work requested. Darrell Dunn

WORK NOTES

Date	Hours	Remarks/Notes
<u>5-15-02</u>		<u>Rothe ST # 359491</u>

ROTHE (210) 648-3131
ISO 9002

INST NO. 28151

RD1
04
CAL ON
MAY 26 2002
DATE DUE CAL
MAY 26 2003

Date	Hours	Part Name	Part Number	Failure	Cost

TRACEABLE TO NIST

WORK SUMMARY

Failure Description

Repair Action

Tech Rothe Cal Hrs. Repair Hrs Parts Cost Temp 70 F Hum. 40 %

Standards Used

Date Picked Up 6/6/2002

Picked Up By Darrell Dunn

48529

ROTHER DEVELOPMENT METROLOGY SERVICES

CALIBRATION DATA: KEITHLEY 614 ELECTROMETER

CUSTOMER: SWRT
 WO NUMBER: 002019884
 SERIAL: 0704936
 CUST ID: 7089

DATE: 26 May 02
 TECH: PJS
 INST NO: 28151

CALIBRATION DATA TAKEN

INCOMING ✓
 OUTGOING ✓

CONDITION OF EQUIPMENT

IN TOLERANCE ✓
 OUT OF TOLERANCE

INPUT CURRENT VERIFICATION

RANGE	INPUT	MIN	READING	MAX
20 pA	NONE	-00.06	<u>00.01</u>	+00.06

DC VOLTMETER ACCURACY

RANGE	INPUT	MIN	READING	MAX
.2 V	0.0 mV	-.00002	<u>.00000</u>	+.00002
	+190 mV	.18983	<u>.18999</u>	.19017
	-190 mV	.18983	<u>.18999</u>	.19017
2 V	+1.9 V	1.8984	<u>1.9000</u>	1.9016
	-1.9 V	1.8984	<u>1.9000</u>	1.9016
20 V	+19 V	18.984	<u>18.998</u>	19.016
	-19 V	18.984	<u>18.997</u>	19.016

DC CURRENT ACCURACY

RANGE	INPUT	MIN	READING	MAX
200 uA	190 uA	189.3	<u>189.9</u>	190.7
200 nA	190 nA	189.0	<u>190.1</u>	191.0
2000 pA	1900 pA	1871	<u>1899</u>	1929

RESISTANCE ACCURACY

RANGE	ACTUAL INPUT	TOLERANCE	READING
20 K Ω (19)	<u>19.00 kΩ</u>	± 11 C	<u>18.99</u>
200 K Ω	<u>190.0 kΩ</u>	± 11 C	<u>189.7</u>
20 M Ω (10)	<u>10.00 MΩ</u>	± 10 C	<u>9.98</u>
20 G Ω (10)	<u>9.99 GΩ</u>	± 2 %	<u>9.95</u>

CHARGE ACCURACY

RANGE	INPUT	MIN	READING	MAX
2 nC	1.0 VDC INTO 1000 pF	0.9500	<u>.9917</u>	1.0500

**SOUTHWEST RESEARCH INSTITUTE
CALIBRATION LABORATORY
MEMORANDUM**

June 4, 2002

To: DARRELL DUNN DIV20 B57

From: Walt Hill, Metrology Group Leader
Institute Calibration Laboratory *WH*

Subject: Status of Calibration Supplier

Manufacturer/Model: KEITHLEY 614

Description: ELECTROMETER

Serial Number: 0704936

Asset Number: 007089

Work Order Number: 444048529

Date Calibrated: May 26, 2002


Supplier: ROTHE DEVELOPMENT, INC., SAN ANTONIO, TX

Remarks: Rothe Cert # 28151:1022430393

- Supplier is on the Approved Suppliers List (ASL) and is accredited.
- Supplier is on the Approved Suppliers List (ASL) but is not accredited.
- There is no known accredited supplier at this time.
- Supplier is not on the Approved Suppliers List (ASL). Please contact Don Dunavant at ext. 2942 if you wish to add the supplier to the ASL.

Please notify the Institute Calibration Laboratory, extension 5215, of any discrepancies with the item or calibration documentation.

Attachment(s) 2



CERT. NUMBER:
29151:1022430393

ROTHE DEVELOPMENT, INC.
METROLOGY SERVICES DIVISION

4614 SINCLAIR RD., SAN ANTONIO, TEXAS 78222 PH:210-648-3131

CERTIFICATE OF CALIBRATION

ISSUED TO:	Southwest Research Institute 1301 6220 Gulebra San Antonio, TX 78284	MFG:	Keithley	CAL DATE:	5/26/2002
		MODEL:	614	DUE DATE:	5/26/2003
		NOMEN:	Electrometer		
		S/N:	0704936		
		CUST. ID:	7099		

CONTROL NO.: 103 - 29151
TECHNICIAN: 4
SPECIFICATIONS: MFG
PROCEDURE: MFG
WORK ORDER: 002019884
CUSTOMER P.O.: 01571R/ST299491/1.20 007
RECEIVED CONDITION: IN TOLERANCE
RETURNED CONDITION: IN TOLERANCE

CALIBRATION PERFORMED AT: RDMSD
CALIBRATION INTERVAL: 12Mo.
TEMPERATURE: 71.0 °F
RELATIVE HUMIDITY: 39 %
DATE RECEIVED: 5/16/2002

COMMENTS:

ATTACHMENTS: CALIBRATION DATA 2 SHEETS

All Calibrations performed at Rothe Development, Inc. Metrology Services Division meet the requirements of ANSI/NCSL Z540-1-1994, ISO/IEC GUIDE 25, and ISO 10012-1, and are traceable to the National Institute of Standards and Technology. The collective uncertainty of the measurement(s) does not exceed 25% (TUR \geq 4:1) of the instrument specification(s) unless noted in the COMMENTS section.

TR#	MFG	MODEL	SERIAL NO.	DUE DATE
00	FLUKE	57002/EE	1605002	01/11/2002
069	FLUKE	5705A	6595002	01/11/2002
75	JPL	NOON	75	07/07/2003
035	PHILIPINE	6500A	59660	01/11/2002

APPROVED BY:

Peter Stemmer

CMS

QCO

DATE