

S O U T H W E S T R E S E A R C H I N S T I T U T E

Department of Quality Assurance
Calibration Laboratory

CERTIFICATE OF CALIBRATION

Issued to: DIV20 B57 RANDALL MANTEUFEL

Device No: 1444

Manufacturer: SARTORIUS

Model: 3808

Nomenclature: ELECTRONIC BALANCE

Serial Number: 3903006

SwRI No:

Remarks

Accuracy: MFGR SPECS

Procedure: SWRI

ENVIRONMENT


Temperature: 0 Humidity: 0 Location: SWRI BLDG. 57

CONCLUSION

Tolerance/Remarks: Received in tolerance, no adjustments made

Calibration was in accord with requirements of MIL-STD-45662A. Measurements are traceable to the National Institute of Standards and Technology. Inspection and test data are on file and available for inspection.


Signed



Calibration Date: 03/03/93

Record Number: 00010818

Next Calibration Due: 09/03/93



S O U T H W E S T R E S E A R C H I N S T I T U T E

Department of Quality Assurance
Calibration Laboratory

Device Serial No: 3903006

Calibration Date: 03/03/93

STANDARDS


Standard No: 1716	Manufacturer: RICE LAKE	Model: 1 KG
Nomenclature: WEIGHT STANDARD		
Serial No: C871	Cal.Due: 07/08/93	Cal.Rec.No: 00000000
Standard No: 2060	Manufacturer: RICE LAKE	Model: 5 KG
Nomenclature: STANDARD WEIGHT		
Serial No: E302	Cal.Due: 01/08/94	Cal.Rec.No: 00010678
Standard No: 2061	Manufacturer: RICE LAKE	Model: 5 KG
Nomenclature: STANDARD WEIGHT		
Serial No: E204	Cal.Due: 01/08/94	Cal.Rec.No: 00010679
Standard No: 2062	Manufacturer: RICE LAKE	Model: 10 KG
Nomenclature: STANDARD WEIGHT		
Serial No: E203	Cal.Due: 01/08/94	Cal.Rec.No: 00010680

BALANCE CALIBRATION VERIFICATION FORM

DATE: 4 MAR 93 MFR.: SARTORIUS MODEL: 3808
 SER. NO. 3903005 RANGE: 30K CALIBRATION DATES
 TEMPERATURE: ✓ HUMIDITY: ✓ LAST: 22 Oct 9 NEXT: 4 Sep 93
 CALIBRATED BY: JERRY A. WHITE BALANCE TOLERANCES
 PRINT NAME
 BALANCE CAL #: 10818 LIN. TOL.: 2g REP. TOL.: 1/9

RANGE VERIFICATION

	RANGE #1			RANGE #2			RANGE #3		
	CALIBRATION POINTS	CALIBRATION POINTS	CALIBRATION POINTS	CALIBRATION POINTS	CALIBRATION POINTS	CALIBRATION POINTS	CALIBRATION POINTS	CALIBRATION POINTS	
DATA PTS.	1	2	3	4	2	3	1	2	3
RUN #1	1K	10K	20K	25K					
RUN #2	1000.0	10.0000	19.9999	24.9999					
RUN #3	1000.0	10.0000	19.9999	24.9999					
MEAN	1000.0	10.0000	19.9999	24.9999					
STD. DEV.	00	00	CV	00					



POSITION GUIDE

SHIFT VERIFICATION

SELF CALIBRATION Y/N INTERNAL: ✓ EXTERNAL: _____

WEIGHT VALUE	PAN POSITION			
	1	2	3	4
6K	4999.8	5000.2	5000.1	5000.7

COMMENTS: _____

SIGNATURE: Jerry A. White

S O U T H W E S T R E S E A R C H I N S T I T U T E

Department of Quality Assurance
Calibration Laboratory

CERTIFICATE OF CALIBRATION

Issued to: DIV20 B57 RANDALL MANTEUFEL

Device No: 1444

Manufacturer: SARTORIUS

Model: 3808

Nomenclature: ELECTRONIC BALANCE

Serial Number: 3903006

SwRI No:

Cal interval 6 Mo.

Remarks

Accuracy: MFGR SPECS

Procedure: SWRI

ENVIRONMENT

Temperature: 0 Humidity: 0 Location: SWRI BLDG. 57

CONCLUSION

Tolerance/Remarks: Received in tolerance, no adjustments made

Temperature and Humidity at Room temperature.

Calibration was in accord with requirements of MIL-STD-45662A. Measurements are traceable to the National Institute of Standards and Technology. Inspection and test data are on file and available for inspection.

Signed _____

Calibration Date: 09/03/93

Cal interval: 6 Months

Record Number: 00012124

Next Calibration Due: 03/03/94



S O U T H W E S T R E S E A R C H I N S T I T U T E

Department of Quality Assurance
Calibration Laboratory

Device Serial No: 3903006

Calibration Date: 09/03/93

STANDARDS

Standard No: 1716	Manufacturer: RICE LAKE	Model: 1 KG
Nomenclature: WEIGHT STANDARD		
Serial No: C871	Cal.Due: 06/30/94	Cal.Rec.No: 00011768
Standard No: 1719	Manufacturer: RICE LAKE	Model: 5 KG
Nomenclature: WEIGHT STANDARD		
Serial No: C874	Cal.Due: 06/30/94	Cal.Rec.No: 00011771
Standard No: 2060	Manufacturer: RICE LAKE	Model: 5 KG
Nomenclature: STANDARD WEIGHT		
Serial No: E302	Cal.Due: 06/30/94	Cal.Rec.No: 00011775
Standard No: 2061	Manufacturer: RICE LAKE	Model: 5 KG
Nomenclature: STANDARD WEIGHT		
Serial No: E204	Cal.Due: 06/30/94	Cal.Rec.No: 00011774
Standard No: 2062	Manufacturer: RICE LAKE	Model: 10 KG
Nomenclature: STANDARD WEIGHT		
Serial No: E203	Cal.Due: 06/30/94	Cal.Rec.No: 00011776

BALANCE CALIBRATION VERIFICATION FORM

DATE: 3 Sep 93 MFGR: SARTON 108 MODEL: 3808 CALIBRATION DATES

SER. NO. 3503006 RANGE: 30K LAST: 4 MAR 93 NEXT: 3 MAR 94

TEMPERATURE: Room HUMIDITY: Room

CALIBRATED BY: Tony A. White

BALANCE CAL #: 12124 LINE. TOL.: .2 REP. TOL.: .1

BALANCE TOLERANCES

RANGE VERIFICATION

	RANGE #1			RANGE #2			RANGE #3		
	CALIBRATION POINTS	CALIBRATION POINTS	CALIBRATION POINTS	CALIBRATION POINTS	CALIBRATION POINTS	CALIBRATION POINTS	CALIBRATION POINTS	CALIBRATION POINTS	
DATA PTS.	1	2	3	1	2	3	1	2	3
RUN #1	25K	20K	10K	1K					
RUN #2	24999.9	19999.9	10000.0	1000.0					
RUN #3	24999.9	19999.9	10000.0	1000.0					
MEAN	24999.9	19999.9	10000.0	1000.0					
STD. DEV.	—	0	—	—					

SHIFT VERIFICATION

SELF CALIBRATION Y/N INTERNAL: INTERNAL EXTERNAL


WEIGHT VALUE	PAN POSITION			
	1	2	3	4
	4999.8	5000.2	5000.7	5000.2

COMMENTS: _____

SIGNATURE: _____

Tony A. White

POSITION GUIDE



S O U T H W E S T R E S E A R C H I N S T I T U T E

Department of Quality Assurance
Calibration Laboratory

CERTIFICATE OF CALIBRATION

Issued to: DIV20 B57 RON GREEN

Device No: 1444

Manufacturer: SARTORIUS

Model: 3808

Nomenclature: ELECTRONIC BALANCE

Serial Number: 3903006

SwRI No:

Cal interval 6 Mo.

Remarks

Accuracy: MFGR SPECS

Procedure: SWRI

ENVIRONMENT

Temperature:

Humidity:

Location: SWRI DIV20 BLDG. 51


CONCLUSION

Tolerance/Remarks: Received in tolerance, no adjustments made

Room Temperature.

Calibration was in accord with requirements of MIL-STD-45662A. Measurements are traceable to the National Institute of Standards and Technology. Inspection and test data are on file and available for inspection.

Signed



Calibration Date: 03/04/94

Cal interval: 6 Months

Record Number: 00013579

Next Calibration Due: 09/04/94

S O U T H W E S T R E S E A R C H I N S T I T U T E

Department of Quality Assurance
Calibration Laboratory

Device Serial No: 3903006

Calibration Date: 03/04/94

STANDARDS

Standard No: 1716	Manufacturer: RICE LAKE	Model: 1 KG
Nomenclature: WEIGHT STANDARD		
Serial No: C871	Cal.Due: 06/30/94	Cal.Rec.No: 00011768
Standard No: 1719	Manufacturer: RICE LAKE	Model: 5 KG
Nomenclature: WEIGHT STANDARD		
Serial No: C874	Cal.Due: 06/30/94	Cal.Rec.No: 00011771
Standard No: 2060	Manufacturer: RICE LAKE	Model: 5 KG
Nomenclature: STANDARD WEIGHT		
Serial No: E302	Cal.Due: 06/30/94	Cal.Rec.No: 00011775
Standard No: 2061	Manufacturer: RICE LAKE	Model: 5 KG
Nomenclature: STANDARD WEIGHT		
Serial No: E204	Cal.Due: 06/30/94	Cal.Rec.No: 00011774
Standard No: 2062	Manufacturer: RICE LAKE	Model: 10 KG
Nomenclature: STANDARD WEIGHT		
Serial No: E203	Cal.Due: 06/30/94	Cal.Rec.No: 00011776

BALANCE CALIBRATION VERIFICATION FORM

DATE: 4 MAR 94 MFGR: SARTORIUS MODEL: 2808
 SER. NO. 3903006 RANGE: _____ CALIBRATION DATES
 TEMPERATURE: Pan HUMIDITY: Pan LAST: 3 SEP 93 NEXT: 4 SEP 94
 CALIBRATED BY: J. WHITE
 BALANCE CAL #: 13579 LINE. TOL.: .2 REP. TOL.: .1

BALANCE TOLERANCES

	RANGE VERIFICATION											
	RANGE #1			RANGE #2			RANGE #3			RANGE #4		
	CALIBRATION POINTS			CALIBRATION POINTS			CALIBRATION POINTS			CALIBRATION POINTS		
DATA PTS.	1	2	3	1	2	3	1	2	3	1	2	3
RUN #1	24999.9	10,000.0	1000.0									
RUN #2	24999.9	10,000.0	1000.0									
RUN #3	25000.0	10,000.0	1000.0									
MEAN	24999.9			10,000.0			1000.0					
STD. DEV.												


SHIFT VERIFICATION

WEIGHT VALUE	PAN POSITION			
	1	2	3	4
500g	499.8	500.3	500	500.1

COMMENTS: _____

SIGNATURE: _____

POSITION GUIDE



SOUTHWEST RESEARCH INSTITUTE

**Department of Quality Assurance
Calibration Laboratory**

**CERTIFICATE OF CALIBRATION
09/15/94**

Issued to: RANDALL MANTEUFEL DIV20 ,B57
Manufacturer: SARTO
Nomenclature: ELECTRONIC BALANCE
Serial Number: 3903006

Asset Number: 001444
Model Number: 3808
SwRI Capital Number: UNKWN

ENVIRONMENTAL CONDITIONS

Temperature: 0.0 F

Relative Humidity: 0 %

CALIBRATION INFORMATION

Location: B51
Procedure Number: SWRI
Remarks: CALIBRATED AT ROOM TEMP AND HUMID


Technician: 7213
Accuracy: MFGR SPECS
Received IN Tolerance

**Calibration was in accordance with requirements of MIL-STD-45662A.
Measurements are traceable to the National Institute of Standards and Tech-
nology. Inspection and test data are on file and available for inspection.**

STANDARDS USED FOR CERTIFICATION

Asset #	Serial #	Mfg	Model #	Nomenclature	Cal Date	Int.	Cal Due
002062	E203	RICE	10 KG	STANDARD WEIGHT	07/19/94	12	07/19/95
002061	E204	RICE	5 KG	STANDARD WEIGHT	07/19/94	12	07/19/95
002060	E302	RICE	5 KG	STANDARD WEIGHT	07/19/94	12	07/19/95
001719	C874	RICE	5 KG	WEIGHT STANDARD	07/20/94	12	07/20/95
001716	C871	RICE	1 KG	WEIGHT STANDARD	07/20/94	12	07/20/95

Certified by :



Calibration Date: 09/14/94
Interval: 6 months
Next Calibration Due: 03/14/95



Work Order: 15131

BALANCE CALIBRATION VERIFICATION FORM

DATE: 14 Sep 94 MFGR: SANTORIUS MODEL: 3808
 SER. NO. 3203006 RANGE: 30K CALIBRATION DATES
 TEMPERATURE: Room HUMIDITY: Room LAST: AMANSY NEXT: 14 MARCH
 CALIBRATED BY: Terry A. White
 BALANCE CAL #: 15131 LINE. TOL.: .2 REP. TOL.: .1
 BALANCE TOLERANCES

RANGE VERIFICATION

	RANGE #1			RANGE #2			RANGE #3		
	CALIBRATION POINTS	CALIBRATION POINTS	CALIBRATION POINTS	CALIBRATION POINTS	CALIBRATION POINTS	CALIBRATION POINTS	CALIBRATION POINTS	CALIBRATION POINTS	
DATA PTS.	1	2	3	1	2	3	1	2	3
RUN #1	25K ₁	10K ₀	1K ₅	/					
RUN #2	24999.9	10000.0	1000.0						
RUN #3	24555.9	10000.0	1000.0						
MEAN									
STD. DEV.									



SHIFT VERIFICATION

SELF CALIBRATION Y/N INTERNAL: INTERNAL EXTERNAL

WEIGHT VALUE	PAN POSITION			
	1	2	3	4
5K	4997.8	5000.1	5000.1	5000.0

COMMENTS:

SIGNATURE: [Signature]

SOUTHWEST RESEARCH INSTITUTE

**Department of Quality Assurance
Calibration Laboratory**

**CERTIFICATE OF CALIBRATION
03/16/95**

Issued to: **RON GREEN** **DIV20** **,B57**
Manufacturer: **SARTO**
Nomenclature: **ELECTRONIC BALANCE**
Serial Number: **3903006**

Asset Number: **001444**
Model Number: **3808**

SwRI/Div. I.D. #: **UNKWN**

ENVIRONMENTAL CONDITIONS

Temperature: **78.0F**

Relative Humidity: **40%**

CALIBRATION INFORMATION

Procedure Number: **SWRI**
Remarks:


Accuracy: **MFGR SPECS**
Received IN Tolerance

**Calibration was in accordance with requirements of MIL-STD-45662A.
Measurements are traceable to the National Institute of Standards and Tech-
nology. Inspection and test data are on file and available for inspection.**

STANDARDS USED FOR CERTIFICATION

Asset #	Serial #	Mfg	Model #	Nomenclature	Cal Date	Int.	Cal Due
002062	E203	RICE	10 KG	STANDARD WEIGHT	07/19/94	12	07/19/95
002061	E204	RICE	5 KG	STANDARD WEIGHT	07/19/94	12	07/19/95
002060	E302	RICE	5 KG	STANDARD WEIGHT	07/19/94	12	07/19/95
001719	C874	RICE	5 KG	WEIGHT STANDARD	07/20/94	12	07/20/95
001716	C871	RICE	1 KG	WEIGHT STANDARD	07/20/94	12	07/20/95

Certified by :



Calibration Date: **03/14/95**
Interval: **6 months**
Next Calibration Due: **09/14/95**

Certificate#: **16723**

BALANCE CALIBRATION VERIFICATION FORM

DATE: 14 MAR 95 MFGR: SARTORIUS MODEL: 3808


SER. NO. 3903006 RANGE: 30K
 266 CALIBRATION DATES
 TEMPERATURE: _____ HUMIDITY: 40 LAST: 14 SEPT 94 NEXT: 14 SEPT 95

CALIBRATED BY: TERRY P. WHITE

BALANCE TOLERANCES

BALANCE CAL #: 16723 LINE. TOL.: .2 REP. TOL.: .1

	RANGE VERIFICATION												POSITION GUIDE
	RANGE #1			RANGE #2			RANGE #3			RANGE #4			
	CALIBRATION POINTS	CALIBRATION POINTS	CALIBRATION POINTS	CALIBRATION POINTS	CALIBRATION POINTS	CALIBRATION POINTS	CALIBRATION POINTS	CALIBRATION POINTS	CALIBRATION POINTS	CALIBRATION POINTS	CALIBRATION POINTS	CALIBRATION POINTS	
DATA PTS.	1	2	3	1	2	3	1	2	3	1	2	3	
RUN #1	25K	10K	1K										
RUN #2	25000.1	10000.1	1000.0										
RUN #3	25000.1	10000.1	1000.0										
MEAN													
STD. DEV.													
SHIFT VERIFICATION													
SELF CALIBRATION Y/N INTERNAL: _____ EXTERNAL: _____													
COMMENTS: _____													
SIGNATURE: <u>Terry P. White</u>													
WEIGHT VALUE													
PAN POSITION													
5K	4899.8	5000.2	4999.9	5000.1	5000.1	5000.1	5000.1	5000.1	5000.1	5000.1	5000.1	5000.1	



SOUTHWEST RESEARCH INSTITUTE

**Department of Quality Assurance
Calibration Laboratory**

**CERTIFICATE OF CALIBRATION
02/14/96**

Issued to: RON GREEN DIV20 ,B57
Manufacturer/Model: SARTO/3808
Nomenclature: ELECTRONIC BALANCE
Serial Number: 3903006
Asset Number: 001444
Notes:

ENVIRONMENTAL CONDITIONS

Temperature: 72.0F

Relative Humidity: 23%

CALIBRATION INFORMATION

Procedure Number: MFGR
Remarks:


Accuracy: MFGR SPECS
Received IN Tolerance

**Calibration was in accordance with requirements of MIL-STD-45662A.
Measurements are traceable to the National Institute of Standards and Tech-
nology. Inspection and test data are on file and available for inspection.**

STANDARDS USED FOR CERTIFICATION

Asset #	Serial #	Mfg	Model #	Nomenclature	Cal Date	Int.	Cal Due
002062	E203	RICE	10 KG	STANDARD WEIGHT	06/23/95	12	06/23/96
002061	E204	RICE	5 KG	STANDARD WEIGHT	06/23/95	12	06/23/96
002060	E302	RICE	5 KG	STANDARD WEIGHT	06/23/95	12	06/23/96
001719	C874	RICE	5 KG	WEIGHT STANDARD	06/23/95	12	06/23/96
001716	C871	RICE	1 KG	WEIGHT STANDARD	06/23/95	12	06/23/96

Certified by :



Certificate#: 202005

Calibration Date: 02/13/96
Interval: 6 months
Next Calibration Due: 08/13/96

BALANCE CALIBRATION VERIFICATION FORM

DATE: 13 Feb 96 MFGR: SAATCHI'S MODEL: 3808


SER. NO. 3902005 RANGE: 30K CALIBRATION DATES

TEMPERATURE: 72 HUMIDITY: 23 LAST: 6 Sept 95 NEXT: 13 AUG 96

CALIBRATED BY: Jerry R. White BALANCE TOLERANCES

BALANCE CAL #: 20205 LINE. TOL: .2 REP. TOL: .1

	RANGE VERIFICATION								
	RANGE #1			RANGE #2			RANGE #3		
	1	2	3	1	2	3	1	2	3
DATA PTS.	25K	10K	1K						
RUN #1	24999.9	10000.0	1000.0						
RUN #2	24999.9	10000.0	1000.0						
RUN #3	24999.9	10000.0	1000.0						
MEAN									
STD. DEV.									



SHIFT VERIFICATION			
WEIGHT VALUE	PAN POSITION		
	1	2	3
5K	5000.2	5000.1	5000.1

SELF CALIBRATION Y/N INTERNAL: _____ EXTERNAL: _____

COMMENTS: _____

SIGNATURE: [Signature]

SOUTHWEST RESEARCH INSTITUTE

Department of Quality Assurance

Calibration Laboratory • 522-5215

WORK ORDER

CERTIFICATE # 24877 ASSET # 001444 DATE 28 MAR 97

ITEM DATA:

Manufacturer Sartorius Model 3808

Description Electronic balance Serial # 3903006

Accessories _____

ACTION REQUESTED cal

CUSTODIAN Div. 20, Ron Fisen

Turned in by: _____ Phone _____

CHARGE # _____ Date Required _____

INSTRUMENT USED ON: NUCLEAR DOD NASA GLP SPPE

OTHER _____

COPY OF CALIBRATION CERTIFICATE Yes No

- CONDITION RECEIVED:
- Out of tolerance, repaired to specifications
 - In tolerance, minor adjustments/repairs made
 - In tolerance, no adjustments/repairs
 - Out of tolerance, adjusted to specifications
 - Received into system, introduced or reactivated
 - Calibration interval
 - Reliability code

ACTION TAKEN: (Calibration/Repair/Parts) cal

CAL ENVIRONMENT:

Temperature 74 °F Humidity 38 %RH

CALIBRATED/REPAIRED:

By [Signature] Cal Procedure CLP-UB-001

Date 28 MAR 97 Accuracy MS

Cal Interval 6 Time to complete: _____

Next Cal due 28 Sep 97 Cal _____ Repair _____

Standards used (Asset#) 2062 2062 2060 1719 1718 1717 1716

DATE COMPLETED 28 MAR 97

DATE PICKED UP _____ PICKED UP BY [Signature]

24877

(202005)

14674

B51

BALANCE CALIBRATION VERIFICATION FORM

DATE: 28 MAR 97 MFGR: Sartorius MODEL: 3808

SER. NO. 3902005 RANGE: 30K CALIBRATION DATES

TEMPERATURE: 24 HUMIDITY: 38 LAST: 13 FEB 96 NEXT: 28 SEPT 97



CALIBRATED BY: Jerry A. White

BALANCE TOLERANCES

BALANCE CAL #: 24877 LINE TOL: 2 REP. TOL: 1

RANGE VERIFICATION

	RANGE #1			RANGE #2			RANGE #3		
	1	2	3	1	2	3	1	2	3
DATA PTS.	30K	20K	5K	5K	1K				
RUN #1	29999.8	19999.9	5000.0	1000.0					
RUN #2	/	/	/	/					
RUN #3	/	/	/	/					
MEAN									
STD. DEV.									



SHIFT VERIFICATION

SELF CALIBRATION Y/N INTERNAL: _____ EXTERNAL: _____

WEIGHT VALUE	PAN POSITION			
	1	2	3	4

COMMENTS: _____

SIGNATURE: _____



Southwest Research Institute
6220 Culebra Road
San Antonio, TX 78238
Department of Quality Assurance
Calibration Laboratory

Certificate of Calibration

28 March 1997

Issued to: RON GREEN DIV20 B57
Manufacturer/Model: SARTO 3808
Description: ELECTRONIC BALANCE
Serial Number: 3903006
Asset Number: 001444

Environmental Conditions

Temperature: 74.0 Deg. F Humidity: 38%

Calibration Information

Calibration was in accordance with requirements of MIL-STD-45662A and ANSI/NCSL Z540-1-1994. Measurements are traceable to the National Institute of Standards and Technology (NIST). This report may not be reproduced except in full without written approval of the originator. Inspection and test data are on file and available for inspection.

Calibration Date: 28 Mar 97 Calibration Procedure: CLCP-WB-001
Interval: 6 months Accuracy: MFGR SPECS
Next Calibration Due: 28 Sep 97 Received: In Tolerance

Remarks:

Standards Used

Asset	MFR	Model	Description	Serial No.	Due Cal
001716	RICE	1 KG	WEIGHT STANDARD	C871	19 Jul 97
001717	RICE	2 KG	WEIGHT STANDARD	C872	19 Jul 97
001718	RICE	2 KG	WEIGHT STANDARD	C873	19 Jul 97
001719	RICE	5 KG	WEIGHT STANDARD	C874	19 Jul 97
002060	RICE	5 KG	STANDARD WEIGHT	E302	19 Jul 97
002061	RICE	5 KG	STANDARD WEIGHT	E204	19 Jul 97
002062	RICE	10 KG	STANDARD WEIGHT	E203	19 Jul 97

Certificate # 24877

Signed: 

SOUTHWEST RESEARCH INSTITUTE
Department of Quality Assurance
Calibration Laboratory • 522-5215

WORK ORDER

CERTIFICATE # 30920 ASSET # 00444 DATE 20 Aug 98

ITEM DATA:
Manufacturer SARTORIUS Model 3808
Description Balance Serial # 3903006
Accessories _____

ACTION REQUESTED cal on site

CUSTODIAN Div. 20 B51 Ron Green

Turned in by: Melissa Hill Phone 2012

CHARGE # 20-04 Date Required _____

INSTRUMENT USED ON: (DOD/NASA) (NUCLEAR) (GLP) (SPPE) (ISO)
 OTHER _____

COPY OF CALIBRATION CERTIFICATE (Yes) (No)

NEW WORK Yes No If yes, an evaluation shall be made to verify capabilities.

By Melissa Hill Date 08-20-98

Work involves proprietary/confidential information or equipment (Yes) (No)

CONDITION RECEIVED: _____ (F) Out of tolerance, repaired to specifications
_____ (G) In tolerance, minor adjustments/repairs made
_____ (J) In tolerance, no adjustments/repairs
_____ (K) Out of tolerance, adjusted to specifications
_____ (S) Received into system, introduced or reactivated

ACTION TAKEN: (Calibration/Repair/Parts) _____
_____ cal

CAL ENVIRONMENT:
Temperature 79 °F Humidity 47 %RH

CALIBRATED/REPAIRED:
By [Signature] Cal Procedure QACP-WI-001
Date 20 AUG 98 Accuracy [Signature]
Cal Interval 65 Reliability Code: _____
Next Cal due 20 FEB 99 Cal Time 1.5 Repair Time _____

Standards used (Asset#) 2060 2061 2062 1719 1716

DATE COMPLETED _____
DATE PICKED UP cal on site PICKED UP BY _____

30920

2012

BALANCE CALIBRATION VERIFICATION FORM

DATE: 20 AUG 98
 BALANCE CAL NO.: 30970
 TEMPERATURE: 79
 HUMIDITY: 47
 BARO. PRESSURE: 14.70

MFGR: SMITHLUS
 MODEL: 380
 SERIAL NO: 332001
 ASSET NO: 1344
 RANGE:
 MASS UNCERTAINTY (U_m):

MFGR BALANCE TOLERANCES
 LINEARITY:
 ECCENTRICITY:
 REPRODUCIBILITY:
 TEMP. DRIFT/°C:
 COMBINED MFR SPECS (U_{mfr}):

TUR = Comb. MFR Specs (U_{mfr}) = 1
 U_{m-stc}

CALIBRATION DATES
 LAST CAL: 9 SEP 97
 NEXT CAL: 20 FEB 99

If TUR < 4:1, U_o = _____

1714

	RANGE VERIFICATION			
	RANGE 1	RANGE 2	RANGE 3	
	CALIBRATION POINTS			CALIBRATION POINTS
DATA POINTS	25K	15K	1K	20K
RUN #1	24999.9	15000.0	1.0000	20000.0
RUN #2				
RUN #2				
MEAN				
STD. DEV.				



SHIFT VERIFICATION

	PAN POSITION			
	1	2	3	4
ACTUAL WEIGHT IN CENTER				
DIFF.				

COMMENTS: _____

 SIGNATURE: [Signature]

MEASURED UNCERTAINTY:

Linearity ± _____ g
 Eccentricity ± _____ g
 Comb. Uncertainty = _____ g
 Rep (σ) = _____ g
 Mass tol. (2σ) = _____ g
 Comb. Uncertainty = _____ g

Std Uncertainty = _____ g
 Std. Uncertainty = _____ g

Expanded Unc (k = 2) = _____ g

SWRI CALIBRATION LAB
 (210) 522-5215
 CAL BY [Signature]
 DUE 20 FEB 99
 SN 332001



Southwest Research Institute
 6220 Culebra Road
 San Antonio, TX 78238
 Department of Quality Assurance
 Calibration Laboratory

ACCREDITED



Certificate #
0972-01

Certificate of Calibration

21 August 1998

Issued to: RON GREEN DIV20 B57
 Manufacturer/Model: SARTORIOUS 3808
 Description: ELECTRONIC BALANCE
 Serial Number: 3903006
 Asset Number: 001444

Calibration Information

Calibration was in accordance with requirements of MIL-STD-45662A and ANSI/NCSL Z540-1-1994. Measurements are traceable to the National Institute of Standards and Technology (NIST). This report may not be reproduced except in full without written approval of the originator. Inspection and test data are on file and available for inspection.



This laboratory is accredited by the American Association for Laboratory Accreditation (A2LA) and the results shown in this calibration certificate have been determined in accordance with the laboratory's terms of accreditation unless stated otherwise in the report. The uncertainty of the calibration was sufficient to determine that the instrument met the manufacturer's specifications.

Temperature: 79.0 Deg. F Humidity: 47 % RH
 Calibration Date: 20 Aug 98 Calibration Procedure: CLCP-WT-001
 Interval: 6 months Received: In Tolerance
 Next Calibration Due: 20 Feb 99

Remarks:

Standards Used

Asset	MFR	Model	Description	Serial No.	Due Cal
002061	RICE LAKE	5KG	WEIGHT STANDARD	E204	26 Jun 99
001716	RICE LAKE	1KG	WEIGHT STANDARD	C871	26 Jun 99
001719	RICE LAKE	5KG	WEIGHT STANDARD	C874	26 Jun 99
002060	RICE LAKE	5KG	WEIGHT STANDARD	E302	26 Jun 99
002062	RICE LAKE	10KG	WEIGHT STANDARD	E203	26 Jun 99

Signed: 
 Title: 



Southwest Research Institute
 6220 Culebra Road
 San Antonio, TX 78238
 Department of Quality Assurance
 Calibration Laboratory

Certificate of Calibration

9 September 1997

Issued to: RON GREEN DIV20 B57
 Manufacturer/Model: SARTO 3808
 Description: ELECTRONIC BALANCE
 Serial Number: 3903006
 Asset Number: 001444

Environmental Conditions

Temperature: 74.0 Deg. F Humidity: 65%

Calibration Information

Calibration was in accordance with requirements of MIL-STD-45662A and ANSI/NCSL Z540-1-1994. Measurements are traceable to the National Institute of Standards and Technology (NIST). This report may not be reproduced except in full without written approval of the originator. Inspection and test data are on file and available for inspection.


Calibration Date: 9 Sep 97 Calibration Procedure: CLCP-WT-001
 Interval: 6 months Accuracy: MFGR SPECS
 Next Calibration Due: 9 Mar 98 Received: In Tolerance

Remarks:

Standards Used

Asset	MFR	Model	Description	Serial No.	Due Cal
001716	RICE LAKE	1KG	WEIGHT STANDARD	C871	21 Jul 98
001719	RICE LAKE	5KG	WEIGHT STANDARD	C874	21 Jul 98
002060	RICE LAKE	5KG	WEIGHT STANDARD	E302	21 Jul 98
002061	RICE LAKE	5KG	WEIGHT STANDARD	E204	21 Jul 98
002062	RICE LAKE	10KG	WEIGHT STANDARD	E203	21 Jul 98

Certificate # 26717

Signed: 

SOUTHWEST RESEARCH INSTITUTE
Department of Quality Assurance
Calibration Laboratory • 522-5215

WORK ORDER

CERTIFICATE # 33419 ASSET # 1444 DATE 19 Feb 99

ITEM DATA:

Manufacturer SANTORIUS Model 3808
Description ELECTRONIC BALANCE Serial # 390300C
Accessories _____

ACTION REQUESTED OnL

CUSTODIAN DIV 20 RIC GREEN

Turned in by: _____ Phone _____

CHARGE # _____ Date Required _____

INSTRUMENT USED ON: (DOD/NASA) (NUCLEAR) (GLP) (SPPE) (ISO)
 OTHER _____

COPY OF CALIBRATION CERTIFICATE (Yes) (No)

NEW WORK Yes No If yes, an evaluation shall be made to verify capabilities.

By _____ Date _____

Work involves proprietary/confidential information or equipment (Yes) (No)

- CONDITION RECEIVED: _____ (F) Out of tolerance, repaired to specifications
_____ (G) In tolerance, minor adjustments/repairs made
_____ (J) In tolerance, no adjustments/repairs
_____ (K) Out of tolerance, adjusted to specifications
_____ (S) Received into system, introduced or reactivated

33419

ACTION TAKEN: (Calibration/Repair/Parts) OnL

CAL ENVIRONMENT:

Temperature 69 °F Humidity 33 %RH

CALIBRATED/REPAIRED:

By [Signature] Cal Procedure QCCP-WT-001
Date 19 Feb 99 Accuracy mk
Cal Interval 1 Reliability Code: _____
Next Cal due 15 April 99 Cal Time 1.5 Repair Time _____
Standards used (Asset#) 2062 2061 2060 1719 1718 1717 1714

DATE COMPLETED 19 Feb 99 QAL OK 517

DATE PICKED UP _____ PICKED UP BY _____

BALANCE CALIBRATION VERIFICATION FORM

DATE 19 Feb 99
 BALANCE CAL NO: 333419
 TEMPERATURE: 69
 HUMIDITY: 33
 BARO. PRESSURE: 14.34



MFGR: Sartorius
 MODEL: 3108
 SERIAL NO: 3501001
 ASSET NO: 1334
 RANGE: _____
 MASS UNCERTAINTY (U_m): _____


MFGR BALANCE TOLERANCES
 LINEARITY: _____
 ECCENTRICITY: _____
 REPRODUCIBILITY: _____
 TEMP. DRIFT/°C: _____
 COMBINED MFR SPECS (U_{mfr}): _____

CALIBRATION DATES
 LAST CAL: 20 AUG 98
 NEXT CAL: 19 AUG 99

TUR = Comb. MFR Specs (U_{mfr}) = 1
 U_{m-sig}

If TUR < 4:1, U₃ = _____

		RANGE VERIFICATION			
		RANGE 1	RANGE 2	RANGE 3	RANGE 4
DATA POINTS	CALIBRATION POINTS	CALIBRATION POINTS	CALIBRATION POINTS	CALIBRATION POINTS	POSITION GUIDE
30K	25	10 ⁵	5		
RUN #1	30000.1	25000.0	15000.0	5000.0	
RUN #2	/				
RUN #3					
MEAN					
STD. DEV.					
SHIFT VERIFICATION					
SELF CALIBRATION Y/N INTERNAL: <input checked="" type="checkbox"/> EXTERNAL: _____					
PAN POSITION					
ACTUAL WEIGHT CENTER	1	2	3	4	
DIFF					

COMMENTS: _____
 SIGNATURE: 

MEASURED UNCERTAINTY:

Linearity = _____ g
 Eccentricity = _____ g
 Comb Uncertainty = _____ g
 Rep (σ) = _____ g
 Mass tol. (2σ) = _____ g
 Comb Uncertainty = _____ g

Std Uncertainty = _____ g
 Std Uncertainty = _____ g

Expanded Unc. (k=2) = _____ g



Southwest Research Institute
6220 Culebra Road
San Antonio, TX 78238
(210) 522-5215
Department of Quality Assurance
Calibration Laboratory

Accredited



Certificate #
0972-01

Certificate of Calibration

22 February 1999

Issued to: RON GREEN DIV20 B57
Manufacturer/Model: SARTORIOUS 3808
Description: ELECTRONIC BALANCE
Serial Number: 3903006
Asset Number: 001444

This certifies the above item was calibrated in compliance with MIL-STD-45662A and ANSI/NCSL Z540-1-1994. Standards used in this calibration, described in the referenced calibration procedure with associated uncertainties or tolerances, are traceable to the National Institute of Standards and Technology (NIST). Supporting documentation relative to traceability is on file and is available for examination upon request. This certificate is not to be reproduced, except in full, without the written approval of the Southwest Research Institute Department of Quality Assurance Calibration Laboratory.

This laboratory is accredited by the American Association for Laboratory Accreditation (A2LA) and the results of this calibration certificate were determined in accordance with the terms of accreditation unless stated otherwise below.

The uncertainty of the calibration was sufficient to determine that the item met the manufacturer's published specifications unless stated otherwise below.


Ambient Conditions: Temperature: 69.0 Degrees Fahrenheit Humidity: 33 % RH

Calibration Date: 19 Feb 99 **Calibration Procedure:** CLCP-WT-001

Condition as Received: IN TOLERANCE

Condition as Released: IN TOLERANCE

Remarks:



Jerry White / Technician

SOUTHWEST RESEARCH INSTITUTE
Department of Quality Assurance
Calibration Laboratory • 522-5215

B51-
19 Aug 99

WORK ORDER

WORK ORDER # 35654 ASSET # 1444 DATE 16 Aug 99

ITEM DATA:

Manufacturer SARTORIUS Model 3808
Description BALANCE Serial # 3903006
Accessories _____

ACTION REQUESTED Cal

CUSTODIAN RON GREEN

Turned in by: _____ Phone _____

CHARGE # 20-0751-006 Date Required _____

INSTRUMENT USED ON: (DOD/NASA) (NUCLEAR) (GLP) (SPPE) (ISO)
 OTHER _____

COPY OF CALIBRATION CERTIFICATE (Yes) (No)

NEW WORK Yes No If yes, an evaluation shall be made to verify capabilities.

By _____ Date _____

Work involves proprietary/confidential information or equipment (Yes) (No)

CONDITION RECEIVED: Out of tolerance
 In tolerance
 Damaged (Contact customer)
 Contact _____ Date _____
 Received into system, introduced or reactivated

ACTION TAKEN: (Calibration/Repair/Parts) cal

35654

CAL ENVIRONMENT:

Temperature 71 °F Humidity 47 %RH

CALIBRATED/REPAIRED

By [Signature] Cal Procedure CLCP-WT-001 AUG 97

Date 19 AUG 99 Accuracy _____

Cal Interval 2 Reliability Code _____

Next Cal Due 15 FEB 00 Cal Time _____ Repair Time _____

Standards used (Asset #) 20064 6160 1715 17 16

DATE COMPLETED 19 AUG 99 Cal on [Signature]

DATE PICKED UP _____ PICKED UP BY _____

BALANCE CALIBRATION VERIFICATION FORM

DATE 15 AUG 99
 BALANCE CAL NO: 25689
 TEMPERATURE 77
 HUMIDITY 47
 BARO PRESSURE 14.25

MFR: SARTORIUS
 MODEL: 340P
 SERIAL NO: 3903001
 ASSET NO: 2444
 RANGE: 30K
 MASS UNCERTAINTY (U_m): _____

MFR BALANCE TOLERANCES

LINEARITY: _____
 ECCENTRICITY: _____
 REPRODUCIBILITY: _____
 TEMP. DRIFT/°C: _____
 COMBINED MFR SPECS (U_{max}): _____



LAST CAL: 19 FEB 99
 NEXT CAL: 19 FEB 00
 TUR = Comb MFR Specs (U_{max}) = 1
 U_{m-95%} _____

If TUR < 4.1, U₃ = _____

2002 61 60 17A 18 17 K

DATA POINTS	RANGE VERIFICATION		
	RANGE 1	RANGE 2	RANGE 3
	CALIBRATION POINTS	CALIBRATION POINTS	CALIBRATION POINTS
RUN #1	30K 15K 10K SK 1K		
RUN #2	1799.7 1499.9 1000.0 500.0 100.0		
RUN #3			
MEAN			
STD. DEV			

POSITION GUIDE



SHIFT VERIFICATION

SELF CALIBRATION Y/N INTERNAL: EXTERNAL: _____

ACTUAL WEIGHT IN CENTER	PAN POSITION			
	1	2	3	4
DIFF				

COMMENTS:

SWRI CALIBRATION LAB

(210) 522-5210

CAL: 8/16/99 BY: DM

DUP: 5 AUG 99 ID: 2444

SIN: 3903001

SIGNATURE: [Signature]

MEASURED UNCERTAINTY:

Linearity = _____ g
 Eccentricity = _____ g
 Comb Uncertainty = _____ g

Std Uncertainty = _____ g
 Std Uncertainty = _____ g

Rep (σ) = _____ g
 Mass tol (2σ) = _____ g
 Comb Uncertainty = _____ g

(σ) = _____ g
 (linearity, eccentricity, reproducibility, mass)



Southwest Research Institute
6220 Culebra Road
San Antonio, TX 78238
(210) 522-5215
Department of Quality Assurance
Calibration Laboratory

Accredited



Certificate #
0972-01

Certificate of Calibration

24 August 1999

Issued to: RON GREEN DIV20 B57
Manufacturer/Model: SARTORIUS 3808
Description: ELECTRONIC BALANCE
Serial Number: 3903006
Asset Number: 001444

This certifies the above item was calibrated in compliance with MIL-STD-45662A and ANSI/NCSL Z540-1-1994. Standards used in this calibration, described in the referenced calibration procedure with associated uncertainties or tolerances, are traceable to the National Institute of Standards and Technology (NIST). Supporting documentation relative to traceability is on file and is available for examination upon request. This certificate is not to be reproduced, except in full, without the written approval of the Southwest Research Institute Department of Quality Assurance Calibration Laboratory.

This laboratory is accredited by the American Association for Laboratory Accreditation (A2LA) and the results of this calibration certificate were determined in accordance with the terms of accreditation unless stated otherwise below.

The uncertainty of the calibration was sufficient to determine that the item met the manufacturer's published specifications unless stated otherwise below.

Ambient Conditions: Temperature: 77.0 Degrees Fahrenheit Humidity: 47 % RH

Calibration Date: 19 Aug 99


Calibration Procedure: CLCP-WT-001 AUG 97

Condition as Received: IN TOLERANCE

Condition as Released: IN TOLERANCE

Remarks:


Approved by:


Jim Patterson, Supervisor or Walt Hill, Metrologist

Certificate # 35654

m:\a2la.rpt Rev date 10 Mar 99

Measurements performed by:


Jerry White, Technician

Page 1 of 1

WORK ORDER 37737

Date Received 2/9/00

Asset No. 001444 Manufacturer SARTORIUS Model 3808
Description ELECTRONIC BALANCE Serial Number 3903006
Accessory Received/Required NONE
Div/CC ID NONE Accessory to Asset No. N/A Accuracy MFG SPECS
Div/CC DIV20 Location B57 Custodian RON GREEN Tel. 5305
Charge/Project No. 20.00751.006 Proprietary/Confidential N Date Required ROUTINE
Work Requested CALIBRATION ONSITE
Receiving Inspection _____
Delivered By N/A Tel. 5305

WORK HISTORY

Date	Start Time	Stop Time	Notes

PARTS

Part Name	Part Number	Cost	Failure Description

WORK SUMMARY

Failure Description _____

Repair Action _____

Cal Procedure _____ Temp 71 F Hum 45 %

Tech JW Cal Hrs. 1.5 Repair Hrs. _____ Part Cost _____

Action Taken OK

Standards Used 2062 2060 2061 1719 1718 1717 1716 1717

Date Cal 9 Feb 00 Int. 6 Mo. Date Due _____ Reliability Code _____

Date Picked Up _____ Picked Up By _____

37737

ELECTRONIC BALANCE CALIBRATION DATA SHEET

WORK ORDER Station 5 DATE 9 Feb 00 TECHNICIAN J. Paul
 MODEL 3807 SERIAL NO. 71 ASSET NO. _____
 LOCATION B51
 AMBIENT: TEMP 71 HUMIDITY 48 BARO PRESS 14.41

1) CALIBRATION CHECK

AS FOUND FULL-CAPACITY INDICATION 29999.0 *adjusted*
 POST-CALIBRATION INDICATION 30000.1 TOLERANCE _____ P/F _____

2) REPEATABILITY

1	10000.1	6	10000.1
2	10000.1	7	10000.1
3	10000.1	8	10000.1
4	10000.1	9	10000.1
5	10000.1	10	10000.1

S.D. _____ TOLERANCE _____ P/F _____

3) OFF-CENTER ERROR

1	9999.7	3	9999.9
2	10000.3	4	10000.3

TOLERANCE _____ P/F _____

4) NON-LINEARITY

TEST POINT	INDICATION	NON-LIN ERROR	TEST WEIGHT S/N
0		—	—
25%	7500.1		
50%	7500.1		
75%	7500.1		
100%	7500.1		

TOLERANCE _____ P/F _____

5) COMMENTS:

SWRI CALIBRATION LAB
 (210) 522-5215
 CAL BY J. Paul
 DUE 9 Feb 00 ID 7444
 SN 578 306



Southwest Research Institute
6220 Culebra Road
San Antonio, TX 78238
(210) 522-5215
Department of Quality Assurance
Calibration Laboratory

Certificate of Calibration

9 February 2000

Issued to: RON GREEN DIV20 B57
Manufacturer/Model: SARTORIUS 3808
Description: ELECTRONIC BALANCE
Serial Number: 3903006
Asset Number: 001444

This certifies the above item was calibrated in compliance with MIL-STD-45662A and ANSI/NCSL Z540-1-1994. Standards used in this calibration, described in the referenced calibration procedure with associated uncertainties or tolerances, are traceable to the National Institute of Standards and Technology (NIST). Supporting documentation relative to traceability is on file and is available for examination upon request. This certificate is not to be reproduced, except in full, without the written approval of the Southwest Research Institute Department of Quality Assurance Calibration Laboratory.

The uncertainty of the calibration was sufficient to determine that the item met the manufacturer's published specifications unless stated otherwise below.

Ambient Conditions: Temperature: 71.0 Degrees Fahrenheit Humidity: 45 % RH

Calibration Date: 9 Feb 00 **Calibration Procedure:** CLCP-WT-001 DEC 99

Condition as Received: IN TOLERANCE

Condition as Released: IN TOLERANCE, ADJUSTED

Remarks:

Approved by:

Jim Patterson, Supervisor or Walt Hill, Metrologist

Certificate # 37737

m:\nona21a.rpt Rev date 13 Apr 99

Measurements performed by:

Jerry White, Technician

Page 1 of 1

SOUTHWEST RESEARCH INSTITUTE

CALIBRATION LABORATORY

6220 CULEBRA ROAD POST OFFICE DRAWER 28510 SAN ANTONIO, TEXAS 78228-0510 TEL (210) 522-5215 FAX (210) 522-3692

Receipt for Work Order 37737 Date Received 2/9/00

Asset No. 001444 Manufacturer SARTORIUS Model 3808

Description ELECTRONIC BALANCE Serial Number 3903006

Accessory Received/Required NONE

Div/CC ID NONE Accessory to Asset No. N/A Accuracy MFG SPECS

Div/CC DIV20 Location B57 Custodian RON GREEN Tel. 5305

Charge/Project No. 20.00751.006 Proprietary/Confidential N Date Required ROUTINE

Work Requested CALIBRATION ONSITE

Receiving Inspection

Delivered By N/A Tel. 5305

CUSTOMER COMMENTS REQUESTED

The Calibration Laboratory staff is committed to good laboratory practice and providing our customers quality calibration services. Please rate how well we are meeting our goal and provide us with your recommendations for improvement.

Quality of service provided	_____	Rating Scale
Turn-around time	_____	4 = Excellent
Certificates	_____	3 = Good
Calibration labels	_____	2 = Acceptable
Recall notification	_____	1 = Needs improvement
Condition of returned equipment	_____	
Overall service	_____	

Please check all that apply: Equipment user Delivery person Custodian

IF WE PLEASE YOU, TELL OTHERS; IF NOT, TELL US.

COMMENTS: _____

Name _____ Tel. _____

WORK ORDER 40067

Date Received 8/7/00

Asset No. 001444 Manufacturer SARTORIUS Model 3808
Description ELECTRONIC BALANCE Serial Number 3903006
Accessory Received/Required NONE
Div/CC ID NONE Accessory to Asset No. N/A
Div/CC DIV20 Location B57 Custodian RON GREEN Tel. 5305
Charge/Project No. 20.00751.006 Proprietary/Confidential N Date Required ROUTINE
Work Requested CALIBRATION ONSITE
Receiving Inspection N/A
Delivered By N/A Tel. 5305

WORK HISTORY

Date	Start Time	Stop Time	Notes
10 AUG 2000	2h		Cal Fail on site / Removal to shop. JW
11 AUG 2000	2h		Trouble shoot / on-site micro sw's JW

PARTS

Part Name	Part Number	Cost	Failure Description

40067

WORK SUMMARY

Failure Description _____

Repair Action _____

Cal Procedure CLCP-WT-001, 12/99 Temp 69 F Hum 60 %

Tech JW Cal Hrs. 1.8 Repair Hrs. 2h Part Cost _____

Action Taken Cal

Standards Used 2062 61 60 1919 18 17 16 15

Date Cal _____ Int. 6 Mo. Date Due _____ Reliability Code _____

Date Picked Up _____ Picked Up By _____

ELECTRONIC BALANCE CALIBRATION DATA SHEET

WORK ORDER H0067 DATE 7/1/2000 TECHNICIAN Qu

MODEL 3903 SERIAL NO. 3903006 ASSET NO. _____

LOCATION D82

AMBIENT: TEMP 24 HUMIDITY 61 BARO PRESS 17.51

1) CALIBRATION CHECK


AS FOUND FULL-CAPACITY INDICATION 30000.9

POST-CALIBRATION INDICATION 30000.0 TOLERANCE _____ P/F _____

2) REPEATABILITY

1	10000.2	6	.2
2	.2	7	.2
3	.2	8	.3
4	.2	9	.2
5	.3	10	.2

S.D. _____ TOLERANCE _____ P/F _____



3) OFF-CENTER ERROR

1		3	
2		4	

TOLERANCE _____ P/F _____

4) NON-LINEARITY

TEST POINT	INDICATION	NON-LIN ERROR	TEST WEIGHT S/N
0	<i>del</i>	—	—
25%			
50%			
75%			
100%			

TOLERANCE _____ P/F _____

5) COMMENTS:

scr to will not pass
check on parts

SOUTHWEST RESEARCH INSTITUTE

CALIBRATION LABORATORY

6220 CULEBRA ROAD POST OFFICE DRAWER 28510 SAN ANTONIO, TEXAS 78228-0510 TEL (210) 522-5215 FAX (210) 522-3692

Receipt for Work Order 40067 Date Received 8/7/00
Asset No. 001444 Manufacturer SARTORIUS Model 3808
Description ELECTRONIC BALANCE Serial Number 3903006
Accessory Received/Required NONE
Div/CC ID NONE Accessory to Asset No. N/A Accuracy MFG SPECS
Div/CC DIV20 Location B57 Custodian RON GREEN Tel. 5305
Charge/Project No. 20.00751.006 Proprietary/Confidential N Date Required ROUTINE
Work Requested CALIBRATION ONSITE
Receiving Inspection N/A
Delivered By N/A Tel. 5305

CUSTOMER COMMENTS REQUESTED

The Calibration Laboratory staff is committed to good laboratory practice and providing our customers quality calibration services. Please rate how well we are meeting our goal and provide us with your recommendations for improvement.

Quality of service provided	_____	Rating Scale
Turn-around time	_____	4 = Excellent
Certificates	_____	3 = Good
Calibration labels	_____	2 = Acceptable
Recall notification	_____	1 = Needs improvement
Condition of returned equipment	_____	
Overall service	_____	

Please check all that apply: Equipment user _____ Delivery person _____ Custodian _____

IF WE PLEASE YOU, TELL OTHERS; IF NOT, TELL US.

COMMENTS: _____

Name _____ Tel. _____

ELECTRONIC BALANCE CALIBRATION DATA SHEET

WORK ORDER 40067 DATE 21 Aug 2000 TECHNICIAN JMS

MODEL 3803 SERIAL NO. 390300L ASSET NO. _____

LOCATION R64

AMBIENT: TEMP 69 HUMIDITY 60 BARO PRESS 14.27

1) CALIBRATION CHECK

AS FOUND FULL-CAPACITY INDICATION 30000.9


POST-CALIBRATION INDICATION 30000.1 TOLERANCE _____ P/F _____

2) REPEATABILITY

1	10000.0	6	10000.0
2	10000.0	7	10000.1
3	10200.0	8	10000.0
4	10000.0	9	10000.0
5	10000.1	10	10000.0

2067
2061
2060
1719
1718
1717
1716
1715 506

S.D. _____ TOLERANCE _____ P/F _____



3) OFF-CENTER ERROR

1		3	
2		4	

TOLERANCE _____ P/F _____

4) NON-LINEARITY

TEST POINT	INDICATION	NON-LIN ERROR	TEST WEIGHT S/N
0	0.0	—	—
25%	7500.1		
50%	7500.0		
75%	7500.1		
100%	7500.0		

TOLERANCE _____ P/F _____

5) COMMENTS:

DID NOT REPLACE TRMS WITH
CORRECT SW. USE SUB.



Southwest Research Institute
6220 Culebra Road
San Antonio, TX 78238
(210) 522-5215
Department of Quality Assurance
Calibration Laboratory



Certificate #
0972-01

Certificate of Calibration

30 August 2000

Issued to: RON GREEN DIV20 B57
Manufacturer/Model: SARTORIUS 3808
Description: ELECTRONIC BALANCE
Serial Number: 3903006
Asset Number: 001444

This certifies the above item was calibrated in compliance with MIL-STD-45662A and ANSI/NC SL Z540-1-1994. The results of this calibration relate only to the individual item as described above. Standards used in this calibration, described in the referenced calibration procedure with associated uncertainties or tolerances, are traceable to the National Institute of Standards and Technology (NIST). Supporting documentation relative to traceability is on file and available for examination upon request. This certificate is not to be reproduced, except in full, without the written approval of the Southwest Research Institute Department of Quality Assurance Calibration Laboratory.

This laboratory is accredited by the American Association for Laboratory Accreditation (A2LA) and the results of this calibration certificate were determined in accordance with the terms of accreditation unless stated otherwise below.

The uncertainty of the calibration was sufficient to determine that the item met the manufacturer's published specifications unless stated otherwise below.

Ambient Conditions: Temperature: 69.0 Degrees Fahrenheit Humidity: 60 % RH

Calibration Date: 28 Aug 00 **Calibration Procedure:** CLCP-WT-001, 12/99

Condition as Received: INOPERATIVE SEE REMARKS

Condition as Released: IN TOLERANCE

Remarks: THE BALANCE TARE SWITCH WAS INOPERATIVE. A TEMPORARY SWITCH WAS INSTALLED PENDING REPLACEMENT SWITCH FROM SARTORIUS.

Approved by:

Jim Patterson, Supervisor, or Walt Hill, Metrologist

Certificate # 40067

m:\a2la.rpt Rev date 22 May 00

Measurements performed by:

Jerry White, Technician

Page 1 of 1

SOUTHWEST RESEARCH INSTITUTE

Calibration Laboratory

WORK ORDER

Processed by JIBARRA at 2:55:52PM on 2/8/01

1 2 3 4 5 6 7 8 9 10 11 12 13 14 15 16 17 18 19 20

Work Order 444042341

Arrived 2/8/01

Asset No. 001444 Manufacturer SARTORIUS

Model 3808

Instrument Type/Class BALANCE

Serial No. 3903006

Accessory No. Calibration Procedure CLCP-WT-001, 12/99

Location B57

Div/Client DIV20

Custodian RON GREEN

Mail Stop B57

Tel. 5305

Charge/Project No. ³⁰75-00851.006

Delivered By / Telephone

IN4CAL

Special Instructions _____

WORK NOTES

Date	Hours	Remarks/Notes
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____

REPAIR PARTS

Date	Hours	Part Name	Part Number	Failure Description	Cost
_____	_____	_____	_____	_____	_____
_____	_____	_____	_____	_____	_____
_____	_____	_____	_____	_____	_____
_____	_____	_____	_____	_____	_____

WORK SUMMARY

Failure Description _____

Repair Action _____

Calibration Procedure CLCP-WT-001, 12/99 Temp 78 F Hum. 61 %

Tech UMonds Totals Cal Hours 1.0 Repair Hours _____ Parts Cost _____

Standards Used 1716, 1717, 1718, 1719, 2060, 2061, 2062, 1715

Date Picked Up _____

Picked Up By _____

42541

ELECTRONIC BALANCE CALIBRATION DATA SHEET

WORK ORDER 44404234/ DATE 27 Feb 01 TECHNICIAN UMorela
 MODEL 3808 SERIAL NO. 3903006 ASSET NO. 1444
 LOCATION DIU 20 Bldg S1
 AMBIENT: TEMP 78 HUMIDITY 66 BARO PRESS 14.25

1) CALIBRATION CHECK

AS FOUND FULL-CAPACITY INDICATION 30000.5g


POST-CALIBRATION INDICATION 30000.0g TOLERANCE _____ P/F _____

2) REPEATABILITY

1	10000.1g	6	10000.1g
2	10000.1g	7	10000.1g
3	10000.1g	8	10000.0g
4	10000.0g	9	10000.1g
5	10000.0g	10	10000.1g

1716, 1717, 1718
1719, 2060, 2061, 2062
1715

S.D. _____ TOLERANCE _____ P/F _____



3) OFF-CENTER ERROR

1		3	
2		4	

TOLERANCE _____ P/F _____

4) NON-LINEARITY

TEST POINT	INDICATION	NON-LIN ERROR	TEST WEIGHT S/N
0	0	—	—
25%	7500.1g		
50%	7500.0g		
75%	7500.0g		
100%	7500.0g		

TOLERANCE _____ P/F _____

5) COMMENTS:

SwRI Cal-Lab By: jaw
 CAL: Aug 28, 00 DUE: Feb 28, 01
 AN: 001444 SN: 3003006



Southwest Research Institute
6220 Culebra Road
San Antonio, TX 78238
(210) 522-5215
Department of Quality Assurance
Calibration Laboratory



Certificate #
0972-01

Certificate of Calibration

28 February 2001

Issued to: RON GREEN DIV20 B57
Manufacturer/Model: SARTORIUS 3808
Description: BALANCE
Serial Number: 3903006
Asset Number: 001444
Work Order Number: 444042341

This certifies the above item was calibrated in compliance with MIL-STD-45662A and ANSI/NC SL Z540-1-1994. The results of this calibration relate only to the individual item as described above. Standards used in this calibration, described in the referenced calibration procedure with associated uncertainties or tolerances, are traceable to the National Institute of Standards and Technology (NIST). Supporting documentation relative to traceability is on file and available for examination upon request. This certificate is not to be reproduced, except in full, without the written approval of the Southwest Research Institute Department of Quality Assurance Calibration Laboratory.

This laboratory is accredited by the American Association for Laboratory Accreditation (A2LA) and the results of this calibration certificate were determined in accordance with the terms of accreditation unless stated otherwise below.

The uncertainty of the calibration was sufficient to determine that the item met the manufacturer's published specifications unless stated otherwise below.

Ambient Conditions: Temperature: 78.0 Degrees Fahrenheit Humidity: 66 % RH

Calibration Date: 27 Feb 01 **Calibration Procedure:** CLCP-WT-001, 12/99

Condition as Received: IN TOLERANCE

Condition as Released: IN TOLERANCE

Remarks:

Approved by:

Walt Hill, Supervisor
Institute Calibration Laboratory

Measurements performed by:

Vince Morales, Technician

SOUTHWEST RESEARCH INSTITUTE

Calibration Laboratory

WORK ORDER

Processed by RCRUZ at 3:48:40PM on 7/30/01

|||||

Work Order **444044579**

Arrived 7/30/01

Asset No. 001444

Manufacturer SARTORIUS

Model 3808

Instrument Type/Class BALANCE

Serial No. 3903006

Accessory No.

Calibration Procedure CLCP-WT-001, 12/99

Location B57

Div/Client DIV20

Custodian RON GREEN

Mail Stop B57

Tel. 5305

Charge/Project No. 00751.006 1.20

Delivered By / Telephone

IN4CAL

Special Instructions _____

WORK NOTES

Date	Hours	Remarks/Notes
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____

REPAIR PARTS

Date	Hours	Part Name	Part Number	Failure Description	Cost
_____	_____	_____	_____	_____	_____
_____	_____	_____	_____	_____	_____
_____	_____	_____	_____	_____	_____
_____	_____	_____	_____	_____	_____

WORK SUMMARY

Failure Description _____

Repair Action _____

Calibration Procedure _____

Temp 74 F

Hum. 60 %

Tech Qm

Totals

Cal Hours 1.5

Repair Hours _____

Parts Cost _____

Standards Used _____

aml or srt

Date Picked Up _____

Picked Up By _____

444579

ELECTRONIC BALANCE CALIBRATION DATA SHEET

WORK ORDER 44579 DATE 7/31/2001 TECHNICIAN Jim

MODEL 3888 SERIAL NO. 390300L ASSET NO. 1444

LOCATION B5

AMBIENT TEMP 74 HUMIDITY 60 BARO PRESS 1432

1) CALIBRATION CHECK

AS FOUND FULL CAPACITY INDICATION 30000.9


POST-CALIBRATION INDICATION 29999.9 TOLERANCE _____ R/F _____

2) REPEATABILITY

1	10000.0	6	9999.9
2	10000.0	7	9999.9
3	10000.0	8	10000.0
4	10000.0	9	10000.0
5	9999.9	10	10000.0

2062
2261
2066
1719
1718
1717
1712
1710

S.D. _____ TOLERANCE _____ R/F _____



3) OFF-CENTER ERROR

1		3	
2		4	

TOLERANCE _____ R/F _____

4) NON-LINEARITY

TEST POINT	INDICATION	NON-LIN ERROR	TEST WEIGHT S/N
0		—	—
25%	7500.1		
50%	7500.0		
75%	7499.9		
100%	7499.9		

TOLERANCE _____ R/F _____



Southwest Research Institute
6220 Culebra Road
San Antonio, TX 78238
(210) 522-5215
Department of Quality Assurance
Calibration Laboratory



Certificate #
0972-01

Certificate of Calibration

1 August 2001

Issued to: RON GREEN DIV20 B57
Manufacturer/Model: SARTORIUS 3808
Description: BALANCE
Serial Number: 3903006
Asset Number: 001444
Work Order Number: 444044579

This certifies the above item was calibrated in compliance with MIL-STD-45662A and ANSI/NCSL Z540-1-1994. The results of this calibration relate only to the individual item as described above. Standards used in this calibration, described in the referenced calibration procedure with associated uncertainties or tolerances, are traceable to the National Institute of Standards and Technology (NIST). Supporting documentation relative to traceability is on file and available for examination upon request. This certificate is not to be reproduced, except in full, without the written approval of the Southwest Research Institute Department of Quality Assurance Calibration Laboratory.

This laboratory is accredited by the American Association for Laboratory Accreditation (A2LA) and the results of this calibration certificate were determined in accordance with the terms of accreditation unless stated otherwise below.

The uncertainty of the calibration was sufficient to determine that the item met the manufacturer's published specifications unless stated otherwise below.

Ambient Conditions: Temperature: 74.0 Degrees Fahrenheit Humidity: 60 % RH

Calibration Date: 31 Jul 01 **Calibration Procedure:** CLCP-WT-001, 12/99

Condition as Received: IN TOLERANCE

Condition as Released: IN TOLERANCE

Remarks:

Approved by:

Walt Hill, Supervisor
Institute Calibration Laboratory

Measurements performed by:

Jerry White, Technician

SOUTHWEST RESEARCH INSTITUTE

Calibration Laboratory

WORK ORDER

Received by MROMERO, 1/22/02 1:15:40PM

Arrived 1/22/02

Work Order **444046884**

Asset No. 001444 Manufacturer SARTORIUS

Model 3808

Equipment Type BALANCE

Serial No. 3903006

Accessory No.

Interval 6 M

Calibration Procedure CLCP-WT-001, 12/99

Location B57

Div/Client DIV20

Custodian RON GREEN

Mail Stop B57

Tel 5305

IN4CAL

Special Instructions _____

Notify before adjustments or repairs. () Provide data with certificate () Certificate Type _____

Charge/Project No. 00751.006 1.20

Requester / Telephone _____

This information is correct for the work requested. _____

WORK NOTES

SWRI CALIBRATION LAB
(210) 522-5275
CAL 31 Jul 2001 BY [Signature]
DUE 1 Jan 2002 ID 7794
SN 3903006

Date	Hours	Remarks/Notes
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____

Date	Hours	Part Name	Part Number	Failure Description	Cost
_____	_____	_____	_____	_____	_____
_____	_____	_____	_____	_____	_____
_____	_____	_____	_____	_____	_____
_____	_____	_____	_____	_____	_____
_____	_____	_____	_____	_____	_____
_____	_____	_____	_____	_____	_____

WORK SUMMARY

Failure Description _____

Repair Action _____

Tech [Signature] Cal Hrs. 1.0 Repair Hrs _____ Parts Cost _____ Temp 71 F Hum. 43 %

Standards Used MIS, 1717, 1719, 6098, 2060, 2061, 2062

Date Picked Up 22 JAN.02

Picked Up By GRK ONITE

444046884



Southwest Research Institute
6220 Culebra Road
San Antonio, TX 78238
(210) 522-5215
Department of Quality Assurance
Calibration Laboratory

Certificate of Calibration

22 January 2002

Issued to: RON GREEN DIV20 B57
Manufacturer/Model: SARTORIUS 3808
Description: BALANCE
Serial Number: 3903006
Asset Number: 001444
Work Order Number: 444046884

This certifies the above item was calibrated in compliance with MIL-STD-45662A and ANSI/NCCL Z540-1-1994. Standards used in this calibration, described in the referenced calibration procedure with associated uncertainties or tolerances, are traceable to the National Institute of Standards and Technology (NIST). Supporting documentation relative to traceability is on file and is available for examination upon request. This certificate is not to be reproduced, except in full, without the written approval of the Southwest Research Institute Department of Quality Assurance Calibration Laboratory.

The uncertainty of the calibration was sufficient to determine that the item met the manufacturer's published specifications unless stated otherwise below.

Ambient Conditions: Temperature: 71.0 Degrees Fahrenheit Humidity: 43 % RH


Calibration Date: 22 Jan 02 **Calibration Procedure:** CLCP-WT-001, 12/99

Condition as Received: IN TOLERANCE


Condition as Returned: IN TOLERANCE

Remarks:

Approved by:


Walt Hill, Supervisor
Institute Calibration Laboratory

Measurements performed by:


Mark Romero, Technician

Southwest Research Institute
Calibration Laboratory
 Calibration Data Sheet

Work Order 444046884	Mfr. Sartorius	Technician Mark A. Romero
Asset No. 1444	Model 3808MP8-1V20	Procedure CLCP-WT-001, 12/99
Serial No. 3903006	Type Balance	Cal Date 22-Jan-02

Location: Bldg. 51

Ambient Conditions: 71 F 43 %RH 14.3 PSIA

Operational Check: Limits +/- : 0.5 g

STD Mass Load	As Found Indication	Instrument Error
30000.0 g	29998.1 g	-1.9 g

Post Calibration Check:


STD Mass Load	Post calibration Indication	Instrument Error	Results
30000.0 g	29999.6 g	-0.4 g	Pass

Repeatability Check: Mass Load: 15000.0 g

1	14999.9 g	6	14999.8 g
2	14999.9 g	7	14999.8 g
3	14999.9 g	8	14999.9 g
4	14999.8 g	9	14999.9 g
5	14999.9 g	10	14999.9 g

Std Deviation	Tolerance
0.05	0.1 g

Off-Centerline Check: Mass Load: 10000.0 g



	Indication	Instrument Error	+/- Limits	Results
1	0.0 g	0.0 g	0.4	Pass
2	0.0 g	0.0 g	0.4	Pass
3	0.4 g	0.4 g	0.4	Pass
4	-0.4 g	-0.4 g	0.4	Pass


Non-Linearity Check: Range: 30000.0 g

STD Mass Load	Indication	Instrument Error	+/- Limits	Results
0.0 g	0.0 g	0.0 g	0.15	Pass
7500.0 g	7500.0 g	0.0 g	0.15	Pass
15000.0 g	7499.9 g	-0.1 g	0.15	Pass
22500.0 g	7500.0 g	0.0 g	0.15	Pass
30000.0 g	7499.9 g	-0.1 g	0.15	Pass

Remarks: Readability is 0.1g. Standards used 2060, 2061, 2062, 6098, 1719, 1717, and 1715.

Southwest Research Institute
 Calibration Laboratory
 Uncertainty Budget

Sartorius 3808MP8 (TI)	Units	Range	Acc. +/- (1)	Resolution
	g	32 500	0.5	0.1
Source of Uncertainty	Value	Distribution	Divisor	Std. Uncert.
Standard weight (2)	0.0702	Rectangular	Sqrt 3	0.0405
Resolution	0.1	Rectangular	Sqrt 3	0.0577
Air buoyancy (3)	0.0325	Rectangular	Sqrt 3	0.0188
Combined Uncertainty	RSS			0.0730
Expanded Uncertainty	$k=2$			0.15
Test Accuracy Ratio (TAR)	TI Acc./STD Acc.			
	7.1	to 1		
Test Uncertainty Ratio (TUR)	TI Acc./Muk=2			
	3.4	to 1		
<p>(1) Combined Uncertainty (2 sigma) of mfg std dev. (0.1g), linearity (0.15g), corner load (0.4g), & Class S1 30kg internal check weight (75mg) tolerances.</p> <p>(2) RSS of combined tolerances for check weights [Class S1 25kg (62mg), 10kg (25mg), (3) 5kg (12mg), 2kg (5mg), and 500g (1.2mg)].</p> <p>(3) No correction is made for air buoyancy. As the span of the weighing machine was adjusted before calibration, the uncertainty limits were estimated to be 1 ppm of the nominal value ie= 32.5 mg.</p>				



Southwest Research Institute
6220 Culebra Road
San Antonio, TX 78238
(210) 522-5215
Department of Quality Assurance
Calibration Laboratory

Certificate of Calibration

18 July 2002

Issued to: RON GREEN DIV20 T1
Manufacturer/Model: SARTORIUS 3808-MP8
Description: BALANCE
Serial Number: 39030006
Asset Number: 001444
Work Order Number: 444049442

Due JAN 17, 03

This certifies the above item was calibrated in compliance with MIL-STD-45662A and ANSI/NC SL Z540-1-1994. Standards used in this calibration, described in the referenced calibration procedure with associated uncertainties or tolerances, are traceable to the National Institute of Standards and Technology (NIST). Supporting documentation relative to traceability is on file and is available for examination upon request. This certificate is not to be reproduced, except in full, without the written approval of the Southwest Research Institute Department of Quality Assurance Calibration Laboratory.

The uncertainty of the calibration was sufficient to determine that the item met the manufacturer's published specifications unless stated otherwise below.

Ambient Conditions: Temperature: 72.0 Degrees Fahrenheit Humidity: 62 % RH


Calibration Date: 17 Jul 02 **Calibration Procedure:** CLCP-WT-001, 12/99

Condition as Received: IN TOLERANCE

Condition as Returned: IN TOLERANCE


Remarks:

Approved by:



Walt Hill, Metrology Group Leader
Institute Calibration Laboratory

Measurements performed by:



Mark Romero, Technician

Southwest Research Institute
 Calibration Laboratory
 Calibration Data Sheet

Work Order 444049442	Mfr. Sartorius	Technician Mark A. Romero
Asset No. 001444	Model 3808-MP8	Procedure CLCP-WT-001, 12/99
Serial No. 39030006	Type Balance	Cal Date 17-Jul-02

Location: Bldg. 51

Ambient Conditions: 72 F 62 %RH 14.32 PSIA

Operational Check: Limits +/- : 0.6 g

STD Mass Load	As Found Indication	Instrument Error
30000.0 g	29998.6 g	-1.4 g


Post Calibration Check:

STD Mass Load	Post calibration Indication	Instrument Error	Results
30000.0 g	29999.8 g	-0.2 g	Pass

Repeatability Check: Mass Load: 10000.0 g

1	10000.1 g	6	10000.2 g
2	10000.1 g	7	10000.2 g
3	10000.0 g	8	10000.1 g
4	10000.0 g	9	10000.1 g
5	10000.0 g	10	10000.2 g
Std Deviation		Tolerance	
0.1 g		0.1 g	

Off-Centerline Check: Mass Load: 10000.0 g



	Indication	Instrument Error	+/- Limits	Results
1	0.0 g	0.0 g	0.4	Pass
2	0.2 g	0.2 g	0.4	Pass
3	0.1 g	0.1 g	0.4	Pass
4	-0.3 g	-0.3 g	0.4	Pass

Non-Linearity Check: Range: 32000.0 g

STD Mass Load	Indication	Instrument Error	+/- Limits	Results
0.0 g	0.0 g	0.0 g	0.2	Pass
8000.0 g	8000.1 g	0.1 g	0.2	Pass
16000.0 g	8000.2 g	0.2 g	0.2	Pass
24000.0 g	7999.8 g	-0.2 g	0.2	Pass
32000.0 g	7999.8 g	-0.2 g	0.2	Pass

Remarks: Readability is 0.1g. Standards used 2060, 2061, 2062, 6098, 1719, 1718, 1716, and 1717.
 Balance located on table that does not give the unit proper support. As weight increases on balance, the table becomes less stable and causes unit to drift.



SOUTHWEST RESEARCH INSTITUTE™

6220 Culebra Road, P.O. Drawer 28510
Institute Quality Systems
Institute Calibration Laboratory
Phone: 210-522-5215 Fax 210-522-3692

Certificate of Calibration

Submitted By: DIV20

Address: T1

Contact: RON GREEN

Manufacturer Model: SARTORIUS 3808-MP8

Description: BALANCE

Serial No: 39030006

Asset No: 001444

Procedure: CLCP-WT-001, 12/99

Work Order: 444051979

Date Issued: Jan 22, 2003

Calibration Date: Jan 21, 2003

****Calibration Due:** Jul 21, 2003

Calibration Location: B51

Environment: Temp. 77.0°F Hum. 45 %RH

***As Found:** IN TOLERANCE

***As Left:** IN TOLERANCE

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 1999 and ANSI/NC SL Z540-1-1994 which are equivalent to relevant requirements of the ISO 9000-1994 series of standards. This certificate may not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. The results of this calibration relate only to the individual instrument described above. This certificate shall not be used to claim product endorsement by the American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government

Uncertainty evaluation includes the item under test and is calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM). The uncertainty represents an expanded uncertainty using a coverage factor of k=2 to approximate a 95% confidence level. The calibration process provides a Test Uncertainty Ratio (TUR) of less than or equal to 25% (4:1) of the test limit unless otherwise stated in remarks or an attachment

*The client has sole responsibility for determination of in/out of tolerance or compliance/noncompliance. An in/out of tolerance opinion is provided for your convenience based only on the Test Instrument (TI) reading(s) and limits as reported. The reported uncertainty relates only to the results at the time of calibration and does not imply any short or long term stability of the TI.

**Calibration interval is determined by the client and does not assure the instrument will remain within tolerance until this date. Any number of factors may cause the instrument to be out of tolerance before the next calibration date.

Remarks: None

Standards Used

Asset	Manufacturer	Model	Description	Cal Due
002060	RICE LAKE	5KG	WEIGHT, CLASS 1	Jun 14, 03
002061	RICE LAKE	5KG	WEIGHT, CLASS 1	Jun 14, 03
002062	RICE LAKE	10KG	WEIGHT, CLASS 1	Jun 14, 03
006098	RICE LAKE	25KG	WEIGHT, CLASS 1	Jun 14, 03
001718	RICE LAKE	2KG	WEIGHT, CLASS 1	Jun 14, 03
001715	RICE LAKE	500G	WEIGHT, CLASS S	Jun 14, 03

Approved by: Walt Hill

Metrology Group Leader

m:\Nona21a1.rpt Rev date 15, August 02

Measurements by: Mark Romero

Metrology Technician

Southwest Research Institute
Calibration Laboratory
 Calibration Data Sheet

Work Order 444051979	Mfr. Sartorius	Technician Mark A. Romero
Asset No. 001444	Model 3808	Procedure CLCP-WT-001, 12/99
Serial No. 39030006	Type Balance	Cal Date 21-Jan-03

Location: Bldg. 51

Ambient Conditions: 77 F 45 %RH 14.22 PSIA

Operational Check: Limits +/- : 1.1 g Uncertainty: 0.1 g

STD Mass Load	As Found Indication	Instrument Error
30000.0 g	30000.8 g	0.8 g

Post Calibration Check:



STD Mass Load	Post Calibration Indication	Instrument Error	Results
30000.0 g	30000.0 g	0.0 g	Pass

Repeatability Check: Mass Load: 10000.0 g

1	10000.1 g	6	10000.0 g
2	10000.1 g	7	10000.0 g
3	10000.1 g	8	10000.0 g
4	10000.0 g	9	10000.0 g
5	10000.0 g	10	10000.0 g

Std Deviation	Tolerance
0.0 g	0.2 g

Off-Centerline Check: Mass Load: 10000.0 g Uncertainty: 0.1 g



	Indication	Instrument Error	+/- Limits	Results
1	0.1 g	0.1 g	0.8	Pass
2	0.2 g	0.2 g	0.8	Pass
3	0.1 g	0.1 g	0.8	Pass
4	0.8 g	0.8 g	0.8	Pass

Non-Linearity Check: Range: 30000.0 g Uncertainty: 0.1 g

STD Mass Load	Indication	Instrument Error	+/- Limits	Results
0.0 g	0.0 g	0.0 g	0.4	Pass
7500.0 g	7500.0 g	0.0 g	0.4	Pass
15000.0 g	7499.8 g	-0.2 g	0.4	Pass
22500.0 g	7500.0 g	0.0 g	0.4	Pass
30000.0 g	7499.8 g	-0.2 g	0.4	Pass

Remarks: Readability is 0.1g. Standards used 2060, 2061, 2062, 6098, 1718, and 1715.