

SOUTHWEST RESEARCH INSTITUTE
Department of Quality Assurance
Calibration Laboratory • 522-5215

WORK ORDER

CERTIFICATE # 27654 ASSET # 001435 DATE 20 Apr 97

ITEM DATA:

Manufacturer Filko Model 8050A
Description DMM Serial # 5005110
Accessories _____

ACTION REQUESTED MI

CUSTODIAN M. P. D. D. D.

Turned in by: _____ Phone 1090

CHARGE # 20-1402-5-71 Date Required _____

INSTRUMENT USED ON: DOD/NASA NUCLEAR GLP SPPE ISO
 OTHER _____

COPY OF CALIBRATION CERTIFICATE Yes No

NEW WORK Yes No If yes, an evaluation shall be made to verify capabilities.
By _____ Date _____

CONDITION RECEIVED: (F) Out of tolerance, repaired to specifications
 (G) In tolerance, minor adjustments/repairs made
 (J) In tolerance, no adjustments/repairs
 (K) Out of tolerance, adjusted to specifications
 (S) Received into system, introduced or reactivated


ACTION TAKEN: (Calibration/Repair/Parts) Roche Dev ST# 311670 4/21
At indicator on when received. Batteries would not charge.
Replaced battery pack

CAL ENVIRONMENT:
Temperature 72 °F Humidity 34 %RH

CALIBRATED/REPAIRED:
By Roche Dev. Sam Antonio, TX Cal Procedure MEF
Date 12/26/97 Accuracy MEF
Cal Interval 12 mos Reliability Code: _____
Next Cal due 12/26/98 Cal Time _____ Repair Time _____
Standards used (Asset#) Vendor

DATE COMPLETED 24 Apr 97
DATE PICKED UP 12/4/97 PICKED UP BY [Signature]

27654



CERT. NUMBER: 50533

ROTHE DEVELOPMENT, INC.
METROLOGY SERVICES DIVISION
4614 SINCLAIR RD., SAN ANTONIO, TEXAS 78222 PH:210-648-3131

CERTIFICATE OF CALIBRATION

ISSUED TO: Southwest Research Institute
6220 Culebra, Bldg. 64/Division 30
San Antonio, TX 78284
522-5460

(30) MFG: Fluke
MODEL: 8050A
NOMEN: DMM
S/N: 5005110
CUST. ID:

CAL DATE: 12/02/1997
DUE DATE: 12/02/1998

CONTROL NO.: 103 - 21872
TECHNICIAN: 4
SPECIFICATIONS: MFG
PROCEDURE: MFG
WORK ORDER: 62631
CUSTOMER P.O.: 02127/ST311670/20-1402571
RECEIVED CONDITION: OUT OF TOLERANCE
RETURNED CONDITION: IN TOLERANCE

CALIBRATION PERFORMED AT: RDMSD
CALIBRATION INTERVAL: 12mos.
TEMPERATURE: 72 °F
RELATIVE HUMIDITY: 34 %

DATE RECEIVED: 11/24/1997

COMMENTS:

ATTACHMENTS: CALIBRATION DATA

All Calibrations performed at Rothe Development, Inc. Metrology Services Division meet the requirements of ANSI/NC SL Z540-1-1994, ISO/IEC GUIDE 25, and ISO 10012-1, and are traceable to the National Institute of Standards and Technology. The collective uncertainty of the measurement(s) does not exceed 25% (TUR_>4:1) of the instrument specification(s) unless noted in the COMMENTS section.

TR#	MFG	MODEL	SERIAL NO.	DUE DATE
20	FLUKE	5700A	4605002	12/09/1997
258	Fluke	5725A	6585002	12/09/1997

APPROVED BY:

CMS

QCO

DATE: 12/03/1997

ROTHE DEVELOPMENT METROLOGY SERVICES

CALIBRATION DATA: FLUKE 8050A DIGITAL MULTIMETER

CUSTOMER: SWRT
 WO NUMBER: 62631
 SERIAL: 5005110
 CUST ID: _____

DATE: 2 Dec 97
 TECH: PJS
 INST NO: 21872

CALIBRATION DATA TAKEN

INCOMING ✓
 OUTGOING _____

CONDITION OF EQUIPMENT

IN TOLERANCE _____
 OUT OF TOLERANCE ✓ *
 * BT operation only

DC VOLTAGE ACCURACY

RANGE	INPUT	MIN	READING	MAX
200 mV	+190.0 mV	189.92	<u>190.01</u>	190.08
	-190.0 mV	189.92	<u>190.00</u>	190.08
2 V	+1.9 V	1.8992	<u>1.9000</u>	1.9008
	-1.9 V	1.8992	<u>1.9001</u>	1.9008
20 V	+19 V	18.992	<u>19.000</u>	19.008
200 V	+190 V	189.92	<u>190.00</u>	190.08
1000 V	+1000 V	999.5	<u>1000.0</u>	1000.5

AC VOLTAGE ACCURACY

RANGE	INPUT	MIN	READING	MAX
2 V	SHORT		<u>.0017</u>	<.0040
200 mV	190.0 mV 100 Hz	188.95	<u>189.83</u>	191.05
	190.0 mV 10 KHz	188.95	<u>189.73</u>	191.05
	190.0 mV 50 KHz	180.20	<u>186.61</u>	199.80
2 V	100.0 mV 100 Hz	.0985	<u>.0997</u>	.1015
	1.9 V 100 Hz	1.8895	<u>1.8994</u>	1.9105
	1.9 V 10 KHz	1.8895	<u>1.8983</u>	1.9105
20 V	1.9 V 50 KHz	1.8020	<u>1.8674</u>	1.9980
	19 V 100 Hz	18.895	<u>18.987</u>	19.105
	19 V 10 KHz	18.895	<u>18.997</u>	19.105
	19 V 50 KHz	18.020	<u>18.891</u>	19.980

200 V	190 V	100 Hz	188.95	<u>189.95</u>	191.05
	100 V	10 KHz	99.40	<u>99.93</u>	100.60
750 V	750 V	100 Hz	745.2	<u>749.8</u>	754.8
	750 V	1 KHz	745.2	<u>748.9</u>	754.8

dB VOLTAGE ACCURACY

RANGE	INPUT	MIN	READING	MAX
200 mV	SHORT	> -75	935-39 -93	
	10 mV 100 Hz	-37.28	<u>-37.78</u>	-38.28
	10 mV 10 KHz	-37.28	<u>-37.78</u>	-38.28
	1.0 V 100 Hz	+2.07	<u>+ 2.22</u>	+2.37

DC CURRENT ACCURACY

RANGE	INPUT	MIN	READING	MAX
200 uA	190 uA	189.41	<u>190.03</u>	190.59
2 mA	1.9 mA	1.8941	<u>1.9000</u>	1.9059
20 mA	19 mA	18.941	<u>18.999</u>	19.059
200 mA	190 mA	189.41	<u>190.12</u>	190.59
2000 mA	1.9 A	1894.1	<u>1901.8</u>	1905.9

RESISTANCE ACCURACY

RANGE	ACTUAL INPUT	TOLERANCE	READING
200 Ω	SHORT	<00.04 C	<u>00.01</u>
	(1) <u>100.01 Ω</u>	±14 C	<u>99.94</u>
2 KΩ	<u>1.0000 k</u>	±12 C	<u>.9993</u>
20 KΩ	<u>10.000 k</u>	±7 C	<u>9.998</u>
200 KΩ	<u>100.00 k</u>	±7 C	<u>99.98</u>
2000 KΩ	<u>1.0001 M</u>	±28 C	<u>999.9</u>
20 MΩ	<u>10.001 M</u>	±28 C	<u>10.002</u>
2 mS	<u>1.0000 k</u>	±15 C	<u>1.0003</u>
200 nS	<u>10.001 M</u>	±70 C	<u>99.97</u>

ROTHE DEVELOPMENT METROLOGY SERVICES

CALIBRATION DATA: FLUKE 8050A DIGITAL MULTIMETER

CUSTOMER: SWRI
 WO NUMBER: 62631
 SERIAL: 5005110
 CUST ID: _____

DATE: 2 Dec 97
 TECH: PJS
 INST NO: 21872

CALIBRATION DATA TAKEN

INCOMING _____
 OUTGOING ✓

CONDITION OF EQUIPMENT

IN TOLERANCE ✓
 OUT OF TOLERANCE _____

DC VOLTAGE ACCURACY

RANGE	INPUT	MIN	READING	MAX
200 mV	+190.0 mV	189.92	<u>190.00</u>	190.08
	-190.0 mV	189.92	<u>189.99</u>	190.08
2 V	+1.9 V	1.8992	<u>1.9000</u>	1.9008
	-1.9 V	1.8992	<u>1.9000</u>	1.9008
20 V	+19 V	18.992	<u>18.999</u>	19.008
200 V	+190 V	189.92	<u>190.00</u>	190.08
1000 V	+1000 V	999.5	<u>1000.0</u>	1000.5

AC VOLTAGE ACCURACY

RANGE	INPUT	MIN	READING	MAX
2 V	SHORT		<u>.0017</u>	<.0040
200 mV	190.0 mV 100 Hz	188.95	<u>189.83</u>	191.05
	190.0 mV 10 KHz	188.95	<u>189.75</u>	191.05
	190.0 mV 50 KHz	180.20	<u>186.64</u>	199.80
2 V	100.0 mV 100 Hz	.0985	<u>.0998</u>	.1015
	1.9 V 100 Hz	1.8895	<u>1.8992</u>	1.9105
	1.9 V 10 KHz	1.8895	<u>1.8986</u>	1.9105
20 V	1.9 V 50 KHz	1.8020	<u>1.8676</u>	1.9980
	19 V 100 Hz	18.895	<u>18.984</u>	19.105
	19 V 10 KHz	18.895	<u>19.015</u>	19.105
	19 V 50 KHz	18.020	<u>18.926</u>	19.980

200 V	190 V	100 Hz	188.95	<u>189.95</u>	191.05
	100 V	10 KHz	99.40	<u>100.03</u>	100.60
750 V	750 V	100 Hz	745.2	<u>749.9</u>	754.8
	750 V	1 KHz	745.2	<u>748.8</u>	754.8

dB VOLTAGE ACCURACY

RANGE	INPUT	MIN	READING	MAX
200 mV	SHORT	> -75	<u>-93</u>	
	10 mV 100 Hz	-37.28	<u>-37.78</u>	-38.28
	10 mV 10 KHz	-37.28	<u>-37.78</u>	-38.28
	1.0 V 100 Hz	+2.07	<u>+2.22</u>	+2.37

DC CURRENT ACCURACY

RANGE	INPUT	MIN	READING	MAX
200 uA	190 uA	189.41	<u>190.02</u>	190.59
2 mA	1.9 mA	1.8941	<u>1.8998</u>	1.9059
20 mA	19 mA	18.941	<u>18.998</u>	19.059
200 mA	190 mA	189.41	<u>190.11</u>	190.59
2000 mA	1.9 A	1894.1	<u>1901.7</u>	1905.9

RESISTANCE ACCURACY

RANGE	ACTUAL INPUT	TOLERANCE	READING
200 Ω	SHORT	<00.04 C	<u>00.01</u>
	(1) <u>100.01 Ω</u>	±14 C	<u>99.94</u>
2 KΩ	<u>1.0000 k</u>	±12 C	<u>999.93</u>
20 KΩ	<u>10.000 k</u>	±7 C	<u>9999</u>
200 KΩ	<u>100.00 k</u>	±7 C	<u>9998</u>
2000 KΩ	<u>1.0001 M</u>	±28 C	<u>1000.0</u>
20 MΩ	<u>10.001 M</u>	±28 C	<u>10.002</u>
2 mS	<u>1.0000 k</u>	±15 C	<u>1.0004</u>
200 nS	<u>10.001 M</u>	±70 C	<u>99.97</u>



Southwest Research Institute
6220 Culebra Road
San Antonio, TX 78238
Department of Quality Assurance
Calibration Laboratory

Certificate of Calibration

4 December 1997

Issued to: DARRELL DUNN DIV20 B57
Manufacturer/Model: FLUKE 8050A
Description: DIGITAL MULTIMETER
Serial Number: 5005110
Asset Number: 001435

Environmental Conditions

Temperature: 72 Deg. F Humidity: 34 % RH

Calibration Information

Calibration was in accordance with requirements of MIL-STD-45662A and ANSI/NC SL Z540-1-1994. Measurements are traceable to the National Institute of Standards and Technology (NIST). This report may not be reproduced except in full without written approval of the originator. Inspection and test data are on file and available for inspection.

The uncertainty of the calibration was sufficient to determine that the instrument met the manufacturer's specifications.

Calibration Date: 2 Dec 97 Calibration Procedure: MFG

Interval: 12 months Received: Out of Tolerance


Next Calibration Due: 2 Dec 98

Remarks: CALIBRATED BY ROTHE DEV., SAN ANTONIO, TX.
ROTHE CERT. #50533.

Out of Tolerance Data

The unit was found to be in an Out of Tolerance condition as indicated below. An evaluation should be made by the user to determine any adverse impact that may have occurred.

BT INDICATOR ON WHEN RECEIVED - BATTERIES WOULD NOT CHARGE.
REPLACED BATTERY PACKS; TOOK DATA. NO ADJUSTMENTS NEEDED.

Signed: 
Title: Calibration Laboratory

Checked By: 
Title: Supervisor

LAST PAGE OF REPORT
Total Pages Printed: 1

Certificate # 27654

SOUTHWEST RESEARCH INSTITUTE
Department of Quality Assurance
Calibration Laboratory • 522-5215

WORK ORDER

CERTIFICATE # 34811 ASSET # 001435 DATE 07 June 99

ITEM DATA:

Manufacturer Fluke Model 8050A
Description Digital Multimeter Serial # 5005110
Accessories _____

ACTION REQUESTED cal

CUSTODIAN Div. 20, Darrell Dunn

Turned in by: _____ Phone 6090

CHARGE # 20-04 Date Required _____

INSTRUMENT USED ON: (DOD/NASA) (NUCLEAR) (GLP) (SPPE) (ISO)
 OTHER _____

COPY OF CALIBRATION CERTIFICATE (Yes) (No)

NEW WORK Yes No If yes, an evaluation shall be made to verify capabilities.

By [Signature] Date 07 June 99

Work involves proprietary/confidential information or equipment (Yes) (No)

- CONDITION RECEIVED:
- (F) Out of tolerance, repaired to specifications
 - (G) In tolerance, minor adjustments/repairs made
 - (J) In tolerance, no adjustments/repairs
 - (K) Out of tolerance, adjusted to specifications
 - (S) Received into system, introduced or reactivated

ACTION TAKEN: (Calibration/Repair/Parts) _____

CAL ENVIRONMENT:
Temperature 75 °F Humidity 42 %RH

CALIBRATED/REPAIRED:
By [Signature] Cal Procedure MEV/CAL 5700A/CAL VER REV 1.1
Date 16 June 99 Accuracy MFS
Cal Interval 12 Reliability Code: 1
Next Cal due 16 June 00 Cal Time 1.0 Repair Time _____
Standards used (Asset#) 7198

DATE COMPLETED 16 June 99
DATE PICKED UP 6/22/99 PICKED UP BY [Signature]

34811

WORK ORDER 39975

Date Received 7/31/00

Asset No. 001435 Manufacturer FLUKE Model 8050A
 Description DIGITAL MULTIMETER Serial Number 5005110
 Accessory Received/Required TEST LEADS
 Div/CC ID NONE Accessory to Asset No. N/A
 Div/CC DIV20 Location B57 Custodian DARRELL DUNN Tel. 6090
 Charge/Project No. 20.00751.006 Proprietary/Confidential _____ Date Required ROUTINE
 Work Requested CALIBRATION
 Receiving Inspection OK
 Delivered By DARRELL DUNN Tel. 6090

WORK HISTORY

Date	Start Time	Stop Time	Notes
			Check Fuse.

PARTS

Part Name	Part Number	Cost	Failure Description

39975

WORK SUMMARY

Failure Description _____
 Repair Action _____
 Cal Procedure CL-251, 9/99 Fluke 8050A 5520 IIR SEP 99 Temp 74 F Hum 40 %
 Tech Vincent Morales Cal Hrs. 1.0 Repair Hrs. _____ Part Cost _____
 Action Taken Calibrated
 Standards Used 6413
 Date Cal 09 Aug 2000 Int. 0 Mo. Date Due Not on Recal Reliability Code 3
 Date Picked Up 8/10/2000 Picked Up By Darrell Dunn

Calibration Results

SwRI Calibration Laboratory

UUT: FLUKE 8050A
 DIGITAL MULTIMETER
 Serial No: 5005110
 Asset No. 001435

Result: **PASS**
 Performed on: 8/9/00 at 09:41:10
 Performed by: Mark Anthony Romero
 Environment: Temp. 74.0°F Humid. 40 %
 Condition F/L: FOUND-LEFT
 Procedure Completed: YES

Notes:

Standards Used

Asset	Mfg	Model	Description	Cal. Date	Due Date
006413	FLUKE	5520A	MULTI-PRODUCT CALIBRATOR	14-Oct-99	14-Oct-00

Test Data

TEST#	STD PARAMETER	TRUE VALUE	UNIT UNDER TEST			ERROR in (% of Tol)	NOTIFY	
			READING	TOLERANCE	UUT ERROR		TUR	USER
DISPLAY TEST								
	overrange indicator							
	Result of Operator Evaluation					PASS		
	00.00							
	Result of Operator Evaluation					PASS		
	.0000							
	Result of Operator Evaluation					PASS		
	0.000							
	Result of Operator Evaluation					PASS		
	00.00							
	Result of Operator Evaluation					PASS		
	000.0							
	Result of Operator Evaluation					PASS		
	0.000							
	Result of Operator Evaluation					PASS		
							
	Result of Operator Evaluation					PASS		
	all segments of the digits illuminated							
	Result of Operator Evaluation					PASS		
	display segments							
	Result of Operator Evaluation					PASS		
	high voltage indicator (HV) on							
	Result of Operator Evaluation					PASS		
	high voltage indicator (HV) off							
	Result of Operator Evaluation					PASS		
	negative sign (-)							
	Result of Operator Evaluation					PASS		
	(dB) and (rel) indicators on							
	Result of Operator Evaluation					PASS		
DC VOLTAGE								
200mV Range								
	15	190.00mV	190.00	80uV	0V	0		
	16	-190.00mV	-190.00	80uV	0V	0		
2V Range								
	17	1.9000V	1.9000	800uV	0V	0		
	18	-1.9000V	-1.9000	800uV	0V	0		

TEST#	STD PARAMETER	TRUE VALUE	UNIT UNDER TEST			ERROR in (% of Tol)	NOTIFY TUR USER
			READING	TOLERANCE	UUT ERROR		
20V Range							
19	19.000V		19.000	8mV	0V	0	
200V Range							
20	190.00V		190.00	80mV	0V	0	
1000V Range							
21	1000.0V		1000.0	500mV	0V	0	
AC VOLTAGE							
200mV Range							
22	190.00mV @ 100Hz		189.83	1.05mV	-170uV	16	
23	190.00mV @ 10kHz		189.76	1.05mV	-240uV	23	
24	190.00mV @ 50kHz		186.91	9.8mV	-3.09mV	32	
2V Range							
25	0.0000V		0.0018	4mV	1.8mV	45	
26	0.1000V @ 100Hz		0.0994	1.5mV	-600uV	40	
27	1.9000V @ 100Hz		1.8989	10.5mV	-1.1mV	10	
28	1.9000V @ 10kHz		1.8980	10.5mV	-2mV	19	
29	1.9000V @ 50kHz		1.8701	98mV	-29.9mV	31	
20V Range							
30	19.000V @ 100Hz		18.980	105mV	-20mV	19	
31	19.000V @ 10kHz		18.998	105mV	-2mV	2	
32	19.000V @ 50kHz		18.924	980mV	-76mV	8	
200V Range							
33	190.00V @ 100Hz		189.88	1.05V	-120mV	11	
34	100.00V @ 10kHz		99.89	600mV	-110mV	18	
750V Range							
35	750.0V @ 100Hz		749.2	4.8V	-800mV	17	
36	750.0V @ 1kHz		748.1	4.8V	-1.9V	40	
dB VOLTAGE TESTS							
UUT reading below -75dBm							
Result of Operator Evaluation							PASS
UUT reading below -75dBm							
dB VOLTAGE TESTS							
UUT reading below -75dBm							
dB VOLTAGE TESTS							
UUT reading below -75dBm							
Result of Operator Evaluation							PASS
37	37.78dBm @ 100Hz		40.00	134.93uV	-2.22dBm	>500	*
38	-37.78dBm @ 100Hz		-37.78	174.224uV	0dBm	0	
39	-37.78dBm @ 100Hz		-37.78	174.224uV	0dBm	0	
40	2.22dBm @ 100Hz		2.22	17.4224mV	0dBm	0	
RESISTANCE							
20 MOhm Range							
41	10.00 MOhm		10.00	28 kOhm	0 Ohm	0	
2000 kOhm Range							
42	1000.0 kOhm		999.9	2.8 kOhm	-100 Ohm	4	
200 kOhm Range							
43	100.00 kOhm		99.97	70 Ohm	-30 Ohm	43	
200nS Range							
44	100.00nS		99.99	700pS	-10pS	1	
20 kOhm Range							
45	10.000 kOhm		9.998	7 Ohm	-2 Ohm	29	
2 kOhm Range							
46	1.0000 kOhm		0.9993	1.2 Ohm	-700 mOhm	58	
200 Ohm Range							
47	0.00 Ohm		0.01	40 mOhm	10 mOhm	25	
48	100.00 Ohm		99.94	140 mOhm	-60 mOhm	43	
2mS Range							
49	1.0000mS		1.0007	1.5uS	700nS	47	
DC CURRENT							
200uA Range							
50	190.00uA		190.01	590nA	10nA	2	
2mA Range							

TEST#	STD PARAMETER	TRUE VALUE	UNIT UNDER TEST			ERROR in (% of Tol)	NOTIFY	
			READING	TOLERANCE	UUT ERROR		TUR	USER
51	1.9000mA		1.8997	5.9uA	-300nA	5		
	20mA Range							
52	19.000mA		18.998	59uA	-2uA	3		
	200mA Range							
53	190.00mA		190.11	590uA	110uA	19		
	2000mA Range							
54	1900.0mA		1901.6	5.9mA	1.6mA	27		
	AC CURRENT							
	20mA Range							
55	19.000mA @ 100Hz		18.973	200uA	-27uA	14		

End of Test Data



Southwest Research Institute
6220 Culebra Road
San Antonio, TX 78238
(210) 522-5215
Department of Quality Assurance
Calibration Laboratory



Certificate #
0972-01

Certificate of Calibration

9 August 2000

Issued to: DARRELL DUNN DIV20 B57
Manufacturer/Model: FLUKE 8050A
Description: DIGITAL MULTIMETER
Serial Number: 5005110
Asset Number: 001435

This certifies the above item was calibrated in compliance with MIL-STD-45662A and ANSI/NC SL Z540-1-1994. The results of this calibration relate only to the individual item as described above. Standards used in this calibration, described in the referenced calibration procedure with associated uncertainties or tolerances, are traceable to the National Institute of Standards and Technology (NIST). Supporting documentation relative to traceability is on file and available for examination upon request. This certificate is not to be reproduced, except in full, without the written approval of the Southwest Research Institute Department of Quality Assurance Calibration Laboratory.

This laboratory is accredited by the American Association for Laboratory Accreditation (A2LA) and the results of this calibration certificate were determined in accordance with the terms of accreditation unless stated otherwise below.

The uncertainty of the calibration was sufficient to determine that the item met the manufacturer's published specifications unless stated otherwise below.

Ambient Conditions: Temperature: 74.0 Degrees Fahrenheit Humidity: 40 % RH

Calibration Date: 9 Aug 00 **Calibration Procedure:** FLUKE 8050A 5520 1YR SEP 99

Condition as Received: IN TOLERANCE

Condition as Released: IN TOLERANCE

Remarks:

Approved by:

Jim Patterson, Supervisor, or Walt Hill, Metrologist

Certificate # 39975

m:\a2la.rpt Rev date 22 May 00

Measurements performed by:

Vince Morales, Technician