

**SOUTHWEST RESEARCH INSTITUTE**

**Department of Quality Assurance  
Calibration Laboratory**

**CERTIFICATE OF CALIBRATION  
06/15/95**

Issued to: JIM PRIKRYL DIV20 ,B189  
Manufacturer: PARR  
Nomenclature: TEMPERATURE CONTROLLER  
Serial Number: 1190  
Notes:

Asset Number: 003940  
Model Number: 4843  
SwRI/Div. I.D. #:

**ENVIRONMENTAL CONDITIONS**

Temperature: 72.0F

Relative Humidity: 42%

**CALIBRATION INFORMATION**

Procedure Number: MFGR  
Remarks: CALIBRATED FOR TEMPERATURE USING "J"  
THERMOCOUPLE ONLY. PSI NOT CALIBRATED.


Accuracy: MFGR SPECS  
Received IN Tolerance

**Calibration was in accordance with requirements of MIL-STD-45662A.  
Measurements are traceable to the National Institute of Standards and Tech-  
nology. Inspection and test data are on file and available for inspection.**

**STANDARDS USED FOR CERTIFICATION**

Asset #	Serial #	Mfg	Model #	Nomenclature	Cal Date	Int.	Cal Due
000182	5200003	FLUKE	5700A	CALIBRATOR	06/07/95	3	09/07/95
002868	311210	KAYE	X0251	ICE POINT REFERENCE	01/03/95	12	01/03/96

Certified by :



Certificate#: 17597

Calibration Date: 06/15/95  
Interval: 6 months  
Next Calibration Due: 12/15/95