

SOUTHWEST RESEARCH INSTITUTE

Department of Quality Assurance
Calibration Laboratory • 522-5215

WORK ORDER

WORK ORDER # 35503 ASSET # 7467 DATE 4 Aug 99

ITEM DATA:

Manufacturer MITUTOYO Model 350-712-30
Description Micrometer Head Serial # 705954
Accessories With Fixture

ACTION REQUESTED _____

CUSTODIAN Simon Hsiung
Turned in by: Jason Mock Phone 5209/6249
1402-674

CHARGE # 20-0751-006 Date Required 6 Aug 99

INSTRUMENT USED ON: (DOD/NASA) (NUCLEAR) (GLP) (SPPE) (ISO)
 OTHER _____

COPY OF CALIBRATION CERTIFICATE (Yes) (No)

NEW WORK Yes No If yes, an evaluation shall be made to verify capabilities.

By Walt Hie Date 4 Aug 99

Work involves proprietary/confidential information or equipment (Yes) (No)

CONDITION RECEIVED: _____ Out of tolerance
 In tolerance
_____ Damaged (Contact customer)
_____ Contact _____ Date _____
 Received into system, introduced or reactivated

ACTION TAKEN: (Calibration/Repair/Parts) Calibrated per procedure

CAL ENVIRONMENT:
Temperature 68 °F Humidity 44 %RH

CALIBRATED/REPAIRED:
By Simon Hsiung Cal Procedure CL-10, 4/24/99
Date 5/4/99 Accuracy 1/20
Cal Interval 6 mos Reliability Code 1
Next Cal Due 5/2000 Cal Time 2.0 Repair Time _____
Standards used (Asset #) 6465

DATE COMPLETED 5 Aug 99
DATE PICKED UP 6/6/99 PICKED UP BY Jason R Mock

35503

