SOUTHWEST RESEARCH INSTITUTE
Department of Quality Assurance
Calibration Laboratory • 522-5215

WORK ORDER

WORK ORDER # 36233 ASSET # 007579 DATE 30 Sept. 99

ITEM DATA:

Manufacturer S.M. Gauge Co. Model 01179-054615 / 0-100 mbar

Description pressure gauge Serial # 991785

Accessories

ACTION REQUESTED Cal

CUSTODIAN DIV. 20, Anne Marie

Turned in by: Mark Ehnstrom Phone 3530

CHARGE # 20.01402.159 Date Required

INSTRUMENT USED ON: ☐ (DOD/NASA) ☑ (NUCLEAR) ☐ (GLP) ☐ (SPPE) ☐ (ISO)
☐ OTHER

COPY OF CALIBRATION CERTIFICATE ☑ (Yes) ☐ (No)

NEW WORK ☐ Yes ☑ No If yes, an evaluation shall be made to verify capabilities.

By [illegible] Date 09-30-99

Work involves proprietary/confidential information or equipment ☐ (Yes) ☑ (No)

CONDITION RECEIVED:
- ___ Out of tolerance
- ___ In tolerance
- ✓ Damaged (Contact customer)
- ___ Contact ___ Date ___
- ___ Received into system, introduced or reactivated

ACTION TAKEN: (Calibration/Repair/Parts)

* Gage Rejected. Sticks badly at about 65 mbar
Could be dangerous to project materials
* Replacement Suggested

CAL ENVIRONMENT:

Temperature ___ °F Humidity ___ %RH

CALIBRATED/REPAIRED:

By [signature] Cal Procedure

Date 10-6-99 Accuracy

Cal Interval Reliability Code

Next Cal Due Cal Time 0.5 hr Repair Time

Standards used (Asset #)

DATE COMPLETED 10-6-99

DATE PICKED UP 10-7-99 PICKED UP BY Anne-Marie LESEUME [signature]

SwRi Form QA-174-0

36233

WORK ORDER HISTORY

DATE	START / STOP	CAL	REP	REMARKS
				20 23
				Sticks
				40 43
				badly
				60 63
				at about 60 mbar
				80 83
				100 97 ? Could be dangerous
				7?. error

TOTAL CAL/REPAIR | | TOTAL HOURS ____________